

CERTIFICATE

on Product Conformity (QAL1)

Number of Certificate: 0000025927_01

Certified AMS: AR500 with ER120 for NO₂, SO₂ and O₃

Manufacturer: Opsis AB
Skytteskogsvägen 16
244 02 Furulund
Sweden

Test Institute: TÜV Rheinland Energie und Umwelt GmbH

**This is to certify that the AMS has been tested
and found to comply with:**

**EN 14211: 2005, EN 14212: 2005, EN 14625: 2005,
EN 15267-1: 2009 and EN 15267-2: 2009**

Certification is awarded in respect of the conditions stated in this certificate
(also see the following pages).

The present certificate replaces Certificate No. 0000025927 of 10 March 2010



- Certified equivalent EN method
- Complying with 2008/50/EC
- TUV approved
- Annual inspection

Publication in the German Federal Gazette
(BAnz.) of 02 March 2012

The certificate is valid until:
11 February 2015

Umweltbundesamt
Dessau, 16 March 2012

TÜV Rheinland Energie und Umwelt GmbH
Köln, 15 March 2012

i. A. Dr. Hans-Joachim Hummel

ppa. Dr. Peter Wilbring

www.umwelt-tuv.de / www.eco-tuv.com
teu@umwelt-tuv.de
Tel. +49 221 806-2756

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Köln

Accreditation according to EN ISO/IEC 17025 and certified according to ISO 9001:2008.

Test report: 936/21211350/B of 07 October 2011
First certification: 12 February 2010
Validity ends: 11 February 2015
Publication: BAnz. 02 March 2012, No. 36, p. 920, chapter IV, No. 2.1

Approved application

The certified AMS is suitable for continuous ambient air monitoring (stationary operation).

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three months field test.

The AMS is approved for a temperature range of +5 °C bis +40 °C.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for ambient air applications under which it will be operated.

Basis of the certification

This certification is based on:

- test report 936/21211350/A dated 26 October 2010 of TÜV Rheinland Energie und Umwelt GmbH
- test report 936/21211350/B dated 7 October 2011 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Environmental Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette (BAnz. 02 March 2012, No. 36, p. 920, chapter IV, No. 2.1, Announcement by UBA from 23 February 2012)

AMS name:

AR500 with ER120 for NO₂, SO₂ and O₃

Manufacturer:

Opsis AB, Furulund, Sweden

Field of application:

For stationary Measuring of the concentrations of Nitrogen Dioxide, Sulphur Dioxide and Ozone in ambient air

Measuring ranges during the suitability test:

Component	Certification range	supplementary measurement ranges	Unit
NO ₂	0 - 400	0 - 1800	µg/m ³
SO ₂	0 - 700	0 - 1000	µg/m ³
O ₃	0 - 360	0 - 500	µg/m ³

Software version:

7.21

Restrictions:

None

Notes:

1. The measuring path length during the suitability test was 320 m.
2. The maintenance interval is four weeks.
3. The equivalence with the reference measurement methods according to the guideline „Demonstration of Equivalence of Ambient Air Monitoring Methods“ has been demonstrated for the components NO₂, SO₂ and O₃.
4. Function tests by external sample gas feeding are possible.
5. The test report is available on the Internet at www.qal1.de.
6. Supplementary report (Demonstration of Equivalence for the component SO₂ according the guideline „Demonstration of Equivalence of Ambient Air Monitoring Methods“) to the announcement of the UBA from 25 January 2010 (BAnz. p. 552, chapter III number 1.1).

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Köln
Report No.: 936/21211350/B dated 7 October 2011

Certified product

This certificate applies to automated measurement systems confirming to the following description:

The measurement system AR500 operates on the basis of the Differential Optical Absorption Spectroscopy (DOAS). The DOAS measuring principle uses the characteristic radiation absorption by gaseous components for quantification of the respective concentrations.

The DOAS monitor AR500 with ER120 consists of a combined emitter/receiver unit, a reflector unit and an analyser. The absorbed light is transferred from the emitter/receiver unit to the analyser via fibre optic cable.

Combined Sender-Receiver Unit ER 120

The combined Emitter-Receiver Unit ER120 comprises the optical components, the xenon light-source and the power supply PS150 for igniting the xenon light-source.

The used high-pressure Xenon lamp is a point light source. The light is generated by ignition of ultra pure Xenon gas at a pressure of approx. 30 bar. The lamp is powered by a stabilised D.C. voltage source and requires a short high-voltage ignition pulse.

The radiation of the lamp includes the ultraviolet, visible and infrared range. The wavelengths are continuously distributed over the entire spectrum, with the exception of some peaks in the near infrared range.

Analyser

The light is led to a spectrometer upon hitting the analyser and then refracted into its wavelength components by an internal grating. The refracted light is then projected onto a rapid scanning slit in front of a photo-multiplier detector, where a selected part of the spectrum is detected. The scanning slit allows separate recordings of all wavelengths by a single detector.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energie und Umwelt GmbH must be notified at the given address on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energie und Umwelt GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the validity of the certificate and on requests of the TÜV Rheinland Energie und Umwelt GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and the validity is also accessible on the internet Address: **qal1.de**.

Certificate:
0000025927_01 / 16 March 2012



Certification of AR500 with ER120 for NO₂, SO₂ and O₃ is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267:

Certificate No. 0000025927: 10 March 2010

Validity of the certificate: 11 February 2015

Test report: 936/21211350/A of 26 October 2010,
TÜV Rheinland Energie und Umwelt GmbH, Köln,

Publication: BAnz. 12 February 2010, No. 24, p.552, chapter III, Nr. 1.1:
Announcement by UBA from 25 January 2010

Supplementary testing according to EN 15267:

Certificate No. 0000025927_01:16 March 2012

Validity of the certificate: 11 February 2015

Test report: 936/21211350/B of 07 October 2011,
TÜV Rheinland Energie und Umwelt GmbH, Köln,

Publication: BAnz. 02 March 2012, No. 36, p. 920, chapter IV, No. 2.1,
Announcement by UBA from 23 February 2012.

Table 1: Total expanded uncertainty with the results of the laboratory test according to EN 14211 (Component NO₂) for system 1329

Device: AR600		Serial No: 1329		nmol/mol		
Component: NO ₂		1h-limit value: 104,6				
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty	
1	Repeatability at zero	≤ 1.0 nmol/mol	0,000	U _{r,z} 0,00	0,0000	
2	Repeatability at concentration ct	≤ 3.0 nmol/mol	2,000	U _{r,ph} 0,04	0,0015	
3	"lack of fit"	≤ 4,0% of measured value	0,800	U _{lfh} 0,48	0,2334	
4	Sensitivity coefficient of sample gas pressure	≤ 8,0 nmol/mol/kPa	0,000	U _{sp} 0,00	0,0000	
5	Sensitivity coefficient of sample gas temperature	≤ 3,0 nmol/mol/K	0,026	U _{gt} 0,04	0,0016	
6	Sensitivity coefficient of surrounding temperature	≤ 3,0 nmol/mol/K	-0,050	U _{gt} -0,06	0,0036	
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	-0,021	U _v -0,07	0,0046	
8a	H2O with concentration 21 nmol/mol	≤ 5.0 nmol/mol	0,000	U _{H2O} 0,00	0,0000	
8b	CO2 with concentration 500 µmol/mol	≤ 5.0 nmol/mol	0,001	U _{H2O,POS}		
8c	O3 with concentration 200 nmol/mol	≤ 2.0 nmol/mol	0,002	or	0,2304	
8d	NH3 with concentration 200 nmol/mol	≤ 5.0 nmol/mol	0,002	U _{H2O,NEG}		
9	Averaging effect	≤ 7,0% of measured value	-0,600	U _{av} -0,36	0,1313	
18	Difference sample/calibration port	≤ 1,0%	0,000	U _{asc} 0,00	0,0000	
21	Converter efficiency	≈ 98%	100,000	U _{CE} 0,00	0,0000	
22	Increase of NO2 concentration due to residence time	≤ 4.0 nmol/mol	0,000	U _{ctr} 0,00	0,0000	
23	Uncertainty calibration gas	≤ 3,0%	2,000	U _{cg} 1,05	1,0941	
combined standard uncertainty				U _c	1,3046	nmol/mol
expanded uncertainty				U _c	2,6092	nmol/mol
expanded uncertainty actual				U _{c,rel}	2,49	%
expanded uncertainty required				U _{req,rel}	15	%

Table 2: Total expanded uncertainty with the results of the laboratory test and field test according to EN 14211 (Component NO₂) for system 1329

Device: AR500		Serial No: 1329		1h-limit value: 104,6		nmol/mol	
Component: NO ₂		Performance characteristic		Criterion		Result	
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability at zero	≤ 1,0 nmol/mol	0,000	u _{r,z} 0,00	0,0000		
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	2,000	u _{r,th} not respected because: u _{r,th} = 0,075 < u _{r,f}	-		
3	"lack of fit"	≤ 4,0% of measured value	0,800	u _{lh} 0,48	0,2334		
4	Sensitivity coefficient of sample gas pressure	≤ 8,0 nmol/mol/kPa	0,000	u _{gp} 0,00	0,0000		
5	Sensitivity coefficient of sample gas temperature	≤ 3,0 nmol/mol/K	0,026	u _{gt} 0,04	0,0016		
6	Sensitivity coefficient of surrounding temperature	≤ 3,0 nmol/mol/K	-0,050	u _{st} -0,06	0,0036		
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	-0,021	u _v -0,07	0,0046		
8a	H2O with concentration 21 nmol/mol	≤ 5,0 nmol/mol	0,000	u _{H2O} 0,00	0,0000		
8b	CO2 with concentration 500 µmol/mol	≤ 5,0 nmol/mol	0,001	u _{CO2, pos}			
8c	O3 with concentration 200 nmol/mol	≤ 2,0 nmol/mol	0,002	or	0,2304		
8d	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,002	u _{NH3, neg}			
9	Averaging effect	≤ 7,0% of measured value	-0,600	u _{av} -0,36	0,1313		
10	Reproducibility under field conditions	≤ 5,0% of the average of 3 Mon.	4,720	u _{rf} 4,94	24,3752		
11	Long term drift at zero level	≤ 5,0 nmol/mol	-1,420	u _{gl,z} -0,82	0,6721		
12	Long term drift at span level	≤ 5,0% of max. of certification range	0,430	u _{gl,th} 0,26	0,0674		
18	Difference sample/calibration port	≤ 1,0%	0,000	u _{Dsc} 0,00	0,0000		
21	Converter efficiency	≥ 98%	100,000	u _{CE} 0,00	0,0000		
22	Increase of NO2 concentration due to residence time	≤ 4,0 nmol/mol	0,000	u _{ctr} 0,00	0,0000		
23	Uncertainty calibration gas	≤ 3,0%	2,000	u _{cg} 1,05	1,0941		
combined standard uncertainty				u _c	7,1546	nmol/mol	
expanded uncertainty				U _c	14,3093	nmol/mol	
expanded uncertainty actual				U _{c,rel}	13,68	%	
expanded uncertainty required				U _{req,rel}	15	%	

Table 3: Total expanded uncertainty with the results of the laboratory test according to EN 14211 (Component NO₂) for system 1330

Device: AR500		Serial No: 1330		nmol/mol	
Component: NO ₂		1h-limit value: 104,6			
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	u _{r,z} 0,02	0,0003
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,900	u _{r,ih} 0,02	0,0004
3	"lack of fit"	≤ 4,0% of measured value	0,000	u _{l,ih} 0,36	0,1313
4	Sensitivity coefficient of sample gas pressure	≤ 8,0 nmol/mol/kPa	0,000	u _{gp} 0,00	0,0000
5	Sensitivity coefficient of sample gas temperature	≤ 3,0 nmol/mol/K	0,000	u _{gt} -0,05	0,0025
6	Sensitivity coefficient of surrounding temperature	≤ 3,0 nmol/mol/K	0,000	u _{st} 0,00	0,0000
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	0,000	u _v 0,24	0,0553
8a	H2O with concentration 21 nmol/mol	≤ 5,0 nmol/mol	0,000	u _{H2O} 0,00	0,0000
8b	CO2 with concentration 500 µmol/mol	≤ 5,0 nmol/mol	0,000	u _{int,pos}	
8c	O3 with concentration 200 nmol/mol	≤ 2,0 nmol/mol	0,000	or	0,1764
8d	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,000	u _{int,neg}	
9	Averaging effect	≤ 7,0% of measured value	0,000	u _{av} -0,18	0,0328
18	Difference sampler/calibration port	≤ 1,0%	0,000	u _{asc} 0,00	0,0000
21	Converter efficiency	≥ 98%	100,000	u _{CE} 0,00	0,0000
22	Increase of NO2 concentration due to residence time	≤ 4,0 nmol/mol	0,000	u _{ct} 0,00	0,0000
23	Uncertainty calibration gas	≤ 3,0%	2,000	u _{cg} 1,05	1,0941
			combined standard uncertainty	u _c	1,2222
			expanded uncertainty	U _c	2,4445
			expanded uncertainty actual	U _{crel}	2,34
			expanded uncertainty required	U _{req,rel}	15

Table 4: Total expanded uncertainty with the results of the laboratory test and field test according to EN 14211 (Component NO₂) for system 1330

Device: AR500		Serial No: 1330		nmol/mol	
Component: NO ₂		1h-limit value: 104,6			
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	U _{r,z}	0,0003
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,900	U _{r,th} not respected because: u _{r,th} = 0,034 < u _{r,f}	-
3	"lack of fit"	≤ 4,0% of measured value	0,600	U _{r,th}	0,1313
4	Sensitivity coefficient of sample gas pressure	≤ 8,0 nmol/mol/kPa	0,000	U _{sp}	0,0000
5	Sensitivity coefficient of sample gas temperature	≤ 3,0 nmol/mol/K	-0,032	U _{gt}	0,0025
6	Sensitivity coefficient of surrounding temperature	≤ 3,0 nmol/mol/K	0,000	U _{st}	0,0000
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	0,073	U _v	0,0553
8a	H2O with concentration 21 nmol/mol	≤ 5,0 nmol/mol	0,000	U _{H2O}	0,0000
8b	CO2 with concentration 500 µmol/mol	≤ 5,0 nmol/mol	0,001	U _{int,pos}	
8c	O3 with concentration 200 nmol/mol	≤ 2,0 nmol/mol	0,002	or	0,1764
8d	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,000	U _{int,neg}	
9	Averaging effect	≤ 7,0% of measured value	-0,300	U _{av}	0,0328
10	Reproducibility under field conditions	≤ 5,0% of the average of 3 Mon.	4,720	U _{r,f}	24,3752
11	Long term drift at zero level	≤ 5,0 nmol/mol	1,620	U _{d,z}	0,8748
12	Long term drift at span level	≤ 5,0% of max. of certification range	0,500	U _{d,th}	0,0912
18	Difference sample/calibration port	≤ 1,0%	0,000	uDsc	0,0000
21	Converter efficiency	≥ 0,98	100,000	uCE	0,0000
22	Increase of NO2 concentration due to residence time	≤ 4,0 nmol/mol	0,000	uctr	0,0000
23	Uncertainty calibration gas	≤ 3,0%	2,000	ucg	1,0941
combined standard uncertainty				U _c	7,1661
expanded uncertainty				U _e	14,3121
expanded uncertainty actual				U _{e,rel}	13,68
expanded uncertainty required				U _{exp,rel}	15

Table 5: Total expanded uncertainty with the results of the laboratory test according to EN 14212 (Component SO₂) for system 1329

Device: AR500		Serial-No.: Gerät 1 (1329)		1h-limit value: 132		nmol/mol	
Component: SO ₂							
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	u _{r,z} 0,02	0,0003		
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,100	u _{r,v} 0,02	0,0003		
3	"lack of fit"	≤ 4,0% of measured value	1,600	u _{lv} 1,22	1,4868		
4	Sensitivity coefficient of sample gas pressure	≤ 3,0 nmol/mol/kPa	0,000	u _{gp} 0,00	0,0000		
5	Sensitivity coefficient of sample gas temperature	≤ 1,0 nmol/mol/K	0,071	u _{gt} 0,54	0,2908		
6	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	-0,030	u _{st} -0,23	0,0523		
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	-0,010	u _v -0,10	0,0103		
8a	H2O with concentration 21 nmol/mol	≤ 10 nmol/mol	0,000	u _{h2o} 0,00	0,0000		
8b	H2S with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,409	u _{h2s,pos}			
8c	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,406				
8d	NO with concentration 500 nmol/mol	≤ 5,0 nmol/mol	-0,604	or	0,1600		
8e	NO2 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,404				
8f	m-Xyol with concentration 1 µmol/mol	≤ 10 nmol/mol	1,421	u _{h2l,neg}			
9	Averaging effect	≤ 7,0% of measured value	-0,100	u _{av} -0,08	0,0058		
18	Difference sample/calibration port	≤ 1,0%	0,000	u _{bc} 0,00	0,0000		
23	Uncertainty calibration gas	≤ 3,0%	2,000	u _{cg} 1,32	1,7424		
combined standard uncertainty						U _c	1,9363
expanded standard uncertainty						U _c	3,8726
expanded uncertainty actual						U _{c,rel}	2,93
expanded uncertainty required						U _{req,rel}	15

Table 6: Total expanded uncertainty with the results of the laboratory test and field test according to EN 14212 (Component SO₂) for system 1329

Device: AR500		Serial-No.: Gerat 1 (1329)		132		nmol/mol	
Component: SO ₂		1h-limit value:					
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	u _{r,z}	0,02	0,0003	
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,100	u _{r,lv}	not respected because, u _{r,lv} = 0,01 < u _{r,f}	-	
3	"lack of fit"	≤ 4,0% of measured value	1,600	u _{lv}	1,22	1,4888	
4	Sensitivity coefficient of sample gas pressure	≤ 3,0 nmol/mol/kPa	0,000	u _{gp}	0,00	0,0000	
5	Sensitivity coefficient of sample gas temperature	≤ 1,0 nmol/mol/K	0,071	u _{gt}	0,54	0,2908	
6	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	-0,030	u _{st}	-0,23	0,0523	
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	-0,010	u _v	-0,10	0,0103	
8a	H2O with concentration 21 nmol/mol	≤ 10 nmol/mol	0,000	u _{h2o}	0,00	0,0000	
8b	H2S with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,409	u _{h2s,pos}			
8c	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,406				
8d	NO with concentration 500 nmol/mol	≤ 5,0 nmol/mol	-0,604				
8e	NO2 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,404	or	0,40	0,1600	
8f	m-Xylol with concentration 1 µmol/mol	≤ 10 nmol/mol	1,421	u _{h2n,neg}			
9	Averaging effect	≤ 7,0% of measured value	-0,100	u _{av}	-0,08	0,0038	
10	Reproducibility under field conditions	≤ 5,0% of the average of 3 Mon.	4,830	u _{r,f}	6,38	40,6483	
11	Long term drift at zero level	≤ 5,0 nmol/mol	-0,920	u _{g,z}	-0,53	0,2821	
12	Long term drift at span level	≤ 5,0% of max. of certification range	1,490	u _{g,lv}	1,14	1,2894	
18	Differenz Proben-/Kalibriergaseingang	≤ 1,0%	0,000	u _{bsc}	0,00	0,0000	
23	Unsicherheit Prüfgas	≤ 3,0%	2,000	u _{cg}	1,32	1,7424	
combined standard uncertainty						u _c	6,7800
expanded standard uncertainty						U _c	13,5600
expanded uncertainty actual						U _{c,rel}	10,27
expanded uncertainty required						U _{req,rel}	15

Table 7: Total expanded uncertainty with the results of the laboratory test according to EN 14212 (Component SO₂) for system 1330

Device: AR500		Serial-No.: Gerät 2 (1330)	132	nmol/mol
Component: SO ₂		1h-limit value:		
No.	Performance characteristic	Criterion	Result	Square of uncertainty
1	Repeatability at zero	≤ 1,0 nmol/mol	0,000	0,0000
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,100	0,0003
3	"lack of fit"	≤ 4,0% of measured value	1,400	1,1384
4	Sensitivity coefficient of sample gas pressure	≤ 3,0 nmol/mol/kPa	0,000	0,0000
5	Sensitivity coefficient of sample gas temperature	≤ 1,0 nmol/mol/K	0,011	0,0070
6	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	-0,060	0,2091
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	0,010	0,0103
8a	H2O with concentration 21 nmol/mol	≤ 10 nmol/mol	0,000	0,0000
8b	H2S with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,503	1,5129
8c	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,203	
8d	NO with concentration 500 nmol/mol	≤ 5,0 nmol/mol	0,202	
8e	NO2 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,401	
8f	m-Xylol with concentration 1 µmol/mol	≤ 10 nmol/mol	0,809	
9	Averaging effect	≤ 7,0% of measured value	0,100	
18	Difference sample/calibration port	≤ 1,0%	0,000	0,0000
23	Uncertainty calibration gas	≤ 3,0%	2,000	1,7424
			combined standard uncertainty	u _c
			expanded uncertainty	2,1509 nmol/mol
			expanded uncertainty actual	U _c
			expanded uncertainty required	4,3017 nmol/mol
				3,26 %
				15 %

Table 8: Total expanded uncertainty with the results of the laboratory test and field test according to EN 14212 (Component SO₂) for system 1330

Device: AR500		Serial-No.: Gerät 2 (1330)		1h-limit value: 132		nmol/mol	
Component: SO ₂		Performance characteristic		Criterion	Result	Uncertainty	Square of uncertainty
1	Repeatability at zero	≤	1,0 nmol/mol	0,000	u _{r,z}	0,00	0,0000
2	Repeatability at concentration ct	≤	3,0 nmol/mol	0,100	u _{r,v}	not respected because, u _{r,v} = 0,01 < u _{r,f}	-
3	"lack of fit"	≤	4,0% of measured value	1,400	u _{lv}	1,07	1,1384
4	Sensitivity coefficient of sample gas pressure	≤	3,0 nmol/mol/kPa	0,000	u _{gp}	0,00	0,0000
5	Sensitivity coefficient of sample gas temperature	≤	1,0 nmol/mol/K	0,011	u _{gt}	0,08	0,0070
6	Sensitivity coefficient of surrounding temperature	≤	1,0 nmol/mol/K	-0,060	u _{st}	-0,46	0,2091
7	Sensitivity coefficient of electrical voltage	≤	0,30 nmol/mol/V	0,010	u _v	0,10	0,0103
8a	H2O with concentration 21 nmol/mol	≤	10 nmol/mol	0,000	u _{h2o}	0,00	0,0000
8b	H2S with concentration 200 nmol/mol	≤	5,0 nmol/mol	0,503	u _{h2s,pos}		
8c	NH3 with concentration 200 nmol/mol	≤	5,0 nmol/mol	0,203	or	1,23	1,5129
8d	NO with concentration 500 nmol/mol	≤	5,0 nmol/mol	0,202			
8e	NO2 with concentration 200 nmol/mol	≤	5,0 nmol/mol	0,401			
8f	m-Xylol with concentration 1 µmol/mol	≤	10 nmol/mol	0,809	u _{h2c,neg}		
9	Averaging effect	≤	7,0% of measured value	0,100	u _{av}	0,08	0,0058
10	Reproducibility under field conditions	≤	5,0% of the average of 3 Mon.	4,830	u _{rf}	6,38	40,6483
11	Long term drift at zero level	≤	5,0 nmol/mol	1,160	u _{l,z}	0,67	0,4485
12	Long term drift at span level	≤	5,0% of max. of certification range	-2,070	u _{l,v}	-1,58	2,4887
18	Differenz Proben-/Kalibrierungsgang	≤	1,0%	0,000	u _{bsc}	0,00	0,0000
23	Unsicherheit Prüfgas	≤	3,0%	2,000	0	1,32	1,7424
				combined standard uncertainty	u _c	u _c	6,9434
				expanded uncertainty	U _c	U _c	13,8869
				expanded uncertainty actual	U _{c,rel}	U _{c,rel}	10,52
				expanded uncertainty required	U _{req,rel}	U _{req,rel}	15

Table 9: Total expanded uncertainty with the results of the laboratory test according to EN 14625 (Component O₃) for system 1329

Device: AR500		Serial No. Gerät 1 (1329)		Measured component: O ₃		120 nmol/mol	
		hourly alert threshold					
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability standard deviation at zero	1,0 nmol/mol	0,200	u _{r,z} 0,04	0,0013		
2	Repeatability standard deviation at ct	3,0 nmol/mol	0,600	u _{r,iv} 0,11	0,0120		
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	0,400	u _{iv} 0,28	0,0768		
4	Variations in sample gas pressure	2,0 nmol/mol/kPa	0,000	u _{gp} 0,00	0,0000		
5	Variations in sample gas temperature	1,0 nmol/mol/K	0,014	u _{gt} 0,15	0,0212		
6	Variations in surrounding temperature	1,0 nmol/mol/K	0,150	u _{st} 0,52	0,2700		
7	Variations in electrical voltage	0,30 nmol/mol/V	-0,010	u _v -0,12	0,0147		
8a	Interference H2O with 21 nmol/mol	10 nmol/mol	0,000	u _{H2O} 0,00	0,0000		
8b	Interference Toluol with 0,5 µmol/mol	5,0 nmol/mol	2,147	u _{int,pos} 1,47	2,1573		
8c	Interference Xylol with 0,5 µmol/mol	5,0 nmol/mol	0,397	or u _{int,neg}			
9	Averaging effect	7,0% of measured value	0,200	u _{av} 0,14	0,0192		
18	Difference sample/calibration port	1,0%	0,000	u _{disc} 0,00	0,0000		
23	Uncertainty test gas	3,0%	2,000	ucg 1,20	1,4400		
Combined standard uncertainty				u _c	2,0031	nmol/mol	
Expanded uncertainty				U _c	4,0062	nmol/mol	
Expanded uncertainty actual				U _{c,rel}	3,34	%	
Expanded uncertainty required				U _{req,rel}	15	%	

Table 10 Total expanded uncertainty with the results of the laboratory test and field test according to EN 14625 (Component O₃) for system 1329

Device: AR500		Serial No. Gerät 1 (1329)		120		
Measured component: O ₃		hourly alert threshold		nmol/mol		
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty	
1	Repeatability standard deviation at zero	1,0 nmol/mol	0,200	u _{r,z}	0,0013	
2	Repeatability standard deviation at ct	3,0 nmol/mol	0,600	u _{r,iv}	-	
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	0,400	u _{iv}	0,0768	
4	Variations in sample gas pressure	2,0 nmol/mol/kPa	0,000	u _{gp}	0,0000	
5	Variations in sample gas temperature	1,0 nmol/mol/K	0,014	u _{gt}	0,0212	
6	Variations in surrounding temperature	1,0 nmol/mol/K	0,150	u _{st}	0,2700	
7	Variations in electrical voltage	0,30 nmol/mol/V	-0,010	u _v	0,0147	
8a	Interference H2O with 21 nmol/mol	10 nmol/mol	0,000	u _{H2O}	0,0000	
8b	Interference Toluol with 0,5 µmol/mol	5,0 nmol/mol	2,147	u _{tol,pos}	2,1573	
8c	Interference Xylol with 0,5 µmol/mol	5,0 nmol/mol	0,397	u _{tol,reg} or u _{tol,reg}		
9	Averaging effect	7,0% of measured value	0,200	u _{av}	0,0192	
10	Reproducibility standard deviation in field	5,0% of average of 3 month	2,410	u _{r,f}	8,3637	
11	Long term drift at zero	5,0 nmol/mol	1,460	u _{cl,z}	0,7105	
12	Long term drift at span level	5,0% of range	-2,450	u _{cl,iv}	2,8812	
18	Difference sample/calibration port	1,0%	0,000	u _{Dsc}	0,0000	
23	Uncertainty test gas	3,0%	2,000	ucg	1,4400	
				Combined standard uncertainty	u _c	3,9945
				Expanded uncertainty	U _c	7,9890
				Expanded uncertainty actual	U _{c,rel}	6,66
				Expanded uncertainty required	U _{req,rel}	15

Table 11: Total expanded uncertainty with the results of the laboratory test according to EN 14625 (Component O₃) for system 1330

Device:		AR500	Serial No.	Gerät 2 (1330)	120	nmol/mol
Measured component:		O ₃	hourly alert threshold			
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty	
1	Repeatability standard deviation at zero	1,0 nmol/mol	0,200	u _{r,z} 0,04	0,0013	
2	Repeatability standard deviation at ct	3,0 nmol/mol	0,400	u _{r,v} 0,07	0,0053	
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	-0,300	u _{lv} -0,21	0,0432	
4	Variations in sample gas pressure	2,0 nmol/mol/kPa	0,000	u _{gp} 0,00	0,0000	
5	Variations in sample gas temperature	1,0 nmol/mol/K	0,007	u _{gt} 0,07	0,0053	
6	Variations in surrounding temperature	1,0 nmol/mol/K	-0,120	u _{st} -0,42	0,1728	
7	Variations in electrical voltage	0,30 nmol/mol/V	0,010	u _v 0,12	0,0147	
8a	Interference H2O with 21 mmol/mol	10 nmol/mol	0,000	u _{H2O} 0,00	0,0000	
8b	Interference Toluol with 0,5 µmol/mol	5,0 nmol/mol	0,396	u _{int,pos} 1,72	2,9416	
8c	Interference Xylol with 0,5 µmol/mol	5,0 nmol/mol	2,574	or u _{int,neg}		
9	Averaging effect	7,0% of measured value	-0,900	u _{av} -0,62	0,3888	
18	Difference sample/calibration port	1,0%	0,000	u _{psc} 0,00	0,0000	
23	Uncertainty test gas	3,0%	2,000	0	1,4400	
		Combined standard uncertainty		u _c		nmol/mol
		Expanded uncertainty		u _c		nmol/mol
		Expanded uncertainty actual		u _{c,rel}	3,73	%
		Expanded uncertainty required		u _{req,rel}	15	%

Table 12 Total expanded uncertainty with the results of the laboratory test and field test according to EN 14625 (Component O₃) for system 1330

Device: AR500		Serial No. Gerat 2 (1330)		120		nmol/mol	
Measured component: O ₃		hourly alert threshold					
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability standard deviation at zero	1.0 nmol/mol	0.200	u _{r,z}	0.0013		
2	Repeatability standard deviation at ct	3.0 nmol/mol	0.400	u _{r,lv}	-		
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	-0.300	u _{lv}	0.0432		
4	Variations in sample gas pressure	2.0 nmol/mol/kPa	0.000	u _{gp}	0.0000		
5	Variations in sample gas temperature	1.0 nmol/mol/K	0.007	u _{gt}	0.0053		
6	Variations in surrounding temperature	1.0 nmol/mol/K	-0.120	u _{st}	0.1728		
7	Variations in electrical voltage	0.30 nmol/mol/V	0.010	u _v	0.0147		
8a	Interference H2O with 21 nmol/mol	10 nmol/mol	0.000	u _{h2o}	0.0000		
8b	Interference Toluol with 0.5 µmol/mol	5.0 nmol/mol	0.396	u _{tol,pos}	2,9416		
8c	Interference Xylol with 0.5 µmol/mol	5.0 nmol/mol	2.574	Of u _{tol,neg}			
9	Averaging effect	7,0% of measured value	-0.900	u _{av}	0.3888		
10	Reproducibility standard deviation in field	5,0% of average of 3 month	2.410	u _{r,f}	8,3637		
11	Long term drift at zero	5.0 nmol/mol	-1.840	u _{dl,z}	1,1285		
12	Long term drift at span level	5.0% of range	2.900	u _{dl,lv}	4,0368		
18	Difference sample/calibration port	1.0%	0.000	u _{Dsc}	0.0000		
23	Uncertainty test gas	3.0%	2.000	0	1,4000		
				Combined standard uncertainty	u _c	4,3054	nmol/mol
				Expanded uncertainty	U _c	8,6109	nmol/mol
				Expanded uncertainty actual	U _{c,rel}	7,18	%
				Expanded uncertainty required	U _{req,rel}	15	%