

CERTIFICATE

of Product Conformity (QAL1)

Certificate number: 0000028730_01

Certified AMS: GCS for HF, N₂O, CO, NO, NO₂, SO₂, HCl, NH₃, H₂O and CO₂

Manufacturer: General Impianti S.r.l.
Via Collefreddo 8/9
60030 Maiolati Spontini (AN)
Italy

Test Institute: TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested and certified
according to the standards**

**EN 15267-1 (2009), EN 15267-2 (2009), EN 15267-3 (2007)
and EN 14181 (2004)**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 16 pages).



Suitability Tested
EN 15267
QAL1 Certified
Regular
Surveillance


www.tuv.com
ID 0000028730

Publication in the German Federal Gazette
(BAnz.) of 29 July 2011

German Federal Environment Agency
Dessau, 22 July 2016

This certificate will expire on:
28 July 2021

TÜV Rheinland Energy GmbH
Cologne, 21 July 2016



Dr. Marcel Langner
Head of Section II 4.1



ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. + 49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00

Certificate:
0000028730_01 / 22 July 2016

Test report: 936/21211855/B of 25 March 2011
Initial certification: 29 July 2011
Expiry date: 28 July 2021
Certificate: renewal (previous certificate 0000028730 dated from 19 August 2011 with validity up to the 28 July 2016)
Publication: BAnz. 29 July 2011, No. 113, page 2725, chapter I No. 4.3

Approved application

The tested AMS is suitable for use at combustion plants according to Directive 2010/75/EU, chapter III (13. BImSchV), at waste incineration plants according to Directive 2010/75/EU, chapter IV (17. BImSchV) and other plants requiring official approval. The measured ranges have been selected considering the wide application range of the AMS.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a more than 12 months field test at a waste incineration plant. For the components HF and N₂O the field test was done at a tunnel kiln plant for firing refractory and acid proof bricks.

The AMS is approved for an ambient temperature range of +5 °C to +40 °C.

The notification of suitability of the AMS, performance testing, and the uncertainty calculation have been effected on the basis of the regulations valid at the time of performance testing. As changes in legal regulations are possible, any potential user should ensure that this AMS is suitable for monitoring the limit value relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Basis of the certification

This certification is based on:

- test report 936/21211855/B of 25 March 2011 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz. 29 July 2011, No 113, p. 2725, chapter I No. 4.3, Announcement by UBA from 15 July 2011:

AMS designation:

GIGAS 10M for HF, N₂O, CO, NO, NO₂, SO₂, HCl, NH₃, H₂O and CO₂

Manufacturer:

General Impianti S.r.l., Moie di Maiolati, Italy

Field of application:

For measurements at plants requiring official approval and plants according to 27th BImSchV

Measuring ranges during the performance test:

Component	Certification range	Supplementary measurement ranges	Unit
HF	0 - 5	0 - 10, 0 - 20	mg/m ³
N ₂ O	0 - 50	0 - 1000	mg/m ³
CO	0 - 75	0 - 300	mg/m ³
SO ₂	0 - 75	0 - 300	mg/m ³
NO	0 - 200	0 - 400	mg/m ³
NO ₂	0 - 100	0 - 200	mg/m ³
HCl	0 - 15	0 - 90	mg/m ³
NH ₃	0 - 15	-	mg/m ³
CO ₂	0 - 20	-	Vol.-%
H ₂ O	0 - 30	-	Vol.-%

Software versions:

Omnisc 7.2 GasCalc: 4.4

Restriction:

The measurement system shall only be operated at plants waste gas humidity does not constantly exceed 30 Vol.-%.

Notes:

1. Wet test gases shall be used for the testing of HF, HCl and NH₃.
2. A six month period has been determined as maintenance interval.
3. Supplementary testing (including the components N₂O and HF, instrument changes and conversion of test results to standard EN 15267-3) on the announcements of the Federal Environment Agency on 12 August 2008 (BAnz. p. 3243, chapter I No 2.3).
4. For the measuring component CO the requirement for the total uncertainty according to EN 15267-3 is not fulfilled.
5. The measuring unit works with wet process gases.

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Cologne
Report No.: 936/21211855/B of 25 March 2011

Publication in the German Federal Gazette: BAnz. 02 March 2012, No. 36, p. 920, chapter V notice 8, Announcement by UBA from 23 February 2012:

8 Notification as regards Federal Environment Agency (UBA) notices of 15 July 2011 (Federal Gazette BAnz. p. 2725, chapter I, number 4.3)

The new name of the GIGAS 10M multi-component measuring system manufactured by General Impianti s.r.l. is:

GCS

Statement of TÜV Rheinland Energie und Umwelt GmbH of 29 September 2011

Publication in the German Federal Gazette: BAnz AT 14.03.2016 B7, chapter V notice 23, Announcement by UBA from 18 February 2016:

23 Notification as regards Federal Environment Agency (UBA) notices of 15 July 2011 (BAnz. page 2725, chapter I number 4.3) and of 23 February 2012 (BAnz. page 920, chapter V notification 8 as well as chapter VI correction 1)

The measuring equipment GCS for CO, NO, NO₂, N₂O, SO₂, HCl, HF, NH₃, H₂O and CO₂ of General Impianti s.r.l. can be installed in cabinet ETA mod. ENUX-121808PR also. For this cabinet the air conditioning system of Kelvin Jet 20 in the case of wall mounting and Stulz 20002207000 at roof mounting can be used.

The new address of the manufacturer is Via Collefreddo 8/9 - 60030 Maiolati Spontini (AT), Italy.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 21 October 2015

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The GCS measuring system is an extractive multiple-component measuring system based on the measuring principle of FTIR spectrometry which measures at high temperatures. It comprises the main components as described below:

Sampling

Sampling probe: General Impianti GL – SRPF (180 °C) coated
Sampling tube: RACO (Length during suitability testing approx. 10 m,
heated to 180 °C)
Heated filter: M&C – FT-H2 (180 °C)

Analyser

FTIR: GIGAS 10M , temperature of the cuvette: 180 °C

Sample gas post-treatment

The following components are installed after the sample gas outlet:

Sample gas cooler: General Impianti – FRIGO GI PELLTIER R
Sample gas pump: KNF – N.814.KTE

Control modules

DAQ module: GL-AnDe
Omron module: GL-TPReg

Calculator

Standard PC of the following minimum requirements:

Operating system: MS Windows XP
Processor: Intel Pentium III, 1 GHz
Primary storage: 512 MB
Hard disk: 40 GB
Interfaces: USB Interface
Network interface RJ 45
Serial Interface RS 232

A Siemens Industry PC with 17" Touch Screen Display has been installed during the suitability test.

Software

Evaluation-Software: GasCalc 4.4 and Omnic 7.2

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certification mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: qal1.de.

Certification of GCS for HF, N₂O, CO, NO, NO₂, SO₂, HCl, NH₃, H₂O and CO₂ is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Basic test

Test report: 936/21206517/A from 08 July 2007
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne
Publication: BAnz. 06 November 2007, No 206, p. 7925, chapter I No 2.1
UBA announcement from 23 September 2007

Test report: 936/21206517/B from 09 November 2007
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne
Publication: BAnz. 07 March 2008, No 38, p. 901, chapter I No 2.3
UBA announcement from 14 February 2008

Test report: 936/21206517/C from 27 February 2008
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne
Publication: BAnz. 03 September 2008, No 133, p. 3242, chapter I No 2.3
UBA announcement from 12 August 2008

Notification

Statement of TÜV Rheinland Energie und Umwelt GmbH of 12 November 2010
Publication: BAnz. 26 January 2011, No 14, p. 294, chapter IV notification 29
UBA announcement from 10 January 2011
(Software changing)

Initial certification according to EN 15267

Certificate No. 0000028730: 19 August 2011
Expiry date of the certificate: 28 July 2011
Test report: 936/21211855/B of 25 March 2011
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAnz. 29 July 2011, No 113, p. 2725, chapter I No 4.3
Announcement by UBA from 15 July 2011

Notification according to EN 15267

Statement of TÜV Rheinland Energie und Umwelt GmbH of 29. September 2011
Publication: BAnz. 2 March 2012, No. 36, p. 920, chapter VI correction 1
Announcement by UBA from 23 February 2012
(correction certification range of NH₃)

Statement of TÜV Rheinland Energie und Umwelt GmbH of 29 September 2011
Publication: BAnz. 2 March 2012, No. 36, p. 920, chapter V notification 8
Announcement by UBA from 23 February 2012
(changing name of AMS)

Statement of TÜV Rheinland Energie und Umwelt GmbH of 21 October 2015
Publication: BAnz AT 14.03.2016 B7, chapter V notification 23,
Announcement by UBA from 18 February 2016
(hardware addition)

Renewal of the certificate

Certificate No. 0000028730_01: 22 July 2016
Expiry date of the certificate: 28 July 2021

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	RSE09/TUV/H1 / RSE09/TUV/H2
Measuring principle	FTIR

Test report

Test laboratory	TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	HF	0 - 5 mg/m ³
---------------------	----	-------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.20 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	0.15 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	0.20 mg/m ³
Uncertainty of cross sensitivity	0.12 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.080 mg/m ³	0.006 (mg/m ³) ²
Lack of fit	u _{lof} -0.052 mg/m ³	0.003 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.066 mg/m ³	0.004 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.084 mg/m ³	0.007 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.051 mg/m ³	0.003 (mg/m ³) ²
Influence of supply voltage	u _v 0.029 mg/m ³	0.001 (mg/m ³) ²
Cross sensitivity (interference)	u _i 0.115 mg/m ³	0.013 (mg/m ³) ²
Influence of sample gas flow	u _p 0.046 mg/m ³	0.002 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.040 mg/m ³	0.002 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 0.20 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 0.40 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 2 mg/m³ 19.9

U in % of the ELV 2 mg/m³ 40.0

U in % of the ELV 2 mg/m³ 30.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	RSE09/TUV/H1 / RSE09/TUV/H2
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	N ₂ O 0 - 50 mg/m ³
---------------------	--

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.31 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	0.93 mg/m ³
Sum of negative CS at reference point	-1.98 mg/m ³
Maximum sum of cross sensitivities	-1.98 mg/m ³
Uncertainty of cross sensitivity	-1.14 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.100 mg/m ³	0.010 (mg/m ³) ²
Lack of fit	u _{lof} 0.231 mg/m ³	0.053 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.231 mg/m ³	0.053 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.808 mg/m ³	0.653 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.321 mg/m ³	0.103 (mg/m ³) ²
Influence of supply voltage	u _v 0.128 mg/m ³	0.016 (mg/m ³) ²
Cross sensitivity (interference)	u _i -1.143 mg/m ³	1.307 (mg/m ³) ²
Influence of sample pressure	u _p 0.225 mg/m ³	0.051 (mg/m ³) ²
Influence of sample gas flow	u _q 0.000 mg/m ³	0.000 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.404 mg/m ³	0.163 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 1.55 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c \cdot k = u_c \cdot 1.96 \quad 3.04 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 20 mg/m³ **15.2**

Requirement of 2000/76/EC and 2001/80/EC

U in % of the ELV 20 mg/m³ **20.0 ****

Requirement of EN 15267-3

U in % of the ELV 20 mg/m³ 15.0

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

A value of 20.0 % was used for this.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	CO 0 - 75 mg/m ³
---------------------	--------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.41 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	3.00 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	3.00 mg/m ³
Uncertainty of cross sensitivity	1.732 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.407 mg/m ³	0.166 (mg/m ³) ²
Lack of fit	u _{lof} -0.404 mg/m ³	0.163 (mg/m ³) ²
Zero drift from field test	u _{d,z} -0.476 mg/m ³	0.227 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.996 mg/m ³	0.992 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.321 mg/m ³	0.103 (mg/m ³) ²
Influence of supply voltage	u _v 0.093 mg/m ³	0.009 (mg/m ³) ²
Cross sensitivity (interference)	u _i 1.732 mg/m ³	3.000 (mg/m ³) ²
Influence of sample gas flow	u _p 0.433 mg/m ³	0.187 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.606 mg/m ³	0.368 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max, j})^2} \quad 2.28 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 4.48 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

U in % of the ELV 50 mg/m³ 9.0

Requirement of EN 15267-3

U in % of the ELV 50 mg/m³ 10.0

U in % of the ELV 50 mg/m³ 7.5

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	NO 0 - 200 mg/m ³
---------------------	---------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 mg/m ³
Sum of negative CS at zero point	-6.80 mg/m ³
Sum of positive CS at reference point	2.60 mg/m ³
Sum of negative CS at reference point	-5.20 mg/m ³
Maximum sum of cross sensitivities	-6.80 mg/m ³
Uncertainty of cross sensitivity	-3.926 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 1.782 mg/m ³	3.176 (mg/m ³) ²
Lack of fit	u _{lof} 1.155 mg/m ³	1.334 (mg/m ³) ²
Zero drift from field test	u _{d,z} -0.808 mg/m ³	0.653 (mg/m ³) ²
Span drift from field test	u _{d,s} -3.002 mg/m ³	9.012 (mg/m ³) ²
Influence of ambient temperature at span	u _t 1.650 mg/m ³	2.723 (mg/m ³) ²
Influence of supply voltage	u _v 0.513 mg/m ³	0.263 (mg/m ³) ²
Cross sensitivity (interference)	u _i -3.926 mg/m ³	15.413 (mg/m ³) ²
Influence of sample gas flow	u _p 1.155 mg/m ³	1.334 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 1.617 mg/m ³	2.613 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	6.04 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	11.84 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC	U in % of the ELV 130.4 mg/m³	9.1
Requirement of EN 15267-3	U in % of the ELV 130.4 mg/m³	20.0
	U in % of the ELV 130.4 mg/m³	15.0

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	NO ₂ 0 - 100 mg/m ³
---------------------	--

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	3.99 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	3.50 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	3.99 mg/m ³
Uncertainty of cross sensitivity	2.304 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.864 mg/m ³	0.746 (mg/m ³) ²
Lack of fit	u _{lof} 0.924 mg/m ³	0.854 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.346 mg/m ³	0.120 (mg/m ³) ²
Span drift from field test	u _{d,s} -1.559 mg/m ³	2.430 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.306 mg/m ³	0.094 (mg/m ³) ²
Influence of supply voltage	u _v 0.289 mg/m ³	0.084 (mg/m ³) ²
Cross sensitivity (interference)	u _i 2.304 mg/m ³	5.307 (mg/m ³) ²
Influence of sample gas flow	u _p 0.577 mg/m ³	0.333 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.808 mg/m ³	0.653 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	3.26 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	6.39 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC	U in % of the ELV 60 mg/m³	10.6
Requirement of EN 15267-3	U in % of the ELV 60 mg/m³	20.0
	U in % of the ELV 60 mg/m³	15.0

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	SO ₂ 0 - 75 mg/m ³
---------------------	---

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.89 mg/m ³
Sum of negative CS at zero point	-0.53 mg/m ³
Sum of positive CS at reference point	3.00 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	3.00 mg/m ³
Uncertainty of cross sensitivity	1.732 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.263 mg/m ³	0.069 (mg/m ³) ²
Lack of fit	u _{lof} -0.572 mg/m ³	0.327 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.563 mg/m ³	0.317 (mg/m ³) ²
Span drift from field test	u _{d,s} 1.212 mg/m ³	1.469 (mg/m ³) ²
Influence of ambient temperature at span	u _t 1.664 mg/m ³	2.769 (mg/m ³) ²
Influence of supply voltage	u _v 0.179 mg/m ³	0.032 (mg/m ³) ²
Cross sensitivity (interference)	u _i 1.732 mg/m ³	3.000 (mg/m ³) ²
Influence of sample gas flow	u _p 0.433 mg/m ³	0.187 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.606 mg/m ³	0.368 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{\max, j})^2} \quad 2.92 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c \cdot k = u_c \cdot 1.96 \quad 5.73 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 50 mg/m³ **11.5**

U in % of the ELV 50 mg/m³ **20.0**

U in % of the ELV 50 mg/m³ 15.0

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	HCl 0 - 15 mg/m ³
---------------------	---------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.49 mg/m ³
Sum of negative CS at zero point	-0.61 mg/m ³
Sum of positive CS at reference point	0.60 mg/m ³
Sum of negative CS at reference point	-0.15 mg/m ³
Maximum sum of cross sensitivities	-0.61 mg/m ³
Uncertainty of cross sensitivity	-0.350 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.144 mg/m ³	0.021 (mg/m ³) ²
Lack of fit	u _{lof} -0.104 mg/m ³	0.011 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.251 mg/m ³	0.063 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.251 mg/m ³	0.063 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.186 mg/m ³	0.035 (mg/m ³) ²
Influence of supply voltage	u _v 0.026 mg/m ³	0.001 (mg/m ³) ²
Cross sensitivity (interference)	u _i -0.350 mg/m ³	0.122 (mg/m ³) ²
Influence of sample gas flow	u _p 0.087 mg/m ³	0.008 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.121 mg/m ³	0.015 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.58 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	1.14 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC	U in % of the ELV 10 mg/m³	11.4
Requirement of EN 15267-3	U in % of the ELV 10 mg/m³	40.0
	U in % of the ELV 10 mg/m³	30.0

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	NH ₃ 0 - 15 mg/m ³
---------------------	---

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.52 mg/m ³
Sum of negative CS at zero point	-0.27 mg/m ³
Sum of positive CS at reference point	0.60 mg/m ³
Sum of negative CS at reference point	-0.15 mg/m ³
Maximum sum of cross sensitivities	0.60 mg/m ³
Uncertainty of cross sensitivity	0.346 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.086 mg/m ³	0.007 (mg/m ³) ²
Lack of fit	u _{lof} 0.165 mg/m ³	0.027 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.147 mg/m ³	0.022 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.251 mg/m ³	0.063 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.173 mg/m ³	0.030 (mg/m ³) ²
Influence of supply voltage	u _v 0.017 mg/m ³	0.000 (mg/m ³) ²
Cross sensitivity (interference)	u _i 0.346 mg/m ³	0.120 (mg/m ³) ²
Influence of sample gas flow	u _b 0.087 mg/m ³	0.008 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.121 mg/m ³	0.015 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 0.54 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c \cdot k = u_c \cdot 1.96 \quad 1.06 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 10 mg/m³ 10.6

Requirement of 2000/76/EC and 2001/80/EC

U in % of the ELV 10 mg/m³ 40.0 **

Requirement of EN 15267-3

U in % of the ELV 10 mg/m³ 30.0

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

A value of 40.0 % was used for this.

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	CO ₂ 0 - 20 Vol.-%
---------------------	----------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 Vol.-%
Sum of negative CS at zero point	0.00 Vol.-%
Sum of positive CS at reference point	0.00 Vol.-%
Sum of negative CS at reference point	0.00 Vol.-%
Maximum sum of cross sensitivities	0.00 Vol.-%
Uncertainty of cross sensitivity	0.000 Vol.-%

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.067 Vol.-%	0.004 (Vol.-%) ²
Lack of fit	u _{lof} -0.104 Vol.-%	0.011 (Vol.-%) ²
Zero drift from field test	u _{d,z} -0.058 Vol.-%	0.003 (Vol.-%) ²
Span drift from field test	u _{d,s} -0.231 Vol.-%	0.053 (Vol.-%) ²
Influence of ambient temperature at span	u _t 0.252 Vol.-%	0.064 (Vol.-%) ²
Influence of supply voltage	u _v 0.026 Vol.-%	0.001 (Vol.-%) ²
Cross sensitivity (interference)	u _i 0.000 Vol.-%	0.000 (Vol.-%) ²
Influence of sample gas flow	u _b 0.115 Vol.-%	0.013 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.162 Vol.-%	0.026 (Vol.-%) ²

* The larger value is used :

"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.42 Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.82 Vol.-%

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the range 20 Vol.-%	4.1
U in % of the range 20 Vol.-%	10.0 **
U in % of the range 20 Vol.-%	7.5

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

A value of 10.0 % was used for this.

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	H ₂ O 0 - 30 Vol.-%
---------------------	-----------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 Vol.-%
Sum of negative CS at zero point	0.00 Vol.-%
Sum of positive CS at reference point	0.00 Vol.-%
Sum of negative CS at reference point	0.00 Vol.-%
Maximum sum of cross sensitivities	0.00 Vol.-%
Uncertainty of cross sensitivity	0.000 Vol.-%

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.208 Vol.-%	0.043 (Vol.-%) ²
Lack of fit	u _{lof} -0.173 Vol.-%	0.030 (Vol.-%) ²
Zero drift from field test	u _{d,z} -0.017 Vol.-%	0.000 (Vol.-%) ²
Span drift from field test	u _{d,s} 0.468 Vol.-%	0.219 (Vol.-%) ²
Influence of ambient temperature at span	u _t 0.172 Vol.-%	0.030 (Vol.-%) ²
Influence of supply voltage	u _v 0.015 Vol.-%	0.000 (Vol.-%) ²
Cross sensitivity (interference)	u _i 0.000 Vol.-%	0.000 (Vol.-%) ²
Influence of sample gas flow	u _b 0.173 Vol.-%	0.030 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.242 Vol.-%	0.059 (Vol.-%) ²

* The larger value is used :

"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.64 Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	1.26 Vol.-%

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the range 30 Vol.-%	4.2
U in % of the range 30 Vol.-%	10.0 **
U in % of the range 30 Vol.-%	7.5

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

A value of 10.0 % was used for this.

*** and RSE09/TUV/H1 / RSE09/TUV/H2