

CONFIRMATION

Notification: 0000040207_03_01
on changes according to EN 15267 regarding certificate 0000040207_03 dated 02 June 2021

AMS designation: Spirant BAM 1100 mit PM_{2,5}-Vorabscheider for PM_{2,5}

Manufacturer: ACOEM Australasia (Ecotech Pty Ltd)
1492 Ferntree Gully Road,
Knoxfield, VIC, 3180
Australia

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**62 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 6.1) and
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 8)**


The company name of Ecotech Pty. Ltd. has changed to ACOEM Australasia.


The Spirant BAM 1100 measuring system with PM_{2,5} pre-separator for suspended particulate matter PM_{2,5} from the company ACOEM Australasia will have a touch-sensitive screen in the future and a USB port will now be installed in the front panel. On the rear panel there is now an electrical connection for a new generation of digital frahling sensors. In this context, the previous temperature sensor (BX-592) will be replaced with the digital model BX-598, or the combined pressure and temperature sensor BX-596 will be replaced with the digital model BX-597A, which also measures humidity.

The current software version of the measuring system is:
81237-05 V.1.1.0 or R9.4.0 (version with touch-sensitive screen).

Statement issued by TÜV Rheinland Energy GmbH dated 17 September 2022

TÜV Rheinland Energy GmbH
Cologne, 25 April 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.