

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000040212_02

Certified AMS: Fidas® 200 S respectively Fidas® 200
for particulate matter PM₁₀ and PM_{2.5}

Manufacturer: PALAS GmbH
Greschbachstraße 3b
76229 Karlsruhe
Germany

Test Institute: TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested and certified
according to the standards
VDI 4202-1 (2010), VDI 4203-3 (2010), EN 12341 (1998), EN 14907 (2005),
Guide to Demonstration of Equivalence of Ambient Air Monitoring Methods (2005),
EN 15267-1 (2009) and EN 15267-2 (2009)**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 10 pages).
The present certificate replaces certificate 0000040212_01 of 30 September 2015



Suitability Tested
Complying with
2008/50/EC
EN 15267
Regular
Surveillance
www.tuv.com
ID 0000040212

Publication in the German Federal Gazette
(BAnz.) of 26 August 2015

German Federal Environment Agency
Dessau, 25 April 2016



Dr. Marcel Langner
Head of Section II 4.1

This certificate will expire on:
31 March 2019

TÜV Rheinland Energy GmbH
Cologne, 24 April 2016



ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu / www.eco-tuv.com
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00

Certificate:
0000040212_02 / 25 April 2016

Test report: 936/21227195/B of 5 October 2015
Initial certification: 01 April 2014
Date of expiry: 31 March 2019
Publication: BAnz AT 26.08.2015 B7, chapter III no. 2.1

Approved application

The certified AMS is suitable for permanent monitoring of suspended particulate matter PM₁₀ and PM_{2.5} in ambient air (stationary operation).

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a sixteen-month field test.

The Version Fidas® 200 S is approved for a temperature range of -20 °C to +50 °C. The Version Fidas® 200 is approved for a temperature range of +5 °C to +40 °C.

The notification of suitability of the AMS, performance testing, and the uncertainty calculation have been effected on the basis of the regulations valid at the time of performance testing. As changes in legal regulations are possible, any potential user should ensure that this AMS is suitable for monitoring the limit value relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for ambient air applications at which it will be installed.

Basis of the certification

This certification is based on:

- test report 936/21227195/B of 5 October 2015 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the on-going surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 26.08.2016 B4, chapter III number 2.1,
Announcement by UBA from 22 July 2015:

Measuring system:

Fidas® 200 S and Fidas® 200 for suspended particulate matter PM₁₀ and PM_{2.5}

Manufacturer:

PALAS GmbH, Karlsruhe

Field of application:

For the continuous parallel measurement of the PM₁₀ and PM_{2.5} fractions in suspended particulate matter in ambient air in stationary application

Measuring ranges during performance testing:

Component	Certification range	Unit
PM ₁₀	0 – 10,000	µg/m ³
PM _{2.5}	0 – 10,000	µg/m ³

Software version:

100380.0014.0001.0001.0011

Restrictions:

None

Notes:

1. The Fidas® 200 S measuring system is also available for indoor installation in temperature controlled environments under the product designation Fidas® 200.
2. The requirements of the guideline "Demonstration of Equivalence of Ambient Air Monitoring Methods" were fulfilled during the first four comparison campaigns of the preliminary test as well as during the six comparison campaigns of the supplementary test for both measured components PM₁₀ and PM_{2.5}.
3. The requirements as related to the variance coefficient R² in accordance with EN 12341 (issue of 1998) were not met by one of the candidates at the location Cologne, summer.
4. The sensitivity of the particle sensor shall be checked once a month with CalDust 1100 or MonoDust1500.
5. The measuring system shall be calibrated regularly on site by means of the gravimetric reference method for PM_{2.5} and PM₁₀ as stipulated in EN 12341 (issue of 2014).
6. The performance test report is available online at www.qal1.de.
7. Supplementary testing (extension of equivalence test, presentation of design changes, new test standard MonoDust1500) to Federal Environment Agency announcement of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 5.1) and 25 February 2015 (BAnz AT 02.04.2015 B5, chapter IV notification 14).

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Cologne
Report no.: 936/21227195/A of 9 March 2015

Publication in the German Federal Gazette: BAnz AT 14.03.2016 B7, chapter V notification 6,
Announcement by UBA from 18 February 2016:

**6 Notification as regards Federal Environment Agency (UBA) notices of
22 July 2015 (BAnz AT 26.08.2015 B4, chapter III number 2.1)**

A mistake regarding the description of the of the IADS-control functions was detected in the manual for the Fidas® 200 S or the Fidas® 200 measuring system for PM₁₀ and PM_{2.5} manufactured by PALAS GmbH. The description should correctly read as follows:

“The temperature of the IADS is controlled as a function of the ambient temperature and humidity (as measured by the weather station). The minimum temperature is 23°C. Moisture compensation is ensured via a dynamic adjustment of the IADS temperature up to a maximum heat capacity of 90 Watt.”

The manufacturer corrected this mistake as of manual version V0140815. Test report 936/21227195/A dated 9 March 2015 issued by TÜV Rheinland Energie und Umwelt GmbH was corrected accordingly and replaced by test report 936/21227195/B dated 5 October 2015.

The measuring system can alternatively be operated with a WS300-UMB weather station. An extended IADS adaptable for lengths between 1.20m and 2.10m is available for the measuring system.

Furthermore, the Fidas® 200 E version of the measuring system may be used with an external sensor.

The current software version is: 100396.0014.0001.0001.0011.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 6 November 2015

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The Fidas® 200 S respectively Fidas® 200 is an optical aerosol spectrometer which determines particle size by means of scattered light analysis according to Lorenz-Mie.

The version Fidas® 200 is the indoor version, the Fidas® 200 S is the outdoor version.

The measuring system tested consists of the Sigma-2 sampling head, a sampling tube with the IADS (Intelligent Aerosol Drying System) moisture compensation module, the Fidas® control unit with integrated aerosol sensor, the compact WS600-UMB weather station, a UMTS-antenna, weatherproof housing (IP 65), corresponding connection lines and cables, a bottle of CalDust 1000 or MonoDust1500 as well as manuals in German.

The particle sample passes through the Sigma-2 sampling head at a flow rate of 4.8 l/min (based on 25 °C and 1013 hPa) and is led into the sampling line which connects the sampling head to the Fidas control unit. The IADS (Intelligent Aerosol Drying System) moisture compensation module is used in order to avoid the possible effects of condensation, especially when ambient air humidity is high. The IADS is regulated with regard to relative humidity and ambient temperature (measured with weather station WS600-UMB). The minimum temperature is 23 °C, the humidity compensation is done via a dynamic adjustment of the IADS Temperature up to a maximum heat output of max. 90 watts. The IADS module is controlled via the Fidas Firmware. After passing through the IADS module, the particle sample is led to the aerosol sensor where the actual measuring is performed. From the aerosol sensor the sample is then led through an absolute filter which can be used, for instance, to further analyse the collected aerosol. The measuring system Fidas® 200 S is complete with an integrated weather station (WS600-UMB) to capture the measured quantities wind velocity, wind direction, amount of precipitation, type of precipitation, temperature, humidity, and pressure. The Fidas® 200 S control unit contains the necessary electronics for operating the measuring system as well as the 2 parallel-connected sample pumps. Should one pump fail, proper operation is secured by the remaining pump.

The Fidas® 200 S measuring system saves data in the RAW format. In order to determine the mass concentration values, the stored raw data have to be converted by means of an evaluation algorithm. A size-dependent and weighted algorithm is used to convert particle size and number to mass concentrations. During performance testing, conversion was performed using the evaluation algorithm PM_ENVIRO_0011.

The measuring system can be operated using either the touch screen at the front side of the instrument or remotely via radio modem using the corresponding software (e.g. TeamViewer). The user can access measurement data and device information, change parameters, and perform tests to monitor the functionality of the measuring system.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and the validity is also accessible on the internet: gal1.de.

Certificate:
0000040212_02 / 25 April 2016

Certification of Fidas® 200 S respectively Fidas® 200 for particulate matter PM₁₀ and PM_{2.5} is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267

Certificate No. 0000040212: 29 April 2014
Validity of the certificate: 31 March 2019

Test report: 936/21218896/A of 20 September 2013
TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAAnz AT 01.04.2014 B12, chapter IV, No. 5.1
Announcement by UBA from 27 February 2014

Notification:

Statement of TÜV Rheinland Energie und Umwelt GmbH of 27 September 2014
Publication: BAAnz AT 2 April 2015 B5, chapter IV notification 14
UBA announcement of 25 February 2015
(New LED, Indoor variant, new display of software)

Supplementary testing according to EN 15267

Certificate No. 0000040212_01: 30 September 2015
Expiry date of the certificate: 31 March 2019

Test report: 936/21227195/A of 9 March 2015
TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAAnz AT 26.08.2015 B4, chapter III number 2.1
Announcement by UBA from 22 July 2015

Notification:

Certificate No. 0000040212_02: 25 April 2016
Expiry date of the certificate: 31 March 2019
Statement of TÜV Rheinland Energie und Umwelt GmbH of 6 November 2015
and test report 936/21227195/B of 5 October 2015
Publication: BAAnz AT 14.03.2016 B7, chapter V notification 6
UBA announcement of 18 February 2016
(Correction of manual, alternative weather station and new software version)

**Results of the equivalence test for systems SN 0111 & SN 0112,
for the measured component PM_{2.5} after correction of slope /intercept,
comparison campaign GER+UK,
evaluation algorithm PM_ENVIRO_0011**

Comparison candidate with reference according to Guide "Demonstration of Equivalence Of Ambient Air Monitoring Methods", January 2010				
Candidate	FIDAS 200 S	SN	SN 0111 & SN 0112	
Status of measured values	Slope corrected	Limit value	30	µg/m ³
		Allowed uncertainty	25	%
All comparisons				
Uncertainty between Reference	0.53			µg/m ³
Uncertainty between Candidates	0.45			µg/m ³
SN 0111 & SN 0112				
Number of data pairs	313			
Slope b	0.999			not significant
Uncertainty of b	0.008			
Ordinate intercept a	-0.190			not significant
Uncertainty of a	0.136			
Expanded meas. uncertainty W _{CM}	9.35			%
All comparisons, ≥18 µg/m³				
Uncertainty between Reference	0.60			µg/m ³
Uncertainty between Candidates	0.80			µg/m ³
SN 0111 & SN 0112				
Number of data pairs	67			
Slope b	0.981			
Uncertainty of b	0.020			
Ordinate intercept a	0.306			
Uncertainty of a	0.630			
Expanded meas. uncertainty W _{CM}	12.51			%
All comparisons, <18 µg/m³				
Uncertainty between Reference	0.51			µg/m ³
Uncertainty between Candidates	0.31			µg/m ³
SN 0111 & SN 0112				
Number of data pairs	246			
Slope b	1.065			
Uncertainty of b	0.023			
Ordinate intercept a	-0.782			
Uncertainty of a	0.224			
Expanded meas. uncertainty W _{CM}	11.34			%

Comparison candidate with reference according to Guide "Demonstration of Equivalence Of Ambient Air Monitoring Methods", January 2010				
Candidate	FIDAS 200 S		SN	SN 0111 & SN 0112
Status of measured values	Slope corrected		Limit value	30 $\mu\text{g}/\text{m}^3$
			Allowed uncertainty	25 %
Cologne, Summer				
Uncertainty between Reference	0.66	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.11	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	81			82
Slope b	1.053			1.050
Uncertainty of b	0.032			0.033
Ordinate intercept a	-0.850			-0.810
Uncertainty of a	0.342			0.357
Expanded meas. uncertainty W_{CM}	10.46	%		10.77 %
Cologne, Winter				
Uncertainty between Reference	0.54	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.52	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	51			50
Slope b	0.991			0.956
Uncertainty of b	0.013			0.013
Ordinate intercept a	0.656			0.645
Uncertainty of a	0.296			0.307
Expanded meas. uncertainty W_{CM}	8.50	%		9.43 %
Bonn				
Uncertainty between Reference	0.62	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.66	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	50			50
Slope b	1.050			1.008
Uncertainty of b	0.024			0.026
Ordinate intercept a	-0.723			-0.471
Uncertainty of a	0.539			0.584
Expanded meas. uncertainty W_{CM}	12.32	%		12.33 %
Bornheim				
Uncertainty between Reference	0.42	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.47	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	45			45
Slope b	1.142			1.115
Uncertainty of b	0.051			0.050
Ordinate intercept a	-1.370			-1.482
Uncertainty of a	0.607			0.607
Expanded meas. uncertainty W_{CM}	22.40	%		17.49 %
Teddington, Winter				
Uncertainty between Reference	0.42	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.52	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	44			44
Slope b	0.964			0.963
Uncertainty of b	0.012			0.011
Ordinate intercept a	-0.004			-0.143
Uncertainty of a	0.223			0.208
Expanded meas. uncertainty W_{CM}	9.46	%		10.01 %
Teddington, Summer				
Uncertainty between Reference	0.25	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.35	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	44			44
Slope b	0.934			0.926
Uncertainty of b	0.020			0.020
Ordinate intercept a	0.461			0.399
Uncertainty of a	0.232			0.229
Expanded meas. uncertainty W_{CM}	11.50	%		13.40 %
All comparisons, $\geq 18 \mu\text{g}/\text{m}^3$				
Uncertainty between Reference	0.60	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.80	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	67			67
Slope b	0.999			0.965
Uncertainty of b	0.020			0.021
Ordinate intercept a	0.134			0.443
Uncertainty of a	0.642			0.65
Expanded meas. uncertainty W_{CM}	12.67	%		13.39 %
All comparisons, $< 18 \mu\text{g}/\text{m}^3$				
Uncertainty between Reference	0.51	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.31	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	248			248
Slope b	1.083			1.052
Uncertainty of b	0.023			0.023
Ordinate intercept a	-0.841			-0.744
Uncertainty of a	0.227			0.226
Expanded meas. uncertainty W_{CM}	13.84	%		9.97 %
All comparisons				
Uncertainty between Reference	0.53	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.45	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	315			315
Slope b	1.014	not significant		0.985
Uncertainty of b	0.008			0.008
Ordinate intercept a	-0.225	not significant		-0.137
Uncertainty of a	0.137			0.137
Expanded meas. uncertainty W_{CM}	9.50	%		10.17 %

**Results of the equivalence test for systems SN 0111 & SN 0112,
for the measured component PM₁₀ after correction of slope /intercept,
comparison campaign GER+UK,
evaluation algorithm PM_ENVIRO_0011**

Comparison candidate with reference according to Guide "Demonstration of Equivalence Of Ambient Air Monitoring Methods", January 2010				
Candidate	FIDAS 200 S	SN	SN 0111 & SN 0112	
Status of measured values	Slope & offset corrected	Limit value	50	µg/m ³
		Allowed uncertainty	25	%
All comparisons				
Uncertainty between Reference	0.58			µg/m ³
Uncertainty between Candidates	0.65			µg/m ³
SN 0111 & SN 0112				
Number of data pairs	316			
Slope b	1.000			not significant
Uncertainty of b	0.009			
Ordinate intercept a	0.010			not significant
Uncertainty of a	0.208			
Expanded measured uncertainty WCM	7.33			%
All comparisons, ≥30 µg/m³				
Uncertainty between Reference	0.68			µg/m ³
Uncertainty between Candidates	1.15			µg/m ³
SN 0111 & SN 0112				
Number of data pairs	44			
Slope b	0.955			
Uncertainty of b	0.034			
Ordinate intercept a	2.060			
Uncertainty of a	1.490			
Expanded measured uncertainty WCM	10.68			%
All comparisons, <30 µg/m³				
Uncertainty between Reference	0.56			µg/m ³
Uncertainty between Candidates	0.55			µg/m ³
SN 0111 & SN 0112				
Number of data pairs	272			
Slope b	1.006			
Uncertainty of b	0.018			
Ordinate intercept a	-0.122			
Uncertainty of a	0.300			
Expanded measured uncertainty WCM	6.63			%

Comparison candidate with reference according to Guide "Demonstration of Equivalence Of Ambient Air Monitoring Methods", January 2010				
Candidate	FIDAS 200 S		SN	SN 0111 & SN 0112
Status of measured values	Slope & offset corrected		Limit value	50 $\mu\text{g}/\text{m}^3$
			Allowed uncertainty	25 %
Cologne, Summer				
Uncertainty between Reference	0.80	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.26	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	81			82
Slope b	1.007			0.990
Uncertainty of b	0.027			0.027
Ordinate intercept a	-0.221			-0.112
Uncertainty of a	0.473			0.471
Expanded measured uncertainty W_{CM}	6.59	%		7.00 %
Cologne, Winter				
Uncertainty between Reference	0.53	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.64	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	51			50
Slope b	1.026			0.990
Uncertainty of b	0.014			0.014
Ordinate intercept a	0.130			0.107
Uncertainty of a	0.385			0.384
Expanded measured uncertainty W_{CM}	8.19	%		5.89 %
Bonn				
Uncertainty between Reference	0.38	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.87	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	50			50
Slope b	1.005			0.968
Uncertainty of b	0.026			0.028
Ordinate intercept a	1.279			1.419
Uncertainty of a	0.792			0.834
Expanded measured uncertainty W_{CM}	10.60	%		9.15 %
Bornheim				
Uncertainty between Reference	0.54	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.84	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	47			47
Slope b	1.086			1.043
Uncertainty of b	0.038			0.038
Ordinate intercept a	-0.555			-0.731
Uncertainty of a	0.707			0.694
Expanded measured uncertainty W_{CM}	16.74	%		9.15 %
Teddington, Winter				
Uncertainty between Reference	0.48	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.73	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	44			44
Slope b	0.963			0.934
Uncertainty of b	0.017			0.016
Ordinate intercept a	-0.195			-0.179
Uncertainty of a	0.426			0.405
Expanded measured uncertainty W_{CM}	10.41	%		15.18 %
Teddington, Summer				
Uncertainty between Reference	0.46	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.54	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	45			45
Slope b	0.912			0.910
Uncertainty of b	0.028			0.029
Ordinate intercept a	1.264			0.868
Uncertainty of a	0.457			0.489
Expanded measured uncertainty W_{CM}	13.68	%		15.62 %
All comparisons, $\geq 30 \mu\text{g}/\text{m}^3$				
Uncertainty between Reference	0.68	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	1.15	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	44			44
Slope b	0.983			0.928
Uncertainty of b	0.035			0.034
Ordinate intercept a	1.474			2.590
Uncertainty of a	1.518			1.50
Expanded measured uncertainty W_{CM}	11.17	%		11.47 %
All comparisons, $< 30 \mu\text{g}/\text{m}^3$				
Uncertainty between Reference	0.56	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.55	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	274			274
Slope b	1.025			0.990
Uncertainty of b	0.018			0.017
Ordinate intercept a	-0.172			-0.102
Uncertainty of a	0.308			0.297
Expanded measured uncertainty W_{CM}	8.05	%		6.99 %
All comparisons				
Uncertainty between Reference	0.58	$\mu\text{g}/\text{m}^3$		
Uncertainty between Candidates	0.65	$\mu\text{g}/\text{m}^3$		
	SN 0111			SN 0112
Number of data pairs	318			318
Slope b	1.016	not significant		0.983
Uncertainty of b	0.009	not significant		0.009
Ordinate intercept a	-0.019	not significant		0.043
Uncertainty of a	0.212	not significant		0.209
Expanded measured uncertainty W_{CM}	8.16	%		8.01 %