

# CONFIRMATION

Notification: 0000040212\_05\_04  
on changes according to EN 15267 regarding certificate 0000040212\_05 dated 12 June 2019

---

**AMS designation:** Fidas 200 S, Fidas 200 and Fidas 200 E for particulate matter PM<sub>10</sub> and PM<sub>2.5</sub>

**Manufacturer:** PALAS GmbH  
Greschbachstrasse 3b  
76229 Karlsruhe  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 31 March 2021  
Publication: BAnz AT 03.05.2021 B9**

**III. Notifications as regards the uniform practice in monitoring  
emissions and ambient air**

**45 Notification as regards Federal Environment Agency (UBA) notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 5.1) and  
of 27 May 2020 (BAnz AT 31.07.2020 B10, chapter II notification 16)**


The Fidas 200 S, Fidas 200 E or Fidas 200 measuring systems from the company Palas GmbH can in future alternatively be equipped with either the switching power supply unit RPS-300-24-C from the company Meanwell or the switching power supply unit TOP-200-124-C from the company Traco.


In the future, the measuring systems will have three additional holes with blind plugs on the back of the control unit. In addition, the measuring system does not have an electromechanical operating hours counter. When using the Pico318-N3350 single-board computer, the HY-070MRLA0-CLTPA1 touch-sensitive screen from HY-LINE can also be used in the future. Alternatively, the panel PC ARCHMI-807AR from Wachendorff or AFL3-W07A-AL from iEi can be used.

Statement issued by TÜV Rheinland Energy GmbH dated 10 November 2020

---

TÜV Rheinland Energy GmbH  
Cologne, 25 May 2021

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.