

CONFIRMATION

Notification: 0000040216_02_02
on changes according to EN 15267 regarding certificate 0000040216_02 dated 01 July 2020

AMS designation: Modell 5030i SHARP for PM₁₀

Manufacturer: Thermo Fisher Scientific
27, Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**89 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 7.3) and
of 29 June 2021 (BAnz AT 05.08.2021 B5, chapter IV notification 7)**


The current software version of the measuring system Model 5030i SHARP with PM₁₀
pre-separator for suspended particulate matter PM₁₀ from Thermo Fisher Scientific is:


V 03.00.05.268

Version V 03.00.03.261 can also be used.

Statement issued by TÜV Rheinland Energy GmbH dated 16 February 2022

TÜV Rheinland Energy GmbH
Cologne, 25 April 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.