

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000040333_01

Certified AMS: AR602Z/NHg for NO, NO₂, SO₂, NH₃ and Hg as well as AR602Z/N for NO, NO₂, SO₂ and NH₃

Manufacturer: Opsis AB
Skytteskogsvägen 16
244 02 Furulund
Sweden

Test Institute: TÜV Rheinland Energie und Umwelt GmbH

**This is to certify that the AMS has been tested
and found to comply with:**

**EN 15267-1: 2009, EN 15267-2: 2009, EN 15267-3: 2007
and EN 14181: 2004**

Certification is awarded in respect of the conditions stated in this certificate
(see also the following pages).

The present certificate replaces Certificate No. 0000040333 of 29 April 2014




Suitability Tested
EN 15267
QAL1 Certified
Regular
Surveillance

www.tuv.com
ID 0000040333


Publication in the German Federal Gazette
(BAnz.) of 5 August 2014

German Federal Environment Agency
Dessau, 9 September 2014


i. A. Dr. Marcel Langner

This certificate will expire on:
31 March 2019

TÜV Rheinland Energie und Umwelt GmbH
Cologne, 8 September 2014


ppa. Dr. Peter Wilbring

www.umwelt-tuv.de / www.eco-tuv.com
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Cologne

Accreditation according to EN ISO/IEC 17025 and certified according to ISO 9001:2008.

Test report:	936/21222333/B of 17 February 2014
Initial certification:	1 April 2014
Expiry date:	31 March 2019
Publication:	BAnz AT 5 August 2014 B11, chapter I, no. 4.2

Approved application

The tested AMS is suitable for use at combustion plants according to Directive 2010/75/EU, chapter III, at waste incineration plants according to Directive 2010/75/EU, chapter IV and other plants requiring official approval. The measured ranges have been selected considering the wide application range of the AMS.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a seven-month field test at a municipal waste incinerator.

The AMS is approved for an ambient temperature range of +5 °C to +40 °C.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Basis of the certification

This certification is based on:

- test report 936/21222333/B of 17 February 2014 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette: BAnz AT 5 August 2014 B11, chapter I, no. 4.2
UBA announcement of 17 July 2014

AMS designation:

AR602Z/NHg for NO, NO₂, SO₂, NH₃ and Hg as well as AR602Z/N for NO, NO₂, SO₂ and NH₃

Manufacturer:

Opsis AB, Furulund, Sweden

Field of application:

For measurements at plants requiring official approval (e.g. Directive 2010/75/EU on industrial emissions, chapters III and IV)

Measuring ranges during the performance test:

Components	Certification ranges	Supplementary ranges	Units
NO	0 - 150*	0 - 500*	mg/m ³
NO ₂	0 - 20*	0 - 500*	mg/m ³
SO ₂	0 - 75*	0 - 500*	mg/m ³
NH ₃	0 - 10*	0 - 50*	mg/m ³
Hg	0 - 45	0 - 100	µg/m ³

*with reference to a measuring path of 1.0 m

Software version:

7.21

Restrictions:

1. The requirement for response time in the performance test according to EN 15267-3 for the component Hg was not fulfilled.
2. During the performance test according to EN 15267-3 the requirement of the enclosure degree of protection was not fulfilled.

Notes:

1. The maintenance interval is three months.
2. The measurement path tested was 1 m.
3. The components NO, NO₂, SO₂ and NH₃ were determined in situ. The component Hg may be included by connecting the external EX060H measurement cell (with a measurement path length of 2 m) and the MX004 Multiplexer modules. The measuring system is then designated as AR602Z/NHg. If the component Hg is not included (AR602Z/N), the light path must remain unchanged.
4. A test gas generator, e.g. HovaCal, must be available for regular span point control of component Hg.
5. SO₂ (displayed as XXX) must be defined in the measuring cell for cross sensitivity compensation of the component Hg.
6. In the laboratory as well as during the field test the length of the heated test gas line was 10 m for the component Hg.
7. When including the component Hg (AR602Z/NHg) the filters in the sampling probe must be checked and, if necessary, changed after revision or malfunctions during waste gas cleaning.
8. Supplementary testing (extension of the maintenance interval) as regards Federal Environmental Agency (UBA) notices of 27 February 2014 (Federal Gazette (BAnz) AT 1 April 2014 B12, chapter I, no. 3.2).

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Cologne
Report no.: 936/21222333/B of 17 February 2014

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The AMS AR602Z/NHg for NO, NO₂, SO₂, NH₃ and Hg as well as AR602Z/N for NO, NO₂, SO₂ and NH₃ is an in-situ measuring system according to the principle of DOAS measurement. The tested measuring system consists of a light source, a receiver, a fibre optic cable and an analyser. The measuring components are determined in the analyser using characteristic radiation absorption in the UV range by gaseous components using Differential Optical Absorption Spectroscopy (DOAS).

The measuring path consists of a light path between a light emitter and a light receiver. The light source in the emitter is a xenon high pressure lamp.

The light beam generated by the emitter is aimed at the receiver. Along its path to the medium, the intensity of the light beam is influenced by dispersion and absorption in molecules and particles.

The light that reaches the receiver is transmitted to the analyser via a fibre optic cable. This cable has the sole purpose of enabling the analyser to be positioned in a place protected from dust, excessive moisture, temperature fluctuations etc.

The measuring system consists of:

- Analyser (AR602Z/N)
- Light emitter unit (EM062)
- Receiver unit (RE062)
- Fibre optic cable (OF60 R3)
- Manual

The module for measuring mercury also comprises:

- Sample gas probe SP2000 (manufacturer M&C) in Opsi yellow
- Heated sample gas pipe with interior diameter of 6 mm (length 10 m)
- Heated sample gas cell with an active measuring path length of 2.0 m, including emitter/receiver unit, converter, suction jet pump, flow monitoring, power pack and temperature control (EX060)
- Multiplexer (MX004)

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energie und Umwelt GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energie und Umwelt GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energie und Umwelt GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: qal1.de.

Certification of AR602Z/NHg for NO, NO₂, SO₂, NH₃ and Hg as well as AR602Z/N for NO, NO₂, SO₂ and NH₃ is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267

Certificate no. 0000040333: 29 April 2014
Expiry date of the certificate: 31 March 2019
Test report: 936/21222333/A of 10 October 2013
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAnz AT 1 April 2014 B12, chapter I, no. 3.2
UBA announcement of 27 February 2014

Supplementary testing according to EN 15267

Certificate no. 0000040333_01: 9 September 2014
Expiry date of the certificate: 31 March 2019
Test report: 936/21222333/B of 17 February 2014
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAnz AT 5 August 2014 B11, chapter I, no. 4.2
UBA announcement of 17 July 2014

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Opsis AB
AMS designation	AR602Z/NHg
Serial number of units under test	1759 / 1760
Measuring principle	UV-DOAS

Test report

Test laboratory	936/21222333/B
Date of report	TÜV Rheinland
	2014-02-17

Measured component

Certification range	Hg	0 - 45 µg/m ³
---------------------	----	--------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 µg/m ³
Sum of negative CS at zero point	-0.50 µg/m ³
Sum of positive CS at span point	1.00 µg/m ³
Sum of negative CS at span point	-1.10 µg/m ³
Maximum sum of cross-sensitivities	1.20 µg/m ³
Uncertainty of cross-sensitivity	0.694 µg/m ³

Calculation of the combined standard uncertainty

Tested parameter

			u^2
Repeatability standard deviation at set point *	u_r	0.450 µg/m ³	0.203 (µg/m ³) ²
Lack of fit	u_{lof}	0.404 µg/m ³	0.163 (µg/m ³) ²
Zero drift from field test	$u_{d,z}$	0.260 µg/m ³	0.068 (µg/m ³) ²
Span drift from field test	$u_{d,s}$	-0.546 µg/m ³	0.298 (µg/m ³) ²
Influence of ambient temperature at span	u_t	0.153 µg/m ³	0.023 (µg/m ³) ²
Influence of supply voltage	u_v	0.208 µg/m ³	0.043 (µg/m ³) ²
Cross-sensitivity (interference)	u_i	0.694 µg/m ³	0.481 (µg/m ³) ²
Influence of sample gas flow	u_p	-0.049 µg/m ³	0.002 (µg/m ³) ²
Uncertainty of reference material at 70% of certification range	u_{rm}	0.364 µg/m ³	0.132 (µg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max, i})^2} \quad 1.19 \text{ µg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 2.33 \text{ µg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 30 µg/m³ **7.8**

Requirement of 2010/75/EU

U in % of the ELV 30 µg/m³ **40.0**

Requirement of EN 15267-3

U in % of the ELV 30 µg/m³ **30.0**

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Opsis AB
AMS designation	AR602Z/N
Serial number of units under test	1759 / 1760
Measuring principle	UV-DOAS

Test report

Test laboratory	936/21222333/B TÜV Rheinland
Date of report	2014-02-17

Measured component

Certification range	NH ₃ 0 - 10 mg/m ³
---------------------	---

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.18 mg/m ³
Sum of negative CS at zero point	-0.10 mg/m ³
Sum of positive CS at span point	0.23 mg/m ³
Sum of negative CS at span point	-0.10 mg/m ³
Maximum sum of cross-sensitivities	0.23 mg/m ³
Uncertainty of cross-sensitivity	0.133 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

			u ²
Repeatability standard deviation at set point *	u _r	0.090 mg/m ³	0.008 (mg/m ³) ²
Lack of fit	u _{lof}	0.040 mg/m ³	0.002 (mg/m ³) ²
Zero drift from field test	u _{d,z}	0.052 mg/m ³	0.003 (mg/m ³) ²
Span drift from field test	u _{d,s}	0.110 mg/m ³	0.012 (mg/m ³) ²
Influence of ambient temperature at span	u _t	0.058 mg/m ³	0.003 (mg/m ³) ²
Influence of supply voltage	u _v	0.071 mg/m ³	0.005 (mg/m ³) ²
Cross-sensitivity (interference)	u _i	0.133 mg/m ³	0.018 (mg/m ³) ²
Influence of sample gas pressure	u _p	0.088 mg/m ³	0.008 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm}	0.081 mg/m ³	0.007 (mg/m ³) ²
Excursion of measurement beam	u _{mb}	0.115 mg/m ³	0.013 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.28 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.55 mg/m ³

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the ELV 10 mg/m ³	5.5
Requirement of EN 15267-3	U in % of the ELV 10 mg/m ³	40.0 **
	U in % of the ELV 10 mg/m ³	30.0

** The EU-directive 2010/75/EU on industrial emissions provides no requirements for this component. The chosen value is recommended by the certification body.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Opsis AB
AMS designation	AR602Z/N
Serial number of units under test	1759 / 1760
Measuring principle	UV-DOAS

Test report

Test laboratory	936/21222333/B
Date of report	TÜV Rheinland
	2014-02-17

Measured component

Certification range	NO	0 - 150 mg/m ³
---------------------	----	---------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at span point	0.00 mg/m ³
Sum of negative CS at span point	0.00 mg/m ³
Maximum sum of cross-sensitivities	0.00 mg/m ³
Uncertainty of cross-sensitivity	0.000 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

			u^2
Repeatability standard deviation at set point *	u_r	0.600 mg/m ³	0.360 (mg/m ³) ²
Lack of fit	u_{lof}	-0.635 mg/m ³	0.403 (mg/m ³) ²
Zero drift from field test	$u_{d,z}$	0.520 mg/m ³	0.270 (mg/m ³) ²
Span drift from field test	$u_{d,s}$	-1.039 mg/m ³	1.080 (mg/m ³) ²
Influence of ambient temperature at span	u_t	0.100 mg/m ³	0.010 (mg/m ³) ²
Influence of supply voltage	u_v	0.123 mg/m ³	0.015 (mg/m ³) ²
Cross-sensitivity (interference)	u_i	0.000 mg/m ³	0.000 (mg/m ³) ²
Influence of sample gas pressure	u_p	0.367 mg/m ³	0.135 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u_{rm}	1.212 mg/m ³	1.470 (mg/m ³) ²
Excursion of measurement beam	u_{mb}	-0.537 mg/m ³	0.288 (mg/m ³) ²

* The larger value is used :
"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)	$u_c = \sqrt{\sum (u_{max, j})^2}$	2.01 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	3.94 mg/m ³

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the ELV 100 mg/m³	3.9
Requirement of EN 15267-3	U in % of the ELV 100 mg/m³	20.0
	U in % of the ELV 100 mg/m³	15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Opsis AB
AMS designation	AR602Z/N
Serial number of units under test	1759 / 1760
Measuring principle	UV-DOAS

Test report

Test laboratory	936/21222333/B
Date of report	TÜV Rheinland
	2014-02-17

Measured component

Certification range	NO ₂	0 - 20 mg/m ³
---------------------	-----------------	--------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.52 mg/m ³
Sum of negative CS at zero point	-0.13 mg/m ³
Sum of positive CS at span point	0.46 mg/m ³
Sum of negative CS at span point	-0.57 mg/m ³
Maximum sum of cross-sensitivities	-0.57 mg/m ³
Uncertainty of cross-sensitivity	-0.329 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

			u ²
Standard deviation from paired measurements under field conditions *	u _D	0.053 mg/m ³	0.003 (mg/m ³) ²
Lack of fit	u _{lof}	0.081 mg/m ³	0.007 (mg/m ³) ²
Zero drift from field test	u _{d,z}	0.150 mg/m ³	0.023 (mg/m ³) ²
Span drift from field test	u _{d,s}	0.162 mg/m ³	0.026 (mg/m ³) ²
Influence of ambient temperature at span	u _t	0.058 mg/m ³	0.003 (mg/m ³) ²
Influence of supply voltage	u _v	0.058 mg/m ³	0.003 (mg/m ³) ²
Cross-sensitivity (interference)	u _i	-0.329 mg/m ³	0.108 (mg/m ³) ²
Influence of sample gas pressure	u _p	0.088 mg/m ³	0.008 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm}	0.162 mg/m ³	0.026 (mg/m ³) ²
Excursion of measurement beam	u _{mb}	0.144 mg/m ³	0.021 (mg/m ³) ²

* The larger value is used :
"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max, j})^2}$	0.48 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.94 mg/m ³

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the range 20 mg/m³	4.7
Requirement of EN 15267-3	U in % of the range 20 mg/m³	20.0
	U in % of the range 20 mg/m³	15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Opsis AB
AMS designation	AR602Z/N
Serial number of units under test	1759 / 1760
Measuring principle	UV-DOAS

Test report

Test laboratory	936/21222333/B
Date of report	TÜV Rheinland
	2014-02-17

Measured component

Certification range	SO ₂	0 - 75 mg/m ³
---------------------	-----------------	--------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 mg/m ³
Sum of negative CS at zero point	-0.27 mg/m ³
Sum of positive CS at span point	0.73 mg/m ³
Sum of negative CS at span point	-1.47 mg/m ³
Maximum sum of cross-sensitivities	-1.47 mg/m ³
Uncertainty of cross-sensitivity	-0.849 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

			u ²
Standard deviation from paired measurements under field conditions *	u _D	0.189 mg/m ³	0.036 (mg/m ³) ²
Lack of fit	u _{lof}	0.271 mg/m ³	0.073 (mg/m ³) ²
Zero drift from field test	u _{d,z}	0.260 mg/m ³	0.068 (mg/m ³) ²
Span drift from field test	u _{d,s}	0.390 mg/m ³	0.152 (mg/m ³) ²
Influence of ambient temperature at span	u _t	0.208 mg/m ³	0.043 (mg/m ³) ²
Influence of supply voltage	u _v	0.085 mg/m ³	0.007 (mg/m ³) ²
Cross-sensitivity (interference)	u _i	-0.849 mg/m ³	0.720 (mg/m ³) ²
Influence of sample gas pressure	u _p	0.184 mg/m ³	0.034 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm}	0.606 mg/m ³	0.368 (mg/m ³) ²
Excursion of measurement beam	u _{mb}	-0.277 mg/m ³	0.077 (mg/m ³) ²

* The larger value is used :
"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max, j})^2}$	1.26 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	2.46 mg/m ³

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the ELV 50 mg/m³	4.9
Requirement of EN 15267-3	U in % of the ELV 50 mg/m³	20.0
	U in % of the ELV 50 mg/m³	15.0