

CONFIRMATION

Notification: 0000040334_01_03
on changes according to EN 15267 regarding certificate 0000040334_01 dated 05 November 2019

AMS designation: MEAC300 for emission evaluation (DAHS)

Manufacturer: SICK AG
Nimburger Straße 11
79276 Reute
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 31 March 2021
Publication: BAnz AT 03.05.2021 B9**

**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air**

**54 Notification as regards Federal Environment Agency (UBA) notices
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter IV number 1.2) and
of 27 May 2020 (BAnz AT 31.07.2020 B10, chapter II notification 17)**

The latest software version of the MEAC300 emission evaluation system manufactured
by SICK AG is:

4.1.34.25

Versions 4.1.34.18 to 4.1.34.21 may no longer be used. With these versions, invalid
measured values are sometimes assigned a wrong status.

In installations with remote data transmission, the new version must be used, as in the
older versions data transmission is sometimes incomplete.

Statement issued by TÜV Rheinland Energy GmbH dated 27 August 2020

TÜV Rheinland Energy GmbH
Cologne, 25 May 2021


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.