

# CONFIRMATION

Notification: 0000040336\_02\_01  
on changes according to EN 15267 regarding certificate 0000040336\_02 dated 31 May 2022

---

**AMS designation:** Air Pollution Monitor 2 (APM-2) for PM<sub>10</sub> and PM<sub>2.5</sub>

**Manufacturer:** Comde-Derenda GmbH  
Kieler Straße 9  
14532 Stahnsdorf  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 21 February 2023  
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**63 Notification as regards Federal Environment Agency (UBA) notices  
of 17 July 2014 (BAnz AT 05.08.2014 B11, chapter III number 2.1) and  
of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 7)**

The Air Pollution Monitor 2 measuring system (APM-2) for suspended particulate matter PM<sub>2.5</sub> and PM<sub>10</sub> from the company Comde-Derenda GmbH is manufactured with a new printed circuit board including a new insulation concept for the photometer block starting with serial number 20218.


The current software version is:


3.12.003

Statement issued by TÜV Rheinland Energy GmbH dated 12 August 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 25 April 2023

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.