

CONFIRMATION

Notification: 0000043107_02_05
on changes according to EN 15267 regarding certificate 0000043107_02 dated 12 June 2019

AMS designation: APDA-372 / APDA-372 E for PM₁₀, PM_{2,5}

Manufacturer: HORIBA Europe GmbH
Hans-Mess-Str. 6
61440 Oberursel /Ts.
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4**

III. Notifications as regards the uniform practice in monitoring emissions and ambient air:

43 Notification as regards Federal Environment Agency (UBA) notices of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter III number 3.1) and of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 16)


The measuring devices APDA-372 or APDA-372 E for PM₁₀ and PM_{2,5} of the company HORIBA Europe GmbH can be equipped in the future alternatively with the weather station HTP-Geber compact with actively ventilated weather protection from the company Thies starting from the software version 100532.0014.0001.0011.


For measuring the LED temperature, the sensor B57861S0103F040 from TDK can be used as an alternative in the future.
For the connection of the weather station and the IADS module the plugs WSV 50 or SV 50 (weather station) and WSV 60 or SV 60 (IADS module) of the company Lumberg can be used alternatively.

The current software version is: 100532.0014.0001.0001.0011
In addition to this version number, the following intermediate version is also valid:
100527.0014.0001.0011.

Statement issued by TÜV Rheinland Energy GmbH dated 28 April 2022

TÜV Rheinland Energy GmbH
Cologne, 01 August 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.