

CONFIRMATION

Notification: 0000043526_02_02
on changes according to EN 15267 regarding certificate 0000043526_02 dated 02 April 2020

AMS designation: PCME STACKFLOW 400 for velocity

Manufacturer: ENVEA UK Ltd. (PCME Ltd.)
ENVEA House, Rose & Crown Road
Swavesey / Cambridge CB24 4RB
United Kingdom

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 31 March 2021
Publication: BAnz AT 03.05.2021 B9**

**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air**


**27 Notification as regards Federal Environment Agency (UBA) notices
of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter II number 1.3) and
of 27 May 2020 (BAnz AT 31.07.2020 B10, chapter II notification 8)**


Due to a component discontinuation of the manufacturer, the preamplifier
of the PCME STACKFLOW 400 measuring system for exhaust gas velocity
manufactured by ENVEA UK Ltd. must be replaced.

The previously used preamplifier type Texas Instruments LME49870 L200 is replaced
by the successor product type Texas Instruments OPA1611.

Statement issued by TÜV Rheinland Energy GmbH dated 25 November 2020

TÜV Rheinland Energy GmbH
Cologne, 25 May 2021


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.