

CONFIRMATION

Notification: 0000050627_02_02
on changes according to EN 15267 regarding certificate 0000050627_02 dated 13 March 2021

AMS designation: PCME STACKFLOW 200 for velocity

Manufacturer: ENVEA UK Ltd. (PCME Ltd.)
ENVEA House, Rose & Crown Road
Swavesey / Cambridge CB24 4RB
United Kingdom

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 09 March 2022
Publication: BAnz AT 11.04.2022 B10**


**VI. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**


**35 Notification as regards Federal Environment Agency (UBA) notices
of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter II number 2.1) and
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 26)**

In the "Calibration Unit" adjustment module of the PCME STACKFLOW 200 measuring
device for exhaust gas velocity from ENVEA UK Ltd., valves from Parker can also be
used.

Statement issued by TÜV Rheinland Energy GmbH dated 07 September 2021

TÜV Rheinland Energy GmbH
Cologne, 12 April 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.