

CONFIRMATION

Notification: 0000053802_02_02
on changes according to EN 15267 regarding certificate 0000053802_02 dated 02 March 2022

AMS designation: ACF5000 for CO, NO, NO₂, N₂O, SO₂, HCl, HF, NH₃, CH₄, CH₂O, TOC, O₂, H₂O and CO₂

Manufacturer: ABB AG
Stierstädter Str. 5
60488 Frankfurt/Main
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4**

III. Notifications as regards the uniform practice in monitoring emissions and ambient air:

1 Notification as regards Federal Environment Agency (UBA) notices of 13 July 2017 (BAnz AT 31.07.2017 B12, chapter I number 2.1) and of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 30)


The zero air conditioner of the company LNI installed in the measuring device ACF5000 for the components O₂, CO, NO, NO₂, N₂O, SO₂, HCl, HF, NH₃, H₂O, CO₂, H₂CO, CH₄ and TOC of the company ABB AG can also be used in the version V05.


The current software version of the FTIR spectrometer is FTE 1.52.

The following software versions remain unchanged:
AMC board: 3.9.4
Syscon: 5.2.46

Statement issued by TÜV Rheinland Energy GmbH dated 28 April 2022

TÜV Rheinland Energy GmbH
Cologne, 01 August 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.