

CERTIFICATE

of product conformity (QAL 1)

Certificate number: 1630664-ts

AMS	Set CEM CERT 7MB1957 monitoring CO, NO, SO ₂ and O ₂
Manufacturer	Siemens AG Östliche Rheinbrückenstraße 50 76187 Karlsruhe Germany

Test institute TÜV SÜD Industrie Service GmbH

This is to certify that the AMS fulfils the requirements of the
DIN EN 15267-1: 2009, DIN EN 15267-2: 2009, DIN EN 15267-3: 2008 and DIN EN 14181: 2004
standards.



Certificate No: 1630664-ts

Publication in the German Federal Gazette
dated 5 March 2013

Certificate validity
until 4 March 2018

Umweltbundesamt
Dessau, 27 March 2013

TÜV SÜD Industrie Service GmbH
Testing laboratory Emission measurement/
calibration
Munich, 26 March 2013

p.p. Dr. Marcel Langner

Dr. Michael Waeber

Certification applies to the conditions listed in this certificate

Test report	1630664 from 15 September 2012
Initial certification	05 March 2013
Certificate validity until	04 March 2018 (5 years)
Publication	BAnz. AT 05 March 2013 B10, chapter I, No. 6.1

Approved application

The AMS tested is suitable for plants in compliance with TA-Luft. The suitability of the AMS for this application was assessed on the basis of a laboratory test and a field test of the modular measuring system Set CEM CERT 7MB1957 lasting over three months at a plant in compliance with the EC Directive 2000/76/EC. The modular measuring system is authorised for the ambient temperature range from +5 °C to +40 °C.

The operator should consult the manufacturer to ensure that the AMS is suitable for the plant where it is being installed.

Certification basis

This certificate is based on the TÜV SÜD Industrie Service GmbH test report 1630664 from 15 September 2012, the suitability publication by the Umweltbundesamt as relevant body as well as on monitoring of the product and the manufacturing process and the publication in the Bundesanzeiger (BAnz. AT 05 March 2013 B10, chapter I, No. 6.1, UBA publication from 12 February 2013).

AMS: Set CEM CERT 7MB1957 monitoring CO, NO, SO₂ and O₂

Manufacturer: Siemens AG, Karlsruhe

Suitability: For plants in compliance with TA-Luft

Measurement ranges in the suitability test:

The suitability test covers the following modules:

System variations	Ultramat labelling	Component 1	Component 2	Component 3	Component 4
Ultramat 23-7MB2358	-Z-T 13	CO	NO	SO ₂	O ₂ paramagnetic
Ultramat 23-7MB2358	-Z-T 23	CO	NO	SO ₂	O ₂ electrochemical
Ultramat 23-7MB2358	-Z-T 33	CO	NO	SO ₂	-

Component	Module	Certification range	Additional measurement ranges
CO	Ultramat 23-7MB2358	0 – 250 mg/m ³	0 – 1250 mg/m ³
NO	Ultramat 23-7MB2358	0 – 400 mg/m ³	0 – 2000 mg/m ³
SO ₂	Ultramat 23-7MB2358	0 – 400 mg/m ³	0 – 2000 mg/m ³
O ₂ paramagnetic	Ultramat 23-7MB2358	0 – 25 Vol.-%	-
O ₂ electrochemical	Ultramat 23-7MB2358	0 – 25 Vol.-%	-

Software versions:

Ultramat 23-7MB2358: 2.14.07
PLC: Set CEM CERT Rev. 1.0

Restrictions:

1. The minimum requirement of the correlation co-efficient for the calibration function R^2 could not be fulfilled for the component NO.
2. The requirement for total uncertainty in the suitability test according to DIN EN 15267-3 was not fulfilled for the components CO and NO and only partly fulfilled for the component SO₂.
3. In the case of the component CO monitoring is only possible from a limit value of 130 mg/m³ upwards. The AMS measurement range should be set according to the valid regulations.
4. The protection provided by enclosure class is only IP 20. If the operating conditions require a higher class the analysis modules shall be incorporated into an analysis cabinet with the relevant protection class.
5. The maintenance interval for the Ultramat 23-7MB2358 module is three months. In the case of an extension of the Set CEM CERT 7MB1957 by adding additional modules the functionality of the respective compilation of the modules should be tested within the framework of the test for proper installation and the maintenance interval should be set.

Notes:

1. The AMS should be operated at an interval of 24 hours for automatic alignment.
2. The analyser should be operated with the activated thermo-AUTOCAL-function.

Test report:

TÜV SÜD Industrie Service GmbH, Munich
Report-No.: 1630664 from 15 September 2012

Certified product

The certificate applies to AMS that comply with the following description:

The entire tested modular AMS consists of the sample gas extraction probe, heated sample hose, a dual-level measurement gas cooler, a measurement gas feeder pump and the multi-component analyser Ultramat 23-7MB2358. The modular AMS measures CO, NO, SO₂ and O₂ according to the principle of non-dispersive-infrared-absorption (NDIR procedure). Either an electrochemical or paramagnetic oxygen measurement cell can be used to measure O₂.

To regulate measurement gas flow there is a measurement gas pump with integrated gas recirculation between the first and second cooler level. In the cooler casing there is another fine filter for separating fine dust. After the measurement gas cooler the gas path separates into two pipe sections, each supplying one analyser module with measurement gas. In each of these pipe sections currents there is a condensation filter directly before the analyser module, which closes the gas path on penetration of any humidity, to protect the analyser. To regulate zero / sample gas there is a three way valve between the first and second cooler level, which can be switched on to automatically align the analyser or can be time controlled using programmable logic controller (PLC).

The entire system is made up of the following components:

Probe

Manufacturer: Bühler Technologies GmbH, D - 40880 Ratingen
Type: GAS 222.20-Cal-twin with ceramic filter, length 100 cm, heated to 180 °C

Heated sampling hose

Manufacturer: Winkler GmbH, D-69126 Heidelberg
Heated Temperature: 180 °C, 2 PTFE connections (ID: 4 mm), heated to 180 °C, length in the suitability test 35 m

Controller

Manufacturer: Siemens AG
Type: SIRIUS, PT 100

Compressor cooler

Manufacturer: M&C TechGroup Germany GmbH, D - 40885 Ratingen
Type: CSS V1-S, dew point at 3°C (2 gas paths)

Measurement gas feeder pump

Manufacturer: Bühler Technologies GmbH, D-40880 Ratingen
Type: P 2.3
Flow: 1-2 l/min

Analyser

Ultramat 23-7MB2358
Software version 2.14.07
Software version PLC Set CEM CERT Rev. 1.0

General comments

This certificate is based on the analyser tested. The manufacturer is responsible for the continuous compliance of the production to the DIN EN 15267 requirements. The manufacturer is obliged to maintain a tested quality management system to control the manufacture of the certified product. Regular monitoring must be conducted on both the product and the quality management systems.

Should the product from the current production series no longer comply with the certified product, the Environmental Service Department of TÜV SÜD Industrie Service GmbH should be informed (Address see footnote).

The certification mark, which appears on the certified product or is used in advertising materials, is presented on page 1 of this certificate.

This document and the certification mark shall remain the property of TÜV SÜD Industrie Service GmbH.

Should the publication be revoked, this certificate will become invalid. This document must be returned when the period of validity has elapsed and at the request of TÜV SÜD Industrie Service GmbH and the certification mark may no longer be used.

The current version of the certificate and its validity can also be viewed on the internet page: **qal1.de**.

The certification of the modular measuring system Set CEM CERT 7MB1957 is based on the following documents and the regular continuous monitoring of the manufacturer's quality management system:

Initial certification to DIN EN 15267:

Certificate No. 1630664-ts	5 March 2013
Certificate validity until	4 March 2018 (5 years)

Test report: 1630664 from 15 September 2012,
TÜV SÜD Industrie Service GmbH
Publication: BAnz. AT 05 March 2013 B10, chapter I No. 6.1
UBA publication from 12 February 2013

Calculation of total uncertainty for QAL1 testing to DIN EN 14181 and DIN EN 15267-3

**Total uncertainty for the measurement component CO in the measurement range
0-250 mg/m³ for modules 1/ 2**

<i>Performance characteristic</i>	<i>Uncertainty</i>	<i>Value of stan- dard uncertainty in mg/m³</i>	<i>Square sum of standard uncer- tainty in (mg/m³)²</i>
Lack-of-fit	u_{lof}	0,678	0,5
Zero point drift	$u_{d,z}$	1,443	2,1
Span point drift	$u_{d,s}$	1,443	2,1
Influence of ambient temperature at span point	u_t	0,781	0,6
Influence of sample gas pressure	u_p	-	-
Influence of sample gas flow	u_f	-0,217	0,0
Influence of voltage	u_v	1,392	1,9
Cross-sensitivity	u_i	5,340	28,5
Standard deviation from paired measurements or repeat standard deviation at span point ¹⁾	u_r	1,656	2,7
Uncertainty of the test gas	u_{tg}	2,021	4,1
Sum		-	42,6
Combined standard uncertainty	$u_c = \sqrt{\sum (u_i)^2}$	6,5	mg/m ³
Expanded uncertainty	$U_{0,95} = 1,96 \times u_c$	12,8	mg/m ³
Relative expanded uncertainty	U	9,8	%ELV
Demanded uncertainty (ELV 130 mg/m ³) to DIN EN 15267-3		7,5	% ELV
Requirement concerning uncertainty fulfilled		no	

¹⁾ here: Standard deviation from paired measurements

**Total uncertainty for the measurement component NO in the measurement range
0-400 mg/m³ for modules 1/ 2**

<i>Performance characteristic</i>	<i>Uncertainty</i>	<i>Value of standard uncertainty in mg/m³</i>	<i>Square sum of standard uncertainty in (mg/m³)²</i>
Lack-of-fit	u_{lof}	-0,393	0,2
Zero point drift	$u_{d,z}$	3,233	10,5
Span point drift	$u_{d,s}$	3,695	13,7
Influence of ambient temperature at span point	u_t	2,177	4,7
Influence of sample gas pressure	u_p	-	-
Influence of sample gas flow	u_f	0,277	0,1
Influence of voltage	u_v	1,688	2,8
Cross-sensitivity	u_i	-8,083	65,3
Standard deviation from paired measurements or repeat standard deviation at span point *)	u_r	1,750	3,1
Uncertainty of the test gas	u_{tg}	3,236	10,5
Sum		-	110,8
Combined standard uncertainty	$u_c = \sqrt{\sum (u_i)^2}$	10,5	mg/m ³
Expanded uncertainty	$U_{0,95} = 1,96 \times u_c$	20,6	mg/m ³
Relative expanded uncertainty	U	15,8	%ELV
Demanded uncertainty (ELV 130,4 mg/ m ³) to DIN EN 15267-3		15,0	%ELV
Requirement concerning uncertainty fulfilled		no	

*) here: Standard deviation from paired measurements

**Total uncertainty for the measurement component SO₂ in the measurement range
0-400 mg/m³ for modules 1/ 2**

<i>Performance characteristic</i>	<i>Uncertainty</i>	<i>Value of standard uncertainty in mg/m³</i>	<i>Square sum of standard uncertainty in (mg/m³)²</i>
Lack-of-fit	<i>u</i> _{lof}	2,102	4,4
Zero point drift	<i>u</i> _{d,z}	6,235	38,9
Span point drift	<i>u</i> _{d,s}	4,85	23,5
Influence of ambient temperature at span point	<i>u</i> _t	6,498	42,2
Influence of sample gas pressure	<i>u</i> _p	-	-
Influence of sample gas flow	<i>u</i> _f	-2,215	4,9
Influence of voltage	<i>u</i> _v	2,217	4,9
Cross-sensitivity	<i>u</i> _i	-6,928	48,0
Standard deviation from paired measurements or repeat standard deviation at span point *)	<i>u</i> _r	2,475	6,1
Uncertainty of the test gas	<i>u</i> _{tg}	3,236	10,5
Sum		-	183,5
Combined standard uncertainty	$u_c = \sqrt{\sum (u_i)^2}$	13,5	mg/m ³
Expanded uncertainty	$U_{0,95} = 1,96 \times u_c$	26,5	mg/m ³
Relative expanded uncertainty	<i>U</i>	13,3	%ELV
Demanded uncertainty (ELV 200 mg/m ³) to DIN EN 15267-3		15,0	%ELV
Requirement concerning uncertainty fulfilled		yes	

*) here: Standard deviation from paired measurements

Total uncertainty for the measurement component O₂ in the measurement range 0-25 Vol.-% (in version with paramagnetic oxygen measurement) for modules 1 / 2

<i>Performance characteristic</i>	<i>Uncertainty</i>	<i>Value of standard uncertainty in Vol.%</i>	<i>Square sum of standard uncertainty in (Vol.%)²</i>
Lack-of-fit	<i>u_{lof}</i>	0,017	0,00
Zero point drift	<i>u_{d,z}</i>	-0,092	0,01
Span point drift	<i>u_{d,s}</i>	-0,081	0,01
Influence of ambient temperature at span point	<i>u_t</i>	0,044	0,00
Influence of sample gas pressure	<i>u_p</i>	-	-
Influence of sample gas flow	<i>u_f</i>	-0,017	0,00
Influence of voltage	<i>u_v</i>	0,051	0,00
Cross-sensitivity	<i>u_i</i>	0,162	0,03
Standard deviation from paired measurements or repeat standard deviation at span point *)	<i>u_r</i>	0,081	0,01
Uncertainty of the test gas	<i>u_{tg}</i>	0,230	0,05
Sum		-	0,11
Combined standard uncertainty	$u_c = \sqrt{\sum (u_i)^2}$	0,33	Vol.%
Expanded uncertainty	$U_{0,95} = 1,96 \times u_c$	0,64	Vol.%
Relative expanded uncertainty	<i>U</i>	2,6	%
Demanded uncertainty (% from CR)		7,5	% from CR
Requirement concerning uncertainty fulfilled		yes	

*) here: Standard deviation from paired measurements

**Total uncertainty for the measurement component CO in the measurement range
0-250 mg/m³ for modules 3/ 4**

<i>Performance characteristic</i>	<i>Uncertainty</i>	<i>Value of standard uncertainty in mg/m³</i>	<i>Square sum of standard uncertainty in (mg/m³)²</i>
Lack-of-fit	u_{lof}	0,678	0,5
Zero point drift	$u_{d,z}$	1,443	2,1
Span point drift	$u_{d,s}$	1,443	2,1
Influence of ambient temperature at span point	u_t	1,285	1,7
Influence of sample gas pressure	u_p	-	-
Influence of sample gas flow	u_f	-0,303	0,1
Influence of voltage	u_v	1,568	2,5
Cross-sensitivity	u_i	5,340	28,5
Standard deviation from paired measurements or repeat standard deviation at span point *)	u_r	1,656	2,7
Uncertainty of the test gas	u_{tg}	2,021	4,1
Sum		-	44,2
Combined standard uncertainty	$u_c = \sqrt{\sum (u_i)^2}$	6,6	mg/m ³
Expanded uncertainty	$U_{0,95} = 1,96 \times u_c$	13,0	mg/m ³
Relative expanded uncertainty	U	10,0	%ELV
Demanded uncertainty (ELV 130 mg/m ³) to DIN EN 15267-3		7,5	%ELV
Requirement concerning uncertainty fulfilled		no	

*) here: Standard deviation from paired measurements

**Total uncertainty for the measurement component NO in the measurement range
0-400 mg/m³ for modules 3/ 4**

<i>Performance characteristic</i>	<i>Uncertainty</i>	<i>Value of standard uncertainty in mg/m³</i>	<i>Square sum of standard uncertainty in (mg/m³)²</i>
Lack-of-fit	<i>u</i> _{lof}	-0,393	0,2
Zero point drift	<i>u</i> _{d,z}	3,233	10,5
Span point drift	<i>u</i> _{d,s}	3,695	13,7
Influence of ambient temperature at span point	<i>u</i> _t	1,712	2,9
Influence of sample gas pressure	<i>u</i> _p	-	-
Influence of sample gas flow	<i>u</i> _f	0,531	0,3
Influence of voltage	<i>u</i> _v	2,824	8,0
Cross-sensitivity	<i>u</i> _i	-8,083	65,3
Standard deviation from paired measurements or repeat standard deviation at span point *)	<i>u</i> _r	1,750	3,1
Uncertainty of the test gas	<i>u</i> _{tg}	3,236	10,5
Sum		-	114,3
Combined standard uncertainty	$u_c = \sqrt{\sum (u_i)^2}$	10,7	mg/m ³
Expanded uncertainty	$U_{0,95} = 1,96 \times u_c$	21,0	mg/m ³
Relative expanded uncertainty	<i>U</i>	16,1	%ELV
Demanded uncertainty (ELV 130,4 mg/ m ³) to DIN EN 15267-3		15,0	%ELV
Requirement concerning uncertainty fulfilled		no	

*) here: Standard deviation from paired measurements

**Total uncertainty for the measurement component SO₂ in the measurement range
0-400 mg/m³ for modules 3/ 4**

<i>Performance characteristic</i>	<i>Uncertainty</i>	<i>Value of standard uncertainty in mg/m³</i>	<i>Square sum of standard uncertainty in (mg/m³)²</i>
Lack-of-fit	<i>u</i> _{lof}	2,102	4,4
Zero point drift	<i>u</i> _{d,z}	6,235	38,9
Span point drift	<i>u</i> _{d,s}	4,85	23,5
Influence of ambient temperature at span point	<i>u</i> _t	9,96	99,2
Influence of sample gas pressure	<i>u</i> _p	-	-
Influence of sample gas flow	<i>u</i> _f	-2,125	4,5
Influence of voltage	<i>u</i> _v	2,564	6,6
Cross-sensitivity	<i>u</i> _i	-6,928	48,0
Standard deviation from paired measurements or repeat standard deviation at span point *)	<i>u</i> _r	2,475	6,1
Uncertainty of the test gas	<i>u</i> _{tg}	3,236	10,5
Sum		-	241,7
Combined standard uncertainty	$u_c = \sqrt{\sum (u_i)^2}$	15,5	mg/m ³
Expanded uncertainty	$U_{0,95} = 1,96 \times u_c$	30,5	mg/m ³
Relative expanded uncertainty	<i>U</i>	15,2	%ELV
Demanded uncertainty (ELV 200 mg/m ³) to DIN EN 15267-3		15,0	%ELV
Requirement concerning uncertainty fulfilled		no	

*) here: Standard deviation from paired measurements

Total uncertainty for the measurement component O₂ in the measurement range 0-25 Vol.-% (in version with electrochemical oxygen measurement) for modules 3/ 4

<i>Performance characteristic</i>	<i>Uncertainty</i>	<i>Value of standard uncertainty in Vol.%</i>	<i>Square sum of standard uncertainty in (Vol.%)²</i>
Lack-of-fit	<i>u</i> _{lof}	0,035	0,00
Zero point drift	<i>u</i> _{d,z}	0,167	0,03
Span point drift	<i>u</i> _{d,s}	0,098	0,01
Influence of ambient temperature at span point	<i>u</i> _t	0,021	0,00
Influence of sample gas pressure	<i>u</i> _p	-	-
Influence of sample gas flow	<i>u</i> _f	-0,029	0,00
Influence of voltage	<i>u</i> _v	0,009	0,00
Cross-sensitivity	<i>u</i> _i	0,167	0,03
Standard deviation from paired measurements or repeat standard deviation at span point *)	<i>u</i> _r	0,056	0,00
Uncertainty of the test gas	<i>u</i> _{tg}	0,230	0,05
Sum		-	0,12
Combined standard uncertainty	$u_c = \sqrt{\sum (u_i)^2}$	0,35	Vol.%
Expanded uncertainty	$U_{0,95} = 1,96 \times u_c$	0,69	Vol.%
Relative expanded uncertainty	<i>U</i>	2,8	%
Demanded uncertainty (% from CR)		7,5	% from CR
Requirement concerning uncertainty fulfilled		yes	

*) here: Standard deviation from paired measurements