

CONFIRMATION

Notification: 0000053811_02_01
on changes according to EN 15267 regarding certificate 0000053811_02 dated 02 March 2022

AMS designation: LasIR for HCl and H₂O

Manufacturer: Unisearch Associates
96 Bradwick Drive
Concord, Ontario / L4K 1K8
Canada

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4**


**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**


**31 Notification as regards Federal Environment Agency (UBA) notices
of 26 February 2018 (BAnz AT 26.03.2018 B8, chapter I number 3.4) and
of 29 June 2021 (BAnz AT 05.08.2021 B5, chapter IV notification 50)**

A new motherboard with the designation PCA-MBH1000 is installed in the LasIR
measuring device for HCl and H₂O from Unisearch Associates Inc.

Statement issued by TÜV Rheinland Energy GmbH dated 17 February 2022

TÜV Rheinland Energy GmbH
Cologne, 01 August 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.