

# CONFIRMATION

Notification: 0000053811\_02\_02  
on changes according to EN 15267 regarding certificate 0000053811\_02 dated 02 March 2022

---

**AMS designation:** LasIR for HCl and H<sub>2</sub>O

**Manufacturer:** Unisearch Associates  
96 Bradwick Drive  
Concord, Ontario / L4K 1K8  
Canada

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 05 July 2023  
Publication: BAnz AT 02.08.2023 B7**

**III. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**27 Notification as regards Federal Environment Agency (UBA) notices  
of 21 February 2018 (BAnz AT 26.03.2018 B8, chapter I number 3.4) and  
of 28 June 2022 (BAnz AT 28.07.2022 B4, chapter III notification 31)**


The control/analysis unit of the LasIR measuring system for HCl and H<sub>2</sub>O from Unisearch Associates Inc. may alternatively be installed in a housing with a depth of 12" instead of 10.5".


The measuring system can alternatively be equipped with the enlarged LCD screen PCA-LCD2000.

Statement issued by TÜV Rheinland Energy GmbH dated 02 February 2023

---

TÜV Rheinland Energy GmbH  
Cologne, 02 August 2023

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.