

CONFIRMATION

Notification: 0000056506_02_04
on changes according to EN 15267 regarding certificate 0000056506_02 dated 12 June 2019

AMS designation: CMM for Hg

Manufacturer: Gasmot Technologies Oy
Mestarintie 6
01730 Vantaa
Finland

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 31 March 2021
Publication: BAnz AT 03.05.2021 B9**

- III. Notifications as regards the uniform practice in monitoring
emissions and ambient air**
- 32 Notification as regards Federal Environment Agency (UBA) notices
of 21 February 2018 (BAnz AT 26.03.2018 B8, chapter I number 2.1) and
of 27 May 2020 (BAnz AT 31.07.2020 B10, chapter II notification 9)**

The latest software version of the CMM measuring system for the component Hg
manufactured by Gasmot Technology Oy is:


1.2050


In addition to the previously used power supply unit, the PSF-125-12 power supply unit
from Powerbox Oy can also be used in the future.

For measuring the pressure of the instrument air, the digital pressure transmitter from
the manufacturer Festo, type SPTE-P10R-S6-V-2.5K, can also be used instead of the
analogue manometer used up to now.

Statement issued by TÜV Rheinland Energy GmbH dated 09 September 2020

TÜV Rheinland Energy GmbH
Cologne, 25 May 2021


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.