

# CONFIRMATION

Notification: 0000062066\_00\_03  
on changes according to EN 15267 regarding certificate 0000062066\_00 dated 02 June 2019

---

**AMS designation:** 43iQ for SO<sub>2</sub>

**Manufacturer:** Thermo Fisher Scientific  
27, Forge Parkway  
Franklin, MA 02038  
USA

---

## German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 21 February 2023  
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**82 Notification as regards Federal Environment Agency (UBA) notices  
of 27 February 2019 (BAnz AT 26.03.2019, chapter III number 2.1) and  
of 28 June 2022 (BAnz AT 28.07.2022 B4, chapter III notification 49)**

The current software version of the 43iQ ambient air measuring system for SO<sub>2</sub> from  
Thermo Fisher Scientific is:

01.06.14.34444


The versions 01.06.13.34160 and 01.06.14.34442 are also available.


To increase operational reliability, a "footprint" has been corrected on the optional  
analog IO board of the measuring system.

Statement issued by TÜV Rheinland Energy GmbH dated 15 September 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 25 April 2023

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.