

CONFIRMATION

Notification: 0000074625_01_01
on changes according to EN 15267 regarding certificate 0000074625_01 dated 03 September 2021

AMS designation: X-CEMS for CO, NO_x, SO₂, O₂ and CO₂

Manufacturer: Emerson Process Management GmbH & Co. OHG
Industriestrasse 1
63594 Hasselroth
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4**

**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**


**9 Notification as regards Federal Environment Agency (UBA) notice
of 29 June 2021 (BAnz AT 05.08.2021 B5, chapter I number 4.1 and
of chapter III correction 3)**


The current software version of the X-CEMS measuring device for CO, NO, NO_x, SO₂,
CO₂ and O₂ from Emerson Process Management GmbH & Co. OHG is
1.7.1.

The light electrode in the mercury UV lamp for the determination of SO₂ has been
discontinued by the manufacturer and replaced by an equivalent successor model.

Statement issued by TÜV Rheinland Energy GmbH dated 10 February 2022

TÜV Rheinland Energy GmbH
Cologne, 01 August 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.