



Certificate number: 3210534-ts



Industrie Service

Certificate

of product conformity (QAL 1)

Certificate number: 3210534-ts

| | |
|---------------------|---|
| AMS | MCA 10-HWIR monitoring CO, NO, SO ₂ , NO ₂ , N ₂ O, HCl, NH ₃ , CH ₄ , CO ₂ , O ₂ , humidity and TOC |
| Manufacturer | Dr. Födisch Umweltmesstechnik AG Zwenkauer Straße 159 04420 Markranstädt Germany |

Test institute TÜV SÜD Industrie Service GmbH

**This is to certify that the AMS has been tested and found to comply with:
DIN EN 15267-1: 2009, DIN EN 15267-2: 2009, DIN EN 15267-3: 2008 and
DIN EN 14181: 2015**

Certification is awarded in respect of conditions stated in this certificate
(the certificate consists of 21 pages)

This certificate replaces the certificate 2600495-ts from
25th May 2018



Certificate No: 3210534-ts

**Publication in the German Federal Gazette
(BAnz) of 26 March 2018**

**This certificate will expire on:
25 May 2025**

Umweltbundesamt
Dessau, 26 May 2020

TÜV SÜD Industrie Service GmbH
Testing laboratory Emission measurement/
calibration
Munich, 25 May 2020

Dr. Marcel Langner
Head of Section II 4.1

Hans-Jörg Eisenberger

| | |
|-----------------------------------|---|
| Test report | 2600495 from 29 November 2017 |
| Initial certification | 26 August 2015 |
| Certificate validity until | 25 August 2025 (5 years) |
| Certificate | Renewal (previous certificate 2600495-ts from 24 May 2018 with validity until 25 August 2020) |
| Publication | BAnz AT 26 March 2018 B4, chapter I, No. 3.1 |

Approved application

The tested AMS is suitable for use at plants according to Directive 2010/75/EU, chapter III (13. BImSchV), plants according to Directive 2010/75/EU chapter IV (17. BImSchV), the 27. BImSchV, TA Luft and other plants requiring official approval.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a field test of the MCA 10-HWIR multi component measuring system lasting over three months at a plant according to Directive 2010/75/EU chapter IV (17. BImSchV).

The suitability of the AMS for application on gas turbine plants (13. BImSchV) was assessed on the basis of a laboratory test and a field test of the MCA 10-HWIR multi component measuring system lasting over three months at a plant in according to Directive 2010/75/EU, chapter III (13. BImSchV).

The measuring system is approved for ambient temperatures between +5 °C to +40 °C.

The AMS publication, the suitability test and the performance of the uncertainty calculations were conducted based on the provisions valid at the time of testing. Due to possible amendments to legal foundations every user should ensure before use of the AMS that it is suitable for monitoring the applicable limit values.

The operator should consult the manufacturer to ensure that the AMS is suitable for the plant where it is being installed.

Basis of the certification

This certificate is based on:

- TÜV SÜD Industrie Service GmbH test report 2600495 from 29 November 2017
- Suitability announcement by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

- Publication in the German Federal Gazette (BAnz AT 26.03.2018 B8, chapter I, No. 3.1, Announcement by UBA from 21 February 2018):

AMS designation: MCA 10-HWIR monitoring CO, NO, SO₂, NO₂, N₂O, HCl, NH₃, CH₄, CO₂, O₂, humidity and TOC

Manufacturer: Dr. Födisch Umweltmesstechnik AG, Markranstädt

Suitability: For plants requiring authorisation and plants in compliance with the 27. BImSchV

Measurement ranges in the suitability test:

| Component | Certification range | Supplementary measurement range | | Unit |
|------------------|---------------------|---------------------------------|----------|-------------------|
| | | | | |
| CO | 0 - 75 | 0 - 300 | 0 - 5000 | mg/m ³ |
| CO ₂ | 0 - 25 | 0 - 50 | - | Vol.-% |
| NO | 0 - 80 and 200 | 0 - 400 | 0 - 3000 | mg/m ³ |
| NO ₂ | 0 - 50 | 0 - 500 | - | mg/m ³ |
| N ₂ O | 0 - 50 | 0 - 3000 | - | mg/m ³ |
| NH ₃ | 0 - 10 | 0 - 50 | 0 - 500 | mg/m ³ |
| SO ₂ | 0 - 75 | 0 - 300 | 0 - 2500 | mg/m ³ |
| HCl | 0 - 15 | 0 - 90 | 0 - 5000 | mg/m ³ |
| H ₂ O | 0 - 40 | - | - | Vol.-% |
| CH ₄ | 0 - 50 | 0 - 500 | - | mg/m ³ |
| TOC | 0 - 15 | 0 - 30 | 0 - 500 | mg/m ³ |
| O ₂ | 0 - 25 | - | - | Vol.-% |

Software versions: MCA 10-HWIR: V 4.00|3.61.|3.62
FID 5.31e

Restrictions:

None

Notes:

1. The AMS determines gas concentrations in moist test gas.
2. The analyser should be operated with the activated thermo-AUTOCAL-function.
3. The AMS should be equipped with additional heating for temperatures of less than 20°C at the point of installation.
4. The AMS should be operated at an interval of 12 h for automatic zero alignment. TOC should be operated at an interval of 24 h for automatic zero und span point alignment.
5. When HCl, NO₂ or NH₃ are applied, automatic zero point alignment shall be conducted by adding zero gas locally at the injector block.
6. When checking and aligning the span points for NO₂, HCl and NH₃ the sample gas is added locally at the injector block.
7. The maintenance interval is six months.

8. The manufacturer's specifications for implementing the air supply to the instruments should be observed.
9. Starting with the serial numbers with the annual code 18, the measuring system is equipped with a certification range of 0 - 80 mg/ m³ for the measuring component NO. The annual code is composed of the first two digits of the serial number and is indicated on the nameplate.
10. Supplementary test (AMS with TOC measuring range 0 to 500 mgC/ m³ and NO certification range 0 to 80 mg/ m³) to the publication of the Federal Environmental Agency dated 18 February 2016 (BAntz AT 14.03.2016 B7, chapter I number 4.3) and of 14 July 2016 (BAntz AT 01.08.2016 B11, chapter V, notification 23)

Test report:

TÜV SÜD Industrie Service GmbH, Munich
Report-No.: 2600495 from 29 November 2017

- Publication in the German Federal Gazette (BAntz AT 26.03.2019 B7, chapter IV, number 35, Announcement by the Federal Environment Agency (UBA) from 27 February 2019):

35 Notification to the announcement from the Federal Environment Agency (UBA) 18 February 2016 (BAntz AT 14.03.2016 B7, chapter I, number 4.3) and from 21 February 2018 (BAntz AT 26.03.2018 B8, chapter I, number 3.1)

The current software version of the MCA 10-HWIR measuring system for the components CO, NO, SO₂, NO₂, N₂O, HCl, NH₃, CH₄, CO₂, O₂, humidity and TOC by Dr. Födisch Umweltmesstechnik AG is:

MCA 10: V 4.00 | 3.61 | 3.62 FID: 5.31G

This includes the version 5.31F for the TOC analyser.

Statement issued by TÜV Rheinland Energy GmbH dated 5 Oktober 2018

Certified product

The certificate applies to AMS that comply with the following description:

The entire tested MCA 10-HWIR multi component AMS consists of the sample gas extraction probe, heated sample hose and the measurement cabinet with analyser. The measurement cabinet is equipped with an air conditioner and an additional cabinet heating. The basic components of the measurement cabinet are:

- Modular analyser MCA 10-HWIR
- Total organic analyser Thermo-FID ES
- Panel-PC P1550 Win7 15"
- PLC control

The MCA 10-HWIR multi component AMS records emissions of CO, NO, NO₂, N₂O, SO₂, HCl, NH₃, CH₄, Total organic, CO₂ and their reference components O₂ and moisture in flue gas. The sample gas is applied hot to the AMS after filtering using an air jet pump, without prior separation of the flue gas moisture.

The following 4 measurement principles are applied:

Dual frequency measurement procedure
Gas filter correlation
Zirconium dioxide measurement cell
Flame ionization detector

The sample gas extraction is conducted through a stainless-steel extraction probe with a PTFE filter heated to 185 °C. A sample gas line heated to 185 °C and fitted with a PTFE seal (internal diameter 6 mm) is attached to the probe. The line is max. 50 metres long. After the heated line the sample gas flows into the gas distributor block inside the MCA 10-HWIR analyser. The connection for zero air, the exhaust duct and the carrier gas line for the air jet pump and connector for heated gas line (inner diameter 4 mm) to FID with length of 0.7 m are also in the gas distributor block.

The entire system consists of the following components:

Probe

Manufacturer: M&C TechGroup Germany GmbH, D - 40885 Ratingen
Type: SP2000-H
Filter: F-T2 150 PTFE filter 2 µm

Heated line

Manufacturer: Winkler GmbH, D-69126 Heidelberg
Heated temperature: 185 °C, PTFE line (ID: 6 mm), length in the suitability test 50 m
Regulator: integrated into the MCA 10-HWIR

Air conditioning system

Manufacturer: Rittal GmbH & Co. KG, Herbronn
Type: Wandanbau-Kühlgerät 1500 W/230VAC
Alternative air conditioning
Manufacturer: Pentair, Straubenhardt
Type: Wandanbau-Kühlgerät S101526G031;1500 W/230VAC

Switch cabinet heating

Manufacturer: Rittal GmbH & Co. KG, Herbronn
Type: SK 3105 / 230VAC / 400 W
Regulator: integrated into the MCA 10-HWIR

Programmable logic control (PLC)

Manufacturer: Panasonic
Software: V 3.62

Panel PC with operating software

Software: MCA10_HID.exe
Version: V 4.00
System requirements: Operating system Windows XP or higher
CPU Pentium II or higher
Memory 500 MB
ROM 5 GB free storage for data storage
Interfaces USB 2.0
Display Mind. 1024*768 Pixel

Analysis system

Manufacturer: Dr. Födisch Umweltmesstechnik AG
System type: MCA 10-HWIR
Software: V 4.00|3.61|3.62|
Measurement principle: CO, NO, NO₂, N₂O, SO₂, HCl, NH₃, CH₄, CO₂, H₂O:
Extractive heat measuring infrared spectroscopy system
O₂: Zirconium dioxide cell

Total organic Analysis system

Manufacturer: SK-Elektronik GmbH / Leverkusen
formerly Mess- und Analysentechnik GmbH / Leverkusen
System type: Thermo-FID ES
Software: 5.31e
Measurement principle: flame ionisation detector

Heated connection FID

Manufacturer: Winkler GmbH, D-69126 Heidelberg
Heated temperature: 185 °C, PTFE line (ID: 4 mm), length 0,7 m
Regulator: integrated into the MCA 10-HWIR

General notes

This certificate is based on the analyser tested. The manufacturer is responsible for the continuous compliance of the production to the DIN EN 15267 requirements. The manufacturer is required to maintain an approved quality management system to control the manufacture of the certified product. Regular monitoring must be conducted on both the product and the quality management systems.

If the product from the current production series no longer comply with the certified product, the Environmental Service Department of TÜV SÜD Industrie Service GmbH must be informed (address see footnote).

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied on the product or used in publicity material for the certified product.

This document and the certification mark shall remain the property of TÜV SÜD Industrie Service GmbH.

Should the publication be revoked, this certificate will become invalid. This document must be returned when the period of validity has elapsed and at the request of TÜV SÜD Industrie Service GmbH and the certification mark may no longer be used.

The current version of the certificate and its expiration is also accessible on the internet at qal1.de.

The certification of the multi component measuring system MCA 10-HWIR is based on the following documents and the regular continuous monitoring of the manufacturer's quality management system:

Initial certification to DIN EN 15267:

| | |
|----------------------------|--------------------------|
| Certificate No. 1729865-ts | 26 August 2015 |
| Certificate validity until | 25 August 2020 (5 years) |

Test report: 1729865 from 10 June 2015,
TÜV SÜD Industrie Service GmbH
Publication: BAnz AT 26 August 2015 B4, chapter I no. 2.2
UBA publication from 22 July 2015

Supplementary test to DIN EN 15267:

| | |
|----------------------------|--------------------------|
| Certificate No. 2422091-ts | 14 March 2016 |
| Certificate validity until | 25 August 2020 (5 years) |

Test report: 2422091 from 20 Oktober 2015,
TÜV SÜD Industrie Service GmbH
Publication: BAnz AT 14 March 2016 B7, chapter I no. 4.3
UBA publication from 18 February 2016

Supplementary test to DIN EN 15267:

Certificate No. 2600495-ts

24 May 2018

Certificate validity until

25 August 2020 (5 years)

Test report: 2600495 from 29 November 2017,

TÜV SÜD Industrie Service GmbH

Publication: BAnz AT 26 March 2018 B8, chapter I no. 3.1

UBA publication from 21 February 2018

Notification:

Comment of TÜV SÜD Industrie Service GmbH from 26 February 2016

Publication: BAnz AT 1 August 2016 B11, chapter V comment 23

UBA publication from 14 July 2016 (change software version)

Notification:

Comment of TÜV SÜD Industrie Service GmbH from 26 February 2016

Publication: BAnz AT 1 August 2016 B11, chapter V comment 23

UBA publication from 14 July 2016 (change software version)

Notification

Comment of TÜV Rheinland Energy GmbH from 05 Oktober 2018

Publication: BAnz AT 26 March 2019 B7, chapter IV, comment 35,

UBA publication from 27 February 2019 (change software version)

Renewal of the certificate in acc. to DIN EN 15267

Certificate number 3210534-ts:

26 May 2020

Expiry of the certificate:

25 May 2025 (5 years)

Calculation of total uncertainty for the measuring system MCA 10-HWIR for QAL1 testing to DIN EN 14181 and DIN EN 15267-3

Total uncertainty for the measurement component O₂ in the measurement range 0-25 Vol.%

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty Vol.%</i> | <i>Square of standard uncertainty (Vol.%)²</i> |
|--|------------------------------|---|---|
| Lack-of-fit | u_{lof} | 0,045 | 0,00203 |
| Zero drift from field test | $u_{\text{d,z}}$ | -0,017 | 0,00029 |
| Span drift from field test | $u_{\text{d,s}}$ | -0,052 | 0,0027 |
| Influence of ambient temperature at span | u_t | 0,017 | 0,0003 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | 0,081 | 0,00656 |
| Influence of supply voltage | u_v | 0,011 | 0,00012 |
| Cross-sensitivity (interference) | u_i | 0,15 | 0,0225 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,01 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,053 | 0,00281 |
| Uncertainty of reference material 1 % by 70% of CR | u_m | 0,10104 | 0,01021 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 0,04751 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 0,21797 | Vol.% |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 0,42722 | Vol.% |
| Relativ expanded uncertainty | U | 1,7 | % CR |
| Permissible uncertainty of EN 15267-3 | (of CR 25 Vol.%) | 7,5 | % CR |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of CR 25 Vol.%) | 10 | % CR |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component CO in the measurement range 0-75 mg/m³

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty mg/m³</i> | <i>Square of standard uncertainty (mg/m³)²</i> |
|--|---------------------------------|--|--|
| Lack-of-fit | u_{lor} | 0,13 | 0,0169 |
| Zero drift from field test | $u_{\text{d,z}}$ | -0,299 | 0,0894 |
| Span drift from field test | $u_{\text{d,s}}$ | -1,083 | 1,1729 |
| Influence of ambient temperature at span | u_t | 0,565 | 0,3192 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | 0,77 | 0,5929 |
| Influence of supply voltage | u_v | 0,18 | 0,0324 |
| Cross-sensitivity (interference) | u_i | -0,225 | 0,0506 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,096 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,44 | 0,1936 |
| Uncertainty of reference material 2 % by 70% of CR | u_m | 0,6062 | 0,3675 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 2,8354 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 1,6839 | mg/m ³ |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 3,3004 | mg/m ³ |
| Relativ expanded uncertainty | U | 6,6 | % ELV |
| Permissible uncertainty of EN 15267-3 | (of ELV 50 mg/m ³) | 7,5 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of ELV 50 mg/m ³) | 10 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component NO in the measurement range 0-200 mg/m³

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty mg/m³</i> | <i>Square of standard uncertainty (mg/m³)²</i> |
|--|------------------------------------|--|--|
| Lack-of-fit | u_{lor} | -0,566 | 0,3204 |
| Zero drift from field test | $u_{\text{d,z}}$ | 0,219 | 0,048 |
| Span drift from field test | $u_{\text{d,s}}$ | -1,801 | 3,2436 |
| Influence of ambient temperature at span | u_t | 1,159 | 1,3433 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | -1,08 | 1,1664 |
| Influence of supply voltage | u_v | 0,699 | 0,4886 |
| Cross-sensitivity (interference) | u_i | 1,42 | 2,0164 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,174 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 2,01 | 4,0401 |
| Uncertainty of reference material 2 % by 70% of CR | u_m | 1,6166 | 2,6134 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 15,2802 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 3,909 | mg/m ³ |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 7,6616 | mg/m ³ |
| Relativ expanded uncertainty | U | 5,9 | % ELV |
| Permissible uncertainty of EN 15267-3 | (of ELV 130,4 mg/m ³) | 15 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of ELV 130,4 mg/m ³) | 20 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

**Total uncertainty for the measurement component NO in the measurement range
0-80 mg/m³**

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty mg/m³</i> | <i>Square of standard uncertainty (mg/m³)²</i> |
|--|-----------------------------------|--|--|
| Lack-of-fit | u_{of} | -0,614 | 0,377 |
| Zero drift from field test | $u_{d,z}$ | 0,721 | 0,520 |
| Span drift from field test | $u_{d,s}$ | 0,693 | 0,480 |
| Influence of ambient temperature at span | u_t | 1,076 | 1,158 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | -0,830 | 0,689 |
| Influence of supply voltage | u_v | 0,116 | 0,013 |
| Cross-sensitivity (interference) | u_i | -1,030 | 1,061 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,216 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,634 | 0,402 |
| Uncertainty of reference material 2 % by 70% of CR | u_m | 0,647 | 0,419 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 5,264 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 2,294 | mg/m ³ |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 4,496 | mg/m ³ |
| Relativ expanded uncertainty | U | 13,7 | % ELV |
| Permissible uncertainty of EN 15267-3 | (of ELV 32,7 mg/m ³) | 15 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of ELV 32,7 mg/m ³) | 20 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component NO₂ in the measurement range 0-50 mg/m³

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty mg/m³</i> | <i>Square of standard uncertainty (mg/m³)²</i> |
|--|---------------------------------|--|--|
| Lack-of-fit | u_{lof} | 0,378 | 0,1429 |
| Zero drift from field test | $u_{d,z}$ | 0,127 | 0,0161 |
| Span drift from field test | $u_{d,s}$ | 0,849 | 0,7208 |
| Influence of ambient temperature at span | u_t | 0,445 | 0,198 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | -0,51 | 0,2601 |
| Influence of supply voltage | u_v | 0,31 | 0,0961 |
| Cross-sensitivity (interference) | u_i | 0,289 | 0,0835 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,05 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,620 | 0,3844 |
| Uncertainty of reference material 2 % by 70% of CR | u_m | 0,4041 | 0,1633 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 2,0652 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 1,4371 | mg/m ³ |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 2,8167 | mg/m ³ |
| Relativ expanded uncertainty | U | 5,6 | % ELV |
| Permissible uncertainty of EN 15267-3 | (of ELV 50 mg/m ³) | 15 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of ELV 50 mg/m ³) | 20 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component N₂O in the measurement range 0-50 mg/m³

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty mg/m³</i> | <i>Square of standard uncertainty (mg/m³)²</i> |
|--|---------------------------------|--|--|
| Lack-of-fit | u_{of} | -0,193 | 0,0372 |
| Zero drift from field test | $u_{\text{d,z}}$ | 0,217 | 0,0471 |
| Span drift from field test | $u_{\text{d,s}}$ | -0,854 | 0,7293 |
| Influence of ambient temperature at span | u_t | 0,493 | 0,243 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | -0,410 | 0,1681 |
| Influence of supply voltage | u_v | 0,163 | 0,0266 |
| Cross-sensitivity (interference) | u_i | 0,361 | 0,1303 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,086 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,47 | 0,2209 |
| Uncertainty of reference material 2 % by 70% of CR | u_m | 0,4041 | 0,1633 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{cb} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 1,7658 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 1,3288 | mg/m ³ |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 2,6044 | mg/m ³ |
| Relativ expanded uncertainty | U | 5,2 | % ELV |
| Permissible uncertainty of EN 15267-3 | (of ELV 50 mg/m ³) | 15 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of ELV 50 mg/m ³) | 20 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component SO₂ in the measurement range 0-75 mg/m³

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty mg/m³</i> | <i>Square of standard uncertainty (mg/m³)²</i> |
|--|---------------------------------|--|--|
| Lack-of-fit | u_{lor} | -0,268 | 0,0718 |
| Zero drift from field test | $u_{d,z}$ | 0,16 | 0,0256 |
| Span drift from field test | $u_{d,s}$ | -1,273 | 1,6205 |
| Influence of ambient temperature at span | u_t | 0,748 | 0,5595 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | -0,424 | 0,1798 |
| Influence of supply voltage | u_v | 0,063 | 0,004 |
| Cross-sensitivity (interference) | u_i | 0,524 | 0,2746 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,102 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,88 | 0,7744 |
| Uncertainty of reference material 2 % by 70% of CR | u_m | 0,6062 | 0,3675 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 3,8777 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 1,9692 | mg/m ³ |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 3,8596 | mg/m ³ |
| Relativ expanded uncertainty | U | 7,7 | % ELV |
| Permissible uncertainty of EN 15267-3 | (of ELV 50 mg/m ³) | 15 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of ELV 50 mg/m ³) | 20 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component HCl in the measurement range 0-15 mg/m³

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty mg/m³</i> | <i>Square of standard uncertainty (mg/m³)²</i> |
|--|---------------------------------|--|--|
| Lack-of-fit | u_{of} | -0,172 | 0,0296 |
| Zero drift from field test | $u_{\text{d,z}}$ | 0,146 | 0,0213 |
| Span drift from field test | $u_{\text{d,s}}$ | 0,251 | 0,063 |
| Influence of ambient temperature at span | u_t | 0,158 | 0,025 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | 0,29 | 0,0841 |
| Influence of supply voltage | u_v | 0,093 | 0,0086 |
| Cross-sensitivity (interference) | u_i | 0,235 | 0,0552 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,055 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,22 | 0,0484 |
| Uncertainty of reference material 5 % by 70% of CR | u_{rm} | 0,3031 | 0,0919 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 0,4271 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 0,6535 | mg/m ³ |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 1,2809 | mg/m ³ |
| Relativ expanded uncertainty | U | 12,8 | % ELV |
| Permissible uncertainty of EN 15267-3 | (of ELV 10 mg/m ³) | 30 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of ELV 10 mg/m ³) | 40 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component NH₃ in the measurement range 0-10 mg/m³

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty mg/m³</i> | <i>Square of standard uncertainty (mg/m³)²</i> |
|--|--------------------------------|--|--|
| Lack-of-fit | u_{lof} | 0,114 | 0,013 |
| Zero drift from field test | $u_{d,z}$ | 0,137 | 0,0188 |
| Span drift from field test | $u_{d,s}$ | 0,171 | 0,0292 |
| Influence of ambient temperature at span | u_t | 0,106 | 0,0112 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | -0,057 | 0,0032 |
| Influence of supply voltage | u_v | 0,124 | 0,0154 |
| Cross-sensitivity (interference) | u_i | -0,117 | 0,0137 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,027 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,14 | 0,0196 |
| Uncertainty of reference material 2 % by 70% of CR | u_m | 0,0808 | 0,0065 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_r | | |
| | | total | 0,1306 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 0,3614 | mg/m ³ |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 0,7083 | mg/m ³ |
| Relativ expanded uncertainty | U | 14,2 | % ELV |
| Permissible uncertainty of EN 15267-3 | (of ELV 5 mg/m ³) | 30 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of ELV 5 mg/m ³) | 40 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component TOC in the measurement range 0-15 mg/m³

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty mg/m³</i> | <i>Square of standard uncertainty (mg/m³)²</i> |
|--|---------------------------------|--|--|
| Lack-of-fit | u_{lof} | 0,062 | 0,0038 |
| Zero drift from field test | $u_{d,z}$ | -0,238 | 0,0566 |
| Span drift from field test | $u_{d,s}$ | 0,222 | 0,0493 |
| Influence of ambient temperature at span | u_t | 0,106 | 0,0112 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | 0,022 | 0,0005 |
| Influence of supply voltage | u_v | 0,06 | 0,0036 |
| Cross-sensitivity (interference) | u_i | 0,222 | 0,0493 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,016 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,1 | 0,01 |
| Uncertainty of reference material 2 % by 70% of CR | u_m | 0,1212 | 0,0147 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | 0,226 | 0,0511 |
| | | total | 0,2501 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 0,5001 | mg/m ³ |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 0,9802 | mg/m ³ |
| Relativ expanded uncertainty | U | 9,8 | % ELV |
| Permissible uncertainty of EN 15267-3 | (of ELV 10 mg/m ³) | 22,5 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of ELV 10 mg/m ³) | 30 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component CH₄ in the measurement range 0-50 mg/m³

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty mg/m³</i> | <i>Square of standard uncertainty (mg/m³)²</i> |
|--|---------------------------------|--|--|
| Lack-of-fit | u_{lof} | -0,28 | 0,0784 |
| Zero drift from field test | $u_{d,z}$ | -0,65 | 0,4225 |
| Span drift from field test | $u_{d,s}$ | -0,866 | 0,75 |
| Influence of ambient temperature at span | u_t | 0,286 | 0,0818 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | 0,13 | 0,0169 |
| Influence of supply voltage | u_v | 0,319 | 0,1018 |
| Cross-sensitivity (interference) | u_i | 0,517 | 0,2673 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,055 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,38 | 0,1444 |
| Uncertainty of reference material 2 % by 70% of CR | u_m | 0,4041 | 0,1633 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 2,0264 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 1,4235 | mg/m ³ |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 2,7901 | mg/m ³ |
| Relativ expanded uncertainty | U | 5,6 | % ELV |
| Permissible uncertainty of EN 15267-3 | (of ELV 50 mg/m ³) | 22,5 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of ELV 50 mg/m ³) | 30 | % ELV |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component CO₂ in the measurement range 0-25 Vol.%

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty Vol.%</i> | <i>Square of standard uncertainty (Vol.%)²</i> |
|--|------------------------------|---|---|
| Lack-of-fit | u_{lof} | 0,143 | 0,02045 |
| Zero drift from field test | $u_{d,z}$ | 0,045 | 0,00203 |
| Span drift from field test | $u_{d,s}$ | 0,172 | 0,02958 |
| Influence of ambient temperature at span | u_t | 0,078 | 0,00608 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | 0,018 | 0,00032 |
| Influence of supply voltage | u_v | 0,009 | 0,00008 |
| Cross-sensitivity (interference) | u_i | -0,186 | 0,0346 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,014 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,03 | 0,0009 |
| Uncertainty of reference material 2 % by 70% of CR | u_{rm} | 0,20207 | 0,04083 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 0,13487 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 0,36725 | Vol.% |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 0,71981 | Vol.% |
| Relativ expanded uncertainty | U | 2,9 | % CR |
| Permissible uncertainty of EN 15267-3 | (of CR 25 Vol.%) | 7,5 | % CR |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of CR 25 Vol.%) | 10 | % CR |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |

Total uncertainty for the measurement component H₂O in the measurement range 0-40 Vol.%

| <i>Performance characteristic</i> | <i>Uncertainty</i> | <i>Value standard uncertainty Vol.%</i> | <i>Square of standard uncertainty (Vol.%)²</i> |
|--|------------------------------|---|---|
| Lack-of-fit | u_{of} | -0,157 | 0,0246 |
| Zero drift from field test | $u_{d,z}$ | 0,014 | 0,0002 |
| Span drift from field test | $u_{d,s}$ | 0,621 | 0,3856 |
| Influence of ambient temperature at span | u_t | 0,19 | 0,0361 |
| Influence of sample gas pressure | u_p | | |
| Influence of sample gas flow | u_f | 0,221 | 0,0488 |
| Influence of supply voltage | u_v | 0,074 | 0,0055 |
| Cross-sensitivity (interference) | u_i | 0 | 0 |
| Repeatability standard deviation at span | $u_r = s_r$ | 0,049 | $u_r < u_d$ |
| Standard deviation from paired measurements under field cond. | $u_d = s_d$ | 0,08 | 0,0064 |
| Uncertainty of reference material 2 % by 70% of CR | u_m | 0,3233 | 0,1045 |
| Excursion of measurement beam | u_{mb} | | |
| Converter efficiency for AMS measuring NOx | u_{ce} | | |
| Variation of response factors (TOC) | u_{rf} | | |
| | | total | 0,6117 |
| Combined standard uncertainty | $u_c = \sqrt{\sum (u_i)^2}$ | 0,7821 | Vol.% |
| Total expanded uncertainty | $U_{0,95} = 1,96 \times u_c$ | 1,5329 | Vol.% |
| Relativ expanded uncertainty | U | 3,8 | % CR |
| Permissible uncertainty of EN 15267-3 | (of CR 40 Vol.%) | 7,5 | % CR |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding EN 15267-3 |
| Permissible uncertainty 13. / 17. BImSchV | (of CR 40 Vol.%) | 10 | % CR |
| Complied with requirements relating to the measurement uncertainty | | yes | regarding 13. / 17. BImSchV |