

## TÜV RHEINLAND ENERGY GMBH



Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>

TÜV-Report: 936/21251100/A  
Cologne, 30 July 2021

[www.umwelt-tuv.de](http://www.umwelt-tuv.de)



[tre-service@de.tuv.com](mailto:tre-service@de.tuv.com)

**TÜV Rheinland Energy GmbH and its Ambient Air Quality department in particular**  
is accredited for the following activities:

- Determination of emissions and ambient air quality affected by air pollutants and odorous substances;
- Inspection of correct installation, functionality and calibration of continuous emission monitoring systems including systems for data evaluation and remote monitoring of emissions,
- Measurements in combustion chambers;
- Performance testing of measuring systems for continuous monitoring of emissions and air quality as well as electronic data evaluation and remote monitoring systems for emissions
- Determination of the stack height and air quality forecasts for hazardous and odorous substances;
- Determination of emissions and ambient air quality affected by noise and vibration, determination of sound power levels and noise measurements at wind turbines;

**according to EN ISO/IEC 17025.**

The accreditation has the DAkkS registration number: D-PL-11120-02-00  
and covers the scope specified in the annex to the certificate.

Reproduction of extracts from this test report is subject to prior written consent.

**TÜV Rheinland Energy GmbH**  
**D - 51105 Köln, Am Grauen Stein,**  
**Tel: + 49 (0) 221 806-5200, fax: 0221 806-1349**

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

**Blank page**

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO2 and NOx,  
Report No.: 936/21251100/A

Page 3 of 293



**Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO2 and NOx**

|                             |  |       |           |
|-----------------------------|--|-------|-----------|
| <b>AMS designation:</b>     | N500   |       |           |
| <b>Manufacturer:</b>        | Teledyne API<br>9970 Carrol Canyon Road<br>San Diego, California, 92131<br>USA   |       |           |
| <b>Test period:</b>         | September 2020 to July 2021  |       |           |
| <b>Date of report:</b>      | 30 July 2021   |       |           |
| <b>Report Number:</b>       | 936/21251100/A   |       |           |
| <b>Editor:</b>              | Dipl.-Ing. Martin Schneider<br>Phone: +49 221 806-1614<br><a href="mailto:martin.schneider@de.tuv.com">martin.schneider@de.tuv.com</a> |       |           |
| <b>Scope of the report:</b> | Report:  | 188   | Pages     |
|                             | Manual   | pages | 189       |
|                             | Manual   | with  | 104 Pages |
|                             | Total  | 293   | Pages     |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

**Blank page**

## Table of contents

|   |    |
|---|----|
| 1. SUMMARY AND CERTIFICATION PROPOSAL.....                      | 13 |
| 1.1 Summary Overview .....                                      | 13 |
| 1.2 Certification proposal.....                                 | 15 |
| 1.3 Summary report on test results.....                         | 17 |
| 2. TASK DEFINITION.....   | 25 |
| 2.1 Nature of the test.....                                     | 25 |
| 2.2 Objectives .....  | 25 |
| 3. DESCRIPTION OF THE AMS TESTED .....                          | 26 |
| 3.2 AMS scope and set-up .....                                  | 28 |
| 3.3 AMS adjustment .....  | 31 |
| 4. TEST PROGRAMME .....   | 32 |
| 4.1 General remarks.....  | 32 |
| 4.2 Laboratory test .....                                       | 33 |
| 4.3 Field test.....   | 33 |
| 5. REFERENCE MEASUREMENT METHOD .....                           | 36 |
| 5.1 Method of measurement.....                                  | 36 |
| 6. TEST RESULTS IN ACCORDANCE WITH VDI 4202, PART 1 (2018)..... | 38 |
| 6.1 7.3 General requirements .....                              | 38 |
| 6.1 7.3.1 Measured value display .....                          | 38 |
| 6.1 7.3.2 Calibration inlet .....                               | 39 |
| 6.1 7.3.3 Easy maintenance .....                                | 40 |
| 6.1 7.3.4 Functional check.....                                 | 41 |
| 6.1 7.3.5 Set-up times and warm-up times.....                   | 42 |

|     |   |    |
|-----|---|----|
| 6.1 | 7.3.6 Instrument design .....   | 43 |
| 6.1 | 7.3.7 Unintended adjustment .....   | 44 |
| 6.1 | 7.3.8 Data output .....   | 45 |
| 6.1 | 7.3.9 Digital interface .....   | 46 |
| 6.1 | 7.3.10 Data transmission protocol.....  | 47 |
| 6.1 | 7.3.11 Measuring range .....  | 48 |
| 6.1 | 7.3.12 Negative output signals.....   | 49 |
| 6.1 | 7.3.13 Failure in the mains voltage .....   | 50 |
| 6.1 | 7.3.14 Operating states.....  | 51 |
| 6.1 | 7.3.15 Switch-over .....  | 52 |
| 6.1 | 7.3.16 Instrument software.....   | 53 |
| 6.1 | 7.4 Requirements on performance characteristics for testing in the laboratory ..... | 54 |
| 6.1 | 7.4.1 General requirements .....  | 54 |
| 6.1 | 7.4.2 Test requirements .....   | 55 |
| 6.1 | 7.4.3 Response time and memory effect.....  | 57 |
| 6.1 | 7.4.4 Short-term drift.....   | 58 |
| 6.1 | 7.4.5 Repeatability standard deviation .....  | 59 |
| 6.1 | 7.4.6 Linearity .....   | 60 |
| 6.1 | 7.4.7 Sensitivity coefficient to sample gas pressure .....                          | 61 |
| 6.1 | 7.4.8 Sensitivity coefficient to sample gas temperature.....                        | 62 |
| 6.1 | 7.4.9 Sensitivity coefficient to surrounding temperature .....                      | 63 |
| 6.1 | 7.4.10 Sensitivity coefficient to electrical voltage.....                           | 64 |
| 6.1 | 7.4.11 Cross-sensitivity.....   | 65 |
| 6.1 | 7.4.12 Averaging effect.....  | 72 |
| 6.1 | 7.4.13 Difference between sample and calibration port.....                          | 76 |
| 6.1 | 7.4.14 Converter efficiency .....   | 77 |

|     |   |     |
|-----|---|-----|
| 6.1 | 7.4.15 Residence time in the analyser .....                                   | 78  |
| 6.1 | 7.5 Requirements on performance characteristics for testing in the field..... | 79  |
| 6.1 | 7.5.1 General requirements .....  | 79  |
| 6.1 | 7.5.2 Location for the field test .....                                       | 80  |
| 6.1 | 7.5.3 Test requirements .....   | 81  |
| 6.1 | 7.5.4 Long-term drift .....   | 82  |
| 6.1 | 7.5.5 Reproducibility standard deviation under field conditions .....         | 83  |
| 6.1 | 7.5.6 Inspection interval .....   | 84  |
| 6.1 | 7.5.7 Availability .....  | 85  |
| 6.1 | 7.5.8 Converter efficiency .....  | 86  |
| 6.1 | 7.6 Type approval and calculation of the measurement uncertainty .....        | 89  |
| 7.  | TEST RESULTS IN ACCORDANCE WITH STANDARD EN 14211 (2012).....                 | 90  |
| 7.1 | 8.4.3 Response time .....   | 90  |
| 7.1 | 8.4.4 Short-term drift .....  | 95  |
| 7.1 | 8.4.5 Repeatability standard deviation .....                                  | 101 |
| 7.1 | 8.4.6 Lack of fit of linearity of the calibration function .....              | 105 |
| 7.1 | 8.4.7 Sensitivity coefficient to sample gas pressure .....                    | 112 |
| 7.1 | 8.4.8 Sensitivity coefficient to sample gas temperature .....                 | 115 |
| 7.1 | 8.4.9 Sensitivity coefficient to surrounding temperature .....                | 118 |
| 7.1 | 8.4.10 Sensitivity coefficient to electrical voltage.....                     | 122 |
| 7.1 | 8.4.11 Interferents.....  | 126 |
| 7.1 | 8.4.12 Averaging test.....  | 130 |
| 7.1 | 8.4.13 Difference sample/calibration port.....                                | 134 |
| 7.1 | 8.4.14 Converter efficiency .....   | 136 |
| 7.1 | 8.4.15 Residence time in the analyser .....                                   | 139 |
| 7.1 | 8.5.4 Long-term drift .....   | 140 |

|     |   |     |
|-----|---|-----|
| 7.1 | 8.5.5 Reproducibility standard deviation for NO <sub>2</sub> under field conditions.....    | 145 |
| 7.1 | 8.5.6 Inspection interval .....   | 149 |
| 7.1 | 8.5.7 Period of availability of the analyser .....  | 150 |
| 7.1 | 8.6 Calculation of the total uncertainty in accordance with standard EN 14211 (2012)<br>152 |     |
| 8.  | RECOMMENDATIONS FOR USE IN PRACTICE .....   | 160 |
| 9.  | BIBLIOGRAPHY .....  | 161 |
|     | ANNEXES.....  | 162 |
| 7.1 | Determination of relative uncertainty between test instruments $W_{bs}$ [8.5.3.2] .....     | 163 |
| 7.1 | Comparison with the standard reference method [8.5.3.3] .....                               | 169 |



## List of tables

|           |   |     |
|-----------|---|-----|
| Table 1:  | Measuring ranges tested .....   | 14  |
| Table 2:  | N500 system technical data (manufacturer's data) .....                                      | 31  |
| Table 3:  | Certification ranges VDI 4202-1 and EN 14211 .....  | 48  |
| Table 4:  | Additional cross-sensitivities for the component NO, system 1 .....                         | 66  |
| Table 5:  | Additional cross-sensitivities for the component NO, system 2 .....                         | 66  |
| Table 6:  | Additional cross-sensitivities for the component NO <sub>2</sub> , system 1 .....           | 66  |
| Table 7:  | Additional cross-sensitivities for the component NO <sub>2</sub> , system 2 .....           | 67  |
| Table 8:  | Additional cross-sensitivities for the component NO, system 1 .....                         | 68  |
| Table 9:  | Additional cross-sensitivities for the component NO, system 2 .....                         | 69  |
| Table 10: | Additional cross-sensitivities for the component NO <sub>2</sub> , system 1 .....           | 70  |
| Table 11: | Additional cross-sensitivities for the component NO <sub>2</sub> , system 2 .....           | 71  |
| Table 12: | Results of the averaging test for NO .....  | 73  |
| Table 13: | Results of the averaging test for NO <sub>2</sub> .....                                     | 73  |
| Table 14: | Individual values of the averaging influence study for NO .....                             | 74  |
| Table 15: | Individual values of the averaging influence study for NO <sub>2</sub> .....                | 75  |
| Table 16: | Individual results for the converter efficiency .....                                       | 88  |
| Table 17: | Response times of the two N500 measuring systems for NO .....                               | 92  |
| Table 18: | Response times of the two N500 measuring systems for NO <sub>2</sub> .....                  | 92  |
| Table 19: | Individual results of the response time for NO .....  | 93  |
| Table 20: | Individual results of the response time for NO <sub>2</sub> .....                           | 94  |
| Table 21: | Results of the short-term drift test, component NO .....                                    | 96  |
| Table 22: | Results of the short-term drift test, component NO <sub>2</sub> .....                       | 96  |
| Table 23: | Results of the short-term drift 1 <sup>st</sup> test gas for NO .....                       | 97  |
| Table 24: | Results of the short-term drift 2 <sup>nd</sup> test gas for NO .....                       | 98  |
| Table 25: | Results of the short-term drift 1 <sup>st</sup> test gas for NO <sub>2</sub> .....          | 99  |
| Table 26: | Results of the short-term drift 2 <sup>nd</sup> test gas for NO <sub>2</sub> .....          | 100 |
| Table 27: | Repeatability standard deviation at zero and span point for NO .....                        | 102 |
| Table 28: | Repeatability standard deviation at zero and span point for NO <sub>2</sub> .....           | 102 |
| Table 29: | Repeatability standard deviation for NO .....   | 103 |
| Table 30: | Repeatability standard deviation for NO <sub>2</sub> .....                                  | 104 |
| Table 31: | Deviation from the analytical function for NO .....   | 107 |
| Table 32: | Deviation from the analytical function for NO <sub>2</sub> .....                            | 107 |
| Table 33: | Results of the lack-of-fit test for NO .....  | 110 |
| Table 34: | Results of the lack-of-fit test for NO <sub>2</sub> .....                                   | 111 |
| Table 35: | Sensitivity coefficient to the sample gas pressure for NO .....                             | 113 |
| Table 36: | Sensitivity coefficient to the sample gas pressure for NO <sub>2</sub> .....                | 113 |
| Table 37: | Individual values of the sensitivity to the sample gas pressure for NO .....                | 113 |
| Table 38: | Individual values of the sensitivity to the sample gas pressure for NO <sub>2</sub> .....   | 114 |
| Table 39: | Sample gas temperature sensitivity coefficient for NO .....                                 | 116 |
| Table 40: | Sample gas temperature sensitivity coefficient for NO <sub>2</sub> .....                    | 116 |
| Table 41: | Individual values of the influence of the sample gas temperature for NO .....               | 116 |
| Table 42: | Individual values of the influence of the sample gas temperature for NO <sub>2</sub> .....  | 117 |
| Table 43: | Sensitivity coefficients to ambient temperature for NO .....                                | 119 |
| Table 44: | Sensitivity coefficients to ambient temperature for NO <sub>2</sub> .....                   | 119 |
| Table 45: | Results of the sensitivity coefficient to the ambient temperature for NO .....              | 120 |
| Table 46: | Results of the sensitivity coefficient to the ambient temperature for NO <sub>2</sub> ..... | 121 |
| Table 47: | Sensitivity coefficient to electrical voltage for NO .....                                  | 123 |
| Table 48: | Sensitivity coefficient to electrical voltage for NO <sub>2</sub> .....                     | 123 |
| Table 49: | Results of the sensitivity coefficient to electrical voltage for NO .....                   | 124 |
| Table 50: | Results of the sensitivity coefficient to electrical voltage for NO <sub>2</sub> .....      | 125 |

|           |  |     |
|-----------|--|-----|
| Table 51: | Interferents in accordance with EN 14211 .....   | 127 |
| Table 52: | Influence of the tested interferents for NO ( $c_t = 500 \pm 50$ nmol/mol) .....                       | 127 |
| Table 53: | Influence of the tested interferents for NO <sub>2</sub> ( $c_t = \text{ca. } 104$ nmol/mol) .....     | 127 |
| Table 54: | Individual results for testing interferents for NO .....   | 128 |
| Table 55: | Individual results for testing interferents for NO <sub>2</sub> .....                                  | 129 |
| Table 56: | Results of the averaging test for NO .....   | 131 |
| Table 57: | Results of the averaging test for NO <sub>2</sub> .....  | 131 |
| Table 58: | Results of the averaging tests for NO .....  | 132 |
| Table 59: | Results of the averaging tests for NO <sub>2</sub> .....   | 133 |
| Table 60: | Results of the difference between sample/calibration inlet for NO .....                                | 134 |
| Table 61: | Results of the difference between sample/calibration inlet for NO <sub>2</sub> .....                   | 134 |
| Table 62: | Results for testing the difference between sample and calibration port for NO ...<br>.....             | 135 |
| Table 63: | Results for testing the difference between sample and calibration port for NO <sub>2</sub> ..<br>..... | 135 |
| Table 64: | Individual results for the converter efficiency .....  | 138 |
| Table 65: | Results for the long-term drift at zero point for NO .....   | 141 |
| Table 66: | Results for the long-term drift at span point for NO .....   | 141 |
| Table 67: | Results for the long-term drift at zero point for NO <sub>2</sub> .....                                | 142 |
| Table 68: | Results for the long-term drift at span point for NO <sub>2</sub> .....                                | 142 |
| Table 69: | Results of the drift test for NO .....   | 143 |
| Table 70: | Results of the drift test for NO <sub>2</sub> .....  | 144 |
| Table 71: | Reproducibility standard deviation for NO <sub>2</sub> based on complete field test data               | 146 |
| Table 72: | Reproducibility standard deviation for NO based on complete field test data.                           | 147 |
| Table 73: | Availability of the N500 measuring system .....  | 151 |
| Table 74: | Relevant performance characteristics and criteria according to EN 14211 ....                           | 153 |
| Table 75: | Expanded uncertainty, laboratory test for NO, system 1 .....   | 156 |
| Table 76: | Expanded uncertainty, laboratory and field tests for NO, system 1 .....                                | 156 |
| Table 77: | Expanded uncertainty, laboratory test for NO, system 2 .....   | 157 |
| Table 78: | Expanded uncertainty, laboratory and field tests for NO, system 2 .....                                | 157 |
| Table 79: | Expanded uncertainty, laboratory test for NO <sub>2</sub> , system 1 .....                             | 158 |
| Table 80: | Expanded uncertainty, laboratory and field tests for NO <sub>2</sub> , system 1 .....                  | 158 |
| Table 81: | Expanded uncertainty, laboratory test for NO <sub>2</sub> , system 2 .....                             | 159 |
| Table 82: | Expanded uncertainty, laboratory and field tests for NO <sub>2</sub> , system 2 .....                  | 159 |
| Table 83: | Uncertainty between test systems ( $W_{bs}$ ), NO .....  | 164 |
| Table 84: | Uncertainty between test systems ( $W_{bs}$ ), NO <sub>2</sub> .....                                   | 164 |
| Table 85: | Expanded measurement uncertainties $W_{CM}$ from the field investigations for NO<br>.....              | 171 |
| Table 86: | Expanded measurement uncertainties $W_{CM}$ from the field investigations for NO <sub>2</sub><br>..... | 171 |
| Table 87: | Comparison of the tested and reference instruments, December, NO .....                                 | 172 |
| Table 88: | Comparison of the tested and reference instruments, January, NO .....                                  | 172 |
| Table 89: | Comparison of the tested and reference instruments, May, NO .....                                      | 173 |
| Table 90: | Comparison of the tested and reference instruments, June, NO .....                                     | 173 |
| Table 91: | Comparison of the tested and reference instruments, December, NO <sub>2</sub> .....                    | 174 |
| Table 92: | Comparison of the tested and reference instruments, January, NO <sub>2</sub> .....                     | 174 |
| Table 93: | Comparison of the tested and reference instruments, May, NO <sub>2</sub> .....                         | 175 |
| Table 94: | Comparison of the tested and reference instruments, June, NO <sub>2</sub> .....                        | 175 |
| Table 95: | Meteorological data (daily averages), November 2020 .....  | 180 |
| Table 96: | Meteorological data (daily averages), December 2020 .....  | 181 |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO2 and NOx,  
Report No.: 936/21251100/A

|            |   |     |
|------------|---|-----|
| Table 97:  | Meteorological data (daily averages), January 2021 .....  | 182 |
| Table 98:  | Meteorological data (daily averages), February 2021 ..... | 183 |
| Table 99:  | Meteorological data (daily averages), May 2021 .....      | 184 |
| Table 100: | Meteorological data (daily averages), June 2021 .....     | 185 |

## List of figures

|            |  |     |
|------------|--|-----|
| Figure 1:  | N500 optical absorption cell .....   | 26  |
| Figure 2:  | Phase shift of increased NO <sub>2</sub> concentration .....                           | 27  |
| Figure 3:  | N500 Analysers .....   | 28  |
| Figure 4:  | Internal components of the model N500 .....  | 29  |
| Figure 5:  | N500 pneumatic flow diagram.....   | 30  |
| Figure 6:  | Software version of the N500 tested instruments .....                                  | 32  |
| Figure 7:  | Measuring station used for conducting the field test.....                              | 34  |
| Figure 8:  | N500 in the measuring station during the field test.....                               | 35  |
| Figure 9:  | N500 test system with measured value display.....                                      | 38  |
| Figure 10: | Diagram illustrating the response time .....   | 91  |
| Figure 11: | Analysis function from the group average values for system 1 for NO .....              | 107 |
| Figure 12: | Analysis function from the group average values for system 2 for NO .....              | 108 |
| Figure 13: | Analysis function from the group average values for system 1 for NO <sub>2</sub> ..... | 108 |
| Figure 14: | Analysis function from the group average values for system 2 for NO <sub>2</sub> ..... | 109 |
| Figure 15: | Test of the averaging effect ( $t_{NO} = t_{zero} = 45$ s.) .....                      | 131 |
| Figure 16: | Reproducibility standard deviation for NO <sub>2</sub> under field conditions .....    | 146 |
| Figure 17: | Reproducibility standard deviation for NO under field conditions .....                 | 148 |
| Figure 18: | Parallel measurements with the test systems, December, NO .....                        | 165 |
| Figure 19: | Parallel measurements with the test systems, January, NO .....                         | 165 |
| Figure 20: | Parallel measurements with the test systems, May, NO .....                             | 166 |
| Figure 21: | Parallel measurements with the test systems, June, NO .....                            | 166 |
| Figure 22: | Parallel measurements with the test systems, December, NO <sub>2</sub> .....           | 167 |
| Figure 23: | Parallel measurements with the test systems, January, NO <sub>2</sub> .....            | 167 |
| Figure 24: | Parallel measurements with the test systems, May, NO <sub>2</sub> .....                | 168 |
| Figure 25: | Parallel measurements with the test systems, June, NO <sub>2</sub> .....               | 168 |
| Figure 26: | Reference vs. tested instrument, December, NO .....                                    | 176 |
| Figure 27: | Reference vs. tested instrument, January, NO .....                                     | 176 |
| Figure 28: | Reference vs. tested instrument, May, NO .....   | 177 |
| Figure 29: | Reference vs. tested instrument, June, NO .....  | 177 |
| Figure 30: | Reference vs. tested instrument, December, NO <sub>2</sub> .....                       | 178 |
| Figure 31: | Reference vs. tested instrument, January, NO <sub>2</sub> .....                        | 178 |
| Figure 32: | Reference vs. tested instrument, May, NO <sub>2</sub> .....                            | 179 |
| Figure 33: | Reference vs. tested instrument, June, NO <sub>2</sub> .....                           | 179 |
| Figure 34: | Certificate of accreditation according to EN ISO/IEC 17025:2005 .....                  | 186 |

## **1. Summary and certification proposal**

### **1.1 Summary Overview**

On behalf of Teledyne API, TÜV Rheinland Energy GmbH carried out the performance test of the N500 AMS for the components NO, NO<sub>2</sub> and NO<sub>x</sub>. The test was performed in respect of the following standards and requirements:

- VDI Guideline 4202 part 1: Performance test, declaration of suitability, and certification of point-related measuring systems for gaseous air pollutants of April 2018
- EN 14211: Ambient air – Standard method for the measurement of the concentration of nitrogen dioxide and nitrogen monoxide by chemiluminescence, of November 2012
- Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods of January 2010

The N500 measuring system determines the components NO, NO<sub>2</sub> and NO<sub>x</sub> directly by means of CAPS (Cavity-Attenuated Phase Shift) spectroscopy and conversion by gas phase titration. Thus, the measuring principle does not correspond to the to chemiluminescence as the EU reference method. Annex 1 presents the field test results of comparison measurements with the reference method for NO<sub>x</sub> (chemiluminescence). Measurement results were evaluated in accordance with the Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods.

In contrast with a CLD analyser, the instrument's measuring principle determines NO<sub>2</sub> and NO directly. The output NO<sub>x</sub> value as the sum of NO and NO<sub>2</sub> is always calculated. This is why NO and NO<sub>2</sub> were considered separately for every test item reported here. NO was tested in accordance with the requirements of EN 14211. For the purpose of testing NO<sub>2</sub> (tested measuring range 0–500 µg/m<sup>3</sup> (0–261 nmol/mol)), test gas target values were adapted to the NO<sub>2</sub> measuring range. This, too, is specified in standard EN 14211. An additional cross-sensitivity test was performed for NO<sub>2</sub>.

The tests were carried out in the laboratory of TÜV Rheinland Energy GmbH and during a three-month field test near Cologne. The measuring ranges are indicated below:

Table 1: Measuring ranges tested

| Measured component | Measuring range in [ $\mu\text{g}/\text{m}^3$ ] <sup>1</sup> | Measuring range in [ppb] or [nmol/mol] |
|--------------------|--|--|
| NO                 | 0–1 200  | 0–962                                  |
| NO <sub>2</sub>    | 0–500  | 0–261                                  |

<sup>1</sup> The specifications refer to 20 °C and 101.3 kPa

During performance testing, the AMS met the requirements specified in standard EN 14211 (2012) and VDI 4202, part 1 (2018).

TÜV Rheinland Energy GmbH therefore recommend the instrument's approval as a performance-tested measuring system for continuous monitoring of air quality affected by nitrogen oxides.

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Page 15 of 293

## 1.2 Certification proposal

Based on the positive results obtained, the following recommendation on the announcement of the AMS as a certified system is put forward:

### AMS designation:

N500 for NO, NO<sub>2</sub> and NO<sub>x</sub>

### Manufacturer:

Teledyne API, San Diego, USA

### Field of application:

For the continuous determination of ambient air concentrations of nitrogen oxides in stationary use.

### Measuring ranges during performance testing:

| Component         | Certification range | Unit              |
|-------------------|---------------------|-------------------|
| Nitrogen monoxide | 0 – 1 200           | µg/m <sup>3</sup> |
| Nitrogen dioxide  | 0 - 500             | µg/m <sup>3</sup> |

### Software version:

Rev. 1.6.0

### Restrictions:

None

### Note:

1. This report on the performance test is available online at [www.gal1.de](http://www.gal1.de).
2. Equivalence with the reference method in accordance with the requirements of the Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods was demonstrated.

### Test Report:

TÜV Rheinland Energy GmbH, Cologne  
Report no. 936/21251100/A of 30 July 2021

Blank page



### 1.3 Summary report on test results

| Performance criterion                | Requirement  | Test result  | Satisfied      | Page |
|--------------------------------------|--|--|----------------|------|
| <b>7 Performance criteria</b>        |  |  |                |      |
| <b>7.3 General requirements</b>      |  |  |                |      |
| 7.3.1 Measured value display         | The measuring system shall have an operative measured value display as part of the instrument.   | The measuring system has an operative measured value display at the instrument front.  | yes            | 38   |
| 7.3.2 Calibration inlet              | The measuring system may have a test gas inlet separate from the sample gas inlet.   | The measuring system has a test gas inlet separate from the sample gas inlet at the instrument back.                                     | yes            | 39   |
| 7.3.3 Easy maintenance               | It should be possible to carry out maintenance work from the outside without major effort.   | Maintenance takes reasonable effort and is possible with standard tools from the outside.  | yes            | 40   |
| 7.3.4 Functional check               | Particular instruments required to this effect shall be considered as part of the measuring system and be applied in the corresponding sub-tests and included in the assessment.   | The tested measuring system does not have internal devices for operating the functional check.   | not applicable | 41   |
| 7.3.5 Set-up times and warm-up times | The instruction manual shall include specifications in this regard.  | Set-up times and warm-up times have been determined.   | yes            | 42   |
| 7.3.6 Instrument design              | The instruction manual shall include specifications in this regard.  | Specifications made in the instruction manual concerning instrument design are complete and correct.                                     | yes            | 43   |
| 7.3.7 Unintended adjustment          | It shall be possible to secure the adjustment of the measuring system against illicit or unintended adjustment during operation. Alternatively, the user manual shall specifically note that the measuring system may only be installed in a secured area. | The measuring system is secured against unintended and unauthorised adjustment of instrument parameters by way of a password.            | yes            | 44   |
| 7.3.8 Data output                    | The output signals shall be provided digitally and/or as analogue signals.   | The measuring signals are offered in analogue (0 – 20 mA, 4 – 20 mA or 0 – 1V, 0 – 5 V, 0 – 10 V) and digital (via TCP/IP, RS 232, USB). | yes            | 45   |
| 7.3.9 Digital interface              | The digital interface shall allow the transmission of output signals, status signals, and others.<br><br>Access to the measuring system shall be secured against unauthorised access.  | Digital transmission of measured values operates correctly.  | yes            | 46   |

| Performance criterion               | Requirement   | Test result  | Satisfied | Page |
|-------------------------------------|---|--|-----------|------|
| 7.3.10 Data transmission protocol   | Shall meet the requirements stipulated in Table 1 of VDI Guideline 4202 part 1.   | The measuring system has an installed Modbus "Bavaria/Hesse" (Bayern/Hessen) transmission protocol as standard. Measured and status signals are transmitted correctly. The configuration is listed in the manual in Appendix A.2   | yes       | 47   |
| 7.3.11 Measuring range              | The upper limit of measurement shall be greater or equal to the upper limit of the certification range.   | By default, the measuring range is set to 0–500 µg/m <sup>3</sup> for NO <sub>2</sub> and 0–1200 µg/m <sup>3</sup> for NO. Other measuring ranges up to a maximum of 0 – 1 are possible.<br><br>The measuring system's upper limit of measurement exceeds the upper limit of the certification range in each case. | yes       | 48   |
| 7.3.12 Negative output signals      | May not be suppressed (live zero).  | The measuring system also provides negative output signals.  | yes       | 49   |
| 7.3.13 Failure in the mains voltage | Uncontrolled emission of operation and calibration gas shall be avoided; instrument parameters shall be secured by buffering against loss; when mains voltage returns, the instrument shall automatically reach the operation mode and start the measurement. | The measuring system is in fault-free operational readiness when the voltage returns and automatically resumes measuring operation after a warm-up phase.  | yes       | 50   |
| 7.3.14 Operating states             | The measuring system shall allow their control by telemetrically transmitted status signals.  | The measuring system provides various ports to ensure comprehensive monitoring and control via an external computer.   | yes       | 51   |
| 7.3.15 Switch-over                  | Switch-over between measurement and functional check and/or calibration shall be possible telemetrically.   | In principle, all necessary work for function control can be monitored and controlled directly on the unit or by telemetric remote control.  | yes       | 52   |
| 7.3.16 Instrument software          | Shall be displayed when switched on. Changes affecting instrument functions shall be communicated to the test laboratory.   | The instrument's software version is displayed. Software changes are communicated to the test laboratory.  | yes       | 53   |

| Performance criterion   | Requirement   | Test result  | Satisfied | Page |
|---|---|--|-----------|------|
| <b>7.4 Requirements on performance characteristics for testing in the laboratory</b>                          |   |  |           |      |
| 7.4.1 General requirements  | The manufacturer's specifications in the instruction manual shall not contradict the results of the performance test.   | Tests were performed using the performance characteristics specified in VDI standard 4202, part 1 (2018) and standard EN 14211 (2012)                            | yes       | 54   |
| 7.4.2 Test requirements   | Has to comply with the requirements set out in VDI standard 4202-1:2018.  | Tests were performed using the performance characteristics specified in VDI standard 4202, part 1 (2018) and standard EN 14211 (2012)                            | yes       | 55   |
| Section 8.4 provides a summary of the evaluation of performance characteristics determined in the laboratory. |   |  |           |      |
| <b>7.5 Requirements on performance characteristics for testing in the field</b>                               |   |  |           |      |
| 7.5.1 General requirements  | Has to comply with the requirements set out in VDI standard 4202-1:2018.  | Tests were performed using the performance characteristics specified in VDI standard 4202, part 1 (2018) and standard EN 14211 (2012)                            | yes       | 79   |
| 7.5.2 Location for the field test   | The monitoring station for the field test is to be chosen according to the requirements of 39 <sup>th</sup> BImSchV such that the expected concentrations of the measured components to be measured correspond to the designated task. The equipment of the monitoring station shall allow the implementation of the field test and shall fulfil all requirements considered to be necessary during measurement planning. | The field test location was selected in compliance with the 39 <sup>th</sup> BImSchV. Details on the location of the measuring station are given in chapter 4.3. | yes       | 80   |
| 7.5.3 Test requirements   | The measuring systems shall be installed in the monitoring station and, after connecting to the existing or separate sampling system, activated properly.<br><br>The adjustments of the measuring system shall meet the specifications of the manufacturer. All adjustments are to be documented in the test report.  | During the field test, the measuring system was operated and serviced according to the manufacturer's instructions.  | yes       | 81   |
| Section 8.5 provides a summary of the evaluation of performance characteristics determined in the laboratory. |   |  |           |      |

| Performance criterion   | Requirement  | Test result   | Satisfied | Page |
|---|--|---|-----------|------|
| <b>8.4 Procedures for determination of the performance characteristics during the laboratory test according to EN 14211</b> |  |   |           |      |
| 8.4.3 Response time   | Rise and fall response time ≤ 180 s each. Difference between rise and fall response time ≤ 10 s.   | The values determined remained considerably below the maximum permissible response time of 180 s at all times. The maximum determined response time for unit 1 was 24 sec for NO and 24 sec for NO <sub>2</sub> . For unit 2 it was 25 sec for NO and 25 sec for NO <sub>2</sub> .  | yes       | 90   |
| 8.4.4 Short-term drift  | The short-term drift at zero must be ≤ 2.0 nmol/mol/12 h.<br>The short-term drift at span level must be ≤ 6.0 nmol/mol/12 h.   | For NO, the value for the short-term drift at zero point was 0.06 nmol/mol for instrument 1; for instrument 2, it was -0.27 nmol/mol.<br>For NO, the value for the short-term drift at span point was 2.36 nmol/mol for instrument 1; for instrument 2, it was 1.11 nmol/mol.<br>For NO <sub>2</sub> , the value for the short-term drift at zero point was -0.03 nmol/mol for instrument 1; for instrument 2, it was -0.26 nmol/mol.<br>For NO <sub>2</sub> , the value for the short-term drift at span point was 2.76 nmol/mol for instrument 1; for instrument 2, it was 2.72 nmol/mol. | yes       | 95   |
| 8.4.5 Repeatability standard deviation  | The performance criteria are as follows: Repeatability standard deviation at zero shall not exceed 1.0 nmol/mol. At a sample gas concentration at the span point it shall not exceed 3.0 nmol/mol. | For NO, the value for the repeatability standard deviation at zero point was 0.13 nmol/mol for instrument 1; for instrument 2 it was 0.13 nmol/mol. Repeatability standard deviation at span point was 1.63 nmol/mol for instrument 1 and 1.34 nmol/mol for instrument 2.<br>For NO <sub>2</sub> , the value for the repeatability standard deviation at zero point was 0.00 nmol/mol for instrument 1 and 0.11 nmol/mol for instrument 2. Repeatability standard deviation at span point was 0.08 nmol/mol for instrument 1 and 0.08 nmol/mol for instrument 2.                            | yes       | 101  |

| Performance criterion                                      | Requirement   | Test result   | Satisfied | Page |
|--|---|---|-----------|------|
| 8.4.6 Lack of fit of linearity of the calibration function | The deviation from the linearity of the calibration function at zero shall not exceed 5.0 nmol/mol. At concentrations above zero, it shall not exceed 4% of the measured value. | <p>Component NO</p> <p>The deviation from the linear regression line for instrument 1 is -0.60 nmol/mol at zero point and no more than 1.72% of the rated value for concentrations above zero. The deviation from the linear regression line for instrument 2 is -0.60 nmol/mol at zero point and no more than 2.24% of the rated value for concentrations above zero.</p> <p>Component NO<sub>2</sub></p> <p>The deviation from the linear regression line for instrument 1 is -0.16 nmol/mol at zero point and no more than 0.85% of the rated value for concentrations above zero. The deviation from the linear regression line for instrument 2 is -0.49 nmol/mol at zero point and no more than 0.72% of the rated value for concentrations above zero.</p> | yes       | 105  |
| 8.4.7 Sensitivity coefficient to sample gas pressure       | The sensitivity coefficient to sample gas pressure shall be $\leq 8.0$ nmol/mol/kPa.  | <p>For NO, the sensitivity coefficient to sample gas pressure was 0.23 nmol/mol/kPa for instrument 1; for instrument 2 it was 0.17 nmol/mol/kPa.</p> <p>For NO<sub>2</sub>, the sensitivity coefficient to sample gas pressure was 0.10 nmol/mol/kPa for instrument 1; for instrument 2 it was 0.08 nmol/mol/kPa.</p>   | yes       | 112  |

| Performance criterion                                    | Requirement  | Test result   | Satisfied | Page |
|--|--|---|-----------|------|
| 8.4.8 Sensitivity coefficient to sample gas temperature  | The sensitivity coefficient to sample gas temperature shall be $\leq 3.0 \mu\text{mol/mol/K}$ .      | For instrument 1, the sensitivity coefficient to sample gas temperature was $0.09 \text{ nmol/mol/K}$ for NO; for NO <sub>2</sub> , it was $0.11 \text{ nmol/mol/K}$ .<br><br>For instrument 2, the sensitivity coefficient to sample gas temperature was $0.01 \text{ nmol/mol/K}$ for NO; for NO <sub>2</sub> , it was $0.01 \text{ nmol/mol/K}$ .  | yes       | 115  |
| 8.4.9 Sensitivity coefficient to surrounding temperature | The sensitivity coefficient to the surrounding temperature shall be $\leq 3.0 \mu\text{mol/mol/K}$ . | The sensitivity coefficient to the surrounding temperature did not exceed the performance criterion specified at $3.0 \text{ nmol/mol/K}$ . For the purpose of uncertainty calculation, the largest value is used for both instruments.<br>For NO, this would be $0.952 \text{ nmol/mol/K}$ for instrument 1 and $0.451 \text{ nmol/mol/K}$ for instrument 2.<br>For NO <sub>2</sub> , this would be $0.231 \text{ nmol/mol/K}$ for instrument 1 and $0.241 \text{ nmol/mol/K}$ for instrument 2. | yes       | 118  |
| 8.4.10 Sensitivity coefficient to electrical voltage     | The sensitivity coefficient to electrical voltage shall not exceed $0.30 \text{ nmol/mol/V}$ .       | At no test item did the sensitivity coefficient to electrical voltage exceed the value of $0.3 \text{ nmol/mol/V}$ specified in standard EN 14211. For the purpose of the uncertainty calculation, the largest value is used for both instruments.<br>For NO, this is $0.01 \text{ nmol/mol/V}$ for instrument 1 and $0.00 \text{ nmol/mol/V}$ for instrument 2.<br>For NO <sub>2</sub> , this is $0.02 \text{ nmol/mol/V}$ for instrument 1 and $0.01 \text{ nmol/mol/V}$ for instrument 2.      | yes       | 122  |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Page 23 of 293

| Performance criterion | Requirement   | Test result  | Satisfied | Page |
|-----------------------|---|--|-----------|------|
| 8.4.11 Interferents   | Interferents at zero and at concentration $c_i$ for NO ( $500 \pm 50$ nmol/mol). Deviations for interferents H <sub>2</sub> O, CO <sub>2</sub> and NH <sub>3</sub> shall not exceed 5.0 nmol/mol. | <p>For NO at zero point the result for the interference were 0.00 nmol/mol (AMS 1) and -0.60 nmol/mol (AMS 2) for H<sub>2</sub>O, 0.40 nmol/mol (AMS 1) and 0.40 nmol/mol (AMS 2) for CO<sub>2</sub> and finally 0.60 nmol/mol (AMS 1) and 1.00 nmol/mol (AMS 2) for NH<sub>3</sub>. For NO<sub>2</sub> at zero point the result for the interference were 0.27 nmol/mol (AMS 1) and 0.00 nmol/mol (AMS 2) for H<sub>2</sub>O, 0.00 nmol/mol (AMS 1) and -0.21 nmol/mol (AMS 2) for CO<sub>2</sub> and finally 0.00 nmol/mol (AMS 1) and -0.16 nmol/mol (AMS 2) for NH<sub>3</sub>.</p> <p>The following results were obtained for the cross-sensitivity at the limit value <math>c_t</math>: NO: 0.40 nmol/mol for instrument 1 and 1.60 nmol/mol for instrument 2; H<sub>2</sub>O: 1.80 nmol/mol for instrument 1 and 0.80 nmol/mol for instrument 2; CO<sub>2</sub> and 1.20 nmol/mol for instrument 1 and 1.40 nmol/mol at NH<sub>3</sub>. The following results were obtained for the cross-sensitivity at the limit value <math>c_t</math>: NO<sub>2</sub>: 0.33 nmol/mol for instrument 1 and 0.22 nmol/mol for instrument 2; H<sub>2</sub>O: 0.70 nmol/mol for instrument 1 and 0.76 nmol/mol for instrument 2; CO<sub>2</sub> and 1.09 nmol/mol for instrument 1 and 0.98 nmol/mol at NH<sub>3</sub>.</p> | yes       | 126  |
| 8.4.12 Averaging test | The averaging effect shall not exceed 7% of the measured value.   | The performance criterion specified by standard EN 14211 is fully satisfied.   | yes       | 130  |

| Performance criterion   | Requirement   | Test result   | Satisfied | Page |
|---|---|---|-----------|------|
| 8.4.13 Difference sample/calibration port   | The difference between the sample and calibration ports shall not exceed 1%.  | The performance criterion specified by standard EN 14211 is fully satisfied.  | yes       | 134  |
| 8.4.14 Converter efficiency   | The converter efficiency shall be at least 98%.   | At a converter efficiency of 99.6%, the performance criterion specified by EN 14211 is fully satisfied.   | yes       | 136  |
| 8.4.15 Residence time in the analyser   | The residence time in the analyser shall not exceed 3.0 s.  | Residence time in the analyser was 1.9 s.   | yes       | 136  |
| <b>8.5 Determination of the performance characteristics during the field test according to EN 14211</b> |   |   |           |      |
| 8.5.4 Long-term drift   | The long-term drift at zero point shall not exceed $\leq 5.0$ nmol/mol. Long-term drift at span level shall not exceed 5% of the certification range. | Maximum long-term drift at zero point $DI_z$ for NO was at -0.92 nmol/mol for instrument 1 and 1.08 nmol/mol for instrument 2. Maximum long-term drift at span point $DI_s$ for NO was at 0.93% for instrument 1 and 1.02% for instrument 2.<br><br>Maximum long-term drift at zero point $DI_z$ for NO <sub>2</sub> was at 0.75 nmol/mol for instrument 1 and 0.87 nmol/mol for instrument 2. Maximum long-term drift at span point $DI_s$ for NO <sub>2</sub> was at 2.29% for instrument 1 and 2.05% for instrument 2. | yes       | 140  |
| 8.5.6 Inspection interval   | The period of unattended operation of the AMS shall be at least 2 weeks.  | The necessary maintenance tasks determine the period of unattended operation. In essence, these include contamination checks, plausibility checks and checks of potential status/error warnings. The external particle filter must be changed depending on the dust load at the measuring location. EN 14211 requires checking of zero and span points at least once every two weeks.   | yes       | 149  |
| 8.5.5 Reproducibility standard deviation for NO <sub>2</sub> under field conditions                     | Reproducibility standard deviation under field conditions shall not exceed 5% of the mean value over a period of three months.                        | The reproducibility standard deviation for NO <sub>2</sub> under field conditions was 3.67% as a percentage of the mean value over the three-month field test period. Thus, the requirements of EN 14211 are satisfied.   | yes       | 145  |
| 8.5.7 Period of availability of the analyser  | Availability of the analyser shall be at least 90%.   | The availability is at 100%. Thus, the requirement of EN 14211 is satisfied.  | yes       | 150  |



## 2. Task definition

### 2.1 Nature of the test

On behalf of the company Teledyne API, a performance test for the measuring system N500 was carried out by TÜV Rheinland Energy GmbH. The test was carried out as a complete performance test.

### 2.2 Objectives

The AMS is designed to determine nitrogen oxide concentrations in ambient air in the following concentration ranges:

| Component         | Certification range | Unit              |
|-------------------|---------------------|-------------------|
| Nitrogen monoxide | 0 – 1 200           | µg/m <sup>3</sup> |
| Nitrogen dioxide  | 0 - 500             | µg/m <sup>3</sup> |

The N500 measuring system measures the components NO, NO<sub>2</sub> and NO<sub>x</sub> by means of direct UV absorption.

The task was to carry out performance testing in line with the applicable standards and taking into consideration the latest developments in the field.

The test was performed on the basis of the following standards:

- VDI Guideline 4202 part 1: Automated measuring systems for air quality monitoring – Performance test, declaration of suitability, and certification of point-related measuring systems for gaseous air pollutants, April 2018
- EN 14211: Ambient air – Standard method for the measurement of the concentration of nitrogen dioxide and nitrogen monoxide by chemiluminescence, of November 2012
- Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods of January 2010

### 3. Description of the AMS tested

#### 3.1 Measuring principle

The N500 analyser is an optical absorption spectrometer that can directly measure NO<sub>2</sub> using the Cavity Attenuated Phase Shift (CAPS) method. The CAPS method uses blue UV light from an LED with a wavelength of 405 nm, a measuring cell with highly reflective mirrors on both sides to extend the optical path, and a vacuum photo detector. All components are integrated in the optical measuring cell, which is located in a thermostated area. This area is heated to 45 °C to prevent moisture on the mirrors and the effects of fluctuating ambient temperatures.

The NO<sub>2</sub> is determined directly by optical absorption. This measurement principle is laid down in Lambert-Beer's law. The absorption (light loss) is directly proportional to the light path and the concentration of the absorbing gas.

$$A = \epsilon l c$$

(A = absorption,  $\epsilon$  = molar absorption coefficient, l = light path length, c = concentration)

Ultraviolet light from the modulating LED is sent into the cell behind mirror A. The light intensity is then measured by the detector, which itself is modulated with a slightly different frequency. The detector sits behind mirror B and measures an exponentially increasing signal when the LED is switched on. When the LED is switched off, there is also a drop in intensity. Since both mirrors are highly reflective at 405 nm (the highest absorption range of NO<sub>2</sub>), the light needs a certain time to reach the maximum in the absence of the absorbing gas. In the presence of NO<sub>2</sub>, the path of light is significantly shortened. This has two effects on the light intensity measured by the detector:

- The maximum of light is smaller
- The maximum of light is reached earlier.




Figure 1: N500 optical absorption cell

This measures a phase shift to the modulating LED. The phase shift is greatest at zero gas and decreases when NO<sub>2</sub> is present.

Both the LED and the detector are modulated to produce the measured signal at a much lower frequency (corresponding to the difference between the modulating frequencies). This signal is then easier for the system's hardware to process.

The system transfers the phase shift into a concentration. Using the CAPS method, the phase shift remains constant for each concentration, even if the LED ages and loses intensity over time.




Figure 2: Phase shift of increased NO<sub>2</sub> concentration

An internal vacuum pump, located downstream from the rest of the instrument's components, draws test gas through the instrument's pneumatic connections. The flow rate is controlled by a flow limiter upstream of the sensor.

After the test gas has passed through a series of filtering and conditioning components, a sensor determines the NO<sub>2</sub> reading. The test gas is titrated alternately with high concentrations of ozone. A sensor then measures a second, higher NO<sub>2</sub> value (NO<sub>x</sub> mode). The NO<sub>x</sub> concentration and the NO concentration are derived from this value.

In summary, the N500 analyser measures NO<sub>2</sub> first directly by optical absorption to obtain an actual reading of the NO<sub>2</sub> concentration. Then a precisely timed pulse rate is used to mix the sample with a high concentration of O<sub>3</sub> to generate NO<sub>x</sub>. The software then subtracts the NO<sub>2</sub> concentration from the output NO<sub>x</sub> concentration to calculate a reading for NO. At the NO concentration, compensation is then made for the titration efficiency, which is then added to the measured NO<sub>2</sub> value to produce the corrected NO<sub>x</sub> reading.

### 3.2 AMS scope and set-up

The main components of the N500 analyser are: an optical cell, a pair of highly reflective mirrors at 405 nm, a light emitting diode (LED) light source and a vacuum photodiode detector.

The LED is behind a mirror at one end of the cell, and the detector is behind the other mirror at the opposite end of the cell. The light emitting diode sends light pulses in the ultraviolet spectrum into the measuring cell. The light is reflected by the mirrors over and over again, creating a very long path length. This long path of light extends the "lifetime" of the photon using a data acquisition system timed to the measurement. Together with an algorithm, the measured absorption is converted into a phase offset from which the NO<sub>2</sub> concentration is calculated.



Figure 3: N500 Analysers

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A




Figure 4: Internal components of the model N500




Figure 5: N500 pneumatic flow diagram

Table 2 contains a list of important technical characteristics of the N500.

Table 2: N500 system technical data (manufacturer's data)

|                       |   |
|-----------------------|---|
| Measured range:       | Maximum 0 - 1 ppm for NO <sub>2</sub><br>Maximum 0 - 1 ppm for NO   |
| Units:                | ppb, ppm, mg/m <sup>3</sup> or µg/m <sup>3</sup>  |
| Measured compounds:   | NO, NO <sub>2</sub> and NO <sub>x</sub>   |
| Sample flow rate      | ~1.0 l/min (during the test)  |
| Outputs:              | <ul style="list-style-type: none"> <li>• Ethernet TCP/IP</li> <li>• Serial interface, RS232 or RS 485</li> <li>• 0 – 5/10 Volt analogue</li> <li>• 4 – 20 mA analogue</li> <li>• USB</li> </ul> |
| Input voltage:        | 230 V or 115 V<br>50Hz or 60 Hz   |
| Power:                | 110 W; 190 W max.   |
| Dimension (l x w x h) | 597 x 732 x 178 mm / ~ 15 kg  |

### 3.3 AMS adjustment

The measuring system was commissioned according to manufacturer instructions. No internal zero adjustment cycles were activated during the performance test.

The N500 analyser firmware processes sample concentration data via a programmed adaptive filter. During operation, the firmware can automatically switch between two different temporal filter lengths depending on the conditions. When measuring stable concentrations, the firmware calculates an average over the last measurements by default. This ensures reliable measurement results. If the filter detects rapid changes in concentration, it reduces the averaging time to allow the analyser to respond more quickly. The adaptive filter is always active and can neither be adjusted nor deactivated by the user.

## 4. Test programme

### 4.1 General remarks

Two identical N500 instruments with the following serial numbers were submitted to performance testing:

Instrument 1: SN 65 and  
Instrument 2: SN 76

The tests were performed with software version 1.6.0.

The test comprised a laboratory test to determine the performance characteristics as well as a field test over a period of several months.

In this report, the heading for each performance criterion cites the requirements according to the relevant standards ([1, 2, 3]) including its chapter number and wording.




Figure 6: Software version of the N500 tested instruments



## 4.2 Laboratory test

The laboratory test was carried out with two identical instruments, type N500, with serial numbers S/N: 65 and SN: 76. Standards [1] and [2] specify the following test programme for the laboratory test:

- Description of instrument functions
- General requirements
- Calibration line fit
- Short-term drift
- Repeatability standard deviation
- Sensitivity to sample gas pressure
- Sensitivity to sample gas temperature
- Sensitivity to surrounding temperature
- Sensitivity to supply voltage
- Cross-sensitivities
- Averaging effect
- Response time
- Difference sample/calibration inlet
- Converter efficiency

The measuring system does not work according to the EU standard reference method (chemiluminescence), but alternately analyses NO<sub>2</sub> and NO directly due to its measuring principle (CAPS). The output for the NO<sub>x</sub> value is always a calculated result. This is why NO and NO<sub>2</sub> were considered separately for every test item reported here. NO was tested in accordance with the requirements of EN 14211. For the purpose of testing NO<sub>2</sub> (tested measuring range 0–500 µg/m<sup>3</sup> (0–261 nmol/mol)), test gas target values were adapted to the NO<sub>2</sub> measuring range. This, too, is specified in standard EN 14211. An additional cross-sensitivity test was performed for NO<sub>2</sub>.

Measured values were recorded using an external data logger.  
Chapters 6 and 7 summarize the results of the laboratory tests.

## 4.3 Field test

The field test in accordance with EN 14211 and VDI 4202-1 was carried out with two identical measuring systems of type N500 from 09 November 2020 to 15 February 2021. Subsequently, the field measurements were continued until 01 July 2021 in order to collect further measurement data for evaluation according to the guide "Demonstration of equivalence of ambient air monitoring methods". The instruments used were identical with those used for laboratory testing. The serial numbers were:

Instrument 1: SN 65  
Instrument 2: SN 76

The following test programme was determined for the field test:

- Long-term drift
- Period of unattended operation
- Availability
- Reproducibility standard deviation
- Converter efficiency (VDI 4202-1:2018)

As with the procedure during the laboratory test, the evaluation of the field test is also listed separately for NO and NO<sub>2</sub> in this report.

Measured values were recorded using an external data logger.

Chapters 6 and 7 summarize the results of the field tests.

To carry out the field test, the measuring systems were installed in a measuring station near Cologne. The measuring station is located in the direct vicinity of a busy, six-lane national motorway. The distance from the measuring station to the nearest roadway is approx. 5.0 m



Figure 7: Measuring station used for conducting the field test

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Page 35 of 293

The measuring instruments were installed in the station, which was air-conditioned to 20 °C, in a 19" rack and connected to a sampling system. Both systems were installed with a sample gas line of 2.5 m through a T-piece at the same connection of the sampling system. This ensured that both systems performed a representative measurement of the same sample air volume. The sampling head is located on the front side of the measuring station at a total height of approx. 3.00 m.



Figure 8: N500 in the measuring station during the field test

## 5. Reference Measurement Method

### 5.1 Method of measurement

#### Test gases used for adjustment purposes during the test

Certified nitrogen monoxide and nitrogen dioxide test gases were used to test the performance parameters. The specified test gases were used during the entire test and, where necessary, were diluted with the help of a (type Hovacal) mass flow controller.

The test gas SN: Test gas bottle (S/N 64616) was traced back by the national EU reference laboratory for ambient air quality (Federal Environment Agency in Langen). Quality assurance of test gases used was based on the traceable test gas in the TRE laboratory.

|   |   |
|---|---|
| <b>Zero gas:</b>  | <b>synthetic air</b>  |
| <b>Test gas NO:</b>                                     | <b>224.1 ppb in N<sub>2</sub></b>                               |
| Number of test gas cylinder:                            | 16782 (Bottle-SN: 64616)  |
| Manufacturer/manufacturing date:                        | Linde / 07.04.2020  |
| Stability guarantee / certified:                        | 12 months   |
| Checking of the certificate by / on:                    | 17.06.2020 / UBA Langen<br>Calibration certificate No. 021-2020 |
| Measurement uncertainty as per calibration certificate: | +/- 4.5 nmol/mol  |
| <b>Test gas NO:</b>                                     | <b>2080 ppb in N<sub>2</sub></b>                                |
| Number of test gas cylinder:                            | 16811   |
| Manufacturer/manufacturing date:                        | Nippon Gases / 11.03.2020                                       |
| Stability guarantee / certified:                        | 24 months   |
| Certificate checked by:                                 | Own laboratory  |
| Rel. uncertainty according to certificate:              | 5%  |
| <b>*Test gas NO<sub>2</sub>:</b>                        | <b>4.6 ppm in synth. air</b>                                    |
| Number of test gas cylinder:                            | 16806   |
| Manufacturer/manufacturing date:                        | Nippon Gases / 07.02.2020                                       |
| Stability guarantee / certified:                        | 12 months   |
| Certificate checked by:                                 | Own laboratory  |
| Rel. uncertainty according to certificate:              | 5%  |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Page 37 of 293

|  |                               |
|--|-------------------------------|
| <b>Test gas NO<sub>2</sub>:</b>            | <b>4.45 ppm in synth. air</b> |
| Number of test gas cylinder:               | 16901                         |
| Manufacturer/manufacturing date:           | Nippon Gases / 11.01.2021     |
| Stability guarantee / certified:           | 12 months                     |
| Certificate checked by:                    | Own laboratory                |
| Rel. uncertainty according to certificate: | 5%                            |

Test gases marked with "\*" were used exclusively for individual test points before the expiry of the stability guarantee.

## 6. Test results in accordance with VDI 4202, part 1 (2018)

### 6.1 7.3 General requirements

#### 6.1 7.3.1 Measured value display

*The measuring system shall have an operative measured value display as part of the instrument.*

### 6.2 Equipment

No additional equipment is required.

### 6.3 Testing

It was checked whether the measuring system has a measured value display.

### 6.4 Evaluation

The measuring system has an operative measured value display at the instrument front.

### 6.5 Assessment

The measuring system has an operative measured value display at the instrument front.

Criterion satisfied? yes

### 6.6 Detailed presentation of test results

Figure 9 shows the measuring system with integrated measured value display.




Figure 9: N500 test system with measured value display

## **6.1**      **7.3.2 Calibration inlet**

*The measuring system may have a test gas inlet separate from the sample gas inlet.*

## **6.2**      **Equipment**

No additional equipment is required.

## **6.3**      **Testing**

We tested whether the instrument includes a test gas inlet separate from the sample gas inlet.

## **6.4**      **Evaluation**

The measuring system has a test gas inlet separate from the sample gas inlet at the instrument back.

## **6.5**      **Assessment**

The measuring system has a test gas inlet separate from the sample gas inlet at the instrument back.

Criterion satisfied?    yes

## **6.6**      **Detailed presentation of test results**

Not applicable in this instance.

## **6.1 7.3.3 Easy maintenance**

*Necessary maintenance of the measuring systems should be possible without large effort, if possible from outside.*

### **6.2 Equipment**

No additional equipment is required.

### **6.3 Testing**

The necessary regular maintenance was performed in accordance with the instruction manual.

### **6.4 Evaluation**

The user is advised to perform the following maintenance activities:

1. Checking the operational status  
The operational status may be monitored and checked by visual inspections of the instrument's display or via an external PC connected to the AMS.
2. Checking the particle filter at the sample gas inlet. The replacement interval of the particle filter depends on the dust content of the ambient air.

### **6.5 Assessment**

Maintenance takes reasonable effort and is possible with standard tools from the outside.

Criterion satisfied? yes

### **6.6 Detailed presentation of test results**

The work on the instruments was carried out during the test based on the work and work procedures described in the manual in chapter 5. Complying with the procedures described in the manual, no difficulties were identified. All maintenance activities were possible without any difficulties using standard tools.



## 6.1 7.3.4 Functional check

*If the operation or the functional check of the measuring system require particular instruments, they shall be considered as part of the measuring system and be applied in the corresponding sub-tests and included in the assessment.*

*The performance of test gas generators, which are part of the measuring system, shall be checked by comparing it to the requirements for test gases used for continuous quality assurance. They have to provide a status signal indicating that they are ready for operation. It must be possible to control them directly or remotely.*

## 6.2 Equipment

Operation manual

## 6.3 Testing

The tested measuring system does not have internal devices for operating the functional check. The current operating status of the measuring system is continuously monitored and any issues will be flagged via a series of different error messages.

The functional check of the instruments was performed using external test gases.

## 6.4 Evaluation

The tested measuring system does not have internal devices for operating the functional check. The current system status is continuously monitored and problems are indicated via a number of different warning and status messages.

External monitoring of the zero and reference point using test gases is possible.

## 6.5 Assessment

The tested measuring system does not have internal devices for operating the functional check.

Criterion satisfied? not applicable

## 6.6 Detailed presentation of test results

Not applicable.

## **6.1 7.3.5 Set-up times and warm-up times**

*The set-up times and warm-up-times shall be specified in the instruction manual.*

## **6.2 Equipment**

Operation manual and additional clock

## **6.3 Testing**

The measuring systems were set up following the manufacturer's instructions. Set-up times and warm-up times were recorded separately.

Necessary constructional measures prior to the installation such as the installation of a sampling system in the analysis room were not taken into account.

## **6.4 Evaluation**

The set-up time is of course dependent on the conditions at the installation site as well as the availability of the power supply at the installation site. As the N500 is a compact analyser, the set-up time consists mainly of:

- Connecting the AMS to supply voltage;
- Connecting the tubing (sampling, discharged air).

Commissioning and changing positions in the laboratory on various occasions (installation in/removal from the climatic chamber) as well as the installation at the field test location resulted in a set-up time of ~30 minutes. Information on the set-up time can be found in the manual in chapter 3.2. on page 18.

When switched on from a completely cold state, the system required approx. 60 minutes for the system to complete the warm-up phase and enter measurement mode. The determined warm-up time is in accordance with the information in the manual. In the manual, the warm-up phase is given as approx. 60 minutes (chapter 2.3.4.1, page 28).

The measuring system has to be installed at a location where it is protected from weather conditions, e.g. in an air-conditioned measurement container.

## **6.5 Assessment**

Set-up times and warm-up times have been determined.

It is possible to operate the measuring system at different locations with limited effort. The set-up time is approx. 0.5 hours and the run-in time approx. one hour, depending on the necessary stabilisation time.

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable.

## 6.1 7.3.6 Instrument design

*The instruction manual shall include specifications of the manufacturer regarding the design of the measuring system. The main elements are:*

*Instrument shape (e.g. bench mounting, rack mounting, free mounting)*

*Mounting position (e.g. horizontal or vertical mounting)*

*Safety requirements*

*Dimensions*

*Weight*

*Power consumption*

*Preventing condensation within the analyser.*

## 6.2 Equipment

Operation manual and a measuring system for recording energy consumption (Gossen Metrawatt) and scales.

## 6.3 Testing

The instrument design of the measuring systems handed over for testing was compared to the description provided in the manual. The energy consumption specified was verified during normal operation in the field test.

## 6.4 Evaluation

The measuring system is intended for horizontal mounting (e.g. on a table or in a rack) sheltered from weather conditions. The temperature at the site of installation must be between 0 °C and 30 °C.

The dimensions and weight of the measuring system correspond to the information provided in the operation manual.

The power requirement of the measuring system is specified by the manufacturer as 110 W. During start-up (warm-up) a short-term consumption of 220 W was recorded. In normal operation, the consumption is approx. 110 watts, as indicated.

## 6.5 Assessment

Specifications made in the instruction manual concerning instrument design are complete and correct.

Criterion satisfied? yes

## 6.6 Detailed presentation of test results

Not required for this performance criterion.

## **6.1 7.3.7 Unintended adjustment**

*It shall be possible to secure the adjustment of the measuring system against illicit or unintended adjustment during operation. Alternatively, the user manual shall specifically note that the measuring system may only be installed in a secured area.*

## **6.2 Equipment**

The test of this criterion did not require any further equipment.

## **6.3 Testing**

The measuring system can be operated via a display at its front with touch panel or via a PC connected to the measuring system directly or via a network.

The instrument provides an internal feature (password protection) to secure it against illicit or unintended adjustment. It is only possible to change parameters or adjust the measuring system after entering the password.

## **6.4 Evaluation**

On entering the correct password, it is possible to change instrument parameters affecting measurement characteristics via the control panel and via an external computer.

## **6.5 Assessment**

The measuring system is secured against unintended and unauthorised adjustment of instrument parameters by way of a password.

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not required for this performance criterion.

## 6.1 7.3.8 Data output

*The output signal shall be provided digitally and/or as analogue signals (e.g. 4 mA to 20 mA).*

## 6.2 Equipment

Analogue Yokogawa data logger, PC

## 6.3 Testing

The measuring system provides the following transmission routes: RS232, RS485, USB, digital and analogue inputs and outputs (optional), TCP/IP network. The measuring system also has the option of outputting analogue signals (optional).

## 6.4 Evaluation

Measured signals are displayed on the back of the instrument as follows:

Analogue: 0 – 20, 2 – 20, 4 – 20 mA or 0 – 1/5/10 V, Concentration range selectable

Digital RS232, USB, digital inputs and outputs, TCP/IP network

## 6.5 Assessment

The measuring signals are offered in analogue (0 – 20 mA, 4 – 20 mA or 0 – 1V, 0 – 5 V, 0 – 10 V) and digital (via TCP/IP, RS 232, USB).

The connection of additional measuring and peripheral devices is possible via corresponding connections on the units.

Criterion satisfied? yes

## 6.6 Detailed presentation of test results

Not required for this performance criterion.

## **6.1 7.3.9 Digital interface**

*The digital interface shall allow the transmission of output signals, status signals, and information like instrument type, measurement range, and measured component and unit. The digital interface shall be described fully in respective standards and guidelines.*

*Access to the measuring system via digital interfaces, e.g. for data transmission, shall be secured against unauthorised access, e.g. by a password.*

## **6.2 Equipment**

PC

## **6.3 Testing**

The measuring system provides the following transmission routes: Modbus, RS232/RS485. The measuring system also has the option of outputting analogue signals (in V).

## **6.4 Evaluation**

Digital measured signals are provided as follows:

Modbus, RS232

Digital output signals were checked. All relevant pieces of information such as measured signals, status signals, measured component, measuring range, unit and instrument information can be transmitted digitally. Access to unit parameters is password protected.

## **6.5 Assessment**

Digital transmission of measured values operates correctly.

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not required for this performance criterion.

## 6.1 7.3.10 Data transmission protocol

*The measuring system shall contain at minimum one data transmission protocol for the digital transmission of the output signal.*

*Every data transmission protocol provided by the manufacturer for the measuring system shall allow the correct transmission of the data and detect errors in the transmission. The data transmission protocol including the used commands is to be documented in the instruction manual. The data transmission protocol shall allow to transmit at minimum the following data:*

*identification of the measuring system*

*identification of measured components*

*Unit*

*output signal with time signature (date and time)*

*operation and error status*

*operating commands for remote control of the measuring systems*

*All data are to be transmitted as clear text (ASCII characters).*

## 6.2 Equipment

PC

## 6.3 Testing

The measuring system has an installed Modbus "Bavaria/Hesse" (*Bayern/Hessen*) transmission protocol as standard. Other data transmission protocols are available in consultation with the manufacturer.

## 6.4 Evaluation

The measuring system has an installed Modbus "Bavaria/Hesse" (*Bayern/Hessen*) transmission protocol as standard. Other data transmission protocols are available in consultation with the manufacturer. Measured and status signals are transmitted correctly.

## 6.5 Assessment

The measuring system has an installed Modbus "Bavaria/Hesse" (*Bayern/Hessen*) transmission protocol as standard. Measured and status signals are transmitted correctly. The configuration is listed in the manual in Appendix A.2

Criterion satisfied? yes

## 6.6 Detailed presentation of test results

Not required for this performance criterion.

## 6.1 7.3.11 Measuring range

*The upper limit of measurement of the measuring systems shall be greater or equal to the upper limit of the certification range.*

## 6.2 Equipment

The test of this criterion did not require any further equipment.

## 6.3 Testing

We compared the upper limit of measurement to the upper limit of the certification range to verify whether the former was larger or equal to the latter.

## 6.4 Evaluation

In theory, it is possible to set the measuring system to measuring ranges of up to 0–20 ppm.

The possible measuring range for NO: 1 ppm  
 The possible measuring range for NO<sub>2</sub>: 1 ppm  
 Upper limit of the certification range for NO: 1200 µg/m<sup>3</sup> (962 ppb or nmol/mol)  
 Upper limit of the certification range for NO<sub>2</sub>: 500 µg/m<sup>3</sup> (261 ppb or nmol/mol)

## 6.5 Assessment

By default, the measuring range is set to 0–500 µg/m<sup>3</sup> for NO<sub>2</sub> and 0–1200 µg/m<sup>3</sup> for NO. Other measuring ranges up to a maximum of 0 – 1 are possible.

The measuring system's upper limit of measurement exceeds the upper limit of the certification range in each case.

Criterion satisfied? yes

## 6.6 Detailed presentation of test results

VDI Guideline 4202, part 1 and standard EN 14211 define the following minimum requirements for the certification ranges of continuous air quality monitoring systems for nitrogen oxides:

Table 3: Certification ranges VDI 4202-1 and EN 14211

| Measured component | CR lower limit       | CR upper limit       | Limit value          | Evaluation period |
|--------------------|----------------------|----------------------|----------------------|-------------------|
|                    | in µg/m <sup>3</sup> | in µg/m <sup>3</sup> | in µg/m <sup>3</sup> |                   |
| Nitrogen dioxide   | 0                    | 500                  | 200                  | 1 h               |
| Nitrogen monoxide  | 0                    | 1 200                | 631.3 <sup>*)</sup>  | 1 h               |

<sup>\*)</sup> No limit value is defined for NO. Standard EN 14211 recommends to use a value of 500 ± 50 nmol/mol instead.



Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Page 49 of 293

## **6.1 7.3.12 Negative output signals**

*Negative output signals or measured values may not be suppressed (live zero).*

### **6.2 Equipment**

The test of this criterion did not require any further equipment.

### **6.3 Testing**

The possibility of displaying negative signals was tested both in the laboratory and in the field test.

### **6.4 Evaluation**

The AMS displays negative values.

### **6.5 Assessment**

The measuring system also provides negative output signals.

Criterion satisfied? yes

### **6.6 Detailed presentation of test results**

Not applicable.

## 6.1 7.3.13 Failure in the mains voltage

*In case of malfunction of the measuring system or failure in the mains voltage uncontrolled emission of operation and calibration gas shall be avoided. The measurement parameters shall be secured by buffering against loss caused by failure in the mains voltage. When mains voltage returns, the instrument shall automatically reach the operation mode and start the measurement according to the operating instructions.*

### 6.2 Equipment

Not required for this performance criterion.

### 6.3 Testing

A simulated failure in the mains voltage served to test whether the instrument remained fully functional and reached operation mode on return of the mains voltage.

### 6.4 Evaluation

Since the measuring systems do not rely on operation and calibration gases, uncontrolled emission of gases is not possible.

Once the measuring system resumes operation after a power failure it is in warm-up mode until it reaches an appropriate operating temperature again. How long it will take up to fully warm up again will depend on the ambient conditions and the temperature of the system when switching it back on again. After completion of the warm-up phase, the measuring system will switch back automatically into the mode which had been active before the failure in mains voltage. An operational status signal is displayed during the warm-up phase.

### 6.5 Assessment

The measuring system is in fault-free operational readiness when the voltage returns and automatically resumes measuring operation after a warm-up phase.

Criterion satisfied? yes

### 6.6 Detailed presentation of test results

Not applicable.

## **6.1 7.3.14 Operating states**

*The measuring system shall allow the control of important operating states by telemetrically transmitted status signals.*

## **6.2 Equipment**

Computer for data acquisition

## **6.3 Testing**

The measuring system has various interfaces such as RS232, RS485, LAN/WLAN or USB and analogue inputs and outputs. For example, the LAN/WLAN interface can be used to establish a simple connection between the analyser and an external PC. This enables telemetrically transferring data, adjusting configurations and displaying the analyser reading on the computer screen. In this mode it is possible to access and operate all the information and features from the analyser display via the computer.

## **6.4 Evaluation**

The measuring system allows for comprehensive monitoring and control via various connectors.

## **6.5 Assessment**

The measuring system provides various ports to ensure comprehensive monitoring and control via an external computer.

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable.

## **6.1 7.3.15 Switch-over**

*Switch-over between measurement and functional check and/or calibration shall be possible telemetrically by computer control or manual intervention.*

## **6.2 Equipment**

Not required for this performance criterion.

## **6.3 Testing**

The measuring system can be monitored and controlled by the operator at the unit or by the telemetric remote control.

## **6.4 Evaluation**

All operating procedures which do not require on-site practical handling may be performed both by the operator on the instrument itself or telemetrically.

## **6.5 Assessment**

In principle, all necessary work for function control can be monitored and controlled directly on the unit or by telemetric remote control.

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable.

## **6.1 7.3.16 Instrument software**

*The measuring system shall be able to display the version of the instrument software.*

### **6.2 Equipment**

Not required for this performance criterion.

### **6.3 Testing**

It was tested whether the software version can be displayed on the instrument. The AMS manufacturer was informed of his obligation to communicate any changes to the instrument software to the test laboratory.

### **6.4 Evaluation**

The current software version is displayed when switching on the instrument. Furthermore, it can be accessed via menu item “configuration” at any time.

The tests were performed with software version 1.6.0.

### **6.5 Assessment**

The instrument’s software version is displayed. Software changes are communicated to the test laboratory.

Criterion satisfied? yes

### **6.6 Detailed presentation of test results**

Figure 6 shows the software version displayed by the measuring system.

## **6.1 7.4 Requirements on performance characteristics for testing in the laboratory**

### **6.1 7.4.1 General requirements**

*The performance characteristics which shall be determined during testing in the laboratory and their related performance criteria for measured components according to the 39<sup>th</sup> BImSchV are given in Table A1 of VDI 4202-1.*

*The certification range for other components is to be defined. Performance criteria are to be defined by drawing from Table A1 of standard VDI 4202-1 (2018). These definitions shall be cleared with the relevant body before testing.*

*The determination of the performance characteristics shall be done according to the procedures de-scribed in Section 8.4.*

### **6.2 Equipment**

Not required for this performance criterion.

### **6.3 Testing**

Tests were performed using the performance characteristics specified in VDI standard 4202, part 1 (2018) and standard EN 14211.

### **6.4 Evaluation**

Not applicable.

### **6.5 Assessment**

Tests were performed using the performance characteristics specified in VDI standard 4202, part 1 (2018) and standard EN 14211 (2012)

Criterion satisfied? yes

### **6.6 Detailed presentation of test results**

Not applicable.

## 6.1 7.4.2 Test requirements

*Before operating the measuring system, the instruction manual of the manufacturer shall be followed particularly with regard to the set-up of equipment and the quality and quantity of the consumable supplies necessary.*

*The measuring system shall be allowed to warm up for the duration specified by the manufacturer before undertaking any tests. If the warm-up time is not specified, a minimum of 4 h applies.*

*If auto-scale or self-correction functions are arbitrary, these functions shall be turned off during the laboratory test.*

*If auto-scale or self-correction functions are not arbitrary but treated as “normal operating conditions”, times and values of the self-correction shall be available for the test laboratory. The values of the auto-zero and auto-drift corrections are subject to the same restrictions as given in the performance characteristics.*

*Before applying test gases to the measuring system, the test gas system shall have been operated for a sufficiently long time in order to stabilize the concentrations applied to the measuring system. The measuring system shall be tested using an impregnated particle filter.*

*Most measuring systems are able to display the output signal as running average of an adjustable period. Some measuring systems adjust the integration time as a function of the frequency of the fluctuations of the concentration of the measured component automatically. These options are typically used for equalisation of the output data. It does not have to be proved that the selected value for the averaging period or the use of an active filter affects the result of testing the averaging period and the response time.*

*The adjustments of the measuring system shall meet the specifications of the manufacturer. All adjustments are to be documented in the test report.*

*For the determination of the various performance characteristics, suitable zero and test gases shall be used.*

*Parameters: During the test for each individual performance characteristic, the values of the following parameters shall be stable within the specified range given in Table 3 of standard VDI 4202-1.*

*Test gas: For the determination of the various performance characteristics, test gases traceable to national or international standards shall be used.*

## 6.2 Equipment

Not required for this performance criterion.

## 6.3 Testing

Tests were performed using the performance characteristics specified in VDI standard 4202, part 1 (2018) and standard EN 14211.

## 6.4 Evaluation

The warm-up time described in the manual was observed.

Neither auto-scale nor self-correction functions were activated during the laboratory test.

The tests were carried out with the particle filters associated with the systems.

As described in point 3.3, the adaptive filters mentioned were active by default throughout the test.

Test gases used comply with the requirements of VDI 4202-1.

### **6.5 Assessment**

Tests were performed using the performance characteristics specified in VDI standard 4202, part 1 (2018) and standard EN 14211 (2012)

Criterion satisfied? yes

### **6.6 Detailed presentation of test results**

Not applicable.



## 6.1 7.4.3 Response time and memory effect

*The response time (rise) of the measuring system shall not exceed 180 s.*

*The response time (fall) of the measuring system shall not exceed 180 s.*

*The difference between the response time (rise) and response time (fall) of the measuring system shall not exceed 10% of response time (rise) or 10 s, whatever value is larger.*

## 6.2 Equipment

Not applicable

## 6.3 Testing

Determination and evaluation of the response time corresponds exactly to determining the response time in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.4.3 Response time.

## 6.4 Evaluation

See chapter 7.1 8.4.3 Response time

## 6.5 Assessment

See chapter 7.1 8.4.3 Response time

Criterion satisfied? yes

## 6.6 Detailed presentation of test results

Not applicable in this instance.

## **6.1 7.4.4 Short-term drift**

*The short-term drift at zero point shall not exceed 2.0 nmol/mol.  
The short-term drift at span point shall not exceed 6.0 nmol/mol.*

## **6.2 Equipment**

Not applicable

## **6.3 Testing**

Determination and evaluation of the short-term drift corresponds exactly to determining the short term drift in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.4.4 Short-term drift.

## **6.4 Evaluation**

See chapter 7.1 8.4.4 Short-term drift

## **6.5 Assessment**

See chapter 7.1 8.4.4 Short-term drift

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable in this instance.

## 6.1 7.4.5 Repeatability standard deviation

*The repeatability standard deviation at zero point shall be  $\leq 1.0$  nmol/mol of the upper limit of the certification range.*

*The repeatability standard deviation at reference point shall not exceed 3.0 nmol/mol.*

## 6.2 Equipment

Not applicable

## 6.3 Testing

Performing and evaluating the repeatability standard deviation at zero point corresponds exactly to determining the repeatability standard deviation in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.4.5 Repeatability standard deviation.

## 6.4 Evaluation

See chapter 7.1 8.4.5 Repeatability standard deviation.

## 6.5 Assessment

See chapter 7.1 8.4.5 Repeatability standard deviation.

Criterion satisfied? yes

## 6.6 Detailed presentation of test results

Not applicable in this instance.

## **6.1 7.4.6 Linearity**

*The correlation between the measured values and the target values must be representable with the aid of a linear analysis function.*

*The deviation from the linearity of the calibration function at zero shall not exceed 5 nmol/mol. At concentrations above zero, it shall not exceed 4% of the measured value.*

## **6.2 Equipment**

Not applicable

## **6.3 Testing**

Performing and evaluating the linearity corresponds exactly to determining the lack of fit in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.4.6 Lack of fit of linearity of the calibration function.

## **6.4 Evaluation**

See chapter 7.1 8.4.6 Lack of fit of linearity of the calibration function.

## **6.5 Assessment**

See chapter 7.1 8.4.6 Lack of fit of linearity of the calibration function.  
Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable in this instance.

## **6.1 7.4.7 Sensitivity coefficient to sample gas pressure**

*The sensitivity coefficient of sample gas pressure at reference point shall not exceed 8.0 (nmol/mol)/kPA.*

## **6.2 Equipment**

Not applicable

## **6.3 Testing**

Performing and evaluating the sensitivity coefficient of sample gas pressure corresponds exactly to determining the sensitivity coefficient to sample gas pressure in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.4.7 Sensitivity coefficient to sample gas pressure.

## **6.4 Evaluation**

See chapter 7.1 8.4.7 Sensitivity coefficient to sample gas pressure

## **6.5 Assessment**

See chapter 7.1 8.4.7 Sensitivity coefficient to sample gas pressure

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable in this instance.

## **6.1 7.4.8 Sensitivity coefficient to sample gas temperature**

*The sensitivity coefficient of sample gas temperature shall not exceed 3.0 (nmol/mol)/kPA.*

### **6.2 Equipment**

Not applicable

### **6.3 Testing**

Performing and evaluating the sensitivity coefficient of sample gas temperature corresponds exactly to determining the sensitivity coefficient to sample gas temperature in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.4.8 Sensitivity coefficient to sample gas temperature.

### **6.4 Evaluation**

See chapter 7.1 8.4.8 Sensitivity coefficient to sample gas temperature

### **6.5 Assessment**

See chapter 7.1 8.4.8 Sensitivity coefficient to sample gas temperature

Criterion satisfied? yes

### **6.6 Detailed presentation of test results**

Not applicable in this instance.

## **6.1 7.4.9 Sensitivity coefficient to surrounding temperature**

*The sensitivity coefficient of surrounding temperature shall not exceed 3.0 (nmol/mol)/kPA.*

## **6.2 Equipment**

Not applicable

## **6.3 Testing**

Performing and evaluating the sensitivity coefficient of surrounding temperature corresponds exactly to determining the sensitivity coefficient to the surrounding temperature in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.4.9 Sensitivity coefficient to surrounding temperature.

## **6.4 Evaluation**

See chapter 7.1 8.4.9 Sensitivity coefficient to surrounding temperature

## **6.5 Assessment**

See chapter 7.1 8.4.9 Sensitivity coefficient to surrounding temperature  
Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable in this instance.

## **6.1 7.4.10 Sensitivity coefficient to electrical voltage**

*The sensitivity coefficient of electrical voltage shall not exceed 0.3 (nmol/mol)/V.*

### **6.2 Equipment**

Not applicable

### **6.3 Testing**

Performing and evaluating the sensitivity coefficient of electrical voltage corresponds exactly to determining the sensitivity coefficient to electrical voltage in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.4.10 Sensitivity coefficient to electrical voltage.

### **6.4 Evaluation**

See chapter 7.1 8.4.10 Sensitivity coefficient to electrical voltage

### **6.5 Assessment**

See chapter 7.1 8.4.10 Sensitivity coefficient to electrical voltage

Criterion satisfied? yes

### **6.6 Detailed presentation of test results**

Not applicable in this instance.



## 6.1 7.4.11 Cross-sensitivity

*The change in the measured value caused by interfering components in the sample gas shall not exceed the requirements of Table A of VDI 4202, part 1 (April 2018), at zero and span point.*

*For measuring principles deviating from EN standards the absolute values of the sum of the positive and the sum of negative deviations caused by interfering components in the sample gas shall not exceed 3% of the upper limit of the certification range at zero and span point. A value  $c_t$  at 70% to 80% of the upper limit of the certification range shall be used as reference point.*

## 6.2 Equipment

Test gases, mixing station

## 6.3 Testing

The measuring system does not work according to the reference method chemiluminescence but measures NO and NO<sub>2</sub> directly. The execution and evaluation for the determination of the cross-sensitivities according to EN 14211 (2012) and VDI 4202-1 (2018) is adapted to the reference method in the selection of the cross-sensitivity components. Therefore, in addition to the three interferents mentioned in EN 14211, other interference components possible in normal ambient air were checked for their influence. As the measuring system analyses the sample air in the UV range, only other interferents with optical absorption lines in the UV range were considered. The additional interfering components checked were SO<sub>2</sub>, H<sub>2</sub>S, m-xylene and toluene.

The procedure and evaluation for determining the cross-sensitivities according to EN 14211 (2012) is described in chapter 7.1 8.4.11 Interferents. The contribution to the overall uncertainty from the cross-sensitivity is also determined from the examination carried out strictly according to EN 14211.

## 6.4 Evaluation

The following table summarises differences with and without interfering components for the zero and reference points of the analysers. The table sums up the positive and negative sums of the deviations. The interferents to be checked according to EN 14211 (H<sub>2</sub>O, CO<sub>2</sub> and NH<sub>3</sub>) are also listed in this evaluation. The level of the test gas concentrations was selected on the basis of VDI 4201-1 (2018). The concentration level of the interferents was selected on the basis of the guideline VDI 4202-1 (2018). For the component SO<sub>2</sub> the 1-hour limit value was chosen.

**Table 4: Additional cross-sensitivities for the component NO, system 1**

AMS: N500  
 Component: NO (certification range = 0 - 962 nmol/mol)

| Interferent                       | AMS 1                     |                     |        |                       |                           |                     |        |                       |
|-----------------------------------|---------------------------|---------------------|--------|-----------------------|---------------------------|---------------------|--------|-----------------------|
|                                   | Zero point                |                     |        |                       | Span point                |                     |        |                       |
|                                   | Nominal value<br>nmol/mol | Reading<br>nmol/mol | %TG    | Deviation<br>nmol/mol | Nominal value<br>nmol/mol | Reading<br>nmol/mol | %TG    | Deviation<br>nmol/mol |
| SO <sub>2</sub> 131 nmol/mol      | 0.4                       | 0.8                 | ≤ 0.50 | 0.40                  | 500.4                     | 500.0               | ≤ 0.50 | -0.40                 |
| H <sub>2</sub> S 200 nmol/mol     | 0.4                       | 0.4                 | ≤ 0.50 | 0.00                  | 503.7                     | 504.5               | ≤ 0.50 | 0.80                  |
| CO <sub>2</sub> 500 µmol/mol      | 0.4                       | 0.8                 | ≤ 0.50 | 0.40                  | 500.0                     | 501.8               | ≤ 0.50 | 1.80                  |
| m-Xylol 1 µmol/mol                | 0.2                       | 1.0                 | ≤ 0.50 | 0.80                  | 503.7                     | 506.1               | ≤ 0.50 | 2.40                  |
| Toluol 0.5 µmol/mol               | 0.6                       | 0.6                 | ≤ 0.50 | 0.00                  | 504.5                     | 504.7               | ≤ 0.50 | 0.20                  |
| H <sub>2</sub> O 19 mmol/mol      | 0.0                       | 0.0                 | ≤ 0.50 | 0.00                  | 506.5                     | 506.9               | ≤ 0.50 | 0.40                  |
| NH <sub>3</sub> 200 nmol/mol      | 0.6                       | 1.2                 | ≤ 0.50 | 0.60                  | 503.4                     | 504.6               | ≤ 0.50 | 1.20                  |
| <b>Sum of positive deviations</b> |                           |                     |        | <b>2.20</b>           |                           |                     |        |                       |
| <b>Sum of negative deviations</b> |                           |                     |        | <b>0.00</b>           |                           |                     |        |                       |
|                                   |                           |                     |        |                       | <b>6.80</b>               |                     |        |                       |
|                                   |                           |                     |        |                       | <b>-0.40</b>              |                     |        |                       |

**Table 5: Additional cross-sensitivities for the component NO, system 2**

AMS: N500  
 Component: NO (certification range = 0 - 962 nmol/mol)

| Interferent                       | AMS 2                     |                     |        |                       |                           |                     |        |                       |
|-----------------------------------|---------------------------|---------------------|--------|-----------------------|---------------------------|---------------------|--------|-----------------------|
|                                   | Zero point                |                     |        |                       | Span point                |                     |        |                       |
|                                   | Nominal value<br>nmol/mol | Reading<br>nmol/mol | %TG    | Deviation<br>nmol/mol | Nominal value<br>nmol/mol | Reading<br>nmol/mol | %TG    | Deviation<br>nmol/mol |
| SO <sub>2</sub> 131 nmol/mol      | 0.6                       | 0.4                 | ≤ 0.50 | -0.20                 | 501.4                     | 500.8               | ≤ 0.50 | -0.60                 |
| H <sub>2</sub> S 200 nmol/mol     | 0.0                       | 0.0                 | ≤ 0.50 | 0.00                  | 504.3                     | 506.7               | ≤ 0.50 | 2.40                  |
| CO <sub>2</sub> 500 µmol/mol      | 0.0                       | 0.4                 | ≤ 0.50 | 0.40                  | 501.4                     | 502.2               | ≤ 0.50 | 0.80                  |
| m-Xylol 1 µmol/mol                | 0.0                       | 1.0                 | ≤ 0.50 | 1.00                  | 504.5                     | 506.7               | ≤ 0.50 | 2.20                  |
| Toluol 0.5 µmol/mol               | 0.0                       | 0.0                 | ≤ 0.50 | 0.00                  | 506.3                     | 506.5               | ≤ 0.50 | 0.20                  |
| H <sub>2</sub> O 19 mmol/mol      | 0.0                       | -0.6                | ≤ 0.50 | -0.60                 | 506.7                     | 508.3               | ≤ 0.50 | 1.60                  |
| NH <sub>3</sub> 200 nmol/mol      | 0.0                       | 1.2                 | ≤ 0.50 | 1.20                  | 506.1                     | 507.5               | ≤ 0.50 | 1.40                  |
| <b>Sum of positive deviations</b> |                           |                     |        | <b>2.60</b>           |                           |                     |        |                       |
| <b>Sum of negative deviations</b> |                           |                     |        | <b>-0.80</b>          |                           |                     |        |                       |
|                                   |                           |                     |        |                       | <b>8.60</b>               |                     |        |                       |
|                                   |                           |                     |        |                       | <b>-0.60</b>              |                     |        |                       |

**Table 6: Additional cross-sensitivities for the component NO<sub>2</sub>, system 1**

AMS: N500  
 Component: NO<sub>2</sub> (certification range = 0 - 261 nmol/mol)

| Interferent                       | AMS 1                     |                     |        |                       |                           |                     |        |                       |
|-----------------------------------|---------------------------|---------------------|--------|-----------------------|---------------------------|---------------------|--------|-----------------------|
|                                   | Zero point                |                     |        |                       | Span point                |                     |        |                       |
|                                   | Nominal value<br>nmol/mol | Reading<br>nmol/mol | %TG    | Deviation<br>nmol/mol | Nominal value<br>nmol/mol | Reading<br>nmol/mol | %TG    | Deviation<br>nmol/mol |
| SO <sub>2</sub> 131 nmol/mol      | -0.2                      | -0.2                | ≤ 0.50 | 0.00                  | 108.1                     | 108.8               | 0.65   | 0.70                  |
| H <sub>2</sub> S 200 nmol/mol     | -0.2                      | -0.2                | ≤ 0.50 | 0.00                  | 106.8                     | 108.2               | 1.31   | 1.40                  |
| CO <sub>2</sub> 500 µmol/mol      | -0.2                      | -0.2                | ≤ 0.50 | 0.00                  | 105.8                     | 106.5               | 0.66   | 0.70                  |
| m-Xylol 1 µmol/mol                | -0.2                      | -0.2                | ≤ 0.50 | 0.00                  | 107.6                     | 109.6               | 1.86   | 2.00                  |
| Toluol 0.5 µmol/mol               | -0.2                      | -0.2                | ≤ 0.50 | 0.00                  | 108.5                     | 108.3               | ≤ 0.50 | -0.20                 |
| H <sub>2</sub> O 19 mmol/mol      | 0.1                       | 0.3                 | ≤ 0.50 | 0.20                  | 106.2                     | 106.5               | ≤ 0.50 | 0.30                  |
| NH <sub>3</sub> 200 nmol/mol      | -0.2                      | -0.2                | ≤ 0.50 | 0.00                  | 107.0                     | 108.1               | 1.03   | 1.10                  |
| <b>Sum of positive deviations</b> |                           |                     |        | <b>0.20</b>           |                           |                     |        |                       |
| <b>Sum of negative deviations</b> |                           |                     |        | <b>0.00</b>           |                           |                     |        |                       |
|                                   |                           |                     |        |                       | <b>6.20</b>               |                     |        |                       |
|                                   |                           |                     |        |                       | <b>-0.20</b>              |                     |        |                       |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Table 7: Additional cross-sensitivities for the component NO<sub>2</sub>, system 2

**AMS:** N500  
**Component:** NO<sub>2</sub> (certification range = 0 - 261 nmol/mol)

| Interferent                       | Nominal value<br>nmol/mol | Zero point          |        |                       |                           | Span point          |        |                       |  |
|-----------------------------------|---------------------------|---------------------|--------|-----------------------|---------------------------|---------------------|--------|-----------------------|--|
|                                   |                           | Reading<br>nmol/mol | %TG    | Deviation<br>nmol/mol | Nominal value<br>nmol/mol | Reading<br>nmol/mol | %TG    | Deviation<br>nmol/mol |  |
| SO <sub>2</sub>                   | 131 nmol/mol              | -1.0                | ≤ 0.50 | 0.00                  | 107.9                     | 108.7               | 0.74   | 0.80                  |  |
| H <sub>2</sub> S                  | 200 nmol/mol              | -0.9                | ≤ 0.50 | -0.10                 | 108.3                     | 109.5               | 1.11   | 1.20                  |  |
| CO <sub>2</sub>                   | 500 µmol/mol              | -0.4                | ≤ 0.50 | -0.30                 | 105.6                     | 106.4               | 0.76   | 0.80                  |  |
| m-Xylol                           | 1 µmol/mol                | -1.0                | ≤ 0.50 | 0.00                  | 107.6                     | 109.5               | 1.77   | 1.90                  |  |
| Toluol                            | 0.5 µmol/mol              | -1.0                | ≤ 0.50 | 0.00                  | 108.4                     | 108.2               | ≤ 0.50 | -0.20                 |  |
| H <sub>2</sub> O                  | 19 mmol/mol               | 0.3                 | ≤ 0.50 | 0.00                  | 106.6                     | 106.8               | ≤ 0.50 | 0.20                  |  |
| NH <sub>3</sub>                   | 200 nmol/mol              | -0.8                | ≤ 0.50 | -0.20                 | 106.9                     | 107.9               | 0.94   | 1.00                  |  |
| <b>Sum of positive deviations</b> |                           |                     |        | <b>0.00</b>           |                           |                     |        | <b>5.90</b>           |  |
| <b>Sum of negative deviations</b> |                           |                     |        | <b>-0.60</b>          |                           |                     |        | <b>-0.20</b>          |  |

## 6.5 Assessment

For measuring principles deviating from EN standards the absolute values of the sum of the positive and the sum of negative deviations caused by interfering components in the sample gas shall not exceed 3% of the upper limit of the certification range at zero and reference point. This would be related to the measuring range 0 – 962 nmol/mol = 28.86 nmol/mol allowed deviation for the component NO. For the component NO<sub>2</sub> the permissible deviation for the measuring range 0 – 261 nmol/mol is 7.83 nmol/mol. These values are safely complied with, as shown in Table 4 to Table 7.

The influence of the cross-sensitivity also fulfils the requirement of VDI 4202-1 (2018) for measurement principles that deviate from the EN standards.

Criterion satisfied? yes

## 6.6 Detailed presentation of test results

The individual values of the cross-sensitivity tests carried out are shown in the following tables.

Table 8: Additional cross-sensitivities for the component NO, system 1

**AMS:** N500  
**Component:** NO (certification range = 0 – 962 mg/m<sup>3</sup>)  
**Measurement date:** 23.10.2020 to 26.10.2020

| System 1         |     |          | Nominal value<br>nmol/mol | Zero point     |                |                |               |
|------------------|-----|----------|---------------------------|----------------|----------------|----------------|---------------|
| Interferent      |     |          |                           | 1.<br>nmol/mol | 2.<br>nmol/mol | 3.<br>nmol/mol | ∅<br>nmol/mol |
| SO <sub>2</sub>  | 131 | nmol/mol | 0.40                      | 0.60           | 0.60           | 1.20           | 0.80          |
| H <sub>2</sub> S | 200 | nmol/mol | 0.40                      | 0.60           | 0.60           | 0.00           | 0.40          |
| CO <sub>2</sub>  | 500 | µmol/mol | 0.40                      | 0.60           | 0.60           | 1.20           | 0.80          |
| m-Xylene         | 1   | µmol/mol | 0.20                      | 0.60           | 1.20           | 1.20           | 1.00          |
| Toluene          | 0.5 | µmol/mol | 0.60                      | 0.60           | 0.60           | 0.60           | 0.60          |
| H <sub>2</sub> O | 19  | mmol/mol | 0.00                      | 0.00           | 0.00           | 0.00           | 0.00          |
| NH <sub>3</sub>  | 200 | nmol/mol | 0.60                      | 1.20           | 1.20           | 1.20           | 1.20          |

| System 1         |     |          | Nominal value<br>nmol/mol | Span point     |                |                |               |
|------------------|-----|----------|---------------------------|----------------|----------------|----------------|---------------|
| Interferent      |     |          |                           | 1.<br>nmol/mol | 2.<br>nmol/mol | 3.<br>nmol/mol | ∅<br>nmol/mol |
| SO <sub>2</sub>  | 131 | nmol/mol | 500.44                    | 500.24         | 500.24         | 499.64         | 500.04        |
| H <sub>2</sub> S | 200 | nmol/mol | 503.65                    | 503.85         | 504.45         | 505.05         | 504.45        |
| CO <sub>2</sub>  | 500 | µmol/mol | 500.00                    | 502.00         | 501.40         | 502.00         | 501.80        |
| m-Xylene         | 1   | µmol/mol | 503.65                    | 506.25         | 505.65         | 506.25         | 506.05        |
| Toluene          | 0.5 | µmol/mol | 504.45                    | 505.05         | 504.45         | 504.45         | 504.65        |
| H <sub>2</sub> O | 19  | mmol/mol | 506.47                    | 507.50         | 508.70         | 504.40         | 506.87        |
| NH <sub>3</sub>  | 200 | nmol/mol | 503.40                    | 505.10         | 504.40         | 504.40         | 504.63        |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Page 69 of 293

Table 9: Additional cross-sensitivities for the component NO, system 2

**AMS:** N500  
**Component:** NO (certification range = 0 – 962 mg/m<sup>3</sup>)  
**Measurement date:** 23.10.2020 to 26.10.2020

| System 2         |     |          | Zero point    |          |          |          |          |
|------------------|-----|----------|---------------|----------|----------|----------|----------|
| Interferent      |     |          | Nominal value | 1.       | 2.       | 3.       | ∅        |
|                  |     |          | nmol/mol      | nmol/mol | nmol/mol | nmol/mol | nmol/mol |
| SO <sub>2</sub>  | 131 | nmol/mol | 0.60          | 0.00     | 0.00     | 1.20     | 0.40     |
| H <sub>2</sub> S | 200 | nmol/mol | 0.00          | 0.00     | 0.00     | 0.00     | 0.00     |
| CO <sub>2</sub>  | 500 | µmol/mol | 0.00          | 0.00     | 0.00     | 1.20     | 0.40     |
| m-Xylene         | 1   | µmol/mol | 0.00          | 1.20     | 0.60     | 1.20     | 1.00     |
| Toluene          | 0.5 | µmol/mol | 0.00          | 0.00     | 0.00     | 0.00     | 0.00     |
| H <sub>2</sub> O | 19  | mmol/mol | 0.00          | -0.60    | -0.60    | -0.60    | -0.60    |
| NH <sub>3</sub>  | 200 | nmol/mol | 0.00          | 1.20     | 1.20     | 1.20     | 1.20     |

| System 2         |     |          | Span point    |          |          |          |          |
|------------------|-----|----------|---------------|----------|----------|----------|----------|
| Interferent      |     |          | Nominal value | 1.       | 2.       | 3.       | ∅        |
|                  |     |          | nmol/mol      | nmol/mol | nmol/mol | nmol/mol | nmol/mol |
| SO <sub>2</sub>  | 131 | nmol/mol | 501.44        | 500.24   | 501.44   | 500.84   | 500.84   |
| H <sub>2</sub> S | 200 | nmol/mol | 504.25        | 506.85   | 506.25   | 506.85   | 506.65   |
| CO <sub>2</sub>  | 500 | µmol/mol | 501.40        | 502.60   | 502.00   | 502.00   | 502.20   |
| m-Xylene         | 1   | µmol/mol | 504.45        | 506.25   | 506.25   | 507.46   | 506.65   |
| Toluene          | 0.5 | µmol/mol | 506.25        | 505.65   | 506.85   | 506.85   | 506.45   |
| H <sub>2</sub> O | 19  | mmol/mol | 506.67        | 508.70   | 506.90   | 509.30   | 508.30   |
| NH <sub>3</sub>  | 200 | nmol/mol | 506.10        | 506.90   | 507.50   | 508.10   | 507.50   |

Table 10: Additional cross-sensitivities for the component NO<sub>2</sub>, system 1

**AMS:** N500  
**Component:** NO<sub>2</sub> (certification range = 0 - 261 nmol/mol)  
**Measurement date:** 23.10.2020 to 28.10.2020

| System 1         |     |          | Zero point                |                |                |                |               |
|------------------|-----|----------|---------------------------|----------------|----------------|----------------|---------------|
| Interferent      |     |          | Nominal value<br>nmol/mol | 1.<br>nmol/mol | 2.<br>nmol/mol | 3.<br>nmol/mol | ∅<br>nmol/mol |
| SO <sub>2</sub>  | 131 | nmol/mol | -0.16                     | -0.16          | -0.16          | -0.16          | -0.16         |
| H <sub>2</sub> S | 200 | nmol/mol | -0.16                     | -0.16          | -0.16          | -0.16          | -0.16         |
| CO <sub>2</sub>  | 500 | μmol/mol | -0.20                     | -0.20          | -0.20          | -0.20          | -0.20         |
| m-Xylene         | 1   | μmol/mol | -0.16                     | -0.16          | -0.16          | -0.16          | -0.16         |
| Toluene          | 0.5 | μmol/mol | -0.16                     | -0.16          | -0.16          | -0.16          | -0.16         |
| H <sub>2</sub> O | 19  | mmol/mol | 0.07                      | 0.30           | 0.30           | 0.30           | 0.30          |
| NH <sub>3</sub>  | 200 | nmol/mol | -0.20                     | -0.20          | -0.20          | -0.20          | -0.20         |

| System 1         |     |          | Span point                |                |                |                |               |
|------------------|-----|----------|---------------------------|----------------|----------------|----------------|---------------|
| Interferent      |     |          | Nominal value<br>nmol/mol | 1.<br>nmol/mol | 2.<br>nmol/mol | 3.<br>nmol/mol | ∅<br>nmol/mol |
| SO <sub>2</sub>  | 131 | nmol/mol | 108.10                    | 108.80         | 108.80         | 108.80         | 108.80        |
| H <sub>2</sub> S | 200 | nmol/mol | 106.82                    | 108.23         | 108.23         | 108.23         | 108.23        |
| CO <sub>2</sub>  | 500 | μmol/mol | 105.77                    | 106.40         | 106.50         | 106.50         | 106.47        |
| m-Xylene         | 1   | μmol/mol | 107.61                    | 109.62         | 109.62         | 109.62         | 109.62        |
| Toluene          | 0.5 | μmol/mol | 108.53                    | 108.48         | 108.48         | 107.83         | 108.26        |
| H <sub>2</sub> O | 19  | mmol/mol | 106.20                    | 106.40         | 106.50         | 106.70         | 106.53        |
| NH <sub>3</sub>  | 200 | nmol/mol | 107.03                    | 108.00         | 108.20         | 108.20         | 108.13        |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Page 71 of 293

Table 11: Additional cross-sensitivities for the component NO<sub>2</sub>, system 2

**AMS:** N500  
**Component:** NO<sub>2</sub> (certification range = 0 - 261 nmol/mol)  
**Measurement date:** 23.10.2020 to 28.10.2020

| System 2         |     |          | Zero point                |                |                |                |               |
|------------------|-----|----------|---------------------------|----------------|----------------|----------------|---------------|
| Interferent      |     |          | Nominal value<br>nmol/mol | 1.<br>nmol/mol | 2.<br>nmol/mol | 3.<br>nmol/mol | ∅<br>nmol/mol |
| SO <sub>2</sub>  | 131 | nmol/mol | -1.03                     | -1.14          | -0.98          | -0.98          | -1.03         |
| H <sub>2</sub> S | 200 | nmol/mol | -0.87                     | -0.98          | -0.98          | -0.98          | -0.98         |
| CO <sub>2</sub>  | 500 | µmol/mol | -0.44                     | -0.70          | -0.70          | -0.70          | -0.70         |
| m-Xylene         | 1   | µmol/mol | -0.98                     | -0.98          | -0.98          | -0.98          | -0.98         |
| Toluene          | 0.5 | µmol/mol | -0.98                     | -0.98          | -0.98          | -0.98          | -0.98         |
| H <sub>2</sub> O | 19  | mmol/mol | 0.33                      | 0.30           | 0.30           | 0.30           | 0.30          |
| NH <sub>3</sub>  | 200 | nmol/mol | -0.82                     | -1.00          | -1.00          | -1.00          | -1.00         |

| System 2         |     |          | Span point                |                |                |                |               |
|------------------|-----|----------|---------------------------|----------------|----------------|----------------|---------------|
| Interferent      |     |          | Nominal value<br>nmol/mol | 1.<br>nmol/mol | 2.<br>nmol/mol | 3.<br>nmol/mol | ∅<br>nmol/mol |
| SO <sub>2</sub>  | 131 | nmol/mol | 107.94                    | 108.64         | 108.80         | 108.80         | 108.75        |
| H <sub>2</sub> S | 200 | nmol/mol | 108.26                    | 110.11         | 110.11         | 108.32         | 109.51        |
| CO <sub>2</sub>  | 500 | µmol/mol | 105.60                    | 106.20         | 106.40         | 106.50         | 106.37        |
| m-Xylene         | 1   | µmol/mol | 107.61                    | 109.46         | 109.46         | 109.46         | 109.46        |
| Toluene          | 0.5 | µmol/mol | 108.43                    | 108.32         | 108.48         | 107.66         | 108.15        |
| H <sub>2</sub> O | 19  | mmol/mol | 106.57                    | 106.70         | 106.70         | 107.00         | 106.80        |
| NH <sub>3</sub>  | 200 | nmol/mol | 106.90                    | 107.80         | 107.80         | 108.00         | 107.87        |

## 6.1 7.4.12 Averaging effect

*The measuring system shall enable hourly averages.*

*The averaging effect shall not exceed 7% of the measured value.*

## 6.2 Equipment

Test gases, mixing station

## 6.3 Testing

The averaging test gives a measure of the uncertainty in the averaged values caused by short-term concentration variations in the sampled air shorter than the time scale of the measurement process in the analyser. In general, the output of an analyser is a result of the determination of a reference concentration (normally zero) and the actual concentration which takes a certain time.

For the determination of the uncertainty due to the averaging, the following concentrations are applied to the analyser and readings are taken at each concentration:

- a constant concentration of NO<sub>2</sub> at a concentration  $c_{t,NO_2}$  which is about twice the hourly limit value; and
- a stepwise varied concentration of NO between zero and 600 nmol/mol (concentration  $c_{t,NO}$ ).

The time span ( $t_c$ ) of the constant NO concentrations shall be at least equal to the time span necessary to obtain four independent display values (corresponding to at least 16 response times) The time interval ( $t_v$ ) of the changed NO concentration shall be at least equal to the time interval ( $t_{NO}$ ) required to obtain four independent display values. The NO concentration must be 45 s followed by the time period ( $t_{zero}$ ) of 45 s for zero concentration. Further:

$c_t$  is the test concentration;

$t_v$  is a time period including a whole number of  $t_{NO}$  and  $t_{zero}$  pairs, and contains a minimum of 3 such pairs.

The change from  $t_{NO}$  to  $t_{zero}$  shall be within 0.5 s. The change from  $t_c$  to  $t_v$  shall be within one response time of the analyser under test.

The averaging effect for NO ( $X_{av}$ ) is:

$$E_{av} = \frac{C_{const}^{av} - 2C_{var}^{av}}{C_{const}^{av}} * 100$$

Where:

$E_{av}$  is the averaging effect (%);

$C_{const}^{av}$  is the average of the at least four independent measurements during the variable concentration period;

$C_{var}^{av}$  is the average value of at least four independent measurements during the period of variable concentration



### The following applies differently for NO<sub>x</sub> measuring instruments

The averaging influence shall be determined for the measurement components NO<sub>2</sub> and NO. The averaging influence for NO is calculated according to EN 14211. The determination of the averaging effect for NO<sub>2</sub> is calculated according to the following formula:

$$E_{av} = \frac{C_{const}^{av} - C_{var}^{av}}{C_{const}^{av}} * 100\%$$

Where:

$E_{av}$  is the averaging effect (%);

$C_{const}^{av}$  is the average of the at least four independent measurements during the variable concentration period;

$C_{var}^{av}$  is the average of the at least four independent measurements during the variable concentration period;

The averaging test was carried out according to the specifications of EN 14211 and VDI 4202-1. With the help of a mass flow controller we applied a step change of the NO concentration between zero and 600 nmol/mol and, at the same time, a constant NO<sub>2</sub> concentration  $C_{t,NO_2}$  of roughly twice the hourly limit value. First, the average was calculated at a constant test gas concentration. Then, a three-way valve served to switch between zero and test gas every 45 s. During that period of alternating test gas application the average was calculated again.

## 6.4 Evaluation

The following averages were determined during the test:

Table 12: Results of the averaging test for NO

|                               | requirement | device 1 |   | device 2 |   |
|-------------------------------|-------------|----------|---|----------|---|
| averaging effect $E_{av}$ [%] | ≤ 7%        | 3.1      | ✓ | -2.9     | ✓ |

This results in the following averaging influences for NO:

System 1: 3.1 %

System 2: -2.9 %

Evaluation for the component NO<sub>2</sub> during the test with NO as described above.

Table 13: Results of the averaging test for NO<sub>2</sub>

|                               | requirement | device 1 |   | device 2 |   |
|-------------------------------|-------------|----------|---|----------|---|
| averaging effect $E_{av}$ [%] | ≤ 7%        | 1.5      | ✓ | 1.7      | ✓ |

This results in the following averaging influences for NO<sub>2</sub>:

System 1: 1.5 %

System 2: 1.7 %

## 6.5 Assessment

The performance criterion of VDI 4202-1 and EN 14211 is fully complied with.

Criterion satisfied? yes

## 6.6 Detailed presentation of test results

In Table 14 and in Table 15 the individual results of the investigations into the averaging influence are shown.

Table 14: Individual values of the averaging influence study for NO

|  |          | device 1   | device 2   |
|--|----------|------------|------------|
|  | time     | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 11:34:00 | 602.0      | 615.5      |
|  | till     |            |            |
|  | 11:53:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 11:54:00 | 298.9      | 311.3      |
|  | till     |            |            |
|  | 12:13:00 |            |            |

|  |          | device 1   | device 2   |
|--|----------|------------|------------|
|  | time     | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 12:24:00 | 605.8      | 617.6      |
|  | till     |            |            |
|  | 12:43:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 12:44:00 | 293.2      | 314.5      |
|  | till     |            |            |
|  | 13:03:00 |            |            |

|  |          | device 1   | device 2   |
|--|----------|------------|------------|
|  | time     | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 13:15:00 | 607.1      | 617.8      |
|  | till     |            |            |
|  | 13:34:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 13:35:00 | 287.0      | 326.0      |
|  | till     |            |            |
|  | 13:54:00 |            |            |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Table 15: Individual values of the averaging influence study for NO<sub>2</sub>

|  |          | device 1   | device 2   |
|--|----------|------------|------------|
|  | time     | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 11:34:00 | 218.5      | 218.5      |
|  | till     |            |            |
|  | 11:53:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 11:54:00 | 215.5      | 215.8      |
|  | till     |            |            |
|  | 12:13:00 |            |            |

|  |          | device 1   | device 2   |
|--|----------|------------|------------|
|  | time     | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 12:24:00 | 218.8      | 220.5      |
|  | till     |            |            |
|  | 12:43:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 12:44:00 | 215.6      | 216.1      |
|  | till     |            |            |
|  | 13:03:00 |            |            |

|  |          | device 1   | device 2   |
|--|----------|------------|------------|
|  | time     | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 13:15:00 | 219.1      | 221.8      |
|  | till     |            |            |
|  | 13:34:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 13:35:00 | 215.6      | 217.3      |
|  | till     |            |            |
|  | 13:54:00 |            |            |

## **6.1 7.4.13 Difference between sample and calibration port**

*If a measuring system, standardly or optionally, possesses a test gas inlet separated from the sample gas inlet, this configuration shall be tested.*

*The difference between the measured values obtained by feeding gas at the sample gas and test gas inlet shall not exceed 1 %.*

### **6.2 Equipment**

Not applicable

### **6.3 Testing**

Determination and evaluation of the difference between sample and calibration port corresponds exactly to determining the difference sample/calibration port in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.4.13 Difference sample/calibration port.

### **6.4 Evaluation**

See chapter 7.1 8.4.13 Difference sample/calibration port

### **6.5 Assessment**

See chapter 7.1 8.4.13 Difference sample/calibration port  
Criterion satisfied? yes

### **6.6 Detailed presentation of test results**

Not applicable in this instance.

## 6.1 7.4.14 Converter efficiency

*In case of measuring systems with a converter, the converter efficiency shall be at least 98 % in the laboratory test.*

## 6.2 Equipment

Not applicable

## 6.3 Testing

Determination and evaluation of the converter efficiency in the laboratory corresponds exactly to determining the converter efficiency in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.4.14 Converter efficiency.

The N500 measuring system does not work according to the EU standard reference method of chemiluminescence. Therefore, the measuring system does not have a standard NO-NO<sub>2</sub> converter. The N500 measuring system analyses NO<sub>2</sub> directly in the UV range. NO is determined by oxidising the sample air with ozone in alternating cycles (gas phase titration). The difference between the two measurements is displayed as the NO value. Even if there is no converter in the measuring system to be tested, the test was carried out as indicated above to show that the NO and NO<sub>2</sub> measurement results are equivalent to the standard reference method.

## 6.4 Evaluation

See chapter 7.1 8.4.14 Converter efficiency

## 6.5 Assessment

See chapter 7.1 8.4.14 Converter efficiency

Criterion satisfied? yes

## 6.6 Detailed presentation of test results

Not applicable in this instance.

## **6.1 7.4.15 Residence time in the analyser**

*If the residence time has influence on the output signal, like for NO<sub>x</sub> and ozone measuring systems, it is necessary to calculate the residence time from the volume flow and the volume of the gas lines and other relevant components of the measuring system and the particle filter casing.*

*In the case of NO<sub>x</sub> and O<sub>3</sub> measurements, the residence time shall not exceed 3 s.*

## **6.2 Equipment**

Not applicable

## **6.3 Testing**

Performing and evaluating the averaging effect corresponds exactly to determining the averaging test in accordance with standard EN 14211 (2012). This is why we refer to Chapter 7.1 8.4.14 Residence time in the analyser.

## **6.4 Evaluation**

See chapter 7.1 8.4.14 Residence time in the analyser

## **6.5 Assessment**

See chapter 7.1 8.4.14

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable in this instance.

## **6.1 7.5 Requirements on performance characteristics for testing in the field**

### **6.1 7.5.1 General requirements**

*The performance characteristics which shall be determined during testing in the field and their related performance criteria for measured components according to 39<sup>th</sup> BImSchV are given in Table A1 of VDI 4202-1 (2018).*

*The certification range for other components is to be defined. Performance criteria are to be defined by drawing from Table A1 of VDI 4202-1 (2018) These definitions shall be cleared with the relevant body before testing.*

*The determination of the performance characteristics shall be done according to the procedures described in Section 8.5 of VDI 4202-1 (2018).*

### **6.2 Equipment**

Not required for this performance criterion.

### **6.3 Testing**

Tests were performed using the performance characteristics specified in VDI standard 4202, part 1 (2018) and standard EN 14211 (2012).

### **6.4 Evaluation**

Not applicable.

### **6.5 Assessment**

Tests were performed using the performance characteristics specified in VDI standard 4202, part 1 (2018) and standard EN 14211 (2012)

Criterion satisfied? yes

### **6.6 Detailed presentation of test results**

Not applicable.

## **6.1 7.5.2 Location for the field test**

*The monitoring station for the field test is to be chosen according to the requirements of 39<sup>th</sup> BImSchV such that the expected concentrations of the measured components to be measured correspond to the designated task. The equipment of the monitoring station shall allow the implementation of the field test and shall fulfil all requirements considered to be necessary during measurement planning.*

## **6.2 Equipment**

Not required for this performance criterion.

## **6.3 Testing**

The field test location was selected in compliance with the 39<sup>th</sup> BImSchV.

## **6.4 Evaluation**

The field test location was selected in compliance with the 39<sup>th</sup> BImSchV. Details on the location of the measuring station are given in chapter 4.3.

## **6.5 Assessment**

The field test location was selected in compliance with the 39<sup>th</sup> BImSchV. Details on the location of the measuring station are given in chapter 4.3.

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable.



## 6.1 7.5.3 Test requirements

*The measuring systems shall be installed in the monitoring station and, after connecting to the existing or separate sampling system, activated properly.*

*The adjustments of the measuring system shall meet the specifications of the manufacturer. All adjustments are to be documented in the test report.*

*The measuring systems shall be maintained during the field test, following the manufacturer's specifications, and shall be checked with suitable test gases regularly.*

*If the measuring system contains auto-scale or self-correction functions and they are treated as "normal operating conditions", these functions shall be turned on during the field test. Values of the self-correction shall be available to the test laboratory.*

*The values of the auto-zero and auto-drift corrections for the inspection interval (long-term drift) are subject to the same restrictions as given in the performance characteristics.*

## 6.2 Equipment

Not required for this performance criterion.

## 6.3 Testing

For the purpose of field testing, the measuring system was mounted in a measuring station and connected to the existing sampling system. The measuring system was then commissioned following the manufacturer's instructions in the manual.

Neither self-correction nor auto-zero functions were activated during the field test.

## 6.4 Evaluation

During the field test, the measuring system was operated and serviced according to the manufacturer's instructions. Neither self-correction nor auto-zero functions were activated.

## 6.5 Assessment

During the field test, the measuring system was operated and serviced according to the manufacturer's instructions.

Criterion satisfied? yes

## 6.6 Detailed presentation of test results

Not applicable.

## **6.1 7.5.4 Long-term drift**

*The long-term drift at zero point shall not exceed 5.0 nmol/mol.*

*The long-term drift at reference point shall not exceed 5 % of the upper limit of the certification range.*

## **6.2 Equipment**

Not applicable

## **6.3 Testing**

Determination and evaluation of the long-term drift corresponds exactly to determining the long term drift in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.5.4 Long-term drift.

## **6.4 Evaluation**

See chapter 7.1 8.5.4 Long-term drift.

## **6.5 Assessment**

See chapter 7.1 8.5.4 Long-term drift.

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable in this instance.

## **6.1 7.5.5 Reproducibility standard deviation under field conditions**

*The standard deviation from paired measurements under field conditions shall be determined with two identical measuring systems by paired measurements in the field test.*

*The standard deviation under field conditions shall not exceed 5% of the mean value over a period of three months.*

## **6.2 Equipment**

Not applicable

## **6.3 Testing**

Performing and evaluating the standard deviation from paired measurements corresponds exactly to determining the reproducibility standard deviation in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.5.5 Reproducibility standard deviation for NO<sub>2</sub> under field conditions.

## **6.4 Evaluation**

See chapter 7.1 8.5.5 Reproducibility standard deviation for NO<sub>2</sub> under field conditions

## **6.5 Assessment**

See chapter 7.1 8.5.5 Reproducibility standard deviation for NO<sub>2</sub> under field conditions  
Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable in this instance.

## **6.1 7.5.6 Inspection interval**

*The inspection interval of the measuring system shall be determined during the field test and specified. The maintenance interval should be three months, if possible, but at least two weeks.*

## **6.2 Equipment**

Not required for this performance criterion.

## **6.3 Testing**

Performing and evaluating the inspection interval corresponds exactly to determining the period of unattended operation in accordance with standard EN 14211 (2012). The reader is therefore referred to chapter 7.1 8.5.6 Inspection interval.

## **6.4 Evaluation**

See chapter 7.1 8.5.6 Inspection interval.

## **6.5 Assessment**

See chapter 7.1 8.5.6 Inspection interval.

Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable in this instance.

## **6.1 7.5.7 Availability**

*The availability of the measuring system shall be determined during the field test and shall be at least 95%.*

## **6.2 Equipment**

Not applicable

## **6.3 Testing**

Determination and evaluation of the availability corresponds exactly to determining the period of availability of the analyser in accordance with standard EN 14211 (2012). The reader is therefore referred to the chapter 7.1 8.5.7 Period of availability of the analyser.

## **6.4 Evaluation**

See chapter 7.1 8.5.7 Period of availability of the analyser

## **6.5 Assessment**

See chapter 7.1 8.5.7 Period of availability of the analyser  
Criterion satisfied? yes

## **6.6 Detailed presentation of test results**

Not applicable in this instance.

## 6.1 7.5.8 Converter efficiency

*At the end of the field test, the converter efficiency shall be at least 95 %.*

## 6.2 Equipment

Test samples, ozonator, NO test gas

## 6.3 Testing

VDI 4202-1:2018 additionally requires the converter efficiency to be tested at the end of the field test. This test was performed following the procedures of EN 14211 for converter efficiency in the laboratory test.

The converter efficiency is determined by measurements with calculated amounts of NO<sub>2</sub>. This can be achieved by means of gas-phase titration of NO to NO<sub>2</sub> with ozone.

The test shall be performed at two concentration levels: at about 50% and about 95% of the maximum of the certification range of NO<sub>2</sub>.

The NO<sub>x</sub> measuring system shall be calibrated applying an NO concentration of about 70% to 80% of the maximum certification range for NO to the NO and NO<sub>x</sub> channels. Both channels need to be adjusted to display the same value. The values shall be recorded.

A known NO concentration at 50% of the NO certification range is applied to the measuring system until a stable signal is obtained. This stable period shall be at least as long as four response times. Four individual readings each are performed at the NO and the NO<sub>x</sub> channel. Then, NO is brought to react with O<sub>3</sub> to produce NO<sub>2</sub>. This mixture containing a constant NO<sub>x</sub> concentration is applied to the measuring system until the output signal has stabilised. This stable period shall be at least four response times of the measuring system under test; the NO concentration after gas-phase titration shall be between 10% and 20% of the original NO concentration. Then, four individual readings each are performed at the NO and the NO<sub>x</sub> channel. Then, the O<sub>3</sub> supply is cut and only NO is applied to the measuring system until the output signal has stabilised again. This stable period shall be at least as long as four response times of the measuring system. After that the average of the four individual measurements at the NO and the NO<sub>x</sub> channel is calculated.

The converter efficiency is calculated as follows:

$$E_{conv} = \left( 1 - \frac{(NO_x)_i - (NO_x)_f}{(NO)_i - (NO)_f} \right) \times 100\%$$

Where:

$E_{conv}$  is the converter efficiency in %;

$(NO_x)_i$  is the average of the four individual measurements at the NO<sub>x</sub> channel at the initial NO<sub>x</sub> concentration;

$(NO_x)_f$  is the average of the four individual measurements at the NO<sub>x</sub> channel at the resulting NO<sub>x</sub> concentration after applying O<sub>3</sub>;

$(NO)_i$  is the average of the four individual measurements at the NO channel at the initial NO concentration;

$(NO)_f$  is the average of the four individual measurements at the NO channel at the resulting NO concentration after applying O<sub>3</sub>;

The lowest value of the two converter efficiencies shall be reported.

The N500 measuring system does not work according to the EU standard reference method of chemiluminescence. Therefore, the measuring system does not have a standard NO-NO<sub>2</sub> converter. The N500 measuring system analyses NO<sub>2</sub> directly in the UV range. NO is determined by oxidising the sample air with ozone in alternating cycles. The difference between the two measurements is displayed as the NO value. Even if there is no converter in the measuring system to be tested, the test was carried out as indicated above to show that the NO and NO<sub>2</sub> measurement results are equivalent to the standard reference method.

The test was performed in compliance with the requirements specified in EN 14211. When applying test gas two NO<sub>2</sub> concentrations in the range of 50% to 95% of the certification range for NO<sub>2</sub> were adjusted by means of gas-phase titration.

The converter efficiency was determined at the end of the field test.

#### 6.4 Evaluation

During testing, the following converter efficiencies were determined for the two N500 measuring systems. The lowest value of the two NO<sub>2</sub> concentration is reported below.

|                                | requirement | device 1 |   | device 2 |   |
|--------------------------------|-------------|----------|---|----------|---|
| converter efficiency $E_c$ [%] | ≥ 95%       | 99.6     | ✓ | 99.5     | ✓ |

#### 6.5 Assessment

The performance criterion specified by standard VDI 4202-1 (2018) is fully satisfied.

Criterion satisfied? yes

#### 6.6 Detailed presentation of test results

Table 16 presents the individual values.

**Table 16: Individual results for the converter efficiency**

|                                 | time     | O <sub>3</sub> [nmol/mol] | NO <sub>2</sub> [nmol/mol] | device 1      |                            | device 2      |                            |
|---------------------------------|----------|---------------------------|----------------------------|---------------|----------------------------|---------------|----------------------------|
|                                 |          |                           |                            | NO [nmol/mol] | NO <sub>x</sub> [nmol/mol] | NO [nmol/mol] | NO <sub>x</sub> [nmol/mol] |
|                                 | 11:14:00 |                           |                            | start         |                            |               |                            |
| O <sub>3</sub> =0, NO=50%       | 11:33:00 | 0.0                       | 0.0                        | 484.0         | 490.1                      | 483.9         | 492.8                      |
|                                 | 11:34:00 | 0.0                       | 0.0                        | 491.3         | 491.5                      | 489.9         | 491.6                      |
|                                 | 11:35:00 | 0.0                       | 0.0                        | 487.6         | 489.7                      | 488.7         | 491.6                      |
|                                 | 11:36:00 | 0.0                       | 0.0                        | 486.4         | 489.7                      | 486.3         | 491.6                      |
| average                         |          | 0.0                       | 0.0                        | 487.3         | 490.2                      | 487.2         | 491.9                      |
| NO <sub>2</sub> = 50%<br>130.75 | 11:58:00 | 130.0                     | 133.5                      | 353.5         | 489.7                      | 357.1         | 492.2                      |
|                                 | 11:59:00 | 130.0                     | 133.2                      | 352.9         | 489.7                      | 356.5         | 491.0                      |
|                                 | 12:00:00 | 130.0                     | 132.1                      | 353.5         | 489.7                      | 356.5         | 491.0                      |
|                                 | 12:01:00 | 130.0                     | 134.5                      | 352.9         | 489.7                      | 356.5         | 491.0                      |
| average                         |          | 130.0                     | 133.3                      | 353.2         | 489.7                      | 356.7         | 491.3                      |
|                                 |          |                           |                            |               |                            |               |                            |
| O <sub>3</sub> =0, NO=50%       | 13:01:00 | 0.0                       | 0.0                        | 484.7         | 491.7                      | 489.8         | 492.5                      |
|                                 | 13:02:00 | 0.0                       | 0.0                        | 492.0         | 491.1                      | 489.2         | 494.3                      |
|                                 | 13:03:00 | 0.0                       | 0.0                        | 488.4         | 491.1                      | 488.0         | 493.7                      |
|                                 | 13:04:00 | 0.0                       | 0.0                        | 487.1         | 494.1                      | 488.0         | 490.7                      |
| average                         |          | 0.0                       | 0.0                        | 488.1         | 492.0                      | 488.7         | 492.8                      |
| NO <sub>2</sub> = 95%<br>248.43 | 13:25:00 | 248.0                     | 252.6                      | 239.2         | 491.7                      | 237.7         | 494.9                      |
|                                 | 13:26:00 | 248.0                     | 250.2                      | 241.6         | 491.7                      | 237.7         | 492.5                      |
|                                 | 13:27:00 | 248.0                     | 251.4                      | 239.8         | 491.1                      | 235.9         | 494.3                      |
|                                 | 13:28:00 | 248.0                     | 253.2                      | 237.9         | 491.1                      | 235.3         | 493.7                      |
| average                         |          | 248.0                     | 251.8                      | 239.6         | 491.4                      | 236.7         | 493.8                      |
| O <sub>3</sub> =0, NO=50%       | 13:38:00 | 0.0                       | 0.0                        | 487.1         | 491.7                      | 483.2         | 494.9                      |



## 6.1 7.6 Type approval and calculation of the measurement uncertainty

*The type approval of the measuring system requires the following:*

- 1) The value of each individual performance characteristic tested in the laboratory shall fulfil the criterion stated in Table A1 of VDI 4202-1 (2018).*
- 2) The expanded uncertainty calculated from the standard uncertainties due to the values of the specific performance characteristics obtained in the laboratory tests shall fulfil the criterion as stated Table C1 of VDI 4202-1 (2018). This criterion is the maximum uncertainty of individual measurements for continuous measurements at the 1-hour limit value. The relevant specific performance characteristics and the calculation procedure are given in Annex F of standard VDI 4202-1 (2018).*
- 3) The value of each individual performance characteristic tested in the laboratory shall fulfil the criterion stated in Table A1 of VDI 4202-1 (2018).*
- 4) The expanded uncertainty calculated from the standard uncertainties due to the values of the specific performance characteristics obtained in the laboratory and field tests shall fulfil the criterion as stated Table C1 of VDI 4202-1 (2018). This criterion is the maximum uncertainty of individual measurements for continuous measurements at the 1-hour limit value. The relevant specific performance characteristics and the calculation procedure are given in Annex F of standard VDI 4202-1 (2018).*

### 6.2 Equipment

Not applicable

### 6.3 Testing

Uncertainty calculation was performed in line with standard EN 14211 (2012) and is presented in 7.1 8.6 Calculation of the total uncertainty in accordance with standard EN 14211 (2012) according to Annex E of EN 14211 (2012).

### 6.4 Evaluation

Uncertainty calculation was performed in line with standard EN 14211 (2012) and is presented in 7.1 8.6 Calculation of the total uncertainty in accordance with standard EN 14211 (2012) according to Annex E of EN 14211 (2012).

### 6.5 Assessment

Uncertainty calculation was performed in line with standard EN 14211 (2012) and is presented in 7.1 8.6 Calculation of the total uncertainty in accordance with standard EN 14211 (2012) according to Annex E of EN 14211 (2012)

Criterion satisfied? yes

### 6.6 Detailed presentation of test results

Not applicable in this instance.

## 7. Test results in accordance with Standard EN 14211 (2012)

### 7.1 8.4.3 Response time

*Rise and fall response time ≤ 180 s each. Difference between rise and fall response time ≤ 10 s.*

### 7.2 Testing

The determination of the response time shall be carried out by applying to the analyser a step function in the concentration from less than 20 % to about 80 % of the maximum of the certification range of NO and vice versa.

The change from zero gas to span gas and vice versa needs to be made almost instantaneously, with the use of a suitable valve. The valve outlet must be mounted directly at the inlet of the meter and both zero and span gas must be supplied with the same excess, which is discharged by means of a T-piece. The gas flows of zero and span gas must be selected in such a way that the dead time in the valve and in the T-piece is negligible compared to the dead time of the meter. The step change is made by switching the valve from zero gas to span gas. This event needs to be timed and is the start ( $t = 0$ ) of the (rise) lag time for the dead time (rise) as shown in Figure 10. When the reading shows 98% of the applied concentration, the span gas can be changed to zero gas again; this event is the start ( $t = 0$ ) of the (fall) lag time. When the reading shows 2% of the applied concentration, the whole cycle as shown in Figure 10 is complete.

The elapsed time (response time) between the start of the step change and reaching 90% of the analyser final stable reading of the applied concentration shall be measured. The whole cycle shall be repeated four times. The average of the four response times (rise) and the average of the four response times (fall) shall be calculated.

The test shall then be repeated with NO<sub>2</sub> at levels from less than 20 % to about 80 % of the maximum of the certification range of NO<sub>2</sub> and vice versa.


The difference in response times shall be calculated according to:

$$t_d = \bar{t}_r - \bar{t}_f$$

Where  $T_d$  is the difference between response time (rise) and response time (fall), in s;  
 $t_r$  is the response time (rise) (average of the four response times - rise), in s;  
 $t_f$  is the response time (fall) (average of the four response times - fall), in s.

$t_r$ ,  $t_f$  and  $t_d$  shall comply with the performance criteria indicated above.

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A



**Key**

- Y analyser response
- X time
- 1 lag time
- 2 rise time
- 3 response time (rise)
- 4 fall time
- 5 response time (fall)

Figure 10: Diagram illustrating the response time

### 7.3 Testing

The test was performed in line with the requirements of EN 14211 mentioned before. An external data logger was used to record data.

## 7.4 Evaluation

Table 17: Response times of the two N500 measuring systems for NO

|                        | requirement  | device 1 |   | device 2 |   |
|------------------------|--------------|----------|---|----------|---|
| average rise $t_r$ [s] | $\leq 180$ s | 24       | ✓ | 25       | ✓ |
| average fall $t_f$ [s] | $\leq 180$ s | 22       | ✓ | 23       | ✓ |
| difference $t_d$ [s]   | $\leq 10$ s  | 2.0      | ✓ | 2.0      | ✓ |

For system 1, the average  $t_r$  for NO was 24 sec, the average  $t_f$  was 22 sec and the average  $t_d$  was 2 sec.

For system 2, the average  $t_r$  for NO was 25 sec, the average  $t_f$  was 23 sec and the average  $t_d$  was 2 sec.

 Table 18: Response times of the two N500 measuring systems for NO<sub>2</sub>

|                        | requirement  | device 1 |   | device 2 |   |
|------------------------|--------------|----------|---|----------|---|
| average rise $t_r$ [s] | $\leq 180$ s | 24       | ✓ | 25       | ✓ |
| average fall $t_f$ [s] | $\leq 180$ s | 22       | ✓ | 22       | ✓ |
| difference $t_d$ [s]   | $\leq 10$ s  | 2.0      | ✓ | 3.0      | ✓ |

For unit 1, the average  $t_r$  for NO<sub>2</sub> was 24 sec, the average  $t_f$  was 22 sec and the average  $t_d$  was 2 sec.

For unit 2, the average  $t_r$  for NO<sub>2</sub> was 25 sec, the average  $t_f$  was 22 sec and the average  $t_d$  was 3 sec.

## 7.5 Assessment

The values determined remained considerably below the maximum permissible response time of 180 s at all times. The maximum determined response time for unit 1 was 24 sec for NO and 24 sec for NO<sub>2</sub>. For unit 2 it was 25 sec for NO and 25 sec for NO<sub>2</sub>.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

Table 19: Individual results of the response time for NO

| 80%             |             | device 1    |               |               |               |              |             |
|-----------------|-------------|-------------|---------------|---------------|---------------|--------------|-------------|
|                 |             | rise        |               |               | fall          |              |             |
| measuring range | 768.00      | 0.0<br>0.00 | 0.9<br>691.20 | 1.0<br>768.00 | 1.0<br>768.00 | 0.1<br>76.80 | 0.0<br>0.00 |
| cycle 1         | t = 0       | 15:14:00    | 15:14:24      | 15:15:00      | 15:20:00      | 15:20:21     | 15:22:00    |
|                 | delta t     |             | 00:00:24      |               |               | 00:00:21     |             |
|                 | delta t [s] |             | 24            |               |               | 21           |             |
| cycle 2         | t = 0       | 15:27:00    | 15:27:24      | 15:28:00      | 15:33:00      | 15:33:23     | 15:34:00    |
|                 | delta t     |             | 00:00:24      |               |               | 00:00:23     |             |
|                 | delta t [s] |             | 24            |               |               | 23           |             |
| cycle 3         | t = 0       | 15:39:00    | 15:39:26      | 15:40:00      | 15:45:00      | 15:45:22     | 15:46:00    |
|                 | delta t     |             | 00:00:26      |               |               | 00:00:22     |             |
|                 | delta t [s] |             | 26            |               |               | 22           |             |
| cycle 4         | t = 0       | 15:51:00    | 15:51:22      | 15:52:00      | 15:57:00      | 15:57:22     | 15:58:00    |
|                 | delta t     |             | 00:00:22      |               |               | 00:00:22     |             |
|                 | delta t [s] |             | 22            |               |               | 22           |             |

| 80%             |             | device 2    |               |               |               |              |             |
|-----------------|-------------|-------------|---------------|---------------|---------------|--------------|-------------|
|                 |             | rise        |               |               | fall          |              |             |
| measuring range | 768.00      | 0.0<br>0.00 | 0.9<br>691.20 | 1.0<br>768.00 | 1.0<br>768.00 | 0.1<br>76.80 | 0.0<br>0.00 |
| cycle 1         | t = 0       | 15:14:00    | 15:14:25      | 15:15:00      | 15:20:00      | 15:20:22     | 15:22:00    |
|                 | delta t     |             | 00:00:25      |               |               | 00:00:22     |             |
|                 | delta t [s] |             | 25            |               |               | 22           |             |
| cycle 2         | t = 0       | 15:27:00    | 15:27:27      | 15:28:00      | 15:33:00      | 15:33:26     | 15:34:00    |
|                 | delta t     |             | 00:00:27      |               |               | 00:00:26     |             |
|                 | delta t [s] |             | 27            |               |               | 26           |             |
| cycle 3         | t = 0       | 15:39:00    | 15:39:24      | 15:40:00      | 15:45:00      | 15:45:21     | 15:46:00    |
|                 | delta t     |             | 00:00:24      |               |               | 00:00:21     |             |
|                 | delta t [s] |             | 24            |               |               | 21           |             |
| cycle 4         | t = 0       | 15:51:00    | 15:51:24      | 15:52:00      | 15:57:00      | 15:57:23     | 15:58:00    |
|                 | delta t     |             | 00:00:24      |               |               | 00:00:23     |             |
|                 | delta t [s] |             | 24            |               |               | 23           |             |

**Table 20: Individual results of the response time for NO<sub>2</sub>**

| 80%             |             | device 1    |               |               |               |              |             |
|-----------------|-------------|-------------|---------------|---------------|---------------|--------------|-------------|
| measuring range | 209.21      | rise        |               |               | fall          |              |             |
|                 |             | 0.0<br>0.00 | 0.9<br>188.28 | 1.0<br>209.21 | 1.0<br>209.21 | 0.1<br>20.92 | 0.0<br>0.00 |
| cycle 1         | t = 0       | 07:46:00    | 07:46:24      | 07:47:00      | 07:52:00      | 07:52:22     | 07:53:00    |
|                 | delta t     |             | 00:00:24      |               |               | 00:00:22     |             |
|                 | delta t [s] |             | 24            |               |               | 22           |             |
| cycle 2         | t = 0       | 08:00:00    | 08:00:24      | 08:01:00      | 08:06:00      | 08:06:23     | 08:07:00    |
|                 | delta t     |             | 00:00:24      |               |               | 00:00:23     |             |
|                 | delta t [s] |             | 24            |               |               | 23           |             |
| cycle 3         | t = 0       | 08:11:00    | 08:11:24      | 08:12:00      | 08:17:00      | 08:17:23     | 08:18:00    |
|                 | delta t     |             | 00:00:24      |               |               | 00:00:23     |             |
|                 | delta t [s] |             | 24            |               |               | 23           |             |
| cycle 4         | t = 0       | 08:23:00    | 08:23:24      | 08:24:00      | 08:29:00      | 08:29:20     | 08:30:00    |
|                 | delta t     |             | 00:00:24      |               |               | 00:00:20     |             |
|                 | delta t [s] |             | 24            |               |               | 20           |             |

| 80%             |             | device 2    |               |               |               |              |             |
|-----------------|-------------|-------------|---------------|---------------|---------------|--------------|-------------|
| measuring range | 209.21      | rise        |               |               | fall          |              |             |
|                 |             | 0.0<br>0.00 | 0.9<br>188.28 | 1.0<br>209.21 | 1.0<br>209.21 | 0.1<br>20.92 | 0.0<br>0.00 |
| cycle 1         | t = 0       | 07:46:00    | 07:46:25      | 07:47:00      | 07:52:00      | 07:52:23     | 07:53:00    |
|                 | delta t     |             | 00:00:25      |               |               | 00:00:23     |             |
|                 | delta t [s] |             | 25            |               |               | 23           |             |
| cycle 2         | t = 0       | 08:00:00    | 08:00:24      | 08:01:00      | 08:06:00      | 08:06:22     | 08:07:00    |
|                 | delta t     |             | 00:00:24      |               |               | 00:00:22     |             |
|                 | delta t [s] |             | 24            |               |               | 22           |             |
| cycle 3         | t = 0       | 08:11:00    | 08:11:25      | 08:12:00      | 08:17:00      | 08:17:22     | 08:18:00    |
|                 | delta t     |             | 00:00:25      |               |               | 00:00:22     |             |
|                 | delta t [s] |             | 25            |               |               | 22           |             |
| cycle 4         | t = 0       | 08:23:00    | 08:23:26      | 08:24:00      | 08:29:00      | 08:29:21     | 08:30:00    |
|                 | delta t     |             | 00:00:26      |               |               | 00:00:21     |             |
|                 | delta t [s] |             | 26            |               |               | 21           |             |

## 7.1 8.4.4 Short-term drift

*Short-term drift at zero shall not exceed 2.0 nmol/mol/12 h.*

*The short-term drift at span level shall not exceed 6.0 nmol/mol/12 h.*

## 7.2 Testing

After the required stabilisation period, the analyser shall be adjusted at zero and span level (around 70% to 80% of the maximum of the certification range). Wait the time equivalent to one independent reading and then record 20 individual measurements, first at zero and then at span concentration. From these 20 measurements, the average is calculated for zero and span level.

The analyser shall be kept running under the laboratory conditions. After a period of 12 h, zero and span gas is fed to the analyser. Wait the time equivalent to one independent reading and then record 20 individual measurements, first at zero and then at span concentration. The averages for zero and span level shall be calculated.

The short-term drift at zero and span level shall be calculated as follows:

$$D_{S,Z} = (C_{Z,2} - C_{Z,1})$$

Where:

$D_{S,Z}$  is the 12-hour drift at zero;

$C_{Z,1}$  is the average concentration of the measurements at zero at the beginning of the drift period;

$C_{Z,2}$  is the average concentration of the measurements at zero at the end of the drift period;

$D_{S,Z}$  shall comply with the performance criterion indicated above.

$$D_{S,S} = (C_{S,2} - C_{S,1}) - D_{S,Z}$$

Where:

$D_{S,S}$  is the 12-hour drift at span;

$C_{S,1}$  is the average concentration of the measurements at span level at the beginning of the drift period;

$C_{S,2}$  is the average concentration of the measurements at span level at the end of the drift period.

$D_{S,S}$  shall comply with the performance criterion indicated above.

### 7.3 Testing

The test was performed in line with the requirements of EN 14211 mentioned before. In compliance with the standard, the test has to be performed with the component NO. Pursuant to EN 14211, the test shall be performed at a concentration level of 70% to 80% of the certification range for NO.

The test was also carried out for the NO<sub>2</sub> component at a concentration level of 70 % to 80 % of the certification range for NO<sub>2</sub>.

### 7.4 Evaluation

In Table 21 and in Table 22 the measured values of the short-term drift are presented.

Table 21: Results of the short-term drift test, component NO

|   | requirements | device 1 |   | device 2 |   |
|---|--------------|----------|---|----------|---|
| average at zero at the beginning [nmol/mol] | -            | -0.54    |   | -0.30    |   |
| average at zero at the end (12h) [nmol/mol] | -            | -0.48    |   | -0.57    |   |
| average at span at the beginning [nmol/mol] | -            | 708.06   |   | 714.83   |   |
| average at span at the end (12h) [nmol/mol] | -            | 710.48   |   | 715.66   |   |
| 12-hour drift at zero $D_{s,z}$ [nmol/mol]  | ≤ 2.0        | 0.06     | ✓ | -0.27    | ✓ |
| 12-hour drift at span $D_{s,s}$ [nmol/mol]  | ≤ 6.0        | 2.36     | ✓ | 1.11     | ✓ |

Table 22: Results of the short-term drift test, component NO<sub>2</sub>

|   | requirements | device 1 |   | device 2 |   |
|---|--------------|----------|---|----------|---|
| average at zero at the beginning [nmol/mol] | -            | -0.13    |   | -0.27    |   |
| average at zero at the end (12h) [nmol/mol] | -            | -0.16    |   | -0.53    |   |
| average at span at the beginning [nmol/mol] | -            | 207.91   |   | 208.74   |   |
| average at span at the end (12h) [nmol/mol] | -            | 210.64   |   | 211.20   |   |
| 12-hour drift at zero $D_{s,z}$ [nmol/mol]  | ≤ 2.0        | -0.03    | ✓ | -0.26    | ✓ |
| 12-hour drift at span $D_{s,s}$ [nmol/mol]  | ≤ 6.0        | 2.76     | ✓ | 2.72     | ✓ |



## 7.5 Assessment

For NO, the value for the short-term drift at zero point was 0.06 nmol/mol for instrument 1; for instrument 2, it was -0.27 nmol/mol.

For NO, the value for the short-term drift at span point was 2.36 nmol/mol for instrument 1; for instrument 2, it was 1.11 nmol/mol.

For NO<sub>2</sub>, the value for the short-term drift at zero point was -0.03 nmol/mol for instrument 1; for instrument 2, it was -0.26 nmol/mol.

For NO<sub>2</sub>, the value for the short-term drift at span point was 2.76 nmol/mol for instrument 1; for instrument 2, it was 2.72 nmol/mol.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

The individual values of the test are shown in Table 23 to Table 26.

Table 23: Results of the short-term drift 1<sup>st</sup> test gas for NO

| at beginning (19.10.2020) |             |             | at beginning (19.10.2020) |              |              |
|---------------------------|-------------|-------------|---------------------------|--------------|--------------|
| zero level                |             |             | span level                |              |              |
|                           | device 1    | device 2    |                           | device 1     | device 2     |
| time                      | [nmol/mol]  | [nmol/mol]  | time                      | [nmol/mol]   | [nmol/mol]   |
| 16:36:00                  | 0.0         | 0.0         | 17:11:00                  | 707.0        | 713.7        |
| 16:37:00                  | -0.6        | 0.0         | 17:12:00                  | 706.4        | 713.7        |
| 16:38:00                  | -0.6        | -0.6        | 17:13:00                  | 705.7        | 713.1        |
| 16:39:00                  | -0.6        | -0.6        | 17:14:00                  | 710.3        | 719.1        |
| 16:40:00                  | -0.6        | 0.0         | 17:15:00                  | 708.8        | 715.5        |
| 16:41:00                  | -0.6        | -0.6        | 17:16:00                  | 707.0        | 713.7        |
| 16:42:00                  | -0.6        | -0.6        | 17:17:00                  | 706.9        | 713.7        |
| 16:43:00                  | -0.6        | -0.6        | 17:18:00                  | 713.3        | 717.9        |
| 16:44:00                  | -0.6        | -0.6        | 17:19:00                  | 707.6        | 714.3        |
| 16:45:00                  | -0.6        | 0.0         | 17:20:00                  | 707.0        | 713.7        |
| 16:46:00                  | -0.6        | 0.0         | 17:21:00                  | 705.7        | 713.1        |
| 16:47:00                  | -0.6        | 0.0         | 17:22:00                  | 710.3        | 716.7        |
| 16:48:00                  | -0.6        | 0.0         | 17:23:00                  | 707.0        | 715.5        |
| 16:49:00                  | 0.0         | -0.6        | 17:24:00                  | 708.2        | 713.7        |
| 16:50:00                  | -0.6        | 0.0         | 17:25:00                  | 704.6        | 710.1        |
| 16:51:00                  | -0.6        | 0.0         | 17:26:00                  | 715.0        | 719.1        |
| 16:52:00                  | -0.6        | 0.0         | 17:27:00                  | 708.8        | 714.9        |
| 16:53:00                  | -0.6        | -0.6        | 17:28:00                  | 706.4        | 714.9        |
| 16:54:00                  | -0.6        | -0.6        | 17:29:00                  | 706.4        | 713.1        |
| 16:55:00                  | -0.6        | -0.6        | 17:30:00                  | 708.8        | 717.3        |
| <b>average</b>            | <b>-0.5</b> | <b>-0.3</b> | <b>average</b>            | <b>708.1</b> | <b>714.8</b> |

Table 24: Results of the short-term drift 2<sup>nd</sup> test gas for NO

| after 12h (20.10.2020) |             |             | after 12h (20.10.2020) |              |              |
|------------------------|-------------|-------------|------------------------|--------------|--------------|
| zero level             |             |             | span level             |              |              |
|                        | device 1    | device 2    |                        | device 1     | device 2     |
| time                   | [nmol/mol]  | [nmol/mol]  | time                   | [nmol/mol]   | [nmol/mol]   |
| 04:36:00               | 0.0         | -0.6        | 05:11:00               | 711.1        | 715.3        |
| 04:37:00               | 0.0         | -0.6        | 05:12:00               | 710.8        | 717.9        |
| 04:38:00               | -0.6        | -0.6        | 05:13:00               | 711.8        | 716.5        |
| 04:39:00               | -0.6        | -0.6        | 05:14:00               | 711.2        | 715.3        |
| 04:40:00               | -0.6        | 0.0         | 05:15:00               | 712.3        | 711.7        |
| 04:41:00               | -0.6        | -0.6        | 05:16:00               | 712.5        | 716.1        |
| 04:42:00               | -0.6        | -0.6        | 05:17:00               | 709.3        | 716.5        |
| 04:43:00               | -0.6        | -0.6        | 05:18:00               | 711.2        | 715.3        |
| 04:44:00               | -0.6        | -0.6        | 05:19:00               | 707.6        | 714.1        |
| 04:45:00               | -0.6        | -0.6        | 05:20:00               | 708.9        | 717.8        |
| 04:46:00               | 0.0         | -0.6        | 05:21:00               | 711.3        | 716.6        |
| 04:47:00               | -0.6        | -0.6        | 05:22:00               | 710.6        | 715.9        |
| 04:48:00               | -0.6        | -0.6        | 05:23:00               | 710.6        | 716.5        |
| 04:49:00               | 0.0         | -0.6        | 05:24:00               | 709.3        | 715.3        |
| 04:50:00               | -0.6        | -0.6        | 05:25:00               | 710.4        | 714.9        |
| 04:51:00               | -0.6        | -0.6        | 05:26:00               | 710.6        | 716.5        |
| 04:52:00               | -0.6        | -0.6        | 05:27:00               | 711.2        | 715.9        |
| 04:53:00               | -0.6        | -0.6        | 05:28:00               | 707.5        | 713.5        |
| 04:54:00               | -0.6        | -0.6        | 05:29:00               | 710.2        | 719.1        |
| 04:55:00               | -0.6        | -0.6        | 05:30:00               | 711.2        | 712.3        |
| <b>average</b>         | <b>-0.5</b> | <b>-0.6</b> | <b>average</b>         | <b>710.5</b> | <b>715.7</b> |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO2 and NOx,  
Report No.: 936/21251100/A

Table 25: Results of the short-term drift 1<sup>st</sup> test gas for NO<sub>2</sub>

| at beginning (15.10.2020) |             |             |
|---------------------------|-------------|-------------|
| zero level                |             |             |
|                           | device 1    | device 2    |
| time                      | [nmol/mol]  | [nmol/mol]  |
| 15:41:00                  | 0.0         | -0.2        |
| 15:42:00                  | 0.0         | -0.2        |
| 15:43:00                  | 0.0         | -0.2        |
| 15:44:00                  | 0.0         | -0.2        |
| 15:45:00                  | -0.2        | -0.2        |
| 15:46:00                  | -0.2        | -0.3        |
| 15:47:00                  | -0.2        | -0.2        |
| 15:48:00                  | -0.2        | -0.2        |
| 15:49:00                  | -0.2        | -0.3        |
| 15:50:00                  | -0.2        | -0.3        |
| 15:51:00                  | -0.2        | -0.3        |
| 15:52:00                  | -0.2        | -0.3        |
| 15:53:00                  | -0.2        | -0.3        |
| 15:54:00                  | -0.2        | -0.3        |
| 15:55:00                  | -0.2        | -0.3        |
| 15:56:00                  | -0.2        | -0.3        |
| 15:57:00                  | -0.2        | -0.3        |
| 15:58:00                  | -0.2        | -0.3        |
| 15:59:00                  | -0.2        | -0.3        |
| 16:00:00                  | -0.2        | -0.3        |
| <b>average</b>            | <b>-0.1</b> | <b>-0.3</b> |

| at beginning (15.10.2020) |              |              |
|---------------------------|--------------|--------------|
| span level                |              |              |
|                           | device 1     | device 2     |
| time                      | [nmol/mol]   | [nmol/mol]   |
| 16:16:00                  | 206.8        | 207.8        |
| 16:17:00                  | 207.0        | 207.8        |
| 16:18:00                  | 207.3        | 208.1        |
| 16:19:00                  | 207.3        | 208.1        |
| 16:20:00                  | 207.5        | 208.3        |
| 16:21:00                  | 207.5        | 208.5        |
| 16:22:00                  | 207.7        | 208.6        |
| 16:23:00                  | 207.7        | 208.6        |
| 16:24:00                  | 207.8        | 208.6        |
| 16:25:00                  | 208.0        | 208.8        |
| 16:26:00                  | 208.0        | 208.8        |
| 16:27:00                  | 208.0        | 208.8        |
| 16:28:00                  | 208.1        | 209.0        |
| 16:29:00                  | 208.3        | 209.0        |
| 16:30:00                  | 208.3        | 209.1        |
| 16:31:00                  | 208.5        | 209.1        |
| 16:32:00                  | 208.5        | 209.3        |
| 16:33:00                  | 208.5        | 209.3        |
| 16:34:00                  | 208.6        | 209.5        |
| 16:35:00                  | 208.8        | 209.6        |
| <b>average</b>            | <b>207.9</b> | <b>208.7</b> |

Table 26: Results of the short-term drift 2<sup>nd</sup> test gas for NO<sub>2</sub>

| after 12h (16.10.2020) |             |             | after 12h (16.10.2020) |              |              |
|------------------------|-------------|-------------|------------------------|--------------|--------------|
| zero level             |             |             | span level             |              |              |
|                        | device 1    | device 2    |                        | device 1     | device 2     |
| time                   | [nmol/mol]  | [nmol/mol]  | time                   | [nmol/mol]   | [nmol/mol]   |
| 03:41:00               | -0.2        | -0.5        | 04:16:00               | 210.1        | 210.6        |
| 03:42:00               | -0.2        | -0.5        | 04:17:00               | 210.3        | 210.8        |
| 03:43:00               | -0.2        | -0.5        | 04:18:00               | 210.3        | 210.8        |
| 03:44:00               | -0.2        | -0.5        | 04:19:00               | 210.4        | 210.9        |
| 03:45:00               | -0.2        | -0.5        | 04:20:00               | 210.4        | 210.9        |
| 03:46:00               | -0.2        | -0.5        | 04:21:00               | 210.4        | 211.1        |
| 03:47:00               | -0.2        | -0.5        | 04:22:00               | 210.4        | 211.1        |
| 03:48:00               | -0.2        | -0.5        | 04:23:00               | 210.6        | 211.1        |
| 03:49:00               | -0.2        | -0.5        | 04:24:00               | 210.6        | 211.1        |
| 03:50:00               | -0.2        | -0.5        | 04:25:00               | 210.6        | 211.1        |
| 03:51:00               | -0.2        | -0.5        | 04:26:00               | 210.6        | 211.2        |
| 03:52:00               | -0.2        | -0.7        | 04:27:00               | 210.8        | 211.2        |
| 03:53:00               | -0.2        | -0.5        | 04:28:00               | 210.8        | 211.4        |
| 03:54:00               | -0.2        | -0.5        | 04:29:00               | 210.8        | 211.4        |
| 03:55:00               | -0.2        | -0.7        | 04:30:00               | 210.9        | 211.6        |
| 03:56:00               | -0.2        | -0.7        | 04:31:00               | 210.9        | 211.6        |
| 03:57:00               | -0.2        | -0.7        | 04:32:00               | 210.9        | 211.6        |
| 03:58:00               | -0.2        | -0.5        | 04:33:00               | 210.9        | 211.4        |
| 03:59:00               | -0.2        | -0.5        | 04:34:00               | 211.1        | 211.6        |
| 04:00:00               | -0.2        | -0.7        | 04:35:00               | 211.1        | 211.6        |
| <b>average</b>         | <b>-0.2</b> | <b>-0.5</b> | <b>average</b>         | <b>210.6</b> | <b>211.2</b> |

## 7.1 8.4.5 Repeatability standard deviation

*The performance criteria are as follows: Repeatability standard deviation at zero shall not exceed 1.0 nmol/mol. At a sample gas concentration at the reference point it shall not exceed 3.0 nmol/mol.*

## 7.2 Test procedure

After waiting the time equivalent of one independent reading, 20 individual measurements both at zero concentration and at an NO test concentration ( $c_t$ ) of  $(500 \pm 50)$  nmol/mol shall be performed.

From these measurements, the repeatability standard deviation ( $s_r$ ) at zero concentration and at concentration  $c_t$  shall be calculated according to:

$$s_r = \sqrt{\frac{\sum (x_i - \bar{x})^2}{n - 1}}$$

Where:

- $s_r$  the repeatability standard deviation;
- $x_i$  the  $i$ th measurement;
- $\bar{x}$  is the average of the 20 measurements;
- $n$  is the number of measurements.

The repeatability standard deviation shall be calculated separately for both series of measurements (zero gas and concentration  $c_t$ ).

$s_r$  shall comply with the performance criterion indicated above, both at zero and at the test gas concentration  $c_t$  of  $(500 \pm 50)$  nmol/mol.

The detection limit, lower detection limit of the measuring system is calculated from the repeatability standard deviation and the slope of the calibration function determined in accordance with chapter 8.4.6 according to the following equation:

$$l_{\text{det}} = 3,3 \cdot \frac{s_{r,z}}{B}$$

Where:

$l_{\text{det}}$  is the detection limit, lower detection limit of the measuring system, in nmol/mol;

$s_{r,z}$  is the repeatability standard deviation at zero, in nmol/mol;

$B$  is the slope of the calibration function according to Annex A based on the data from 8.4.6.

### 7.3 Testing

The test was performed in line with the requirements of EN 14211 mentioned before. In compliance with the standard, the test has to be performed with the component NO. Pursuant to EN 14211, the test shall be performed at a concentration level of 500 nmol/mol NO.

The test was also carried out for the component NO<sub>2</sub> at the zero point and at a span concentration at the level of the 1-h limit value of NO<sub>2</sub> (104 nmol/mol).

### 7.4 Evaluation

In Table 27 and in Table 28 the results of the repeatability standard deviation test are presented.

Table 27: Repeatability standard deviation at zero and span point for NO

|   | requirement | device 1 |   | device 2 |   |
|---|-------------|----------|---|----------|---|
| repeatability standard deviation $s_{r,z}$ at zero [nmol/mol]   | $\leq 1.0$  | 0.13     | ✓ | 0.13     | ✓ |
| repeatability standard deviation $s_{r,ct}$ at $c_t$ [nmol/mol] | $\leq 3.0$  | 1.63     | ✓ | 1.34     | ✓ |
| detection limit [nmol/mol]                                      |             | 0.45     |   | 0.45     |   |

Table 28: Repeatability standard deviation at zero and span point for NO<sub>2</sub>

|   | requirement | device 1 |   | device 2 |   |
|---|-------------|----------|---|----------|---|
| repeatability standard deviation $s_{r,z}$ at zero [nmol/mol]   | $\leq 1.0$  | 0.00     | ✓ | 0.11     | ✓ |
| repeatability standard deviation $s_{r,ct}$ at $c_t$ [nmol/mol] | $\leq 3.0$  | 0.08     | ✓ | 0.08     | ✓ |
| detection limit [nmol/mol]                                      |             | 0.00     |   | 0.36     |   |

### 7.5 Assessment

For NO, the value for the repeatability standard deviation at zero point was 0.13 nmol/mol for instrument 1; for instrument 2 it was 0.13 nmol/mol. Repeatability standard deviation at span point was 1.63 nmol/mol for instrument 1 and 1.34 nmol/mol for instrument 2.

For NO<sub>2</sub>, the value for the repeatability standard deviation at zero point was 0.00 nmol/mol for instrument 1 and 0.11 nmol/mol for instrument 2. Repeatability standard deviation at span point was 0.08 nmol/mol for instrument 1 and 0.08 nmol/mol for instrument 2.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

In Table 29 and in Table 30 the results of the individual measurements are presented.

Table 29: Repeatability standard deviation for NO

| zero level |            |            |
|------------|------------|------------|
|            | device 1   | device 2   |
| time       | [nmol/mol] | [nmol/mol] |
| 17:17:00   | 0.0        | 0.0        |
| 17:18:00   | -0.6       | -0.6       |
| 17:19:00   | -0.6       | -0.6       |
| 17:20:00   | -0.6       | -0.6       |
| 17:21:00   | -0.6       | -0.6       |
| 17:22:00   | -0.6       | -0.6       |
| 17:23:00   | -0.6       | -0.6       |
| 17:24:00   | -0.6       | -0.6       |
| 17:25:00   | -0.6       | -0.6       |
| 17:26:00   | -0.6       | -0.6       |
| 17:27:00   | -0.6       | -0.6       |
| 17:28:00   | -0.6       | -0.6       |
| 17:29:00   | -0.6       | -0.6       |
| 17:30:00   | -0.6       | -0.6       |
| 17:31:00   | -0.6       | -0.6       |
| 17:32:00   | -0.6       | -0.6       |
| 17:33:00   | -0.6       | -0.6       |
| 17:34:00   | -0.6       | -0.6       |
| 17:35:00   | -0.6       | -0.6       |
| 17:36:00   | -0.6       | -0.6       |
| average    | -0.6       | -0.6       |

| c <sub>t</sub> level |            |            |
|----------------------|------------|------------|
|                      | device 1   | device 2   |
| time                 | [nmol/mol] | [nmol/mol] |
| 17:42:00             | 515.9      | 521.3      |
| 17:43:00             | 513.5      | 523.7      |
| 17:44:00             | 513.5      | 523.1      |
| 17:45:00             | 512.9      | 524.3      |
| 17:46:00             | 512.9      | 523.1      |
| 17:47:00             | 512.9      | 523.7      |
| 17:48:00             | 514.1      | 522.5      |
| 17:49:00             | 512.3      | 525.5      |
| 17:50:00             | 517.7      | 523.7      |
| 17:51:00             | 513.5      | 523.1      |
| 17:52:00             | 511.7      | 522.5      |
| 17:53:00             | 514.1      | 519.5      |
| 17:54:00             | 515.3      | 522.5      |
| 17:55:00             | 513.5      | 524.3      |
| 17:56:00             | 513.5      | 523.7      |
| 17:57:00             | 515.3      | 524.3      |
| 17:58:00             | 514.7      | 523.7      |
| 17:59:00             | 512.9      | 521.3      |
| 18:00:00             | 517.7      | 522.5      |
| 18:01:00             | 515.3      | 524.3      |
| average              | 514.1      | 523.1      |

Table 30: Repeatability standard deviation for NO<sub>2</sub>

| zero level |            |            |
|------------|------------|------------|
|            | device 1   | device 2   |
| time       | [nmol/mol] | [nmol/mol] |
| 10:02:00   | -0.2       | -1.0       |
| 10:03:00   | -0.2       | -1.0       |
| 10:04:00   | -0.2       | -1.0       |
| 10:05:00   | -0.2       | -1.0       |
| 10:06:00   | -0.2       | -1.1       |
| 10:07:00   | -0.2       | -1.1       |
| 10:08:00   | -0.2       | -1.1       |
| 10:09:00   | -0.2       | -1.1       |
| 10:10:00   | -0.2       | -1.1       |
| 10:11:00   | -0.2       | -1.1       |
| 10:12:00   | -0.2       | -1.1       |
| 10:13:00   | -0.2       | -1.1       |
| 10:14:00   | -0.2       | -1.1       |
| 10:15:00   | -0.2       | -1.1       |
| 10:16:00   | -0.2       | -1.1       |
| 10:17:00   | -0.2       | -1.3       |
| 10:18:00   | -0.2       | -1.3       |
| 10:19:00   | -0.2       | -1.3       |
| 10:20:00   | -0.2       | -1.3       |
| 10:21:00   | -0.2       | -1.3       |
| average    | -0.2       | -1.2       |

| c <sub>t</sub> level |            |            |
|----------------------|------------|------------|
|                      | device 1   | device 2   |
| time                 | [nmol/mol] | [nmol/mol] |
| 10:27:00             | 104.1      | 104.1      |
| 10:28:00             | 104.1      | 104.1      |
| 10:29:00             | 104.1      | 104.1      |
| 10:30:00             | 104.1      | 104.1      |
| 10:31:00             | 104.1      | 104.1      |
| 10:32:00             | 104.1      | 104.1      |
| 10:33:00             | 104.1      | 104.1      |
| 10:34:00             | 103.9      | 103.9      |
| 10:35:00             | 103.9      | 103.9      |
| 10:36:00             | 103.9      | 103.9      |
| 10:37:00             | 103.9      | 103.9      |
| 10:38:00             | 103.9      | 103.9      |
| 10:39:00             | 103.9      | 103.9      |
| 10:40:00             | 103.9      | 103.9      |
| 10:41:00             | 103.9      | 103.9      |
| 10:42:00             | 103.9      | 103.9      |
| 10:43:00             | 103.9      | 103.9      |
| 10:44:00             | 103.9      | 103.9      |
| 10:45:00             | 103.9      | 103.9      |
| 10:46:00             | 103.9      | 103.9      |
| average              | 104.0      | 104.0      |



## 7.1 8.4.6 Lack of fit of linearity of the calibration function

*The deviation from the linearity of the calibration function at zero shall not exceed 5 nmol/mol. At concentrations above zero, it shall not exceed 4% of the measured value.*

### 7.2 Test procedure

The lack of fit of linearity of the calibration function of the analyser shall be tested over the range of 0% to 95% of the maximum of the certification range of NO, using at least six concentrations (including the zero point). The analyser shall be adjusted at a concentration of about 90% of the maximum of the certification range. At each concentration (including zero) at least five individual measurements shall be performed.

The concentrations shall be applied in the following sequence: 80%, 40%, 0%, 60%, 20% and 95%. After each change in concentration, at least four response times shall be taken into account before the next measurement is performed.

The regression function and the deviations are calculated in accordance with Annex A of standard EN 14211. The deviations from the linear regression function shall comply with the performance criterion specified above.

Establishment of the regression line:

A linear regression function in the form of  $Y_i = A + B * X_i$  is made through calculation of the following formula:

$$Y_i = a + B(X_i - X_z)$$

For the regression calculation, all measuring points (including zero) are taken into account. The total number of measuring points is equal to the number of concentration levels (at least six including zero) times the number of repetitions (at least five) at a particular concentration level.

The coefficient a is obtained from:

$$a = \sum Y_i / n$$

Where:

- a is the average value of the Y-values;
- $Y_i$  is the individual Y-value;
- N is the number of measuring points.

The coefficient B is obtained from:

$$B = \left( \sum Y_i (X_i - X_z) \right) / \sum (X_i - X_z)^2$$

Where:

- $X_z$  is the average of the x-values  $\left( = \sum (X_i / n) \right)$
- $X_i$  is the individual x-value.

Is the individual x-value. The function  $Y_i = a + B (X_i - X_z)$  is converted to  $Y_i = A + B * X_i$  through the calculation of A:

$$A = a - B * X_z$$

The residuals of the averages of each calibration point (including the zero point) are calculated as follows.

The average of each calibration point (including the zero point) at one and the same concentration c is calculated according to:

$$(Y_a)_c = \sum (Y_i)_c / m$$

Where:

$(Y_a)_c$  is the average y-value at concentration level c;

$(Y_i)_c$  is the individual y-value at concentration level c;

M is the number of repetitions at one and the same concentration level c;

The residual of each average ( $r_c$ ) at each concentration level is calculated according to:

$$r_c = (Y_a)_c - (A + B \times c)$$

Each residual to a value relative to its own concentration level c is expressed in % as:

$$r_{c,rel} = \frac{r_c}{c} \times 100\%$$

### 7.3 Testing

The test was carried out for the component NO according to the previously mentioned test specifications of EN 14211.

In addition, the test for the component NO<sub>2</sub> was carried out in the certification range 0 - 261 nmol/mol.

### 7.4 Evaluation

The following linear regressions were established:

In Figure 11 to Figure 14 the results of the group average tests are shown in summary form for NO and NO<sub>2</sub>.

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Table 31: Deviation from the analytical function for NO

|   | requirements | device 1 |   | device 2 |   |
|---|--------------|----------|---|----------|---|
| largest value of the relative residuals $r_{max}$ [%] | $\leq 4.0$   | 1.72     | ✓ | 2.24     | ✓ |
| residual at zero $r_z$ [nmol/mol]                     | $\leq 5.0$   | -0.60    | ✓ | -0.60    | ✓ |

Table 32: Deviation from the analytical function for NO<sub>2</sub>

|   | requirements | device 1 |   | device 2 |   |
|---|--------------|----------|---|----------|---|
| largest value of the relative residuals $r_{max}$ [%] | $\leq 4.0$   | 0.85     | ✓ | 0.72     | ✓ |
| residual at zero $r_z$ [nmol/mol]                     | $\leq 5.0$   | -0.16    | ✓ | -0.49    | ✓ |




Figure 11: Analysis function from the group average values for system 1 for NO




Figure 12: Analysis function from the group average values for system 2 for NO




Figure 13: Analysis function from the group average values for system 1 for NO<sub>2</sub>




Figure 14: Analysis function from the group average values for system 2 for NO<sub>2</sub>

## 7.5 Assessment

### Component NO

The deviation from the linear regression line for instrument 1 is -0.60 nmol/mol at zero point and no more than 1.72% of the rated value for concentrations above zero. The deviation from the linear regression line for instrument 2 is -0.60 nmol/mol at zero point and no more than 2.24% of the rated value for concentrations above zero.

### Component NO<sub>2</sub>

The deviation from the linear regression line for instrument 1 is -0.16 nmol/mol at zero point and no more than 0.85% of the rated value for concentrations above zero. The deviation from the linear regression line for instrument 2 is -0.49 nmol/mol at zero point and no more than 0.72% of the rated value for concentrations above zero.

The residuals from the ideal regression line do not exceed the limit values required by standard EN 14211.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

The individual values of the test can be found in Table 33 and in Table 34.

**Table 33: Results of the lack-of-fit test for NO**

|             |           | device 1 [nmol/mol] |                 | device 2 [nmol/mol] |                 |
|-------------|-----------|---------------------|-----------------|---------------------|-----------------|
| time        | level [%] | actual value $y_i$  | set value $x_i$ | actual value $y_i$  | set value $x_i$ |
| 08:00:00    | 80        | 756.37              | 768.00          | 772.01              | 768.00          |
| 08:01:00    | 80        | 755.17              | 768.00          | 768.40              | 768.00          |
| 08:02:00    | 80        | 753.37              | 768.00          | 766.59              | 768.00          |
| 08:03:00    | 80        | 756.37              | 768.00          | 772.01              | 768.00          |
| 08:04:00    | 80        | 753.37              | 768.00          | 765.39              | 768.00          |
| average     |           | 754.93              |                 | 768.88              |                 |
| $r_{c,rel}$ |           | -0.76               |                 | -0.22               |                 |
| 08:08:00    | 40        | 386.00              | 384.00          | 397.43              | 384.00          |
| 08:09:00    | 40        | 387.21              | 384.00          | 397.43              | 384.00          |
| 08:10:00    | 40        | 389.01              | 384.00          | 396.83              | 384.00          |
| 08:11:00    | 40        | 389.61              | 384.00          | 397.43              | 384.00          |
| 08:12:00    | 40        | 389.61              | 384.00          | 397.43              | 384.00          |
| average     |           | 388.29              |                 | 397.31              |                 |
| $r_{c,rel}$ |           | 1.19                |                 | 1.97                |                 |
| 08:16:00    | 0         | -0.60               | 0.00            | -0.60               | 0.00            |
| 08:17:00    | 0         | -0.60               | 0.00            | -0.60               | 0.00            |
| 08:18:00    | 0         | -0.60               | 0.00            | -0.60               | 0.00            |
| 08:19:00    | 0         | -0.60               | 0.00            | -0.60               | 0.00            |
| 08:20:00    | 0         | -0.60               | 0.00            | -0.60               | 0.00            |
| average     |           | -0.60               |                 | -0.60               |                 |
| $r_z$       |           |                     |                 |                     |                 |
| 08:24:00    | 60        | 583.21              | 576.00          | 596.44              | 576.00          |
| 08:25:00    | 60        | 582.01              | 576.00          | 595.24              | 576.00          |
| 08:26:00    | 60        | 579.00              | 576.00          | 582.61              | 576.00          |
| 08:27:00    | 60        | 585.02              | 576.00          | 595.84              | 576.00          |
| 08:28:00    | 60        | 581.41              | 576.00          | 595.24              | 576.00          |
| average     |           | 582.13              |                 | 593.07              |                 |
| $r_{c,rel}$ |           | 1.72                |                 | 2.24                |                 |
| 08:32:00    | 20        | 198.41              | 192.00          | 200.82              | 192.00          |
| 08:33:00    | 20        | 197.81              | 192.00          | 200.22              | 192.00          |
| 08:34:00    | 20        | 197.21              | 192.00          | 200.22              | 192.00          |
| 08:35:00    | 20        | 197.21              | 192.00          | 199.62              | 192.00          |
| 08:36:00    | 20        | 197.21              | 192.00          | 199.62              | 192.00          |
| average     |           | 197.57              |                 | 200.10              |                 |
| $r_{c,rel}$ |           | 1.24                |                 | 0.41                |                 |
| 08:40:00    | 95        | 899.47              | 912.00          | 897.07              | 912.00          |
| 08:41:00    | 95        | 901.27              | 912.00          | 905.48              | 912.00          |
| 08:42:00    | 95        | 899.47              | 912.00          | 904.28              | 912.00          |
| 08:43:00    | 95        | 895.26              | 912.00          | 906.69              | 912.00          |
| 08:44:00    | 95        | 896.46              | 912.00          | 903.08              | 912.00          |
| average     |           | 898.39              |                 | 903.32              |                 |
| $r_{c,rel}$ |           | -0.41               |                 | -1.11               |                 |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Table 34: Results of the lack-of-fit test for NO<sub>2</sub>

|                    |           | device 1 [nmol/mol]         |                          | device 2 [nmol/mol]         |                          |
|--------------------|-----------|-----------------------------|--------------------------|-----------------------------|--------------------------|
| time               | level [%] | actual value y <sub>i</sub> | set value x <sub>i</sub> | actual value y <sub>i</sub> | set value x <sub>i</sub> |
| 09:02:00           | 80        | 210.27                      | 209.21                   | 211.41                      | 209.21                   |
| 09:03:00           | 80        | 210.27                      | 209.21                   | 211.41                      | 209.21                   |
| 09:04:00           | 80        | 210.43                      | 209.21                   | 211.41                      | 209.21                   |
| 09:05:00           | 80        | 210.59                      | 209.21                   | 211.57                      | 209.21                   |
| 09:06:00           | 80        | 210.59                      | 209.21                   | 211.74                      | 209.21                   |
| average            |           | 210.43                      |                          | 211.51                      |                          |
| r <sub>c,rel</sub> |           | -0.42                       |                          | -0.33                       |                          |
| 09:12:00           | 40        | 104.07                      | 104.60                   | 104.56                      | 104.60                   |
| 09:13:00           | 40        | 104.07                      | 104.60                   | 104.56                      | 104.60                   |
| 09:14:00           | 40        | 104.07                      | 104.60                   | 104.56                      | 104.60                   |
| 09:15:00           | 40        | 104.07                      | 104.60                   | 104.56                      | 104.60                   |
| 09:16:00           | 40        | 104.07                      | 104.60                   | 104.56                      | 104.60                   |
| average            |           | 104.07                      |                          | 104.56                      |                          |
| r <sub>c,rel</sub> |           | -0.85                       |                          | -0.72                       |                          |
| 09:22:00           | 0         | -0.16                       | 0.00                     | -0.49                       | 0.00                     |
| 09:23:00           | 0         | -0.16                       | 0.00                     | -0.49                       | 0.00                     |
| 09:24:00           | 0         | -0.16                       | 0.00                     | -0.49                       | 0.00                     |
| 09:25:00           | 0         | -0.16                       | 0.00                     | -0.49                       | 0.00                     |
| 09:26:00           | 0         | -0.16                       | 0.00                     | -0.49                       | 0.00                     |
| average            |           | -0.16                       |                          | -0.49                       |                          |
| r <sub>z</sub>     |           |                             |                          |                             |                          |
| 09:32:00           | 60        | 156.93                      | 156.90                   | 157.58                      | 156.90                   |
| 09:33:00           | 60        | 157.09                      | 156.90                   | 157.58                      | 156.90                   |
| 09:34:00           | 60        | 157.09                      | 156.90                   | 157.74                      | 156.90                   |
| 09:35:00           | 60        | 157.25                      | 156.90                   | 157.74                      | 156.90                   |
| 09:36:00           | 60        | 157.25                      | 156.90                   | 157.74                      | 156.90                   |
| average            |           | 157.12                      |                          | 157.68                      |                          |
| r <sub>c,rel</sub> |           | -0.65                       |                          | -0.69                       |                          |
| 09:42:00           | 20        | 51.55                       | 52.30                    | 51.71                       | 52.30                    |
| 09:43:00           | 20        | 51.55                       | 52.30                    | 51.71                       | 52.30                    |
| 09:44:00           | 20        | 51.55                       | 52.30                    | 51.71                       | 52.30                    |
| 09:45:00           | 20        | 51.55                       | 52.30                    | 51.71                       | 52.30                    |
| 09:46:00           | 20        | 51.55                       | 52.30                    | 51.71                       | 52.30                    |
| average            |           | 51.55                       |                          | 51.71                       |                          |
| r <sub>c,rel</sub> |           | -0.46                       |                          | -0.32                       |                          |
| 09:52:00           | 95        | 253.01                      | 248.43                   | 253.82                      | 248.43                   |
| 09:53:00           | 95        | 253.01                      | 248.43                   | 253.82                      | 248.43                   |
| 09:54:00           | 95        | 253.01                      | 248.43                   | 253.99                      | 248.43                   |
| 09:55:00           | 95        | 253.01                      | 248.43                   | 253.99                      | 248.43                   |
| 09:56:00           | 95        | 253.01                      | 248.43                   | 253.99                      | 248.43                   |
| average            |           | 253.01                      |                          | 253.92                      |                          |
| r <sub>c,rel</sub> |           | 0.73                        |                          | 0.66                        |                          |

## 7.1 8.4.7 Sensitivity coefficient to sample gas pressure

*The sensitivity coefficient to sample gas pressure shall be  $\leq 8.0$  nmol/mol/kPa.*

## 7.2 Test procedures

Measurements are taken at a concentration of about 70% to 80% of the maximum of the certification range of NO at an absolute pressure of about  $(80 \pm 0.2)$  kPa and at an absolute pressure of about  $(110 \pm 0.2)$  kPa. At each pressure after waiting the time equivalent to one independent reading, three individual measurements are recorded. From these measurements, the averages at each pressure are calculated.

Measurements at different pressures shall be separated by at least four response times.

The sensitivity coefficient to sample gas pressure is calculated as follows.

$$b_{sp} = \left| \frac{(C_{P2} - C_{P1})}{(P_2 - P_1)} \right|$$

Where:

$b_{sp}$  is the sample gas pressure sensitivity coefficient;

$C_{P1}$  is the average concentration of the measurements at sampling gas pressure  $P_1$ ;

$C_{P2}$  is the average concentration of the measurements at sampling gas pressure  $P_2$ ;

$P_1$  is the minimum sampling gas pressure  $P_1$ ;

$P_2$  is the maximum sampling gas pressure  $P_2$ .

$b_{sp}$  shall comply with the performance criterion indicated above.

In compliance with the standard, the test has to be performed with the component NO. Pursuant to EN 14211, the test shall be performed at a concentration level of 70% to 80% of the certification range for NO.

The test was also carried out for the NO<sub>2</sub> component at a concentration level of 70 % to 80 % of the certification range for NO<sub>2</sub>.

## 7.3 Testing

The test was performed in line with the requirements of EN 14211 mentioned before.

Negative pressure was produced by reducing the test gas volume fed by means of blocking the sample gas line. For the positive pressure test, the AMS was connected to a sample gas source. The test gas volume generated was set at a higher rate than the volume sucked in by the analyser. The excess supply was diverted via a tee. The positive pressure was produced by blocking the bypass line. The test gas pressure was determined with the help of a pressure sensor located in the sample gas path.

Individual measurements were performed at concentrations around 70% to 80% of the maximum certification range and sample gas pressures of 80 kPa and 110 kPa.



## 7.4 Evaluation

The following sensitivity coefficients to sample gas pressure were determined:

Table 35: Sensitivity coefficient to the sample gas pressure for NO

|  | requirement | device 1 |   | device 2 |   |
|--|-------------|----------|---|----------|---|
| sensitivity coeff. sample gas pressure $b_{gp}$ [nmol/mol/kPa] | $\leq 8.0$  | 0.23     | ✓ | 0.17     | ✓ |

Table 36: Sensitivity coefficient to the sample gas pressure for NO<sub>2</sub>

|  | requirement | device 1 |   | device 2 |   |
|--|-------------|----------|---|----------|---|
| sensitivity coeff. sample gas pressure $b_{gp}$ [nmol/mol/kPa] | $\leq 8.0$  | 0.10     | ✓ | 0.08     | ✓ |

## 7.5 Assessment

For NO, the sensitivity coefficient to sample gas pressure was 0.23 nmol/mol/kPa for instrument 1; for instrument 2 it was 0.17 nmol/mol/kPa.

For NO<sub>2</sub>, the sensitivity coefficient to sample gas pressure was 0.10 nmol/mol/kPa for instrument 1; for instrument 2 it was 0.08 nmol/mol/kPa.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

Table 37: Individual values of the sensitivity to the sample gas pressure for NO

| time             | pressure [kPa] | concentration | device 1   | device 2   |
|------------------|----------------|---------------|------------|------------|
|                  |                |               | [nmol/mol] | [nmol/mol] |
| 09:54:00         | 80             | 710.00        | 708.87     | 712.48     |
| 09:55:00         | 80             | 710.00        | 705.87     | 713.08     |
| 09:56:00         | 80             | 710.00        | 706.47     | 711.28     |
| average $C_{P1}$ |                |               | 707.07     | 712.28     |
| 10:10:00         | 110            | 710.00        | 714.29     | 716.69     |
| 10:11:00         | 110            | 710.00        | 713.68     | 717.89     |
| 10:12:00         | 110            | 710.00        | 713.68     | 717.29     |
| average $C_{P2}$ |                |               | 713.88     | 717.29     |

Table 38: Individual values of the sensitivity to the sample gas pressure for NO<sub>2</sub>

| time                    | pressure [kPa] | concentration | device 1   | device 2   |
|-------------------------|----------------|---------------|------------|------------|
|                         |                |               | [nmol/mol] | [nmol/mol] |
| 08:57:00                | 80             | 200.00        | 206.63     | 207.60     |
| 08:58:00                | 80             | 200.00        | 206.95     | 208.24     |
| 08:59:00                | 80             | 200.00        | 206.63     | 208.56     |
| average C <sub>P1</sub> |                |               | 206.74     | 208.13     |
| 09:23:00                | 110            | 200.00        | 209.52     | 210.49     |
| 09:24:00                | 110            | 200.00        | 209.52     | 210.49     |
| 09:25:00                | 110            | 200.00        | 210.17     | 210.81     |
| average C <sub>P2</sub> |                |               | 209.74     | 210.60     |

## 7.1 8.4.8 Sensitivity coefficient to sample gas temperature

*The sensitivity coefficient to sample gas temperature shall be  $\leq 3.0 \mu\text{mol/mol/K}$ .*

## 7.2 Test procedures

Measurements shall be performed at sample gas temperatures of  $T_{G,1} = 0 \text{ }^\circ\text{C}$  and  $T_{G,2} = 30 \text{ }^\circ\text{C}$ . The sensitivity coefficient to sample gas temperature is determined at a concentration of around 70% to 80% of the maximum certification range. Wait the time equivalent to one independent measurement and record three individual measurements at each temperature.

The sample gas temperature, measured at the inlet of the analyser, shall be held constant for at least 30 minutes.

The sensitivity coefficient to sample gas temperature is calculated as follows:

$$b_{gt} = \frac{(C_{GT,2} - C_{GT,1})}{(T_{G,2} - T_{G,1})}$$

Where:

$b_{gt}$  is the sample gas temperature sensitivity coefficient;

$C_{GT,1}$  is the average concentration of the measurements at sample gas temperature  $T_{G,1}$ ;

$C_{GT,2}$  is the average concentration of the measurements at sample gas temperature  $T_{G,2}$ ;

$T_{G,1}$  is the sample gas temperature  $T_{G,1}$ ;

$T_{G,2}$  is the sample gas temperature  $T_{G,2}$ ;

$b_{gt}$  shall comply with the performance criterion indicated above.

In compliance with the standard, the test has to be performed with the component NO. Pursuant to EN 14211, the test shall be performed at a concentration level of 70% to 80% of the certification range for NO.

The test was also carried out for the NO<sub>2</sub> component at a concentration level of 70 % to 80 % of the certification range for NO<sub>2</sub>.

## 7.3 Testing

The test was performed in line with the requirements of EN 14211 mentioned before.

For the purpose of this test, the test gas mixture was led through a 50 m tube-bundle which was situated in a climatic chamber. The measuring systems were installed directly upstream of the climatic chamber. The end of the tube-bundle was led out of the climatic chamber and connected to the measuring systems. The feed line outside of the climatic chamber was isolated; a thermometer was used to monitor the temperature of the test gas directly upstream of the measuring system. The temperature of the climatic chamber was adjusted so that the gas temperature directly upstream of the analysers was exactly 0 °C. For the purpose of testing a gas temperature of 30 °C, gas was led through a heated line instead of the tube bundle in the climatic chamber.

## 7.4 Evaluation

Table 39: Sample gas temperature sensitivity coefficient for NO

|   | requirement | device 1 |   | device 2 |   |
|---|-------------|----------|---|----------|---|
| sensitivity coeff. sample gas temperature $b_{gt}$ [nmol/mol/K] | $\leq 3.0$  | 0.09     | ✓ | 0.11     | ✓ |

Table 40: Sample gas temperature sensitivity coefficient for NO<sub>2</sub>

|   | requirement | device 1 |   | device 2 |   |
|---|-------------|----------|---|----------|---|
| sensitivity coeff. sample gas temperature $b_{gt}$ [nmol/mol/K] | $\leq 3.0$  | 0.01     | ✓ | 0.01     | ✓ |

## 7.5 Assessment

For instrument 1, the sensitivity coefficient to sample gas temperature was 0.09 nmol/mol/K for NO; for NO<sub>2</sub>, it was 0.11 nmol/mol/K.

For instrument 2, the sensitivity coefficient to sample gas temperature was 0.01 nmol/mol/K for NO; for NO<sub>2</sub>, it was 0.01 nmol/mol/K.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

Table 41: Individual values of the influence of the sample gas temperature for NO

|                    |           |               | device 1   | device 2   |
|--------------------|-----------|---------------|------------|------------|
| time               | temp [°C] | concentration | [nmol/mol] | [nmol/mol] |
| 10:12:00           | 0         | 710.00        | 708.27     | 708.27     |
| 10:13:00           | 0         | 710.00        | 708.87     | 711.28     |
| 10:14:00           | 0         | 710.00        | 708.27     | 710.08     |
| average $C_{GT,1}$ |           |               | 708.47     | 709.88     |
| 13:12:00           | 30        | 710.00        | 711.88     | 713.68     |
| 13:13:00           | 30        | 710.00        | 710.68     | 711.88     |
| 13:14:00           | 30        | 710.00        | 711.28     | 714.29     |
| average $C_{GT,2}$ |           |               | 711.28     | 713.28     |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Table 42: Individual values of the influence of the sample gas temperature for NO<sub>2</sub>

| time                      | temp [°C] | concentration | device 1   | device 2   |
|---------------------------|-----------|---------------|------------|------------|
|                           |           |               | [nmol/mol] | [nmol/mol] |
| 10:19:00                  | 0         | 200.00        | 203.42     | 203.58     |
| 10:20:00                  | 0         | 200.00        | 203.42     | 203.74     |
| 10:21:00                  | 0         | 200.00        | 203.58     | 204.07     |
| average C <sub>GT,1</sub> |           |               | 203.47     | 203.80     |
| 13:20:00                  | 30        | 200.00        | 203.74     | 204.23     |
| 13:21:00                  | 30        | 200.00        | 203.91     | 204.07     |
| 13:22:00                  | 30        | 200.00        | 203.74     | 204.23     |
| average C <sub>GT,2</sub> |           |               | 203.80     | 204.18     |

## 7.1 8.4.9 Sensitivity coefficient to surrounding temperature

*The sensitivity coefficient to surrounding temperature shall be  $\leq 3.0$   $\mu\text{mol/mol/K}$ .*

## 7.2 Test procedures

The sensitivity of the analyser readings to the surrounding temperature shall be determined by performing measurements at the following temperatures within the specifications of the manufacturer:

- 1) at the minimum temperature  $T_{\min} = 0$  °C;
- 2) at the temperature  $T_1 = 20$  °C;
- 3) at the maximum temperature  $T_{\max} = 30$  °C.

For these tests, a climate chamber is necessary.

In compliance with the standard, the test has to be performed with the component NO. Pursuant to EN 14211, the test shall be performed at a concentration level of 70% to 80% of the certification range for NO.

At each temperature setting after waiting the time equivalent to one independent measurement, three individual measurements at zero and at span shall be recorded.

The sequence of test temperatures is as follows:

$T_1, T_{\min}, T_1$  and  $T_1, T_{\max}, T_1$

At the first temperature ( $T_1$ ), the analyser shall be adjusted at zero and at span level (70% to 80% of the maximum of the certification range). Then three individual measurements are recorded after waiting the time equivalent to one independent reading at  $T_1$ , at  $T_{\min}$  and again at  $T_1$ . This measurement procedure shall be repeated at the temperature sequence of  $T_1, T_{\max}$  and at  $T_1$ .

In order to exclude any possible drift due to factors other than temperature, the measurements at  $T_1$  are averaged, which is taken into account in the following formula for calculation of the sensitivity coefficient for temperature dependence:

$$b_{st} = \left| \frac{x_T - \frac{x_1 + x_2}{2}}{T_S - T_{S,0}} \right|$$

Where:

$b_{st}$  is the surrounding temperature sensitivity coefficient;

$x_T$  is the average of the measurements at  $T_{\min}$  or  $T_{\max}$ ;

$x_1$  is the first average of the measurements at  $T_1$ ;

$x_2$  is the second average of the measurements at  $T_1$ ;

$T_S$  is the surrounding temperature in the laboratory;

$T_{S,0}$  is the average of the surrounding temperatures at set point.

For reporting the surrounding temperature dependence the higher value is taken of the two calculations of the temperature dependence at  $T_{S,1}$  and  $T_{S,2}$ .

$b_{st}$  shall comply with the performance criterion indicated above.

### 7.3 Testing

The test was performed in line with the requirements of EN 14211 mentioned before.  
The test was also carried out for the NO<sub>2</sub> component at a concentration level of 70 % to 80 % of the certification range for NO<sub>2</sub>.

### 7.4 Evaluation

The following sensitivity coefficients to surrounding temperature have been determined:

Table 43: Sensitivity coefficients to ambient temperature for NO

|  | requirements | device 1 |   | device 2 |   |
|--|--------------|----------|---|----------|---|
| sensitivity coefficient at 0 °C for zero level [nmol/mol/K]  | ≤ 3.0        | 0.000    | ✓ | 0.040    | ✓ |
| sensitivity coefficient at 30 °C for zero level [nmol/mol/K] | ≤ 3.0        | 0.010    | ✓ | 0.010    | ✓ |
| sensitivity coefficient at 0 °C for span level [nmol/mol/K]  | ≤ 3.0        | 0.030    | ✓ | 0.075    | ✓ |
| sensitivity coefficient at 30 °C for span level [nmol/mol/K] | ≤ 3.0        | 0.952    | ✓ | 0.451    | ✓ |

Table 44: Sensitivity coefficients to ambient temperature for NO<sub>2</sub>

|  | requirements | device 1 |   | device 2 |   |
|--|--------------|----------|---|----------|---|
| sensitivity coefficient at 0 °C for zero level [nmol/mol/K]  | ≤ 3.0        | 0.003    | ✓ | 0.029    | ✓ |
| sensitivity coefficient at 30 °C for zero level [nmol/mol/K] | ≤ 3.0        | 0.011    | ✓ | 0.016    | ✓ |
| sensitivity coefficient at 0 °C for span level [nmol/mol/K]  | ≤ 3.0        | 0.231    | ✓ | 0.241    | ✓ |
| sensitivity coefficient at 30 °C for span level [nmol/mol/K] | ≤ 3.0        | 0.019    | ✓ | 0.024    | ✓ |

As can be seen in Table 43 and in Table 44, the sensitivity coefficient of the ambient temperature at the zero and span point meets the performance requirements.

### 7.5 Assessment

The sensitivity coefficient to the surrounding temperature  $b_{st}$  did not exceed the performance criterion specified at 3.0 nmol/mol/K. For the purpose of uncertainty calculation, the largest value  $b_{st}$  is used for both instruments. For NO, this would be 0.952 nmol/mol/K for instrument 1 and 0.451 nmol/mol/K for instrument 2.

For NO<sub>2</sub>, this would be 0.231 nmol/mol/K for instrument 1 and 0.241 nmol/mol/K for instrument 2.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

The individual test results are presented in Table 45 and in Table 46.

Table 45: Results of the sensitivity coefficient to the ambient temperature for NO

| date                                  | zero level |           |                        |                        | span level |           |                        |                        |
|---------------------------------------|------------|-----------|------------------------|------------------------|------------|-----------|------------------------|------------------------|
|                                       | time       | temp [°C] | device 1<br>[nmol/mol] | device 2<br>[nmol/mol] | time       | temp [°C] | device 1<br>[nmol/mol] | device 2<br>[nmol/mol] |
| 03.11.2020                            | 08:55:00   | 20        | 0.0                    | 0.0                    | 09:23:00   | 20        | 707.7                  | 708.3                  |
| 03.11.2020                            | 08:56:00   | 20        | 0.0                    | 0.0                    | 09:24:00   | 20        | 708.9                  | 706.5                  |
| 03.11.2020                            | 08:57:00   | 20        | 0.0                    | 0.0                    | 09:25:00   | 20        | 704.1                  | 706.5                  |
| average ( $X_{1(TS1)}$ )              |            |           | 0.0                    | 0.0                    |            |           | 706.9                  | 707.1                  |
| 03.11.2020                            | 16:12:00   | 0         | 0.0                    | 0.6                    | 16:24:00   | 0         | 708.9                  | 708.9                  |
| 03.11.2020                            | 16:13:00   | 0         | 0.0                    | 1.2                    | 16:25:00   | 0         | 707.7                  | 708.3                  |
| 03.11.2020                            | 16:14:00   | 0         | 0.0                    | 0.6                    | 16:26:00   | 0         | 707.7                  | 705.9                  |
| average ( $X_{Ts,1}$ )                |            |           | 0                      | 0.8                    |            |           | 708.1                  | 707.7                  |
| 04.11.2020                            | 08:27:00   | 20        | 0.0                    | 0.0                    | 08:40:00   | 20        | 708.9                  | 711.9                  |
| 04.11.2020                            | 08:28:00   | 20        | 0.0                    | 0.0                    | 08:41:00   | 20        | 708.3                  | 708.3                  |
| 04.11.2020                            | 08:29:00   | 20        | 0.0                    | 0.0                    | 08:42:00   | 20        | 707.1                  | 713.7                  |
| average ( $X_{2(TS1)} = X_{1(TS2)}$ ) |            |           | 0.0                    | 0.0                    |            |           | 708.1                  | 711.3                  |
| 04.11.2020                            | 16:38:00   | 30        | 0.0                    | 0.0                    | 16:44:00   | 30        | 716.1                  | 714.3                  |
| 04.11.2020                            | 16:39:00   | 30        | 0.0                    | 0.0                    | 16:45:00   | 30        | 717.9                  | 714.3                  |
| 04.11.2020                            | 16:40:00   | 30        | 0.0                    | 0.0                    | 16:46:00   | 30        | 719.1                  | 717.3                  |
| average ( $X_{Ts,2}$ )                |            |           | 0.0                    | 0.0                    |            |           | 717.7                  | 715.3                  |
| 05.11.2020                            | 07:56:00   | 20        | 0.6                    | 0.6                    | 08:09:00   | 20        | 708.3                  | 714.3                  |
| 05.11.2020                            | 07:57:00   | 20        | 0.0                    | 0.0                    | 08:10:00   | 20        | 708.9                  | 707.7                  |
| 05.11.2020                            | 07:58:00   | 20        | 0.0                    | 0.0                    | 08:11:00   | 20        | 707.7                  | 708.9                  |
| average ( $X_{2(TS2)}$ )              |            |           | 0.2                    | 0.2                    |            |           | 708.3                  | 710.3                  |



Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Page 121 of 293

Table 46: Results of the sensitivity coefficient to the ambient temperature for NO<sub>2</sub>

| date  | zero level |           |                        |                        | span level |           |                        |                        |
|---|------------|-----------|------------------------|------------------------|------------|-----------|------------------------|------------------------|
|   | time       | temp [°C] | device 1<br>[nmol/mol] | device 2<br>[nmol/mol] | time       | temp [°C] | device 1<br>[nmol/mol] | device 2<br>[nmol/mol] |
| 03.11.2020  | 08:55:00   | 20        | 0.0                    | 0.2                    | 09:14:00   | 20        | 204.6                  | 204.4                  |
| 03.11.2020  | 08:56:00   | 20        | 0.0                    | 0.0                    | 09:15:00   | 20        | 205.5                  | 205.2                  |
| 03.11.2020  | 08:57:00   | 20        | 0.0                    | 0.0                    | 09:16:00   | 20        | 206.5                  | 206.2                  |
| average (X <sub>1(TS1)</sub> )                          |            |           | 0.0                    | 0.1                    |            |           | 205.5                  | 205.3                  |
| 03.11.2020  | 16:12:00   | 0         | 0.0                    | -0.5                   | 16:18:00   | 0         | 200.3                  | 200.0                  |
| 03.11.2020  | 16:13:00   | 0         | 0.0                    | -0.7                   | 16:19:00   | 0         | 199.2                  | 199.2                  |
| 03.11.2020  | 16:14:00   | 0         | 0.0                    | -0.7                   | 16:20:00   | 0         | 199.3                  | 200.0                  |
| average (X <sub>TS,1</sub> )                            |            |           | 0                      | -0.6                   |            |           | 199.6                  | 199.7                  |
| 04.11.2020  | 08:27:00   | 20        | 0.2                    | 0.0                    | 08:34:00   | 20        | 202.3                  | 203.1                  |
| 04.11.2020  | 08:28:00   | 20        | 0.2                    | -0.2                   | 08:35:00   | 20        | 202.9                  | 203.6                  |
| 04.11.2020  | 08:29:00   | 20        | 0.0                    | -0.2                   | 08:36:00   | 20        | 203.6                  | 204.7                  |
| average (X <sub>2(TS1)</sub> ) = (X <sub>1(TS2)</sub> ) |            |           | 0.1                    | -0.1                   |            |           | 202.9                  | 203.8                  |
| 04.11.2020  | 16:38:00   | 30        | 0.0                    | 0.0                    | 16:45:00   | 30        | 201.8                  | 202.8                  |
| 04.11.2020  | 16:39:00   | 30        | 0.0                    | 0.0                    | 16:46:00   | 30        | 203.1                  | 203.9                  |
| 04.11.2020  | 16:40:00   | 30        | 0.0                    | 0.0                    | 16:47:00   | 30        | 205.2                  | 205.9                  |
| average (X <sub>TS,2</sub> )                            |            |           | 0.0                    | 0.0                    |            |           | 203.4                  | 204.2                  |
| 05.11.2020  | 07:56:00   | 20        | 0.2                    | -0.2                   | 08:03:00   | 20        | 202.8                  | 203.3                  |
| 05.11.2020  | 07:57:00   | 20        | 0.2                    | -0.2                   | 08:04:00   | 20        | 203.1                  | 203.9                  |
| 05.11.2020  | 07:58:00   | 20        | 0.0                    | -0.3                   | 08:05:00   | 20        | 204.4                  | 205.0                  |
| average (X <sub>2(TS2)</sub> )                          |            |           | 0.1                    | -0.2                   |            |           | 203.4                  | 204.1                  |

## 7.1 8.4.10 Sensitivity coefficient to electrical voltage

*The sensitivity coefficient to electrical voltage shall not exceed 0.3 nmol/mol/V.*

### 7.2 Test procedures

The sensitivity coefficient of electrical voltage shall be determined at both ends of the voltage range specified by the manufacturer,  $V_1$  and  $V_2$ , at zero concentration and at a concentration around 70% to 80% of the maximum of the certification range of NO. After waiting the time equivalent to one independent measurement, three individual measurements at each voltage and concentration level shall be recorded.

The sensitivity coefficient to electrical voltage in accordance with EN 14211 is calculated as follows:

$$b_v = \left| \frac{(C_{V2} - C_{V1})}{(V_2 - V_1)} \right|$$

Where:

$b_v$  is the voltage sensitivity coefficient,

$C_{V1}$  is the average concentration reading of the measurements at voltage  $V_1$

$C_{V2}$  is the average concentration reading of the measurements at voltage  $V_2$

$V_1$  is the minimum voltage  $V_{\min}$

$V_2$  is the maximum voltage  $V_{\max}$

For reporting the dependence on voltage, the higher value of the result at zero and span level shall be taken.

$b_v$  shall comply with the performance criterion indicated above.

### 7.3 Testing

For the purpose of determining the sensitivity coefficient to electrical voltage, a transformer was looped into the measuring system's voltage supply. Test gases were applied to the zero and reference point at various voltages.

The test was performed in line with the requirements of EN 14211 mentioned before.

The test was also carried out for the NO<sub>2</sub> component at a concentration level of 70 % to 80 % of the certification range for NO<sub>2</sub>.

### 7.4 Evaluation

The following sensitivity coefficients to electrical voltage have been determined:

**Table 47: Sensitivity coefficient to electrical voltage for NO**

|  | requirement | device 1 |   | device 2 |   |
|--|-------------|----------|---|----------|---|
| sensitivity coeff. of voltage $b_v$ at zero level [nmol/mol/V] | $\leq 0.3$  | 0.01     | ✓ | 0.00     | ✓ |
| sensitivity coeff. of voltage $b_v$ at span level [nmol/mol/V] | $\leq 0.3$  | 0.00     | ✓ | 0.00     | ✓ |

**Table 48: Sensitivity coefficient to electrical voltage for NO<sub>2</sub>**

|  | requirement | device 1 |   | device 2 |   |
|--|-------------|----------|---|----------|---|
| sensitivity coeff. of voltage $b_v$ at zero level [nmol/mol/V] | $\leq 0.3$  | 0.01     | ✓ | 0.01     | ✓ |
| sensitivity coeff. of voltage $b_v$ at span level [nmol/mol/V] | $\leq 0.3$  | 0.02     | ✓ | 0.01     | ✓ |

## 7.5 Assessment

At no test item did the sensitivity coefficient to electrical voltage  $b_v$  exceed the value of 0.3 nmol/mol/V specified in standard EN 14211. For the purpose of uncertainty calculation, the largest  $b_v$  is used for both instruments. For NO, this is 0.01 nmol/mol/V for instrument 1 and 0.00 nmol/mol/V for instrument 2.

For NO<sub>2</sub>, this is 0.02 nmol/mol/V for instrument 1 and 0.01 nmol/mol/V for instrument 2.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

Table 49: Results of the sensitivity coefficient to electrical voltage for NO

| time                            | voltage [V] | concentration | device 1   | device 2   |
|---------------------------------|-------------|---------------|------------|------------|
|                                 |             |               | [nmol/mol] | [nmol/mol] |
| 11:37:00                        | 207         | 0             | -0.60      | 0.00       |
| 11:38:00                        | 207         | 0             | -0.60      | 0.00       |
| 11:39:00                        | 207         | 0             | 0.00       | 0.00       |
| average C <sub>V1</sub> at zero |             |               | -0.40      | 0.00       |
| 11:47:00                        | 253         | 0             | 0.00       | 0.00       |
| 11:48:00                        | 253         | 0             | 0.00       | 0.00       |
| 11:49:00                        | 253         | 0             | 0.00       | -0.60      |
| average C <sub>V2</sub> at zero |             |               | 0.00       | -0.20      |
| 12:07:00                        | 207         | 710.00        | 708.27     | 710.68     |
| 12:08:00                        | 207         | 710.00        | 707.67     | 710.08     |
| 12:09:00                        | 207         | 710.00        | 708.21     | 709.48     |
| average C <sub>V1</sub> at Span |             |               | 708.05     | 710.08     |
| 12:17:00                        | 253         | 710.00        | 708.27     | 710.68     |
| 12:18:00                        | 253         | 710.00        | 707.67     | 710.08     |
| 12:19:00                        | 253         | 710.00        | 708.87     | 710.08     |
| average C <sub>V2</sub> at Span |             |               | 708.27     | 710.28     |

Table 50: Results of the sensitivity coefficient to electrical voltage for NO<sub>2</sub>

| time                            | voltage [V] | concentration | device 1   | device 2   |
|---------------------------------|-------------|---------------|------------|------------|
|                                 |             |               | [nmol/mol] | [nmol/mol] |
| 11:37:00                        | 207         | 0             | 0.16       | 0.16       |
| 11:38:00                        | 207         | 0             | 0.16       | 0.16       |
| 11:39:00                        | 207         | 0             | 0.16       | 0.00       |
| average C <sub>V1</sub> at zero |             |               | 0.16       | 0.11       |
| 11:47:00                        | 253         | 0             | 0.00       | -0.16      |
| 11:48:00                        | 253         | 0             | -0.16      | -0.16      |
| 11:49:00                        | 253         | 0             | -0.16      | -0.16      |
| average C <sub>V2</sub> at zero |             |               | -0.11      | -0.16      |
| 11:50:00                        | 207         | 205.00        | 205.37     | 206.52     |
| 11:51:00                        | 207         | 205.00        | 205.37     | 206.84     |
| 11:52:00                        | 207         | 205.00        | 205.21     | 206.03     |
| average C <sub>V1</sub> at Span |             |               | 205.32     | 206.46     |
| 11:53:00                        | 253         | 205.00        | 206.19     | 207.01     |
| 11:54:00                        | 253         | 205.00        | 206.03     | 207.17     |
| 11:55:00                        | 253         | 205.00        | 206.03     | 207.17     |
| average C <sub>V2</sub> at Span |             |               | 206.08     | 207.11     |

## 7.1 8.4.11 Interferents

*Interferents at zero and at concentration  $c_t$  for NO ( $500 \pm 50$  nmol/mol). Deviations for interferents  $H_2O$ ,  $CO_2$  and  $NH_3$  shall not exceed 5.0 nmol/mol.*

## 7.2 Test procedures

The analyser response to certain interferents shall be tested. The interferents can give a positive or negative response. The test shall be performed at zero and at an NO test concentration ( $c_t$ ) of ( $500 \pm 50$ ) nmol/mol.

The concentration of the mixtures of the test gases with the interferent shall have an expanded uncertainty of  $\leq 5\%$  and shall be traceable to nationally accepted standards. The interferents to be tested and their respective concentrations are given in Table 51. The influence of each interferent shall be determined separately. The concentration of the measured variable shall be corrected for the dilution flux due to the addition of the interfering component (e.g. water vapour).

After setting the system at zero and at span level, a mixture of zero gas and the interfering component to be tested is applied at the concentration given in Table 51. With this mixture, one independent measurement of NO followed by two individual measurements of NO shall be carried out. This procedure shall be repeated with a mixture of the measured variable at the concentration  $c_t$  and the interfering component to be investigated. The influence quantities at zero and concentration  $c_t$  are calculated from:

$$X_{\text{int},z} = x_z$$

$$X_{\text{int},c_t} = x_{c_t} - c_t$$

Where:

$X_{\text{int},z}$  is the influence quantity of the interferent at zero;

$x_z$  is the average of the measurements of NO at zero;

$X_{\text{int},c_t}$  is the influence quantity of the interferent at concentration  $c_t$ ;

$x_{c_t}$  is the average of the measurements of NO at concentration  $c_t$

$c_t$  is the applied concentration at the one-hour limit value.

The influence quantities of the interferents shall comply with the performance criteria indicated above, both at zero and at concentration  $c_t$ .

## 7.3 Testing

The test was performed in line with the requirements of EN 14211 mentioned before. The instruments are adjusted at zero and at the concentration  $c_t$  (500 ppb). Zero and test gas with the various interfering components were then applied. The interferents listed in Table 51 were applied in the concentrations indicated. As required by standard EN 14211, the measured NO<sub>x</sub> concentration shall be used instead of the NO concentration when testing the interferent  $NH_3$ .

The test was also carried out for the component NO<sub>2</sub> at the zero point and at a span concentration at the level of the 1-h limit value of NO<sub>2</sub> (104 nmol/mol).

Since the measuring system does not operate according to the specified EU reference procedure, the influence of other possible interfering components in normal ambient air was checked in addition to the 3 interfering components specified in EN 14211. The results of the additional test are presented in chapter 6.1 7.4.11 Cross-sensitivity.

Table 51: Interferents in accordance with EN 14211

| Interferent      | Value        |
|------------------|--------------|
| H <sub>2</sub> O | 19 mmol/mol  |
| CO <sub>2</sub>  | 500 µmol/mol |
| NH <sub>3</sub>  | 200 nmol/mol |

## 7.4 Evaluation

The following overview presents the influence quantities of each interfering substance. When determining the influence of moisture, the dilution effect which occurs inside the test gas generation system was also taken into account.

Table 52: Influence of the tested interferents for NO (c<sub>t</sub> = 500±50 nmol/mol)

|  | requirements   | device 1 |   | device 2 |   |
|--|----------------|----------|---|----------|---|
| influence quantity interferent H <sub>2</sub> O at zero [nmol/mol/V]           | ≤ 5.0 nmol/mol | 0.00     | ✓ | -0.60    | ✓ |
| influence quantity interferent H <sub>2</sub> O at c <sub>t</sub> [nmol/mol/V] | ≤ 5.0 nmol/mol | 0.40     | ✓ | 1.60     | ✓ |
| influence quantity interferent CO <sub>2</sub> at zero [nmol/mol/V]            | ≤ 5.0 nmol/mol | 0.40     | ✓ | 0.40     | ✓ |
| influence quantity interferent CO <sub>2</sub> at c <sub>t</sub> [nmol/mol/V]  | ≤ 5.0 nmol/mol | 1.80     | ✓ | 0.80     | ✓ |
| influence quantity interferent NH <sub>3</sub> at zero [nmol/mol/V]            | ≤ 5.0 nmol/mol | 0.60     | ✓ | 1.00     | ✓ |
| influence quantity interferent NH <sub>3</sub> at c <sub>t</sub> [nmol/mol/V]  | ≤ 5.0 nmol/mol | 1.20     | ✓ | 1.40     | ✓ |

Table 53: Influence of the tested interferents for NO<sub>2</sub> (c<sub>t</sub> = ca. 104 nmol/mol)

|  | requirements   | device 1 |   | device 2 |   |
|--|----------------|----------|---|----------|---|
| influence quantity interferent H <sub>2</sub> O at zero [nmol/mol/V]           | ≤ 5.0 nmol/mol | 0.27     | ✓ | 0.00     | ✓ |
| influence quantity interferent H <sub>2</sub> O at c <sub>t</sub> [nmol/mol/V] | ≤ 5.0 nmol/mol | 0.33     | ✓ | 0.22     | ✓ |
| influence quantity interferent CO <sub>2</sub> at zero [nmol/mol/V]            | ≤ 5.0 nmol/mol | 0.00     | ✓ | -0.21    | ✓ |
| influence quantity interferent CO <sub>2</sub> at c <sub>t</sub> [nmol/mol/V]  | ≤ 5.0 nmol/mol | 0.70     | ✓ | 0.76     | ✓ |
| influence quantity interferent NH <sub>3</sub> at zero [nmol/mol/V]            | ≤ 5.0 nmol/mol | 0.00     | ✓ | -0.16    | ✓ |
| influence quantity interferent NH <sub>3</sub> at c <sub>t</sub> [nmol/mol/V]  | ≤ 5.0 nmol/mol | 1.09     | ✓ | 0.98     | ✓ |

## 7.5 Assessment

For NO at zero point the result for the interference were 0.00 nmol/mol (AMS 1) and -0.60 nmol/mol (AMS 2) for H<sub>2</sub>O, 0.40 nmol/mol (AMS 1) and 0.40 nmol/mol (AMS 2) for CO<sub>2</sub> and finally 0.60 nmol/mol (AMS 1) and 1.00 nmol/mol (AMS 2) for NH<sub>3</sub>. For NO<sub>2</sub> at zero point the result for the interference were 0.27 nmol/mol (AMS 1) and 0.00 nmol/mol (AMS 2) for H<sub>2</sub>O, 0.00 nmol/mol (AMS 1) and -0.21 nmol/mol (AMS 2) for CO<sub>2</sub> and finally 0.00 nmol/mol (AMS 1) and -0.16 nmol/mol (AMS 2) for NH<sub>3</sub>.

The following results were obtained for the cross-sensitivity at the limit value  $c_t$ : NO: 0.40 nmol/mol for instrument 1 and 1.60 nmol/mol for instrument 2; H<sub>2</sub>O: 1.80 nmol/mol for instrument 1 and 0.80 nmol/mol for instrument 2; CO<sub>2</sub> and 1.20 nmol/mol for instrument 1 and 1.40 nmol/mol at NH<sub>3</sub>. The following results were obtained for the cross-sensitivity at the limit value  $c_t$ : NO<sub>2</sub>: 0.33 nmol/mol for instrument 1 and 0.22 nmol/mol for instrument 2; H<sub>2</sub>O: 0.70 nmol/mol for instrument 1 and 0.76 nmol/mol for instrument 2; CO<sub>2</sub> and 1.09 nmol/mol for instrument 1 and 0.98 nmol/mol at NH<sub>3</sub>.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

Table 54 and Table 55 present the individual test results.

Table 54: Individual results for testing interferences for NO

|  | without interferences |          |          | with interferences |          |          |
|--|-----------------------|----------|----------|--------------------|----------|----------|
|  | time                  | device 1 | device 2 | time               | device 1 | device 2 |
| zero gas + H <sub>2</sub> O<br>(19 mmol/mol)       | 13:14:00              | 0.00     | 0.00     | 13:29:00           | 0.00     | -0.60    |
|  | 13:15:00              | 0.00     | 0.00     | 13:30:00           | 0.00     | -0.60    |
|  | 13:16:00              | 0.00     | 0.00     | 13:31:00           | 0.00     | -0.60    |
|  | average $x_z$         | 0.00     | 0.00     | average $x_z$      | 0.00     | -0.60    |
| test gas $c_t$ + H <sub>2</sub> O<br>(19 mmol/mol) | 13:45:00              | 504.45   | 508.06   | 14:00:00           | 507.46   | 508.66   |
|  | 13:46:00              | 506.85   | 507.46   | 14:01:00           | 508.66   | 506.85   |
|  | 13:47:00              | 508.06   | 504.45   | 14:02:00           | 504.45   | 509.26   |
|  | average $x_{ct}$      | 506.45   | 506.65   | average $x_{ct}$   | 506.85   | 508.26   |
| zero gas + CO <sub>2</sub><br>(500 µmol/mol)       | 08:14:00              | 0.60     | 0.00     | 08:24:00           | 0.60     | 0.00     |
|  | 08:15:00              | 0.60     | 0.00     | 08:25:00           | 0.60     | 0.00     |
|  | 08:16:00              | 0.00     | 0.00     | 08:26:00           | 1.20     | 1.20     |
|  | average $x_z$         | 0.40     | 0.00     | average $x_z$      | 0.80     | 0.40     |
| test gas $c_t$ + CO <sub>2</sub><br>(500 µmol/mol) | 08:34:00              | 500.24   | 501.44   | 08:44:00           | 502.04   | 502.65   |
|  | 08:35:00              | 500.24   | 502.04   | 08:45:00           | 501.44   | 502.04   |
|  | 08:36:00              | 499.64   | 500.84   | 08:46:00           | 502.04   | 502.04   |
|  | average $x_{ct}$      | 500.04   | 501.44   | average $x_{ct}$   | 501.84   | 502.24   |
| zero gas + NH <sub>3</sub><br>(200 nmol/mol)       | 09:45:00              | 0.60     | 0.00     | 09:55:00           | 1.20     | 1.20     |
|  | 09:46:00              | 0.60     | 0.00     | 09:56:00           | 1.20     | 1.20     |
|  | 09:47:00              | 0.60     | 0.60     | 09:57:00           | 1.20     | 1.20     |
|  | average $x_z$         | 0.60     | 0.20     | average $x_z$      | 1.20     | 1.20     |
| test gas $c_t$ + NH <sub>3</sub><br>(200 nmol/mol) | 10:15:00              | 503.25   | 506.25   | 10:25:00           | 505.05   | 506.85   |
|  | 10:16:00              | 503.25   | 505.65   | 10:26:00           | 504.45   | 507.46   |
|  | 10:17:00              | 503.85   | 506.25   | 10:27:00           | 504.45   | 508.06   |
|  | average $x_{ct}$      | 504.65   | 506.05   | average $x_{ct}$   | 504.65   | 507.46   |

NO<sub>x</sub> reading with interference from NH<sub>3</sub>



Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Table 55: Individual results for testing interferences for NO<sub>2</sub>

|  | without interferences |          |          | with interferences |          |          |
|--|-----------------------|----------|----------|--------------------|----------|----------|
|  | time                  | device 1 | device 2 | time               | device 1 | device 2 |
| zero gas + H <sub>2</sub> O<br>(19 mmol/mol)       | 13:14:00              | 0.16     | 0.33     | 13:29:00           | 0.33     | 0.33     |
|  | 13:15:00              | 0.00     | 0.33     | 13:30:00           | 0.33     | 0.33     |
|  | 13:16:00              | 0.00     | 0.33     | 13:31:00           | 0.33     | 0.33     |
|  | average $x_z$         | 0.05     | 0.33     | average $x_z$      | 0.33     | 0.33     |
| test gas $c_t$ + H <sub>2</sub> O<br>(19 mmol/mol) | 14:35:00              | 106.03   | 106.52   | 14:51:00           | 106.36   | 106.68   |
|  | 14:36:00              | 106.19   | 106.52   | 14:52:00           | 106.52   | 106.68   |
|  | 14:37:00              | 106.36   | 106.68   | 14:53:00           | 106.68   | 107.01   |
|  | average $x_{ct}$      | 106.19   | 106.58   | average $x_{ct}$   | 106.52   | 106.79   |
| zero gas + CO <sub>2</sub><br>(500 µmol/mol)       | 11:24:00              | -0.16    | -0.33    | 11:40:00           | -0.16    | -0.65    |
|  | 11:25:00              | -0.16    | -0.49    | 11:41:00           | -0.16    | -0.65    |
|  | 11:26:00              | -0.16    | -0.49    | 11:42:00           | -0.16    | -0.65    |
|  | average $x_z$         | -0.16    | -0.44    | average $x_z$      | -0.16    | -0.65    |
| test gas $c_t$ + CO <sub>2</sub><br>(500 µmol/mol) | 11:54:00              | 105.71   | 105.38   | 12:09:00           | 106.36   | 106.19   |
|  | 11:55:00              | 105.71   | 105.71   | 12:10:00           | 106.52   | 106.36   |
|  | 11:56:00              | 105.87   | 105.71   | 12:11:00           | 106.52   | 106.52   |
|  | average $x_{ct}$      | 105.76   | 105.60   | average $x_{ct}$   | 106.47   | 106.36   |
| zero gas + NH <sub>3</sub><br>(200 nmol/mol)       | 12:25:00              | -0.16    | -0.82    | 12:40:00           | -0.16    | -0.98    |
|  | 12:26:00              | -0.16    | -0.82    | 12:41:00           | -0.16    | -0.98    |
|  | 12:27:00              | -0.16    | -0.82    | 12:42:00           | -0.16    | -0.98    |
|  | average $x_z$         | -0.16    | -0.82    | average $x_z$      | -0.16    | -0.98    |
| test gas $c_t$ + NH <sub>3</sub><br>(200 nmol/mol) | 12:55:00              | 106.85   | 106.68   | 13:10:00           | 107.99   | 107.83   |
|  | 12:56:00              | 107.01   | 107.01   | 13:11:00           | 108.15   | 107.83   |
|  | 12:57:00              | 107.17   | 107.01   | 13:12:00           | 108.15   | 107.99   |
|  | average $x_{ct}$      | 108.10   | 106.90   | average $x_{ct}$   | 108.10   | 107.88   |

NO<sub>x</sub> reading with interference from NH<sub>3</sub>

## 7.1 8.4.12 Averaging test

*The averaging effect shall not exceed 7% of the measured value.*

### 7.2 Test conditions

The averaging test gives a measure of the uncertainty in the averaged values caused by short-term concentration variations in the sampled air shorter than the time scale of the measurement process in the analyser. In general, the output of an analyser is a result of the determination of a reference concentration (normally zero) and the actual concentration which takes a certain time.

For the determination of the uncertainty due to the averaging, the following concentrations are applied to the analyser and readings are taken at each concentration:

- a constant concentration of NO<sub>2</sub> at a concentration  $c_{t,NO_2}$  which is about twice the hourly limit value; and
- a stepwise varied concentration of NO between zero and 600 nmol/mol (concentration  $c_{t,NO}$ ).

The time period ( $t_c$ ) of the constant NO concentration shall be at least equal to a period necessary to obtain four independent readings (which is equal to at least sixteen response times). The time period ( $t_v$ ) of the varying NO concentration shall be at least equal to a period to obtain four independent readings. The time period ( $t_{NO}$ ) for the NO concentration shall be 45 s followed by a period ( $t_{zero}$ ) of 45 s of zero concentration. Further:

$c_t$  is the test concentration;

$t_v$  is a time period including a whole number of  $t_{NO}$  and  $t_{zero}$  pairs, and contains a minimum of 3 such pairs.

The change from  $t_{NO}$  to  $t_{zero}$  shall be within 0.5 s. The change from  $t_c$  to  $t_v$  shall be within one response time of the analyser under test.

The averaging effect ( $E_{av}$ ) is calculated according to:

$$E_{av} = \frac{C_{const}^{av} - 2C_{var}^{av}}{C_{const}^{av}} * 100$$

Where:

$E_{av}$  is the averaging effect (%);

$C_{const}^{av}$  is the average of the at least four independent measurements during the variable concentration period;

$C_{var}^{av}$  is the average of the at least four independent measurements during the variable concentration period;




Figure 15: Test of the averaging effect ( $t_{NO} = t_{zero} = 45$  s.)

### 7.3 Testing

The averaging test was performed in compliance with the requirements specified in EN 14211. With the help of a mass flow controller we applied a step change of the NO concentration between zero and 600 nmol/mol and, at the same time, a constant NO<sub>2</sub> concentration  $c_{t,NO_2}$  of roughly twice the hourly limit value. First, the average was calculated at a constant test gas concentration. Then, a three-way valve served to switch between zero and test gas every 45 s. During that period of alternating test gas application the average was calculated again.

The test was also carried out for the component NO<sub>2</sub> between zero and a span concentration at the level of the 1-h limit value of NO<sub>2</sub> (104 nmol/mol).

### 7.4 Evaluation

The following averages were determined during the test:

Table 56: Results of the averaging test for NO

|                               | requirement | device 1 | device 2 |
|-------------------------------|-------------|----------|----------|
| averaging effect $E_{av}$ [%] | $\leq 7\%$  | 3.1 ✓    | -2.9 ✓   |

This results in the following averaging effects:

System 1: 3.1%

System 2: -2.9 %

Table 57: Results of the averaging test for NO<sub>2</sub>

|                               | requirement | device 1 | device 2 |
|-------------------------------|-------------|----------|----------|
| averaging effect $E_{av}$ [%] | $\leq 7\%$  | 0.7 ✓    | 3.2 ✓    |

This results in the following averaging effects:

System 1: 0.7%

System 2: 3.2%

## 7.5 Assessment

The performance criterion specified by standard EN 14211 is fully satisfied.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

In Table 58 and in Table 59 the individual results of the averaging tests are presented.

Table 58: Results of the averaging tests for NO

|  |                  | device 1   | device 2   |
|--|------------------|------------|------------|
|  | time             | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 11:34:00         | 602.0      | 615.5      |
|  | till<br>11:53:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 11:54:00         | 298.9      | 311.3      |
|  | till<br>12:13:00 |            |            |

|  |                  | device 1   | device 2   |
|--|------------------|------------|------------|
|  | time             | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 12:24:00         | 605.8      | 617.6      |
|  | till<br>12:43:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 12:44:00         | 293.2      | 314.5      |
|  | till<br>13:03:00 |            |            |

|  |                  | device 1   | device 2   |
|--|------------------|------------|------------|
|  | time             | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 13:15:00         | 607.1      | 617.8      |
|  | till<br>13:34:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 13:35:00         | 287.0      | 326.0      |
|  | till<br>13:54:00 |            |            |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO2 and NOx,  
Report No.: 936/21251100/A

Table 59: Results of the averaging tests for NO<sub>2</sub>

|  |          | device 1   | device 2   |
|--|----------|------------|------------|
|  | time     | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 07:48:00 | 101.9      | 102.0      |
|  | till     |            |            |
|  | 08:07    |            |            |
| average variable concentration<br>$C_{av,c}$ | 08:08:00 | 53.0       | 49.7       |
|  | till     |            |            |
|  | 08:27:00 |            |            |

|  |          | device 1   | device 2   |
|--|----------|------------|------------|
|  | time     | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 08:39:00 | 102.8      | 102.9      |
|  | till     |            |            |
|  | 08:58:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 08:59:00 | 50.9       | 49.2       |
|  | till     |            |            |
|  | 09:18:00 |            |            |

|  |          | device 1   | device 2   |
|--|----------|------------|------------|
|  | time     | [nmol/mol] | [nmol/mol] |
| average constant concentration<br>$C_{av,c}$ | 09:25:00 | 103.3      | 103.4      |
|  | till     |            |            |
|  | 09:44:00 |            |            |
| average variable concentration<br>$C_{av,c}$ | 09:45:00 | 49.1       | 50.3       |
|  | till     |            |            |
|  | 10:04:00 |            |            |

## 7.1 8.4.13 Difference sample/calibration port

*The difference between sample and calibration port shall not exceed 1.0%.*

### 7.2 Test procedures

If the analyser has different ports for feeding sample gas and calibration gas, the difference in response of the analyser to feeding through the sample or calibration port shall be tested. The test shall be carried out by feeding the analyser with a test gas with a concentration of 70% to 80% of the maximum of the certification range of NO through the sample port. The test shall consist of one independent measurement followed by two individual measurements. After a period of at least four response times, the test shall be repeated using the calibration port. The difference shall be calculated according to:

$$\Delta x_{SC} = \frac{x_{sam} - x_{cal}}{c_t} \times 100$$

Where:

$\Delta x_{SC}$  is the difference sample/calibration port;

$x_{sam}$  is the average of the measured concentration using the sample port;

$x_{cal}$  is the average of the measured concentration using the calibration port;

$c_t$  is the concentration of the test gas;

$\Delta_{SC}$  shall comply with the performance criterion indicated above.

### 7.3 Testing

The test was performed in compliance with the requirements specified in EN 14211. During the test, the gas path was switched between sample gas and span gas inlet using a three-way valve.

### 7.4 Evaluation

During the test, the following differences between sample and calibration port were determined:

Table 60: Results of the difference between sample/calibration inlet for NO

|  | requirement | device 1 |   | device 2 |   |
|--|-------------|----------|---|----------|---|
| difference sample/calibration port $\Delta x_{cs}$ [%] | $\leq 1\%$  | 0.14     | ✓ | 0.06     | ✓ |

Table 61: Results of the difference between sample/calibration inlet for NO<sub>2</sub>

|  | requirement | device 1 |   | device 2 |   |
|--|-------------|----------|---|----------|---|
| difference sample/calibration port $\Delta x_{cs}$ [%] | $\leq 1\%$  | -0.14    | ✓ | -0.33    | ✓ |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

## 7.5 Assessment

The performance criterion specified by standard EN 14211 is fully satisfied.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

Not applicable

Table 62: Results for testing the difference between sample and calibration port for NO

|                  |          | device 1   | device 2   |
|------------------|----------|------------|------------|
|                  | time     | [nmol/mol] | [nmol/mol] |
| calibration port | 10:17:00 | 711.3      | 714.9      |
|                  | 10:18:00 | 710.7      | 714.3      |
|                  | 10:19:00 | 711.3      | 713.7      |
| sample port      | 10:31:00 | 710.1      | 714.3      |
|                  | 10:32:00 | 710.1      | 714.3      |
|                  | 10:33:00 | 710.1      | 713.1      |

Table 63: Results for testing the difference between sample and calibration port for NO<sub>2</sub>

|                  |          | device 1   | device 2   |
|------------------|----------|------------|------------|
|                  | time     | [nmol/mol] | [nmol/mol] |
| calibration port | 10:47:00 | 207.0      | 208.5      |
|                  | 10:48:00 | 206.7      | 207.8      |
|                  | 10:49:00 | 206.2      | 207.7      |
| sample port      | 10:57:00 | 207.0      | 208.6      |
|                  | 10:58:00 | 206.7      | 208.6      |
|                  | 10:59:00 | 207.0      | 208.6      |

## 7.1 8.4.14 Converter efficiency

*The converter efficiency should be at least 98%.*

### 7.2 Test procedures

The converter efficiency is determined by measurements with calculated amounts of NO<sub>2</sub>. This can be achieved by means of gas-phase titration of NO to NO<sub>2</sub> with ozone.

The test shall be performed at two concentration levels: at about 50% and about 95% of the maximum of the certification range of NO<sub>2</sub>.

The NO<sub>x</sub> measuring system shall be calibrated applying an NO concentration of about 70% to 80% of the maximum certification range for NO to the NO and NO<sub>x</sub> channels. Both channels need to be adjusted to display the same value. The values shall be recorded.

A known NO concentration at 50% of the NO certification range is applied to the measuring system until a stable signal is obtained. This stable period shall be at least as long as four response times. Four individual readings each are performed at the NO and the NO<sub>x</sub> channel. Then, NO is brought to react with O<sub>3</sub> to produce NO<sub>2</sub>. This mixture containing a constant NO<sub>x</sub> concentration is applied to the measuring system until the output signal has stabilised. This stable period shall be at least four response times of the measuring system under test; the NO concentration after gas-phase titration shall be between 10% and 20% of the original NO concentration. Then, four individual readings each are performed at the NO and the NO<sub>x</sub> channel. Then, the O<sub>3</sub> supply is cut and only NO is applied to the measuring system until the output signal has stabilised again. This stable period shall be at least as long as four response times of the measuring system. After that the average of the four individual measurements at the NO and the NO<sub>x</sub> channel is calculated.

The converter efficiency is calculated as follows:

$$E_{conv} = \left( 1 - \frac{(NO_x)_i - (NO_x)_f}{(NO)_i - (NO)_f} \right) \times 100\%$$

Where:

$E_{conv}$  is the converter efficiency in %;

$(NO_x)_i$  is the average of the four individual measurements at the NO<sub>x</sub> channel at the initial NO<sub>x</sub> concentration;

$(NO_x)_f$  is the average of the four individual measurements at the NO<sub>x</sub> channel at the resulting NO<sub>x</sub> concentration after applying O<sub>3</sub>;

$(NO)_i$  is the average of the four individual measurements at the NO channel at the initial NO concentration;

$(NO)_f$  is the average of the four individual measurements at the NO channel at the resulting NO concentration after applying O<sub>3</sub>;

The lowest value of the two converter efficiencies shall be reported.



Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

### 7.3 Testing

The N500 measuring system does not work according to the EU standard reference method of chemiluminescence. Therefore, the measuring system does not have a standard NO-NO<sub>2</sub> converter. The N500 measuring system analyses NO<sub>2</sub> directly in the UV range. NO is determined by oxidising the sample air with ozone in alternating cycles. The difference between the two measurements is displayed as NO value. Even if there is no converter in the measuring system to be tested, the test was carried out as indicated above to show that the NO and NO<sub>2</sub> measurements are equivalent to the standard reference method.

The test was performed in compliance with the requirements specified in EN 14211. When applying test gas two NO<sub>2</sub> concentrations in the range of 50% to 95% of the certification range for NO<sub>2</sub> were adjusted by means of gas-phase titration.

The converter efficiency was determined in the laboratory.

### 7.4 Evaluation

During testing, the following converter efficiencies were determined for the two N500 measuring systems. The lowest value of the two NO<sub>2</sub> concentration is reported below.

|                                | requirement | device 1 |   | device 2 |   |
|--------------------------------|-------------|----------|---|----------|---|
| converter efficiency $E_c$ [%] | $\geq 98\%$ | 99.6     | ✓ | 99.8     | ✓ |

### 7.5 Assessment

At a converter efficiency of 99.6%, the performance criterion specified by EN 14211 is fully satisfied.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

Table 64 presents the individual values.

Table 64: Individual results for the converter efficiency

|                                 |          | device 1                  |                            |               |                            | device 2      |                            |
|---------------------------------|----------|---------------------------|----------------------------|---------------|----------------------------|---------------|----------------------------|
|                                 | time     | O <sub>3</sub> [nmol/mol] | NO <sub>2</sub> [nmol/mol] | NO [nmol/mol] | NO <sub>x</sub> [nmol/mol] | NO [nmol/mol] | NO <sub>x</sub> [nmol/mol] |
|                                 | 12:30:00 | s t a r t                 |                            |               |                            |               |                            |
| O <sub>3</sub> =0, NO=50%       | 12:47:00 | 0.0                       | 0.0                        | 477.4         | 490.0                      | 479.8         | 489.4                      |
|                                 | 12:48:00 | 0.0                       | 0.0                        | 475.6         | 488.2                      | 479.8         | 489.4                      |
|                                 | 12:49:00 | 0.0                       | 0.0                        | 475.6         | 488.2                      | 479.8         | 489.4                      |
|                                 | 12:50:00 | 0.0                       | 0.0                        | 475.6         | 488.2                      | 479.8         | 489.4                      |
| average                         |          | 0.0                       | 0.0                        | 476.0         | 488.7                      | 479.8         | 489.4                      |
| NO <sub>2</sub> = 50%<br>130.75 | 13:11:00 | 130.0                     | 130.5                      | 350.5         | 488.2                      | 354.7         | 490.0                      |
|                                 | 13:12:00 | 130.0                     | 130.0                      | 349.9         | 488.2                      | 354.1         | 488.8                      |
|                                 | 13:13:00 | 130.0                     | 129.3                      | 350.5         | 488.2                      | 354.1         | 488.8                      |
|                                 | 13:14:00 | 130.0                     | 129.3                      | 349.9         | 488.2                      | 354.1         | 488.8                      |
| average                         |          | 130.0                     | 129.7                      | 350.2         | 488.2                      | 354.3         | 489.1                      |
|                                 |          |                           |                            |               |                            |               |                            |
| O <sub>3</sub> =0, NO=50%       | 14:15:00 | 0.0                       | 0.0                        | 477.4         | 488.2                      | 479.8         | 490.6                      |
|                                 | 14:16:00 | 0.0                       | 0.0                        | 475.6         | 488.2                      | 479.8         | 490.0                      |
|                                 | 14:17:00 | 0.0                       | 0.0                        | 475.6         | 491.2                      | 479.8         | 487.0                      |
|                                 | 14:18:00 | 0.0                       | 0.0                        | 475.6         | 488.8                      | 481.0         | 490.0                      |
| average                         |          | 0.0                       | 0.0                        | 476.0         | 489.1                      | 480.1         | 489.4                      |
| NO <sub>2</sub> = 95%<br>248.43 | 14:39:00 | 248.0                     | 248.5                      | 238.7         | 488.8                      | 236.3         | 488.8                      |
|                                 | 14:40:00 | 248.0                     | 248.5                      | 236.9         | 488.2                      | 234.5         | 490.6                      |
|                                 | 14:41:00 | 248.0                     | 248.6                      | 235.1         | 488.2                      | 233.9         | 490.0                      |
|                                 | 14:42:00 | 248.0                     | 248.7                      | 239.9         | 491.2                      | 233.3         | 487.0                      |
| average                         |          | 248.0                     | 248.6                      | 237.6         | 489.1                      | 234.5         | 489.1                      |
| O <sub>3</sub> =0, NO=50%       | 14:55:00 | 0.0                       | 0.0                        | 484.0         | 488.8                      | 483.4         | 488.2                      |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Page 139 of 293

## **7.1 8.4.15 Residence time in the analyser**

*The residence time in the analyser shall be  $\leq 3.0$  s.*

## **7.2 Test procedures**

The residence time inside the analyser shall be calculated on the basis of the flow and the volumes of the tubing and other relevant components inside the analyser.

## **7.3 Testing**

The gas volume of the N500 measuring system is approximately 30.0 ml from the sample inlet to the measuring cell. The typical sample gas flow is at 1.0 l/min. This results in a residence time in the analyser of 1.9 s.

## **7.4 Evaluation**

Not applicable.

## **7.5 Assessment**

Residence time in the analyser was 1.9 s.

Criterion satisfied? yes

## **7.6 Detailed presentation of test results**

Not applicable.

## 7.1 8.5.4 Long-term drift

*The long-term drift at zero point shall not exceed  $\leq 5.0$  nmol/mol.  
Long-term drift at span level shall not exceed 5% of the certification range.*

## 7.2 Test procedures

After each bi-weekly zero and span check, the drift of the analysers under test shall be calculated at zero and at span following the procedures as given below. If the drift compared to the initial calibration exceeds one of the performance criteria for drift at zero or span level, the “period of unattended operation” equals the number of weeks until the observation of the infringement, minus two weeks. For further (uncertainty) calculations, the values for “long term drift” are the values for zero and span drift over the period of unattended operation.

At the beginning of the drift period, five individual measurements were performed at zero and span level following the calibration (after waiting the time equivalent to a single independent reading).

The long-term drift is calculated as follows:

$$D_{L,Z} = (C_{Z,1} - C_{Z,0})$$

Where:

$D_{L,Z}$  is the drift at zero;

$C_{Z,0}$  is the average concentration of the measurements at zero at the beginning of the drift period;

$C_{Z,1}$  is the average concentration of the measurements at zero at the end of the drift period;

$D_{L,Z}$  shall comply with the performance criterion indicated above.

$$D_{L,S} = \frac{(C_{S,1} - C_{S,0}) - D_{L,Z}}{C_{S,1}} \times 100$$

Where:

$D_{L,S}$  is the drift at span concentration  $c_i$ ;

$C_{S,0}$  is the average concentration of the measurements at span level at the beginning of the drift period;

$C_{S,1}$  is the average concentration of the measurements at span level at the end of the drift period;

$D_{L,S}$  shall comply with the performance criterion indicated above.

### 7.3 Testing

For the purpose of this test, test gas was applied every other week. Table 65 and Table 66 report the measured values for bi-weekly test gas applications. In compliance with the standard, the test has to be performed with the component NO. Pursuant to EN 14211, the test shall be performed at a concentration level of 70% to 80% of the certification range for NO.

The test was also carried out for the NO<sub>2</sub> component at a concentration level of 70 % to 80 % of the certification range for NO<sub>2</sub>.

### 7.4 Evaluation

Table 65: Results for the long-term drift at zero point for NO

|  |            | requirement | Device 1 |   | Device 2 |   |
|--|------------|-------------|----------|---|----------|---|
| average start Cz <sub>1</sub> at zero [nmol/mol] | 09.11.2020 | ≤ 5.0       | --       | ✓ | --       | ✓ |
| long term drift DL,z at zero [nmol/mol]          | 23.11.2020 | ≤ 5.0       | 0.04     | ✓ | -0.36    | ✓ |
| long term drift DL,z at zero [nmol/mol]          | 07.12.2020 | ≤ 5.0       | -0.80    | ✓ | -0.48    | ✓ |
| long term drift DL,z at zero [nmol/mol]          | 18.12.2020 | ≤ 5.0       | -0.92    | ✓ | -0.36    | ✓ |
| long term drift DL,z at zero [nmol/mol]          | 04.01.2021 | ≤ 5.0       | 0.88     | ✓ | 0.48     | ✓ |
| long term drift DL,z at zero [nmol/mol]          | 18.01.2021 | ≤ 5.0       | 0.76     | ✓ | 0.60     | ✓ |
| long term drift DL,z at zero [nmol/mol]          | 01.02.2021 | ≤ 5.0       | 0.64     | ✓ | 1.08     | ✓ |
| long term drift DL,z at zero [nmol/mol]          | 15.02.2021 | ≤ 5.0       | 0.76     | ✓ | 0.84     | ✓ |

Table 66: Results for the long-term drift at span point for NO

|  |            | requirement | Device 2 1 |   | Device 2 |   |
|--|------------|-------------|------------|---|----------|---|
| average start Cs <sub>1</sub> at span [nmol/mol] | 09.11.2020 | ≤ 5 %       | --         | ✓ | --       | ✓ |
| long term drift DL,s at span [nmol/mol]          | 23.11.2020 | ≤ 5 %       | 0.13       | ✓ | 0.07     | ✓ |
| long term drift DL,s at span [nmol/mol]          | 07.12.2020 | ≤ 5 %       | -0.24      | ✓ | -0.17    | ✓ |
| long term drift DL,s at span [nmol/mol]          | 18.12.2020 | ≤ 5 %       | -0.53      | ✓ | -0.49    | ✓ |
| long term drift DL,s at span [nmol/mol]          | 04.01.2021 | ≤ 5 %       | 0.28       | ✓ | 0.33     | ✓ |
| long term drift DL,s at span [nmol/mol]          | 18.01.2021 | ≤ 5 %       | 0.43       | ✓ | 0.33     | ✓ |
| long term drift DL,s at span [nmol/mol]          | 01.02.2021 | ≤ 5 %       | 0.66       | ✓ | 0.74     | ✓ |
| long term drift DL,s at span [nmol/mol]          | 15.02.2021 | ≤ 5 %       | 0.93       | ✓ | 1.02     | ✓ |

**Table 67: Results for the long-term drift at zero point for NO<sub>2</sub>**

|   |            | requirement | Device 1 |   | Device 2 |   |
|---|------------|-------------|----------|---|----------|---|
| average start Cz <sub>1</sub> at zero [nmol/mol]    | 09.11.2020 | ≤ 5.0       | --       | ✓ | --       | ✓ |
| long term drift D <sub>L,z</sub> at zero [nmol/mol] | 23.11.2020 | ≤ 5.0       | 0.19     | ✓ | 0.17     | ✓ |
| long term drift D <sub>L,z</sub> at zero [nmol/mol] | 07.12.2020 | ≤ 5.0       | 0.43     | ✓ | 0.07     | ✓ |
| long term drift D <sub>L,z</sub> at zero [nmol/mol] | 18.12.2020 | ≤ 5.0       | -0.03    | ✓ | -0.05    | ✓ |
| long term drift D <sub>L,z</sub> at zero [nmol/mol] | 04.01.2021 | ≤ 5.0       | 0.75     | ✓ | 0.87     | ✓ |
| long term drift D <sub>L,z</sub> at zero [nmol/mol] | 18.01.2021 | ≤ 5.0       | 0.31     | ✓ | 0.47     | ✓ |
| long term drift D <sub>L,z</sub> at zero [nmol/mol] | 01.02.2021 | ≤ 5.0       | 0.09     | ✓ | 0.57     | ✓ |
| long term drift D <sub>L,z</sub> at zero [nmol/mol] | 15.02.2021 | ≤ 5.0       | 0.73     | ✓ | 0.47     | ✓ |

**Table 68: Results for the long-term drift at span point for NO<sub>2</sub>**

|   |            | requirement | Device 2 1 |   | Device 2 |   |
|---|------------|-------------|------------|---|----------|---|
| average start Cs <sub>1</sub> at span [nmol/mol]    | 09.11.2020 | ≤ 5 %       | --         | ✓ | --       | ✓ |
| long term drift D <sub>L,s</sub> at span [nmol/mol] | 23.11.2020 | ≤ 5 %       | 1.26       | ✓ | 1.22     | ✓ |
| long term drift D <sub>L,s</sub> at span [nmol/mol] | 07.12.2020 | ≤ 5 %       | 2.09       | ✓ | 1.81     | ✓ |
| long term drift D <sub>L,s</sub> at span [nmol/mol] | 18.12.2020 | ≤ 5 %       | -0.04      | ✓ | 0.74     | ✓ |
| long term drift D <sub>L,s</sub> at span [nmol/mol] | 04.01.2021 | ≤ 5 %       | 2.29       | ✓ | 2.05     | ✓ |
| long term drift D <sub>L,s</sub> at span [nmol/mol] | 18.01.2021 | ≤ 5 %       | 1.96       | ✓ | 1.46     | ✓ |
| long term drift D <sub>L,s</sub> at span [nmol/mol] | 01.02.2021 | ≤ 5 %       | 1.26       | ✓ | 1.58     | ✓ |
| long term drift D <sub>L,s</sub> at span [nmol/mol] | 15.02.2021 | ≤ 5 %       | 1.22       | ✓ | 1.95     | ✓ |

## 7.5 Assessment

Maximum long-term drift at zero point D<sub>l,z</sub> for NO was at -0.92 nmol/mol for instrument 1 and 1.08 nmol/mol for instrument 2. Maximum long-term drift at span point D<sub>l,s</sub> for NO was at 0.93% for instrument 1 and 1.02% for instrument 2.

Maximum long-term drift at zero point D<sub>l,z</sub> for NO<sub>2</sub> was at 0.75 nmol/mol for instrument 1 and 0.87 nmol/mol for instrument 2. Maximum long-term drift at span point D<sub>l,s</sub> for NO<sub>2</sub> was at 2.29% for instrument 1 and 2.05% for instrument 2.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

The individual test results for the determination of the long-term drift are shown in Table 69 and in Table 70.

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO2 and NOx,  
Report No.: 936/21251100/A

Table 69: Results of the drift test for NO

| Zero Concentration                              |          |              |              |
|---|----------|--------------|--------------|
| Date  | Time     | Device 1     | Device 2     |
|   |          | [nmol/mol]   | [nmol/mol]   |
| 09.11.2020                                      | 13:54:00 | 0.60         | 0.60         |
|   | 13:55:00 | 0.60         | 0.60         |
|   | 13:56:00 | 0.60         | 0.60         |
|   | average  | 0.60         | 0.60         |
|   | 13:58:00 | 0.60         | 0.60         |
|   | 13:59:00 | 0.60         | 0.60         |
|   | 14:00:00 | 0.60         | 0.60         |
|   | average  | 0.60         | 0.60         |
|   | 14:02:00 | 1.20         | 0.60         |
|   | 14:03:00 | 0.60         | 0.00         |
|   | 14:04:00 | 0.00         | 0.60         |
|   | average  | 0.60         | 0.40         |
|   | 14:06:00 | 0.00         | 0.00         |
|   | 14:07:00 | 0.00         | 0.00         |
|   | 14:08:00 | -0.60        | 0.00         |
|   | average  | -0.20        | 0.00         |
|   | 14:10:00 | 0.00         | 0.00         |
| 14:11:00  | 0.00     | 0.60         |              |
| 14:12:00  | 0.00     | 0.00         |              |
| average   | 0.00     | 0.20         |              |
| <b>Average field start <math>c_{z,0}</math></b> |          | <b>0.32</b>  | <b>0.36</b>  |
| 23.11.2020                                      | 10:10:00 | 0.60         | 0.00         |
|   | 10:11:00 | 0.60         | 0.00         |
|   | 10:12:00 | 0.00         | 0.00         |
|   | 10:13:00 | 0.00         | 0.00         |
|   | 10:14:00 | 0.60         | 0.00         |
| aver. $C_{z,1}$                                 |          | <b>0.36</b>  | <b>0.00</b>  |
| 07.12.2020                                      | 14:01:00 | 0.00         | 0.00         |
|   | 14:02:00 | -0.60        | -0.60        |
|   | 14:03:00 | -0.60        | -0.60        |
|   | 14:04:00 | -1.20        | 0.00         |
|   | 14:05:00 | 0.00         | 0.60         |
| aver. $C_{z,1}$                                 |          | <b>-0.48</b> | <b>-0.12</b> |
| 18.12.2020                                      | 11:17:00 | -0.60        | 0.00         |
|   | 11:18:00 | -0.60        | 0.00         |
|   | 11:19:00 | -0.60        | 0.00         |
|   | 11:20:00 | -0.60        | 0.00         |
|   | 11:21:00 | -0.60        | 0.00         |
| aver. $C_{z,1}$                                 |          | <b>-0.60</b> | <b>0.00</b>  |
| 04.01.2021                                      | 14:19:00 | 1.20         | 0.60         |
|   | 14:20:00 | 0.00         | 0.00         |
|   | 14:21:00 | 1.80         | 1.20         |
|   | 14:22:00 | 1.20         | 1.20         |
|   | 14:23:00 | 1.80         | 1.20         |
| aver. $C_{z,1}$                                 |          | <b>1.20</b>  | <b>0.84</b>  |
| 18.01.2021                                      | 12:11:00 | 1.20         | 0.60         |
|   | 12:12:00 | 1.20         | 1.20         |
|   | 12:13:00 | 0.60         | 1.20         |
|   | 12:14:00 | 1.20         | 0.60         |
|   | 12:15:00 | 1.20         | 1.20         |
| aver. $C_{z,1}$                                 |          | <b>1.08</b>  | <b>0.96</b>  |
| 01.02.2021                                      | 10:00:00 | 0.60         | 1.20         |
|   | 10:01:00 | 0.60         | 1.80         |
|   | 10:02:00 | 1.80         | 1.20         |
|   | 10:03:00 | 1.20         | 1.20         |
|   | 10:04:00 | 0.60         | 1.80         |
| aver. $C_{z,1}$                                 |          | <b>0.96</b>  | <b>1.44</b>  |
| 15.02.2021                                      | 10:29:00 | 0.60         | 0.60         |
|   | 10:30:00 | 1.20         | 1.20         |
|   | 10:31:00 | 0.60         | 1.20         |
|   | 10:32:00 | 1.20         | 1.20         |
|   | 10:33:00 | 1.80         | 1.80         |
| aver. $C_{z,1}$                                 |          | <b>1.08</b>  | <b>1.20</b>  |

| $C_t$ -Concentration                            |          |               |               |
|---|----------|---------------|---------------|
| Date  | Time     | Device 1      | Device 2      |
|   |          | [nmol/mol]    | [nmol/mol]    |
| 09.11.2020                                      | 14:50:00 | 720.30        | 720.90        |
|   | 14:51:00 | 720.30        | 720.30        |
|   | 14:52:00 | 719.70        | 719.70        |
|   | average  | 720.10        | 720.30        |
|   | 14:54:00 | 720.90        | 720.90        |
|   | 14:55:00 | 720.30        | 720.90        |
|   | 14:56:00 | 720.30        | 720.30        |
|   | average  | 720.50        | 720.70        |
|   | 14:58:00 | 720.30        | 720.30        |
|   | 14:59:00 | 720.90        | 720.90        |
|   | 15:00:00 | 720.30        | 720.90        |
|   | average  | 720.50        | 720.70        |
|   | 15:02:00 | 719.70        | 719.70        |
|   | 15:03:00 | 720.30        | 720.30        |
|   | 15:04:00 | 720.90        | 720.90        |
|   | average  | 720.30        | 720.30        |
|   | 15:06:00 | 720.30        | 720.30        |
| 15:07:00  | 719.70   | 719.70        |               |
| 15:08:00  | 720.30   | 720.30        |               |
| average   | 720.10   | 720.10        |               |
| <b>Average field start <math>C_{s,0}</math></b> |          | <b>720.30</b> | <b>720.42</b> |
| 23.11.2020                                      | 10:32:00 | 720.90        | 719.70        |
|   | 10:33:00 | 720.30        | 719.10        |
|   | 10:34:00 | 722.10        | 720.30        |
|   | 10:35:00 | 722.10        | 722.10        |
|   | 10:36:00 | 720.90        | 721.50        |
| aver. $C_{s,1}$                                 |          | <b>721.26</b> | <b>720.54</b> |
| 07.12.2020                                      | 14:11:00 | 717.30        | 718.50        |
|   | 14:12:00 | 717.90        | 718.50        |
|   | 14:13:00 | 717.30        | 719.10        |
|   | 14:14:00 | 718.50        | 719.10        |
|   | 14:15:00 | 717.90        | 718.50        |
| aver. $C_{s,1}$                                 |          | <b>717.78</b> | <b>718.74</b> |
| 18.12.2020                                      | 11:28:00 | 714.90        | 717.30        |
|   | 11:29:00 | 716.10        | 717.30        |
|   | 11:30:00 | 716.10        | 717.30        |
|   | 11:31:00 | 716.10        | 716.70        |
|   | 11:32:00 | 714.90        | 714.30        |
| aver. $C_{s,1}$                                 |          | <b>715.62</b> | <b>716.58</b> |
| 04.01.2021                                      | 14:41:00 | 722.70        | 723.30        |
|   | 14:42:00 | 723.90        | 723.90        |
|   | 14:43:00 | 723.90        | 723.30        |
|   | 14:44:00 | 723.30        | 723.90        |
|   | 14:45:00 | 722.10        | 722.10        |
| aver. $C_{s,1}$                                 |          | <b>723.18</b> | <b>723.30</b> |
| 18.01.2021                                      | 12:33:00 | 723.30        | 722.70        |
|   | 12:34:00 | 723.90        | 723.30        |
|   | 12:35:00 | 723.90        | 722.70        |
|   | 12:36:00 | 725.10        | 723.90        |
|   | 12:37:00 | 724.50        | 724.50        |
| aver. $C_{s,1}$                                 |          | <b>724.14</b> | <b>723.42</b> |
| 01.02.2021                                      | 10:22:00 | 724.50        | 727.50        |
|   | 10:23:00 | 726.30        | 726.90        |
|   | 10:24:00 | 725.70        | 726.30        |
|   | 10:25:00 | 726.30        | 727.50        |
|   | 10:26:00 | 725.70        | 726.30        |
| aver. $C_{s,1}$                                 |          | <b>725.70</b> | <b>726.90</b> |
| 15.02.2021                                      | 10:41:00 | 728.10        | 726.30        |
|   | 10:42:00 | 728.10        | 729.30        |
|   | 10:43:00 | 727.50        | 728.70        |
|   | 10:44:00 | 727.50        | 729.30        |
|   | 10:45:00 | 728.10        | 729.90        |
| aver. $C_{s,1}$                                 |          | <b>727.86</b> | <b>728.70</b> |

**Table 70: Results of the drift test for NO<sub>2</sub>**

| Zero Concentration                              |                 |             |             |
|---|-----------------|-------------|-------------|
| Date  | Time            | Device 1    | Device 2    |
|   |                 | [nmol/mol]  | [nmol/mol]  |
| 09.11.2020                                      | 13:54:00        | 0.20        | 0.20        |
|   | 13:55:00        | 0.20        | 0.20        |
|   | 13:56:00        | 0.20        | 0.20        |
|   | average         | 0.20        | 0.20        |
|   | 13:58:00        | 0.20        | 0.20        |
|   | 13:59:00        | 0.20        | 0.20        |
|   | 14:00:00        | 0.20        | 0.20        |
|   | average         | 0.20        | 0.20        |
|   | 14:02:00        | 0.30        | 0.20        |
|   | 14:03:00        | 0.20        | 0.00        |
|   | 14:04:00        | 0.20        | 0.00        |
|   | average         | 0.23        | 0.07        |
|   | 14:06:00        | -0.20       | 0.00        |
|   | 14:07:00        | 0.00        | 0.20        |
|   | 14:08:00        | 0.00        | 0.00        |
|   | average         | -0.07       | 0.07        |
|   | 14:10:00        | -0.20       | 0.20        |
| 14:11:00  | -0.20           | 0.20        |             |
| 14:12:00  | -0.30           | 0.00        |             |
| average   | -0.23           | 0.13        |             |
| <b>Average field start <math>c_{s,0}</math></b> |                 | <b>0.07</b> | <b>0.13</b> |
| 23.11.2020                                      | 10:10:00        | 0.30        | 0.20        |
|   | 10:11:00        | 0.30        | 0.20        |
|   | 10:12:00        | 0.30        | 0.20        |
|   | 10:13:00        | 0.20        | 0.20        |
|   | 10:14:00        | 0.20        | 0.70        |
|   | aver. $c_{s,1}$ | <b>0.26</b> | <b>0.30</b> |
| 07.12.2020                                      | 14:01:00        | 0.30        | 0.20        |
|   | 14:02:00        | 0.50        | 0.20        |
|   | 14:03:00        | 0.50        | 0.20        |
|   | 14:04:00        | 0.50        | 0.20        |
|   | 14:05:00        | 0.70        | 0.20        |
|   | aver. $c_{s,1}$ | <b>0.50</b> | <b>0.20</b> |
| 18.12.2020                                      | 11:17:00        | 0.20        | 0.20        |
|   | 11:18:00        | 0.20        | 0.20        |
|   | 11:19:00        | 0.00        | 0.00        |
|   | 11:20:00        | 0.00        | 0.00        |
|   | 11:21:00        | -0.20       | 0.00        |
|   | aver. $c_{s,1}$ | <b>0.04</b> | <b>0.08</b> |
|   |                 |             |             |
| 04.01.2021                                      | 14:19:00        | 0.80        | 1.00        |
|   | 14:20:00        | 1.00        | 0.80        |
|   | 14:21:00        | 0.80        | 1.30        |
|   | 14:22:00        | 0.80        | 1.10        |
|   | 14:23:00        | 0.70        | 0.80        |
|   | aver. $c_{s,1}$ | <b>0.82</b> | <b>1.00</b> |
| 18.01.2021                                      | 12:11:00        | 0.50        | 0.50        |
|   | 12:12:00        | 0.30        | 0.80        |
|   | 12:13:00        | 0.30        | 0.50        |
|   | 12:14:00        | 0.50        | 0.50        |
|   | 12:15:00        | 0.30        | 0.70        |
|   | aver. $c_{s,1}$ | <b>0.38</b> | <b>0.60</b> |
| 01.02.2021                                      | 10:00:00        | 0.30        | 0.20        |
|   | 10:01:00        | -0.20       | 0.80        |
|   | 10:02:00        | 0.30        | 0.80        |
|   | 10:03:00        | 0.20        | 0.70        |
|   | 10:04:00        | 0.20        | 1.00        |
|   | aver. $c_{s,1}$ | <b>0.16</b> | <b>0.70</b> |
| 15.02.2021                                      | 10:29:00        | 1.00        | 0.20        |
|   | 10:30:00        | 0.70        | 0.50        |
|   | 10:31:00        | 0.80        | 1.00        |
|   | 10:32:00        | 0.70        | 0.50        |
|   | 10:33:00        | 0.80        | 0.80        |
|   | aver. $c_{s,1}$ | <b>0.80</b> | <b>0.60</b> |

| C <sub>r</sub> -Concentration                   |                 |               |               |
|---|-----------------|---------------|---------------|
| Date  | Time            | Device 1      | Device 2      |
|   |                 | [nmol/mol]    | [nmol/mol]    |
| 09.11.2020                                      | 14:22:00        | 203.30        | 202.60        |
|   | 14:23:00        | 202.80        | 202.60        |
|   | 14:24:00        | 203.30        | 203.10        |
|   | average         | 203.13        | 202.77        |
|   | 14:26:00        | 202.80        | 202.60        |
|   | 14:27:00        | 203.60        | 203.40        |
|   | 14:28:00        | 203.30        | 203.10        |
|   | average         | 203.23        | 203.03        |
|   | 14:30:00        | 203.40        | 203.30        |
|   | 14:31:00        | 203.60        | 203.30        |
|   | 14:32:00        | 203.30        | 203.10        |
|   | average         | 203.43        | 203.23        |
|   | 14:34:00        | 203.30        | 203.40        |
|   | 14:35:00        | 203.40        | 203.30        |
|   | 14:36:00        | 203.60        | 203.40        |
|   | average         | 203.43        | 203.37        |
|   | 14:38:00        | 203.30        | 203.60        |
|   | 14:39:00        | 203.40        | 203.40        |
|   | 14:40:00        | 203.60        | 203.40        |
|   | average         | 203.43        | 203.47        |
| <b>Average field start <math>c_{s,0}</math></b> |                 | <b>203.33</b> | <b>203.17</b> |
| 23.11.2020                                      | 10:23:00        | 206.40        | 206.20        |
|   | 10:24:00        | 206.20        | 206.00        |
|   | 10:25:00        | 205.90        | 205.70        |
|   | 10:26:00        | 205.90        | 205.50        |
|   | 10:27:00        | 206.20        | 205.90        |
|   | aver. $c_{s,1}$ | <b>206.12</b> | <b>205.86</b> |
| 07.12.2020                                      | 14:28:00        | 207.50        | 207.30        |
|   | 14:29:00        | 208.00        | 207.30        |
|   | 14:30:00        | 208.30        | 206.80        |
|   | 14:31:00        | 208.30        | 206.70        |
|   | 14:32:00        | 208.50        | 206.80        |
|   | aver. $c_{s,1}$ | <b>208.12</b> | <b>206.98</b> |
| 18.12.2020                                      | 11:41:00        | 203.40        | 205.20        |
|   | 11:42:00        | 203.10        | 204.70        |
|   | 11:43:00        | 203.60        | 204.70        |
|   | 11:44:00        | 203.40        | 204.40        |
|   | 11:45:00        | 202.60        | 204.20        |
|   | aver. $c_{s,1}$ | <b>203.22</b> | <b>204.64</b> |
|   |                 |               |               |
| 04.01.2021                                      | 14:30:00        | 208.60        | 208.50        |
|   | 14:31:00        | 208.60        | 208.10        |
|   | 14:32:00        | 208.60        | 208.60        |
|   | 14:33:00        | 209.50        | 207.80        |
|   | 14:34:00        | 209.00        | 208.50        |
|   | aver. $c_{s,1}$ | <b>208.86</b> | <b>208.30</b> |
| 18.01.2021                                      | 12:22:00        | 207.70        | 206.80        |
|   | 12:23:00        | 207.50        | 207.30        |
|   | 12:24:00        | 207.70        | 206.20        |
|   | 12:25:00        | 208.00        | 206.80        |
|   | 12:26:00        | 207.70        | 206.20        |
|   | aver. $c_{s,1}$ | <b>207.72</b> | <b>206.66</b> |
| 01.02.2021                                      | 10:12:00        | 205.70        | 207.00        |
|   | 10:13:00        | 206.40        | 206.80        |
|   | 10:14:00        | 206.20        | 206.80        |
|   | 10:15:00        | 205.90        | 207.30        |
|   | 10:16:00        | 205.90        | 207.20        |
|   | aver. $c_{s,1}$ | <b>206.02</b> | <b>207.02</b> |
| 15.02.2021                                      | 10:54:00        | 206.20        | 208.00        |
|   | 10:55:00        | 206.40        | 207.50        |
|   | 10:56:00        | 206.80        | 208.00        |
|   | 10:57:00        | 206.70        | 207.70        |
|   | 10:58:00        | 206.80        | 207.30        |
|   | aver. $c_{s,1}$ | <b>206.58</b> | <b>207.70</b> |



## 7.1 8.5.5 Reproducibility standard deviation for NO<sub>2</sub> under field conditions

*Reproducibility standard deviation under field conditions shall not exceed 5% of the mean value over a period of three months.*

### 7.2 Test procedures

The reproducibility standard deviation under field conditions is calculated from the measured hourly averaged data during the three-month period.

The difference  $\Delta x_{f,i}$  for each (ith) parallel measurement is calculated from:

$$\Delta x_{f,i} = x_{f,1,i} - x_{f,2,i}$$

Where:

$\Delta x_{f,i}$  is the ith difference in a parallel measurement;

$x_{f,1,i}$  is the ith measurement result of analyser 1;

$x_{f,2,i}$  is the ith measurement result of analyser 2;

The reproducibility standard deviation under field conditions is calculated according to:

$$s_{r,f} = \frac{\left( \sqrt{\frac{\sum_{i=1}^n \Delta x_{f,i}^2}{2 * n}} \right)}{c_f} \times 100$$

Where:

$s_{r,f}$  is the reproducibility standard deviation under field conditions (%);

$n$  is the number of parallel measurements;

$c_f$  is the average concentration of nitrogen dioxide measured during the field test;

The reproducibility standard deviation under field conditions,  $s_{r,f}$ , shall comply with the performance criterion indicated above.

### 7.3 Testing

The reproducibility standard deviation under field conditions was calculated from the hourly NO<sub>2</sub> averages over the field test period according to the equation stated above.

## 7.4 Evaluation

Table 71: Reproducibility standard deviation for NO<sub>2</sub> based on complete field test data

| reproducibility standard deviation in field                             |            |             |
|---|------------|-------------|
| no. of measurements (1h- average)                                       | [n]        | 2354        |
| average of both analyzers (3 month)                                     | [nmol/mol] | 13.06       |
| standard deviation from paired measurements                             | [nmol/mol] | 0.479       |
| <b>reproducibility standard deviation in field <math>S_{r,f}</math></b> | <b>[%]</b> | <b>3.67</b> |
| requirement   | ≤ 5.0 %    | ✓           |




Figure 16: Reproducibility standard deviation for NO<sub>2</sub> under field conditions

The reproducibility standard deviation under field conditions was at 3.67% of the average for NO<sub>2</sub>. This value is also used for the calculation of the total uncertainty according to EN 14211.

## 7.5 Assessment

The reproducibility standard deviation for NO<sub>2</sub> under field conditions was 3.67% as a percentage of the mean value over the three-month field test period. Thus, the requirements of EN 14211 are satisfied.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

For the sake of clarity, the results of the reproducibility standard deviation for NO are also given here. According to EN 14211, the evaluation of the reproducibility standard deviation is carried out in the field with NO<sub>2</sub>. Therefore, the following data are listed for information only.

Table 72: Reproducibility standard deviation for NO based on complete field test data

| reproducibility standard deviation in field                             |            |             |
|---|------------|-------------|
| no. of measurements (1h- average)                                       | [n]        | 2354        |
| average of both analyzers (3 month)                                     | [nmol/mol] | 36.62       |
| standard deviation from paired measurements                             | [nmol/mol] | 0.398       |
| <b>reproducibility standard deviation in field <math>S_{r,f}</math></b> | <b>[%]</b> | <b>1.09</b> |
| requirement   | ≤ 5.0 %    | ✓           |




Figure 17: Reproducibility standard deviation for NO under field conditions

## 7.1 8.5.6 Inspection interval

*The period of unattended operation of the AMS shall be at least 2 weeks.*

## 7.2 Equipment

Not required for this performance criterion.

## 7.3 Testing

With regard to this minimum requirement, the maintenance tasks required in a specific period and the length of that period for the correct functioning of the measuring system were identified. Furthermore, in determining the maintenance interval, the drift determined for zero and span point in accordance with 7.1 8.5.4 Long-term drift have been taken into consideration.

## 7.4 Evaluation

Over the entire period of the field test, no unacceptable drift was observed. The maintenance interval is thus determined by the necessary maintenance works.

During the three month field test period, maintenance is generally limited to contamination and plausibility checks and potential status/error messages. Naturally, the frequency of filter replacement will depend on the ambient dust concentration at the site of installation. Chapter 5 of the manual and chapter 8 of this report provide information about tasks to be performed in the maintenance interval.

## 7.5 Assessment

The necessary maintenance tasks determine the period of unattended operation. In essence, these include contamination checks, plausibility checks and checks of potential status/error warnings. The external particle filter must be changed depending on the dust load at the measuring location. EN 14211 requires checking of zero and span points at least once every two weeks.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

Not applicable in this instance.

## 7.1 8.5.7 Period of availability of the analyser

*Availability of the analyser shall be at least 90%.*

### 7.2 Test procedures

The correct operation of the analysers shall be checked at least every fourteen days. It is recommended to perform this check every day during the first fourteen days. These checks consist of plausibility checks on the measured values, as well as, when available, on status signals and other relevant parameters. Time, duration and nature of any malfunctioning shall be logged.

The total time period with useable measuring data is the period during the field test during which valid measuring data of the ambient air concentrations are obtained. In this time period, the time needed for calibrations, conditioning of sample systems and filters and maintenance shall not be included.

The availability of the analyser is calculated as:

$$A_a = \frac{t_u}{t_t} * 100$$

Where:

$A_a$  is the availability of the analyser (%);

$t_u$  is the total time period with validated measuring data;

$t_t$  is the time period of the field test minus the time for calibration, conditioning and maintenance,  $t_u$  and  $t_t$  shall be expressed in the same units.

The availability shall comply with the performance criterion indicated above.

### 7.3 Testing

Using the equation given above, the availability was calculated from the total period of the field test and the outage times which have occurred during this period.

#### Evaluation

Outage times which have occurred during the field test are listed in Table 73.

Table 73: Availability of the N500 measuring system

|  |   | System 1 | System 2 |
|--|---|----------|----------|
| Operation time                                 | h | 2354     | 2354     |
| Outage time                                    | h | 0        | 0        |
| Maintenance time                               | h | 12       | 12       |
| Actual operating time:                         | h | 2342     | 2342     |
| Actual operating time incl. maintenance times: | h | 2354     | 2354     |
| Availability                                   | % | 100      | 100      |

Maintenance times were caused by daily test gas feeding for the purpose of determining the drift behaviour and the maintenance interval and by times needed for replacing the Teflon filter built into the sample gas path.

### 7.5 Assessment

The availability is at 100%. Thus, the requirement of EN 14211 is satisfied.

Criterion satisfied? yes

### 7.6 Detailed presentation of test results

Not applicable.

## **7.1 8.6 Calculation of the total uncertainty in accordance with standard EN 14211 (2012)**

*The type approval of the analyser consists of the following steps:*

- 1) The value of each individual performance characteristic tested in the laboratory shall fulfil the criterion stated in Table E.1 of standard EN 14211.*
- 2) The expanded uncertainty calculated from the standard uncertainties due to the values of the specific performance characteristics obtained in the laboratory tests shall fulfil the criterion as stated in Annex I of Directive 2008/50/EC (15% for fixed measurements or 25% for indicative measurements). This criterion is the maximum uncertainty of individual measurements for continuous measurements at the 1-hour limit value. The relevant specific performance characteristics and the calculation procedure are given in Annex E of standard EN 14211.*
- 3) The value of each of the individual performance characteristics tested in the field shall fulfil the criterion stated in Table E.1 of EN 14211.*
- 4) The expanded uncertainty calculated from the standard uncertainties due to the values of the specific performance characteristics obtained in the laboratory and field tests shall fulfil the criterion as stated in Annex I of Directive 2008/50/EC (15% for fixed measurements or 25% for indicative measurements). This criterion is the maximum uncertainty of individual measurements for continuous measurements at the 1-hour limit value. The relevant specific performance characteristics and the calculation procedure are given in Annex E of standard EN 14211.*

### **7.2 Equipment**

Calculation of the total uncertainty in accordance with standard EN 14211 (2012), Annex E

### **7.3 Testing**

At the end of the performance test, the total uncertainties were calculated from the values obtained during the test.

### **7.4 Evaluation**

Regarding 1) The value of each performance characteristic tested in the laboratory tests fulfils the criterion stated in Table E.1 of EN 14211.

Regarding 2) The expanded uncertainty calculated from the standard uncertainties due to the values of the specific performance characteristics obtained in the laboratory tests fulfils the criterion as stated.

Regarding 3) The value of each performance characteristic tested in the field tests fulfils the criterion stated in Table E.1 of EN 14211.

Regarding 4) The expanded uncertainty calculated from the standard uncertainties due to the values of the specific performance characteristics obtained in the laboratory and field tests fulfils the criterion as stated.

### **7.5 Assessment**

The requirement regarding the expanded uncertainty of the measuring system is complied with.

Criterion satisfied? yes



## 7.6 Detailed presentation of test results

Table 74 summarises the results for items 1 and 3.

Table 75 and Table 77 contain the results regarding item 2.

Table 76 and Table 78 contain the results regarding item 4.

Table 74: Relevant performance characteristics and criteria according to EN 14211

| Performance characteristic   | Performance criterion   | Test result  | Satisfied | Page |
|--|---|--|-----------|------|
| 8.4.5 Repeatability standard deviation at zero                               | $\leq 1.0$ nmol/mol   | S <sub>r</sub> System 1: 0.13 nmol/mol (NO)<br>S <sub>r</sub> System 2: 0.13 nmol/mol (NO)<br><br>S <sub>r</sub> System 1: 0.00 nmol/mol (NO <sub>2</sub> )<br>S <sub>r</sub> System 2: 0.11 nmol/mol (NO <sub>2</sub> )   | yes       | 101  |
| 8.4.5 Repeatability standard deviation at concentration level c <sub>i</sub> | $\leq 3.0$ nmol/mol   | S <sub>r</sub> System 1: 1.63 nmol/mol (NO)<br>S <sub>r</sub> System 2: 1.34 nmol/mol (NO)<br><br>S <sub>r</sub> System 1: 0.08 nmol/mol (NO <sub>2</sub> )<br>S <sub>r</sub> System 2: 0.08 nmol/mol (NO <sub>2</sub> )   | yes       | 101  |
| 8.4.6 "Lack-of-fit" (deviation from the linear regression)                   | Largest deviation from the linear regression function > 0, $\leq 4.0\%$ of the reading<br>Residual at zero: $\leq 5$ nmol/mol | X <sub>i,z</sub> System 1: ZP -0.60 nmol/mol<br>X <sub>i</sub> System 1: RP 1.72% (NO)<br>X <sub>i,z</sub> System 2: ZP -0.60 nmol/mol<br>X <sub>i</sub> System 2: RP 2.24% (NO)<br><br>X <sub>i,z</sub> System 1: ZP -0.16 nmol/mol (NO <sub>2</sub> )<br>X <sub>i</sub> System 1: RP 0.85% (NO <sub>2</sub> )<br>X <sub>i,z</sub> System 2: ZP -0.49 nmol/mol (NO <sub>2</sub> )<br>X <sub>i</sub> System 2: RP 0.72% (NO <sub>2</sub> ) | yes       | 105  |
| 8.4.7 Sensitivity coefficient of sample gas pressure                         | $\leq 8.0$ nmol/mol/kPa   | b <sub>gp</sub> System 1: 0.23 nmol/mol/kPa (NO)<br>b <sub>gp</sub> System 2: 0.17 nmol/mol/kPa (NO)<br><br>b <sub>gp</sub> System 1: 0.10 nmol/mol/kPa (NO <sub>2</sub> )<br>b <sub>gp</sub> System 2: 0.08 nmol/mol/kPa (NO <sub>2</sub> )   | yes       | 112  |
| 8.4.8 Sensitivity coefficient of sample gas temperature                      | $\leq 3.0$ nmol/mol/K   | b <sub>gt</sub> System 1: 0.09 nmol/mol/K (NO)<br>b <sub>gt</sub> System 2: 0.11 nmol/mol/K (NO)<br><br>b <sub>gt</sub> System 1: 0.01 nmol/mol/K (NO <sub>2</sub> )<br>b <sub>gt</sub> System 2: 0.01 nmol/mol/K (NO <sub>2</sub> )   | yes       | 115  |
| 8.4.9 Sensitivity coefficient of surrounding temperature                     | $\leq 3.0$ nmol/mol/K   | b <sub>st</sub> System 1: 0.952 nmol/mol/K (NO)<br>b <sub>st</sub> System 2: 0.451 nmol/mol/K (NO)<br><br>b <sub>st</sub> System 1: 0.231 nmol/mol/K (NO <sub>2</sub> )<br>b <sub>st</sub> System 2: 0.241 nmol/mol/K (NO <sub>2</sub> )   | yes       | 118  |

| Performance characteristic   | Performance criterion   | Test result  | Satisfied | Page |
|--|---|--|-----------|------|
| 8.4.10 Sensitivity coefficient of electrical voltage                 | $\leq 0.3$ nmol/mol/V   | b <sub>v</sub> System 1: RP 0.01 nmol/mol/V (NO)<br>b <sub>v</sub> System 2: RP 0.00 nmol/mol/V (NO)<br><br>b <sub>v</sub> System 1: RP 0.02 nmol/mol/V (NO <sub>2</sub> )<br>b <sub>v</sub> System 2: RP 0.01 nmol/mol/V (NO <sub>2</sub> )   | yes       | 122  |
| 8.4.11 Interferent at zero and at concentration level c <sub>t</sub> | H <sub>2</sub> O<br>$\leq 5.0$ nmol/mol<br>CO <sub>2</sub><br>$\leq 5.0$ nmol/mol<br>NH <sub>3</sub><br>$\leq 5.0$ nmol/mol | NO duct - H <sub>2</sub> O<br>System 1: ZP 0.00 nmol/mol / RP 0.40 nmol/mol<br>System 2: ZP -0.60 nmol/mol / RP 1.60 nmol/mol<br>NO duct - CO <sub>2</sub><br>System 1: ZP 0.40 nmol/mol / RP 1.80 nmol/mol<br>System 2: ZP 0.40 nmol/mol / RP 0.80 nmol/mol<br>NO duct - NH <sub>3</sub><br>System 1: ZP 0.60 nmol/mol / RP 1.20 nmol/mol<br>System 2: ZP 1.00 nmol/mol / RP 1.40 nmol/mol<br>NO <sub>2</sub> duct - H <sub>2</sub> O<br>System 1: ZP 0.27 nmol/mol / RP 0.33 nmol/mol<br>System 2: ZP 0.00 nmol/mol / RP 0.22 nmol/mol<br>NO <sub>2</sub> duct - CO <sub>2</sub><br>System 1: ZP 0.00 nmol/mol / RP 0.70 nmol/mol<br>System 2: ZP -0.21 nmol/mol / RP 0.76 nmol/mol<br>NO <sub>2</sub> duct - NH <sub>3</sub><br>System 1: ZP 0.00 nmol/mol / RP 1.09 nmol/mol<br>System 2: ZP -0.16 nmol/mol / RP 0.98 nmol/mol | yes       | 126  |
| 8.4.12 Averaging effect  | $\leq 7.0\%$ of the measured value  | E <sub>av</sub> System 1: 3.1% NO<br>E <sub>av</sub> System 2: -2.9% NO<br><br>E <sub>av</sub> System 1: 0.7% NO <sub>2</sub><br>E <sub>av</sub> System 2: 3.2% NO <sub>2</sub>  | yes       | 130  |
| 8.4.13 Difference sample/calibration port                            | $\leq 1.0\%$  | Δ <sub>SC</sub> System 1: 0.14% NO<br>Δ <sub>SC</sub> System 2: 0.06% NO<br><br>Δ <sub>SC</sub> System 1: -0.14% NO <sub>2</sub><br>Δ <sub>SC</sub> System 2: -0.33% NO <sub>2</sub>   | yes       | 134  |
| 8.4.3 Response time (rise)   | $\leq 180$ s  | t <sub>r</sub> System 1: 24 s (NO)<br>t <sub>r</sub> System 2: 25 s (NO)<br><br>t <sub>r</sub> System 1: 24 s (NO <sub>2</sub> )<br>t <sub>r</sub> System 2: 25 s (NO <sub>2</sub> )   | yes       | 90   |
| 8.4.3 Response time (fall)   | $\leq 180$ s  | t <sub>f</sub> System 1: 22 s (NO)<br>t <sub>f</sub> System 2: 23 s (NO)<br><br>t <sub>f</sub> System 1: 22 s (NO <sub>2</sub> )<br>t <sub>f</sub> System 2: 22 s (NO <sub>2</sub> )   | yes       | 90   |

| Performance characteristic                                      | Performance criterion                                | Test result   | Satisfied | Page |
|---|--|---|-----------|------|
| 8.4.3 Difference between the rise and fall response time        | ≤ 10 s   | t <sub>d</sub> System 1: 2 s (NO)<br>t <sub>d</sub> System 2: 2 s (NO)<br><br>t <sub>d</sub> System 1: 2 s (NO <sub>2</sub> )<br>t <sub>d</sub> System 2: 3 s (NO <sub>2</sub> )  | yes       | 90   |
| 8.4.14 Converter efficiency                                     | ≥ 98%  | E <sub>conv</sub> System 1: 99.6%<br>E <sub>conv</sub> System 2: 99.8%  | yes       | 136  |
| 8.4.15 Residence time   | ≤ 3.0 s  | System 1: 1.9 s<br>System 2: 1.9 s  | yes       | 139  |
| 8.5.7 Availability of the measuring system                      | > 90%  | A <sub>a</sub> System 1: 100%<br>A <sub>a</sub> System 2: 100%  | yes       | 150  |
| 8.5.5 Reproducibility standard deviation under field conditions | ≤ 5.0% of the average over a period of 3 months      | S <sub>r,f</sub> System 1: 3.67% NO <sub>2</sub><br>S <sub>r,f</sub> System 2: 3.67% NO <sub>2</sub>  | yes       | 145  |
| 8.5.4 Long-term drift at zero point                             | ≤ 5.0 nmol/mol                                       | C <sub>z</sub> System 1: -0.92 nmol/mol NO<br>C <sub>z</sub> System 2: 1.08 nmol/mol NO<br><br>C <sub>z</sub> System 1: 0.75 nmol/mol NO <sub>2</sub><br>C <sub>z</sub> System 2: 0.87 nmol/mol NO <sub>2</sub>                     | yes       | 140  |
| 8.5.4 Long-term drift at span                                   | ≤ 5.0% of the upper limit of the certification range | C <sub>s</sub> System 1: max. 0.93% NO<br>C <sub>s</sub> System 2: max. 1.02% NO<br><br>C <sub>s</sub> System 1: max. 2.29% NO <sub>2</sub><br>C <sub>s</sub> System 2: max. 2.05% NO <sub>2</sub>                                  | yes       | 140  |
| 8.4.4 Short-term drift at zero                                  | ≤ 2.0 nmol/mol over 12 h                             | D <sub>s,z</sub> System 1: 0.06 nmol/mol (NO)<br>D <sub>s,z</sub> System 2: -0.27 nmol/mol (NO)<br><br>D <sub>s,z</sub> System 1: -0.03 nmol/mol (NO <sub>2</sub> )<br>D <sub>s,z</sub> System 2: -0.26 nmol/mol (NO <sub>2</sub> ) | yes       | 95   |
| 8.4.4 Short-term drift at span                                  | ≤ 6.0 nmol/mol over 12 h                             | D <sub>s,s</sub> System 1: 2.36 nmol/mol (NO)<br>D <sub>s,s</sub> System 2: 1.11 nmol/mol (NO)<br><br>D <sub>s,s</sub> System 1: 2.76 nmol/mol (NO <sub>2</sub> )<br>D <sub>s,s</sub> System 2: 2.72 nmol/mol (NO <sub>2</sub> )    | yes       | 95   |

**Table 75: Expanded uncertainty, laboratory test for NO, system 1**

| Measuring device:                    |  | N500                     |        | Serial-No.:         |      | 65                            |          |
|--------------------------------------|--|--------------------------|--------|---------------------|------|-------------------------------|----------|
| Measured component:                  |  | NO                       |        | 1h-limit value:     |      | 104.6 nmol/mol                |          |
| No.                                  | Performance characteristic   | Performance criterion    | Result | Partial uncertainty |      | Square of partial uncertainty |          |
| 1                                    | Repeatability standard deviation at zero                             | ≤ 1.0 nmol/mol           | 0.130  | $u_{r,z}$           | 0.02 | 0.0004                        |          |
| 2                                    | Repeatability standard deviation at 1h-limit value                   | ≤ 3.0 nmol/mol           | 1.630  | $u_{r,1h}$          | 0.05 | 0.0029                        |          |
| 3                                    | "lack of fit" at 1h-limit value                                      | ≤ 4.0% of measured value | 1.720  | $u_{l,1h}$          | 1.04 | 1.0789                        |          |
| 4                                    | Sensitivity coefficient of sample gas pressure at 1h-limit value     | ≤ 8.0 nmol/mol/kPa       | 0.230  | $u_{sp}$            | 0.58 | 0.3396                        |          |
| 5                                    | Sensitivity coefficient of sample gas temperature at 1h-limit value  | ≤ 3.0 nmol/mol/K         | 0.090  | $u_{gt}$            | 0.23 | 0.0527                        |          |
| 6                                    | Sensitivity coefficient of surrounding temperature at 1h-limit value | ≤ 3.0 nmol/mol/K         | 0.952  | $u_{st}$            | 2.43 | 5.9012                        |          |
| 7                                    | Sensitivity coefficient of electrical voltage at 1h-limit value      | ≤ 0.30 nmol/mol/V        | 0.010  | $u_v$               | 0.04 | 0.0015                        |          |
| 8a                                   | Interferent H <sub>2</sub> O with 19 mmol/mol                        | ≤ 10 nmol/mol (Zero)     | 0.000  | $u_{H_2O}$          | 0.06 | 0.0038                        |          |
|                                      |  | ≤ 10 nmol/mol (Span)     | 0.400  |                     |      |                               |          |
| 8b                                   | Interferent CO <sub>2</sub> with 500 µmol/mol                        | ≤ 5.0 nmol/mol (Zero)    | 0.400  | $u_{int,pos}$       | 0.82 | 0.6667                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)    | 1.800  |                     |      |                               |          |
| 8c                                   | Interferent NH <sub>3</sub> mit 200 nmol/mol                         | ≤ 5.0 nmol/mol (Zero)    | 0.600  | $u_{int,neg}$       | 0.82 | 0.6667                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)    | 1.200  |                     |      |                               |          |
| 9                                    | Averaging effect   | ≤ 7.0% of measured value | 3.100  | $u_{av}$            | 1.87 | 3.5048                        |          |
| 18                                   | Difference sample/calibration port                                   | ≤ 1.0%                   | 0.140  | $u_{s,c}$           | 0.15 | 0.0214                        |          |
| 21                                   | Converter efficiency   | ≥ 98                     | 99.60  | $u_{ec}$            | 0.42 | 0.1751                        |          |
| 23                                   | Uncertainty of test gas  | ≤ 3.0%                   | 2.000  | $u_{cg}$            | 1.05 | 1.0941                        |          |
| Combined standard uncertainty        |  |                          |        | $u_c$               |      | 3.5842                        | nmol/mol |
| Expanded uncertainty                 |  |                          |        | U                   |      | 7.1685                        | nmol/mol |
| Relative expanded uncertainty        |  |                          |        | W                   |      | 6.85                          | %        |
| Maximum allowed expanded uncertainty |  |                          |        | $W_{req}$           |      | 15                            | %        |

**Table 76: Expanded uncertainty, laboratory and field tests for NO, system 1**

| Measuring device:                    |  | N500                                  |        | Serial-No.:         |   | 65                            |          |
|--------------------------------------|--|---------------------------------------|--------|---------------------|---|-------------------------------|----------|
| Measured component:                  |  | NO                                    |        | 1h-limit value:     |   | 104.6 nmol/mol                |          |
| No.                                  | Performance characteristic   | Performance criterion                 | Result | Partial uncertainty |   | Square of partial uncertainty |          |
| 1                                    | Repeatability standard deviation at zero                             | ≤ 1.0 nmol/mol                        | 0.130  | $u_{r,z}$           | 0.02  | 0.0004                        |          |
| 2                                    | Repeatability standard deviation at 1h-limit value                   | ≤ 3.0 nmol/mol                        | 1.630  | $u_{r,1h}$          | not considered, as $\sqrt{2} \cdot u_{r,1h} = 0,07 < u_{r,f}$ | -                             |          |
| 3                                    | "lack of fit" at 1h-limit value                                      | ≤ 4.0% of measured value              | 1.720  | $u_{l,1h}$          | 1.04  | 1.0789                        |          |
| 4                                    | Sensitivity coefficient of sample gas pressure at 1h-limit value     | ≤ 8.0 nmol/mol/kPa                    | 0.230  | $u_{sp}$            | 0.58  | 0.3396                        |          |
| 5                                    | Sensitivity coefficient of sample gas temperature at 1h-limit value  | ≤ 3.0 nmol/mol/K                      | 0.090  | $u_{gt}$            | 0.23  | 0.0527                        |          |
| 6                                    | Sensitivity coefficient of surrounding temperature at 1h-limit value | ≤ 3.0 nmol/mol/K                      | 0.952  | $u_{st}$            | 2.43  | 5.9012                        |          |
| 7                                    | Sensitivity coefficient of electrical voltage at 1h-limit value      | ≤ 0.30 nmol/mol/V                     | 0.010  | $u_v$               | 0.04  | 0.0015                        |          |
| 8a                                   | Interferent H <sub>2</sub> O with 19 mmol/mol                        | ≤ 10 nmol/mol (Zero)                  | 0.000  | $u_{H_2O}$          | 0.06  | 0.0038                        |          |
|                                      |  | ≤ 10 nmol/mol (Span)                  | 0.400  |                     |   |                               |          |
| 8b                                   | Interferent CO <sub>2</sub> with 500 µmol/mol                        | ≤ 5.0 nmol/mol (Zero)                 | 0.400  | $u_{int,pos}$       | 0.82  | 0.6667                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)                 | 1.800  |                     |   |                               |          |
| 8c                                   | Interferent NH <sub>3</sub> mit 200 nmol/mol                         | ≤ 5.0 nmol/mol (Zero)                 | 0.600  | $u_{int,neg}$       | 0.82  | 0.6667                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)                 | 1.200  |                     |   |                               |          |
| 9                                    | Averaging effect   | ≤ 7.0% of measured value              | 3.100  | $u_{av}$            | 1.87  | 3.5048                        |          |
| 10                                   | Reproducibility standard deviation under field conditions            | ≤ 5.0% of average over 3 months       | 3.670  | $u_{r,f}$           | 3.84  | 14.7365                       |          |
| 11                                   | Long term drift at zero level  | ≤ 5.0 nmol/mol                        | -0.920 | $u_{d,z}$           | -0.53   | 0.2821                        |          |
| 12                                   | Long term drift at span level  | ≤ 5.0% of max. of certification range | 0.930  | $u_{d,1h}$          | 0.56  | 0.3154                        |          |
| 18                                   | Difference sample/calibration port                                   | ≤ 1.0%                                | 0.140  | $u_{s,c}$           | 0.15  | 0.0214                        |          |
| 21                                   | Converter efficiency   | ≥ 98                                  | 99.60  | $u_{ec}$            | 0.42  | 0.1751                        |          |
| 23                                   | Uncertainty of test gas  | ≤ 3.0%                                | 2.000  | $u_{cg}$            | 1.05  | 1.0941                        |          |
| Combined standard uncertainty        |  |                                       |        | $u_c$               |   | 5.3080                        | nmol/mol |
| Expanded uncertainty                 |  |                                       |        | U                   |   | 10.6160                       | nmol/mol |
| Relative expanded uncertainty        |  |                                       |        | W                   |   | 10.15                         | %        |
| Maximum allowed expanded uncertainty |  |                                       |        | $W_{req}$           |   | 15                            | %        |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO2 and NOx,  
Report No.: 936/21251100/A

Table 77: Expanded uncertainty, laboratory test for NO, system 2

| Measuring device:                    |  | N500                     |        | Serial-No.:          |       | 76                            |          |
|--------------------------------------|--|--------------------------|--------|----------------------|-------|-------------------------------|----------|
| Measured component:                  |  | NO                       |        | 1h-limit value:      |       | 104.6 nmol/mol                |          |
| No.                                  | Performance characteristic   | Performance criterion    | Result | Partial uncertainty  |       | Square of partial uncertainty |          |
| 1                                    | Repeatability standard deviation at zero                             | ≤ 1.0 nmol/mol           | 0.130  | u <sub>r,z</sub>     | 0.02  | 0.0005                        |          |
| 2                                    | Repeatability standard deviation at 1h-limit value                   | ≤ 3.0 nmol/mol           | 1.340  | u <sub>r,1h</sub>    | 0.05  | 0.0021                        |          |
| 3                                    | "lack of fit" at 1h-limit value                                      | ≤ 4.0% of measured value | 2.240  | u <sub>i,1h</sub>    | 1.35  | 1.8299                        |          |
| 4                                    | Sensitivity coefficient of sample gas pressure at 1h-limit value     | ≤ 8.0 nmol/mol/kPa       | 0.170  | u <sub>sp</sub>      | 0.43  | 0.1856                        |          |
| 5                                    | Sensitivity coefficient of sample gas temperature at 1h-limit value  | ≤ 3.0 nmol/mol/K         | 0.110  | u <sub>gt</sub>      | 0.28  | 0.0788                        |          |
| 6                                    | Sensitivity coefficient of surrounding temperature at 1h-limit value | ≤ 3.0 nmol/mol/K         | 0.451  | u <sub>st</sub>      | 1.15  | 1.3244                        |          |
| 7                                    | Sensitivity coefficient of electrical voltage at 1h-limit value      | ≤ 0.30 nmol/mol/V        | 0.000  | u <sub>v</sub>       | 0.00  | 0.0000                        |          |
| 8a                                   | Interferent H <sub>2</sub> O with 19 nmol/mol                        | ≤ 10 nmol/mol (Zero)     | -0.600 | u <sub>H2O</sub>     | -0.11 | 0.0116                        |          |
|                                      |  | ≤ 10 nmol/mol (Span)     | 1.600  |                      |       |                               |          |
| 8b                                   | Interferent CO <sub>2</sub> with 500 µmol/mol                        | ≤ 5.0 nmol/mol (Zero)    | 0.400  | u <sub>int,pos</sub> | 0.90  | 0.8171                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)    | 0.800  |                      |       |                               |          |
| 8c                                   | Interferent NH <sub>3</sub> mit 200 nmol/mol                         | ≤ 5.0 nmol/mol (Zero)    | 1.000  | or                   | 0.90  | 0.8171                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)    | 1.400  |                      |       |                               |          |
| 9                                    | Averaging effect   | ≤ 7.0% of measured value | -2.900 | u <sub>av</sub>      | -1.75 | 3.0672                        |          |
| 18                                   | Difference sample/calibration port                                   | ≤ 1.0%                   | 0.060  | u <sub>1,sc</sub>    | 0.06  | 0.0039                        |          |
| 21                                   | Converter efficiency   | ≥ 98                     | 99.80  | u <sub>EC</sub>      | 0.21  | 0.0438                        |          |
| 23                                   | Uncertainty of test gas  | ≤ 3.0%                   | 2.000  | u <sub>cg</sub>      | 1.05  | 1.0941                        |          |
| Combined standard uncertainty        |  |                          |        | u <sub>c</sub>       |       | 2.9089                        | nmol/mol |
| Expanded uncertainty                 |  |                          |        | U                    |       | 5.8177                        | nmol/mol |
| Relative expanded uncertainty        |  |                          |        | W                    |       | 5.56                          | %        |
| Maximum allowed expanded uncertainty |  |                          |        | W <sub>req</sub>     |       | 15                            | %        |

Table 78: Expanded uncertainty, laboratory and field tests for NO, system 2

| Measuring device:                    |  | N500                                  |        | Serial-No.:          |   | 76                            |          |
|--------------------------------------|--|---------------------------------------|--------|----------------------|---|-------------------------------|----------|
| Measured component:                  |  | NO                                    |        | 1h-limit value:      |   | 104.6 nmol/mol                |          |
| No.                                  | Performance characteristic   | Performance criterion                 | Result | Partial uncertainty  |   | Square of partial uncertainty |          |
| 1                                    | Repeatability standard deviation at zero                             | ≤ 1.0 nmol/mol                        | 0.130  | u <sub>r,z</sub>     | 0.02  | 0.0005                        |          |
| 2                                    | Repeatability standard deviation at 1h-limit value                   | ≤ 3.0 nmol/mol                        | 1.340  | u <sub>r,1h</sub>    | not considered, as $\sqrt{2} \cdot u_{r,1h} = 0,06 < u_{r,f}$ |                               | -        |
| 3                                    | "lack of fit" at 1h-limit value                                      | ≤ 4.0% of measured value              | 2.240  | u <sub>i,1h</sub>    | 1.35  | 1.8299                        |          |
| 4                                    | Sensitivity coefficient of sample gas pressure at 1h-limit value     | ≤ 8.0 nmol/mol/kPa                    | 0.170  | u <sub>sp</sub>      | 0.43  | 0.1856                        |          |
| 5                                    | Sensitivity coefficient of sample gas temperature at 1h-limit value  | ≤ 3.0 nmol/mol/K                      | 0.110  | u <sub>gt</sub>      | 0.28  | 0.0788                        |          |
| 6                                    | Sensitivity coefficient of surrounding temperature at 1h-limit value | ≤ 3.0 nmol/mol/K                      | 0.451  | u <sub>st</sub>      | 1.15  | 1.3244                        |          |
| 7                                    | Sensitivity coefficient of electrical voltage at 1h-limit value      | ≤ 0.30 nmol/mol/V                     | 0.000  | u <sub>v</sub>       | 0.00  | 0.0000                        |          |
| 8a                                   | Interferent H <sub>2</sub> O with 19 nmol/mol                        | ≤ 10 nmol/mol (Zero)                  | -0.600 | u <sub>H2O</sub>     | -0.11   | 0.0116                        |          |
|                                      |  | ≤ 10 nmol/mol (Span)                  | 1.600  |                      |   |                               |          |
| 8b                                   | Interferent CO <sub>2</sub> with 500 µmol/mol                        | ≤ 5.0 nmol/mol (Zero)                 | 0.400  | u <sub>int,pos</sub> | 0.90  | 0.8171                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)                 | 0.800  |                      |   |                               |          |
| 8c                                   | Interferent NH <sub>3</sub> mit 200 nmol/mol                         | ≤ 5.0 nmol/mol (Zero)                 | 1.000  | or                   | 0.90  | 0.8171                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)                 | 1.400  |                      |   |                               |          |
| 9                                    | Averaging effect   | ≤ 7.0% of measured value              | -2.900 | u <sub>av</sub>      | -1.75   | 3.0672                        |          |
| 10                                   | Reproducibility standard deviation under field conditions            | ≤ 5.0% of average over 3 months       | 3.670  | u <sub>r,f</sub>     | 3.84  | 14.7365                       |          |
| 11                                   | Long term drift at zero level  | ≤ 5.0 nmol/mol                        | 1.080  | u <sub>d,z,z</sub>   | 0.62  | 0.3888                        |          |
| 12                                   | Long term drift at span level  | ≤ 5.0% of max. of certification range | 1.020  | u <sub>d,1,1h</sub>  | 0.62  | 0.3794                        |          |
| 18                                   | Difference sample/calibration port                                   | ≤ 1.0%                                | 0.060  | u <sub>1,sc</sub>    | 0.06  | 0.0039                        |          |
| 21                                   | Converter efficiency   | ≥ 98                                  | 99.80  | u <sub>EC</sub>      | 0.21  | 0.0438                        |          |
| 23                                   | Uncertainty of test gas  | ≤ 3.0%                                | 2.000  | u <sub>cg</sub>      | 1.05  | 1.0941                        |          |
| Combined standard uncertainty        |  |                                       |        | u <sub>c</sub>       |   | 4.8951                        | nmol/mol |
| Expanded uncertainty                 |  |                                       |        | U                    |   | 9.7902                        | nmol/mol |
| Relative expanded uncertainty        |  |                                       |        | W                    |   | 9.36                          | %        |
| Maximum allowed expanded uncertainty |  |                                       |        | W <sub>req</sub>     |   | 15                            | %        |

**Table 79: Expanded uncertainty, laboratory test for NO<sub>2</sub>, system 1**

| Measuring device:                    |  | N500                     |        | Serial-No.:          |       | 65                            |          |
|--------------------------------------|--|--------------------------|--------|----------------------|-------|-------------------------------|----------|
| Measured component:                  |  | NO <sub>2</sub>          |        | 1h-limit value:      |       | 104.6 nmol/mol                |          |
| No.                                  | Performance characteristic   | Performance criterion    | Result | Partial uncertainty  |       | Square of partial uncertainty |          |
| 1                                    | Repeatability standard deviation at zero                             | ≤ 1.0 nmol/mol           | 0.000  | u <sub>r,z</sub>     | 0.00  | 0.0000                        |          |
| 2                                    | Repeatability standard deviation at 1h-limit value                   | ≤ 3.0 nmol/mol           | 0.080  | u <sub>r,1h</sub>    | 0.01  | 0.0002                        |          |
| 3                                    | "lack of fit" at 1h-limit value                                      | ≤ 4.0% of measured value | 0.850  | u <sub>l,1h</sub>    | 0.51  | 0.2635                        |          |
| 4                                    | Sensitivity coefficient of sample gas pressure at 1h-limit value     | ≤ 8.0 nmol/mol/kPa       | 0.100  | u <sub>sp</sub>      | 0.91  | 0.8206                        |          |
| 5                                    | Sensitivity coefficient of sample gas temperature at 1h-limit value  | ≤ 3.0 nmol/mol/K         | 0.010  | u <sub>st</sub>      | 0.09  | 0.0082                        |          |
| 6                                    | Sensitivity coefficient of surrounding temperature at 1h-limit value | ≤ 3.0 nmol/mol/K         | 0.231  | u <sub>st</sub>      | 2.09  | 4.3787                        |          |
| 7                                    | Sensitivity coefficient of electrical voltage at 1h-limit value      | ≤ 0.30 nmol/mol/V        | 0.020  | u <sub>v</sub>       | 0.27  | 0.0727                        |          |
| 8a                                   | Interferent H <sub>2</sub> O with 19 nmol/mol                        | ≤ 10 nmol/mol (Zero)     | 0.270  | u <sub>H2O</sub>     | 0.25  | 0.0606                        |          |
|                                      |  | ≤ 10 nmol/mol (Span)     | 0.330  |                      |       |                               |          |
| 8b                                   | Interferent CO <sub>2</sub> with 500 µmol/mol                        | ≤ 5.0 nmol/mol (Zero)    | 0.000  | u <sub>int,pos</sub> | 1.03  | 1.0599                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)    | 0.700  |                      |       |                               |          |
| 8c                                   | Interferent NH <sub>3</sub> mit 200 nmol/mol                         | ≤ 5.0 nmol/mol (Zero)    | 0.000  | u <sub>int,neg</sub> | 1.03  | 1.0599                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)    | 1.090  |                      |       |                               |          |
| 9                                    | Averaging effect   | ≤ 7.0% of measured value | 0.700  | u <sub>av</sub>      | 0.42  | 0.1787                        |          |
| 18                                   | Difference sample/calibration port                                   | ≤ 1.0%                   | -0.140 | u <sub>ASC</sub>     | -0.15 | 0.0214                        |          |
| 21                                   | Converter efficiency   | ≥ 98                     | 99.60  | u <sub>EC</sub>      | 0.42  | 0.1751                        |          |
| 23                                   | Uncertainty of test gas  | ≤ 3.0%                   | 2.000  | u <sub>CG</sub>      | 1.05  | 1.0941                        |          |
| Combined standard uncertainty        |  |                          |        | u <sub>c</sub>       |       | 2.8520                        | nmol/mol |
| Expanded uncertainty                 |  |                          |        | U                    |       | 5.7040                        | nmol/mol |
| Relative expanded uncertainty        |  |                          |        | W                    |       | 5.45                          | %        |
| Maximum allowed expanded uncertainty |  |                          |        | W <sub>req</sub>     |       | 15                            | %        |

**Table 80: Expanded uncertainty, laboratory and field tests for NO<sub>2</sub>, system 1**

| Measuring device:                    |  | N500                                  |        | Serial-No.:          |   | 65                            |          |
|--------------------------------------|--|---------------------------------------|--------|----------------------|---|-------------------------------|----------|
| Measured component:                  |  | NO <sub>2</sub>                       |        | 1h-limit value:      |   | 104.6 nmol/mol                |          |
| No.                                  | Performance characteristic   | Performance criterion                 | Result | Partial uncertainty  |   | Square of partial uncertainty |          |
| 1                                    | Repeatability standard deviation at zero                             | ≤ 1.0 nmol/mol                        | 0.000  | u <sub>r,z</sub>     | 0.00  | 0.0000                        |          |
| 2                                    | Repeatability standard deviation at 1h-limit value                   | ≤ 3.0 nmol/mol                        | 0.080  | u <sub>r,1h</sub>    | not considered, as $\sqrt{2} \cdot u_{r,1h} = 0,01 < u_{r,f}$ | -                             |          |
| 3                                    | "lack of fit" at 1h-limit value                                      | ≤ 4.0% of measured value              | 0.850  | u <sub>l,1h</sub>    | 0.51  | 0.2635                        |          |
| 4                                    | Sensitivity coefficient of sample gas pressure at 1h-limit value     | ≤ 8.0 nmol/mol/kPa                    | 0.100  | u <sub>sp</sub>      | 0.91  | 0.8206                        |          |
| 5                                    | Sensitivity coefficient of sample gas temperature at 1h-limit value  | ≤ 3.0 nmol/mol/K                      | 0.010  | u <sub>st</sub>      | 0.09  | 0.0082                        |          |
| 6                                    | Sensitivity coefficient of surrounding temperature at 1h-limit value | ≤ 3.0 nmol/mol/K                      | 0.231  | u <sub>st</sub>      | 2.09  | 4.3787                        |          |
| 7                                    | Sensitivity coefficient of electrical voltage at 1h-limit value      | ≤ 0.30 nmol/mol/V                     | 0.020  | u <sub>v</sub>       | 0.27  | 0.0727                        |          |
| 8a                                   | Interferent H <sub>2</sub> O with 19 nmol/mol                        | ≤ 10 nmol/mol (Zero)                  | 0.270  | u <sub>H2O</sub>     | 0.25  | 0.0606                        |          |
|                                      |  | ≤ 10 nmol/mol (Span)                  | 0.330  |                      |   |                               |          |
| 8b                                   | Interferent CO <sub>2</sub> with 500 µmol/mol                        | ≤ 5.0 nmol/mol (Zero)                 | 0.000  | u <sub>int,pos</sub> | 1.03  | 1.0599                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)                 | 0.700  |                      |   |                               |          |
| 8c                                   | Interferent NH <sub>3</sub> mit 200 nmol/mol                         | ≤ 5.0 nmol/mol (Zero)                 | 0.000  | u <sub>int,neg</sub> | 1.03  | 1.0599                        |          |
|                                      |  | ≤ 5.0 nmol/mol (Span)                 | 1.090  |                      |   |                               |          |
| 9                                    | Averaging effect   | ≤ 7.0% of measured value              | 0.700  | u <sub>av</sub>      | 0.42  | 0.1787                        |          |
| 10                                   | Reproducibility standard deviation under field conditions            | ≤ 5.0% of average over 3 months       | 3.670  | u <sub>r,f</sub>     | 3.84  | 14.7365                       |          |
| 11                                   | Long term drift at zero level  | ≤ 5.0 nmol/mol                        | 0.750  | u <sub>d,l,z</sub>   | 0.43  | 0.1875                        |          |
| 12                                   | Long term drift at span level  | ≤ 5.0% of max. of certification range | 2.290  | u <sub>d,l,1h</sub>  | 1.38  | 1.9126                        |          |
| 18                                   | Difference sample/calibration port                                   | ≤ 1.0%                                | -0.140 | u <sub>ASC</sub>     | -0.15   | 0.0214                        |          |
| 21                                   | Converter efficiency   | ≥ 98                                  | 99.60  | u <sub>EC</sub>      | 0.42  | 0.1751                        |          |
| 23                                   | Uncertainty of test gas  | ≤ 3.0%                                | 2.000  | u <sub>CG</sub>      | 1.05  | 1.0941                        |          |
| Combined standard uncertainty        |  |                                       |        | u <sub>c</sub>       |   | 4.9970                        | nmol/mol |
| Expanded uncertainty                 |  |                                       |        | U                    |   | 9.9940                        | nmol/mol |
| Relative expanded uncertainty        |  |                                       |        | W                    |   | 9.55                          | %        |
| Maximum allowed expanded uncertainty |  |                                       |        | W <sub>req</sub>     |   | 15                            | %        |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Table 81: Expanded uncertainty, laboratory test for NO<sub>2</sub>, system 2

| Measuring device:                    |  | N500                     |        | Serial-No.:          |        | 76                            |  |
|--------------------------------------|--|--------------------------|--------|----------------------|--------|-------------------------------|--|
| Measured component:                  |  | NO <sub>2</sub>          |        | 1h-limit value:      |        | 104.6 nmol/mol                |  |
| No.                                  | Performance characteristic   | Performance criterion    | Result | Partial uncertainty  |        | Square of partial uncertainty |  |
| 1                                    | Repeatability standard deviation at zero                             | ≤ 1.0 nmol/mol           | 0.110  | u <sub>r,z</sub>     | 0.02   | 0.0003                        |  |
| 2                                    | Repeatability standard deviation at 1h-limit value                   | ≤ 3.0 nmol/mol           | 0.080  | u <sub>r,1h</sub>    | 0.01   | 0.0002                        |  |
| 3                                    | "lack of fit" at 1h-limit value                                      | ≤ 4.0% of measured value | 0.720  | u <sub>l,1h</sub>    | 0.43   | 0.1891                        |  |
| 4                                    | Sensitivity coefficient of sample gas pressure at 1h-limit value     | ≤ 8.0 nmol/mol/kPa       | 0.080  | u <sub>sp</sub>      | 0.72   | 0.5252                        |  |
| 5                                    | Sensitivity coefficient of sample gas temperature at 1h-limit value  | ≤ 3.0 nmol/mol/K         | 0.010  | u <sub>gt</sub>      | 0.09   | 0.0082                        |  |
| 6                                    | Sensitivity coefficient of surrounding temperature at 1h-limit value | ≤ 3.0 nmol/mol/K         | 0.241  | u <sub>st</sub>      | 2.18   | 4.7661                        |  |
| 7                                    | Sensitivity coefficient of electrical voltage at 1h-limit value      | ≤ 0.30 nmol/mol/V        | 0.010  | u <sub>v</sub>       | 0.13   | 0.0182                        |  |
| 8a                                   | Interferent H <sub>2</sub> O with 19 nmol/mol                        | ≤ 10 nmol/mol (Zero)     | 0.000  | u <sub>H2O</sub>     | 0.16   | 0.0267                        |  |
|                                      |  | ≤ 10 nmol/mol (Span)     | 0.220  |                      |        |                               |  |
| 8b                                   | Interferent CO <sub>2</sub> with 500 µmol/mol                        | ≤ 5.0 nmol/mol (Zero)    | -0.210 | u <sub>int,pos</sub> | 1.00   | 0.9999                        |  |
|                                      |  | ≤ 5.0 nmol/mol (Span)    | 0.760  |                      |        |                               |  |
| 8c                                   | Interferent NH <sub>3</sub> mit 200 nmol/mol                         | ≤ 5.0 nmol/mol (Zero)    | -0.160 | or                   | 1.00   | 0.9999                        |  |
|                                      |  | ≤ 5.0 nmol/mol (Span)    | 0.980  |                      |        |                               |  |
| 9                                    | Averaging effect   | ≤ 7.0% of measured value | 3.200  | u <sub>av</sub>      | 1.93   | 3.7346                        |  |
| 18                                   | Difference sample/calibration port                                   | ≤ 1.0%                   | -0.330 | u <sub>18c</sub>     | -0.35  | 0.1191                        |  |
| 21                                   | Converter efficiency   | ≥ 98                     | 99.800 | u <sub>EC</sub>      | 0.21   | 0.0438                        |  |
| 23                                   | Uncertainty of test gas  | ≤ 3.0%                   | 2.000  | u <sub>CG</sub>      | 1.05   | 1.0941                        |  |
| Combined standard uncertainty        |  |                          |        | u <sub>c</sub>       | 3.3950 | nmol/mol                      |  |
| Expanded uncertainty                 |  |                          |        | U                    | 6.7900 | nmol/mol                      |  |
| Relative expanded uncertainty        |  |                          |        | W                    | 6.49   | %                             |  |
| Maximum allowed expanded uncertainty |  |                          |        | W <sub>req</sub>     | 15     | %                             |  |

Table 82: Expanded uncertainty, laboratory and field tests for NO<sub>2</sub>, system 2

| Measuring device:                    |  | N500                                  |        | Serial-No.:          |   | 76                            |   |
|--------------------------------------|--|---------------------------------------|--------|----------------------|---|-------------------------------|---|
| Measured component:                  |  | NO <sub>2</sub>                       |        | 1h-limit value:      |   | 104.6 nmol/mol                |   |
| No.                                  | Performance characteristic   | Performance criterion                 | Result | Partial uncertainty  |   | Square of partial uncertainty |   |
| 1                                    | Repeatability standard deviation at zero                             | ≤ 1.0 nmol/mol                        | 0.110  | u <sub>r,z</sub>     | 0.02  | 0.0003                        |   |
| 2                                    | Repeatability standard deviation at 1h-limit value                   | ≤ 3.0 nmol/mol                        | 0.080  | u <sub>r,1h</sub>    | not considered, as $\sqrt{2} \cdot u_{r,1h} = 0,01 < u_{r,f}$ |                               | - |
| 3                                    | "lack of fit" at 1h-limit value                                      | ≤ 4.0% of measured value              | 0.720  | u <sub>l,1h</sub>    | 0.43  | 0.1891                        |   |
| 4                                    | Sensitivity coefficient of sample gas pressure at 1h-limit value     | ≤ 8.0 nmol/mol/kPa                    | 0.080  | u <sub>sp</sub>      | 0.72  | 0.5252                        |   |
| 5                                    | Sensitivity coefficient of sample gas temperature at 1h-limit value  | ≤ 3.0 nmol/mol/K                      | 0.010  | u <sub>gt</sub>      | 0.09  | 0.0082                        |   |
| 6                                    | Sensitivity coefficient of surrounding temperature at 1h-limit value | ≤ 3.0 nmol/mol/K                      | 0.241  | u <sub>st</sub>      | 2.18  | 4.7661                        |   |
| 7                                    | Sensitivity coefficient of electrical voltage at 1h-limit value      | ≤ 0.30 nmol/mol/V                     | 0.010  | u <sub>v</sub>       | 0.13  | 0.0182                        |   |
| 8a                                   | Interferent H <sub>2</sub> O with 19 nmol/mol                        | ≤ 10 nmol/mol (Zero)                  | 0.000  | u <sub>H2O</sub>     | 0.16  | 0.0267                        |   |
|                                      |  | ≤ 10 nmol/mol (Span)                  | 0.220  |                      |   |                               |   |
| 8b                                   | Interferent CO <sub>2</sub> with 500 µmol/mol                        | ≤ 5.0 nmol/mol (Zero)                 | -0.210 | u <sub>int,pos</sub> | 1.00  | 0.9999                        |   |
|                                      |  | ≤ 5.0 nmol/mol (Span)                 | 0.760  |                      |   |                               |   |
| 8c                                   | Interferent NH <sub>3</sub> mit 200 nmol/mol                         | ≤ 5.0 nmol/mol (Zero)                 | -0.160 | or                   | 1.00  | 0.9999                        |   |
|                                      |  | ≤ 5.0 nmol/mol (Span)                 | 0.980  |                      |   |                               |   |
| 9                                    | Averaging effect   | ≤ 7.0% of measured value              | 3.200  | u <sub>av</sub>      | 1.93  | 3.7346                        |   |
| 10                                   | Reproducibility standard deviation under field conditions            | ≤ 5.0% of average over 3 months       | 3.670  | u <sub>r,f</sub>     | 3.84  | 14.7365                       |   |
| 11                                   | Long term drift at zero level  | ≤ 5.0 nmol/mol                        | 0.870  | u <sub>d,z</sub>     | 0.50  | 0.2523                        |   |
| 12                                   | Long term drift at span level  | ≤ 5.0% of max. of certification range | 2.050  | u <sub>d,1h</sub>    | 1.24  | 1.5327                        |   |
| 18                                   | Difference sample/calibration port                                   | ≤ 1.0%                                | -0.330 | u <sub>18c</sub>     | -0.35   | 0.1191                        |   |
| 21                                   | Converter efficiency   | ≥ 98                                  | 99.800 | u <sub>EC</sub>      | 0.21  | 0.0438                        |   |
| 23                                   | Uncertainty of test gas  | ≤ 3.0%                                | 2.000  | u <sub>CG</sub>      | 1.05  | 1.0941                        |   |
| Combined standard uncertainty        |  |                                       |        | u <sub>c</sub>       | 5.2959  | nmol/mol                      |   |
| Expanded uncertainty                 |  |                                       |        | U                    | 10.5919   | nmol/mol                      |   |
| Relative expanded uncertainty        |  |                                       |        | W                    | 10.13   | %                             |   |
| Maximum allowed expanded uncertainty |  |                                       |        | W <sub>req</sub>     | 15  | %                             |   |

## 8. Recommendations for use in practice

### Work in the maintenance interval

The tested measuring systems require regular performance of the following tasks:

- Visual inspections/telemetric inspections on a regular basis
- Checking the status of the system
- No error messages
- Performing zero and span checks using suitable test gas every two weeks in accordance with standard EN 14211;

Other than that, follow the manufacturer's instructions indicated in the user manual.

Environmental Protection/Air Pollution Control



---

Dipl.-Ing. Martin Schneider



---

Dipl.-Ing. Guido Baum

Cologne, 30 July 2021  
936/21251100/A



## 9. Bibliography

- [1] VDI 4202 part 1: Performance test, declaration of suitability, and certification of point-related measuring systems for gaseous air pollutants of April 2018
- [2] European standard EN 14211 Ambient air - Standard method for the measurement of the concentration of nitrogen dioxide and nitrogen monoxide by chemiluminescence, August 2012
- [3] Directive 2008/50/EG of the European Parliament and of the Council of 21 May 2008 on ambient air quality and cleaner air for Europe.
- [4] Guide, Demonstration of Equivalence of Ambient Air Monitoring Methods, English version dated January 2010

## Annexes

# Annex 1:

## Additional test criteria specified in the Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods

The field test which was carried out according to VDI 4202-1 and EN 14211 was carried out from 09 November 2020 to 15 February 2021.

However, the entire field setup was installed from November 2020 until the end of June 2021. During the entire period, a performance-tested CLD reference analyser of the type Horiba APNA 370 (SN 43286610022) was running in parallel to the N500 systems. This reference analyser was connected to the same sampling mast as the two test instruments. The length of the test gas line to the analyser was approx. 2 m.

With this system, additional tests were carried out according to the guide Demonstration of Equivalence of Ambient Air Monitoring Methods.

To this effect, four measurement periods, each lasting one month, were selected over a period of six months to compare the N500 systems with an already performance-tested SRM system. The measurement periods were selected in such a way that different environmental conditions (temperature, humidity, ambient pressure, NO<sub>2</sub> concentration) could be included in the evaluation of the measuring system. Therefore, the months of December 2020, January 2021, May 2021 and June 2021 were evaluated in accordance with the guide Demonstration of Equivalence of Ambient Air Monitoring Methods.

Before the start of each evaluated period, the adjustment of the test instruments and the reference system was checked. In the case of the Horiba APNA 370 reference analyser, the converter efficiency was also regularly checked and remained at > 98 % throughout the entire test.

The assessment was carried out separately for NO and NO<sub>2</sub>.

## 7.1 Determination of relative uncertainty between test instruments $W_{bs}$ [8.5.3.2]

*The relative uncertainty between the test items  $W_{bs}$  shall be determined in accordance with point 8.5.3.2 of the guidance document "Demonstration of Equivalence of Ambient Air Monitoring Methods".*

## 7.2 Equipment

Not required for this performance criterion.

## 7.3 Testing

The test was conducted in the field test in four different months (December, January, May and June) at the field test site in Cologne. Different seasons and different test gas concentrations were taken into account.

## 7.4 Evaluation

According to point 8.5.3.2 of the Demonstration of Equivalence of Ambient Air Monitoring Methods guide:

The uncertainty between the test items  $W_{bs}$  must be  $\leq 5\%$ . As no other reference value is given, the 1 hour limit value for NO<sub>2</sub> (200  $\mu\text{g}/\text{m}^3 = 104.6$  ppb) (2008/50/EC) was used as the reference value. Thus, the maximum uncertainty between the test items may be a maximum of 5.2 ppb for NO<sub>2</sub>.

For the component NO, the  $c_t$  value for NO (631.3  $\mu\text{g}/\text{m}^3 = 505$  ppb) was used as a criterion. This means that the maximum uncertainty between the test items for NO must not exceed 25 ppb.

An uncertainty above 5.2 ppb or 25 ppb between the two test items is an indication that the performance of one or both systems is inadequate and equivalence cannot be explained.

The uncertainty between the test items is determined individually for each test month.

The uncertainty between the test items  $W_{bs}$  is calculated from the differences of all 1 hour mean values of the test specimens operated in parallel according to the following equation:

$$W_{bs}^2 = \frac{\sum_{i=1}^n (y_{i,1} - y_{i,2})^2}{2ny^2}$$

Where:  $y_{i,1}$  and  $y_{i,2}$  = Results of the parallel measurements of individual 1h-values  $i$   
 $y$  = Average of the test item readings  
 $n$  = Number of 1h values

## 7.5 Assessment

The uncertainty between the test items  $W_{bs}$  remained below the required maximum values with a maximum of 1.440 ppb for NO and 1.846 ppb for NO<sub>2</sub>.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

Table 83 and Table 84 list the calculated values for the uncertainties  $W_{bs}$  between test items. Corresponding charts are provided in Figure 18 to Figure 25.

Table 83: Uncertainty between test systems ( $W_{bs}$ ), NO

| Tested instruments | Measurement month | Number of measurements | Uncertainty $W_{bs}$ |
|--------------------|-------------------|------------------------|----------------------|
| Component NO       |                   |                        |                      |
| SN 65 / SN 76      | December          | 744                    | 0.063                |
| SN 65 / SN 76      | January           | 744                    | 0.003                |
| SN 65 / SN 76      | May               | 744                    | 0.063                |
| SN 65 / SN 76      | June              | 720                    | 0.063                |

Table 84: Uncertainty between test systems ( $W_{bs}$ ), NO<sub>2</sub>

| Tested instruments        | Measurement month | Number of measurements | Uncertainty $W_{bs}$ |
|---------------------------|-------------------|------------------------|----------------------|
| Component NO <sub>2</sub> |                   |                        |                      |
| SN 65 / SN 76             | December          | 744                    | 0.022                |
| SN 65 / SN 76             | January           | 744                    | 0.018                |
| SN 65 / SN 76             | May               | 744                    | 0.021                |
| SN 65 / SN 76             | June              | 720                    | 0.020                |




Figure 18: Parallel measurements with the test systems, December, NO




Figure 19: Parallel measurements with the test systems, January, NO




Figure 20: Parallel measurements with the test systems, May, NO




Figure 21: Parallel measurements with the test systems, June, NO




Figure 22: Parallel measurements with the test systems, December, NO<sub>2</sub>




Figure 23: Parallel measurements with the test systems, January, NO<sub>2</sub>




Figure 24: Parallel measurements with the test systems, May, NO<sub>2</sub>




Figure 25: Parallel measurements with the test systems, June, NO<sub>2</sub>



## 7.1 Comparison with the standard reference method [8.5.3.3]

*The equivalence of the test items to the reference method shall be demonstrated in accordance with point 8.5.3.3 of the Demonstration of Equivalence of Ambient Air Monitoring Methods guide. The highest calculated expanded uncertainty of the test items is to be compared with the requirements for the data quality objectives according to the EU Directive [3].*

## 7.2 Equipment

Additional equipment as described in chapter 4 of this report was used for this test.

## 7.3 Testing

The test was conducted in the field test in four different months (December, January, May and June) at the field test site in Cologne. Different seasons and different test gas concentrations were taken into account.

Four measurements were carried out, each lasting one month. The measurements were spread over half a year. The concentrations measured were related to the ambient conditions.

## 7.4 Evaluation

In order to assess comparability of the tested instruments  $y$  with the reference method  $x$ , a linear relationship  $y_i = a + bx_i$  between the measured values of both methods is assumed. The correlation between the mean values of the reference instruments and the test instruments is established by means of orthogonal regression.

For analysis, the result uncertainty  $w_{c-s}$  of the test samples from the comparison with the reference method is described according to the following equation, which describes  $w_{c-s}$  as a function of the sample gas concentration  $x_i$ .

$$w_{c-s}^2(y_i) = \frac{RSS}{(n-2)} - u^2(x_i) + [a + (b-1)x_i]^2$$

Where RSS = is the sum of the (relative) residuals from orthogonal regression

$u(x_i)$  = is the random uncertainty of the reference method

The algorithms for calculating axis intercept  $a$  and slope  $b$  as well as their variance by means of orthogonal regression are described in detail in the annex to [4].

The sum of (relative) residuals RSS is calculated according to the following equation:

$$RSS = \sum_{i=1}^n (y_i - a - bx_i)^2$$

[Item 8.5.3] For all data sets, the combined uncertainty of the test items  $w_{c,CM}$  is calculated by combining the contributions from 8.5.3.1 and 8.5.3.2 according to the following equation:

$$w_{c,CM}^2(y_i) = \frac{u_{c-s}^2(y_i)}{y_i^2}$$

For each data set the uncertainty  $w_{c,CM}$  at the 1 h limit value for NO<sub>2</sub> (here 200 µg/m<sup>3</sup> = 104.6 ppb) is calculated, where  $y_i$  is used as the concentration at the limit value. For NO, the uncertainty is based on the 1 h limit value (631.3 µg/m<sup>3</sup> = 505 ppb).

[point 8.5.3.4] For each data set, the expanded relative uncertainty of the results of the test items is calculated by multiplying  $w_{c,CM}$  by a coverage factor  $k$  according to the following equation:

$$W_{CM} = k \cdot w_{CM}$$

In practice,  $k$  is specified at  $k=2$  for large  $n$ .

The largest resulting uncertainty  $W_{CM}$  is compared and evaluated with the data quality requirements for air quality measurements according to the EU Directive. The established expanded relative uncertainty  $W_{cm}$  was at 15% for NO and NO<sub>2</sub>

## 7.5 Assessment

The determined uncertainties  $W_{CM}$  are below the defined expanded relative uncertainty  $W_{cm}$  of 15 % without applying correction factors for all data sets considered.

Criterion satisfied? yes

## 7.6 Detailed presentation of test results

In Table 85 and in Table 86 a summary presentation and evaluation of the expanded measurement uncertainties  $W_{CM}$  from the field test investigations are given.

Table 87 to Table 94 show the results of the evaluations of the individual data sets.

Table 85: Expanded measurement uncertainties  $W_{CM}$  from the field investigations for NO

| Month    | Component | Limit value<br>ppb | Slope b<br>(ppb)/(ppb) | Axis intercept a<br>ppb | $u_{c,s}$ at the limit value<br>ppb | $w_{CM}$<br>% | $W_{CM}$<br>% | $W_{CM} \leq W_{dqo}$<br>( $W_{dqo} = 15\%$ ) |
|----------|-----------|--------------------|------------------------|-------------------------|-------------------------------------|---------------|---------------|---|
| December | NO        | 505                | 0.99                   | -0.43                   | 6.59                                | 1.30          | 2.61          | yes   |
| January  | NO        | 505                | 1.00                   | -0.94                   | 0.93                                | 0.18          | 0.37          | yes   |
| May      | NO        | 505                | 0.98                   | -0.71                   | 9.33                                | 1.85          | 3.69          | yes   |
| June     | NO        | 505                | 0.97                   | -1.35                   | 18.64                               | 3.69          | 7.38          | yes   |

Table 86: Expanded measurement uncertainties  $W_{CM}$  from the field investigations for NO<sub>2</sub>

| Month    | Component       | Limit value<br>ppb | Slope b<br>(ppb)/(ppb) | Axis intercept a<br>ppb | $u_{c,s}$ at the limit value<br>ppb | $w_{CM}$<br>% | $W_{CM}$<br>% | $W_{CM} \leq W_{dqo}$<br>( $W_{dqo} = 15\%$ ) |
|----------|-----------------|--------------------|------------------------|-------------------------|-------------------------------------|---------------|---------------|---|
| December | NO <sub>2</sub> | 104.6              | 1.00                   | 0.00                    | 0.25                                | 0.29          | 0.57          | yes   |
| January  | NO <sub>2</sub> | 104.6              | 1.00                   | -0.04                   | 0.48                                | 0.46          | 0.91          | yes   |
| May      | NO <sub>2</sub> | 104.6              | 1.00                   | -1.63                   | 1.47                                | 1.41          | 2.81          | yes   |
| June     | NO <sub>2</sub> | 104.6              | 1.00                   | -1.67                   | 1.82                                | 1.74          | 3.48          | yes   |

Table 87: Comparison of the tested and reference instruments, December, NO

| Comparison candidate with reference according to guidance "Demonstration of Equivalence Of Ambient Air Monitoring Methods" |                   |                     |               |                    |
|--|-------------------|---------------------|---------------|--------------------|
| Device   | N500              | SN                  | SN 65 / SN 76 |                    |
| Testside / month   | Cologne, December | Limit value         | 505           | ppb                |
| Component  | NO                | allowed uncertainty | 15            | %                  |
| Results of the regression analysis   |                   |                     |               |                    |
| Slope b  | <b>0.99</b>       |                     |               | <b>significant</b> |
| Uncertainty of b   | <b>0.00</b>       |                     |               |                    |
| Ordinate intercept a   | <b>-0.43</b>      |                     |               | <b>significant</b> |
| Uncertainty of a   | <b>0.03</b>       |                     |               |                    |
| Results of the equivalence test  |                   |                     |               |                    |
| Deviation at limit value   | <b>-6.55</b>      |                     |               | <b>ppb</b>         |
| Uncertainty $u_{c_s}$ at limit value   | <b>6.59</b>       |                     |               | <b>ppb</b>         |
| Combined measurement uncertainty $w_{CM}$  | <b>1.30</b>       |                     |               | <b>%</b>           |
| Expanded uncertainty $W_{CM}$  | <b>2.61</b>       |                     |               | <b>%</b>           |
| Status equivalence test  | <b>passed</b>     |                     |               |                    |

Table 88: Comparison of the tested and reference instruments, January, NO

| Comparison candidate with reference according to guidance "Demonstration of Equivalence Of Ambient Air Monitoring Methods" |                  |                     |               |                    |
|--|------------------|---------------------|---------------|--------------------|
| Device   | N500             | SN                  | SN 65 / SN 76 |                    |
| Testside / month   | Cologne, January | Limit value         | 505           | ppb                |
| Component  | NO               | allowed uncertainty | 15            | %                  |
| Results of the regression analysis   |                  |                     |               |                    |
| Slope b  | <b>1.00</b>      |                     |               | <b>significant</b> |
| Uncertainty of b   | <b>0.00</b>      |                     |               |                    |
| Ordinate intercept a   | <b>-0.94</b>     |                     |               | <b>significant</b> |
| Uncertainty of a   | <b>0.03</b>      |                     |               |                    |
| Results of the equivalence test  |                  |                     |               |                    |
| Deviation at limit value   | <b>0.54</b>      |                     |               | <b>ppb</b>         |
| Uncertainty $u_{c_s}$ at limit value   | <b>0.93</b>      |                     |               | <b>ppb</b>         |
| Combined measurement uncertainty $w_{CM}$  | <b>0.18</b>      |                     |               | <b>%</b>           |
| Expanded uncertainty $W_{CM}$  | <b>0.37</b>      |                     |               | <b>%</b>           |
| Status equivalence test  | <b>passed</b>    |                     |               |                    |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Table 89: Comparison of the tested and reference instruments, May, NO

| Comparison candidate with reference according to guidance "Demonstration of Equivalence Of Ambient Air Monitoring Methods" |               |                     |               |                    |
|--|---------------|---------------------|---------------|--------------------|
| Device   | N500          | SN                  | SN 65 / SN 76 |                    |
| Testside / month   | Cologne, May  | Limit value         | 505           | ppb                |
| Component  | NO            | allowed uncertainty | 15            | %                  |
| Results of the regression analysis   |               |                     |               |                    |
| Slope b  | <b>0.98</b>   |                     |               | <b>significant</b> |
| Uncertainty of b   | <b>0.00</b>   |                     |               |                    |
| Ordinate intercept a   | <b>-0.71</b>  |                     |               | <b>significant</b> |
| Uncertainty of a   | <b>0.05</b>   |                     |               |                    |
| Results of the equivalence test  |               |                     |               |                    |
| Deviation at limit value   | <b>-9.28</b>  |                     |               | <b>ppb</b>         |
| Uncertainty $u_{c,s}$ at limit value   | <b>9.33</b>   |                     |               | <b>ppb</b>         |
| Combined measurement uncertainty $w_{CM}$  | <b>1.85</b>   |                     |               | <b>%</b>           |
| Expanded uncertainty $W_{CM}$  | <b>3.69</b>   |                     |               | <b>%</b>           |
| Status equivalence test  | <b>passed</b> |                     |               |                    |

Table 90: Comparison of the tested and reference instruments, June, NO

| Comparison candidate with reference according to guidance "Demonstration of Equivalence Of Ambient Air Monitoring Methods" |               |                     |               |                    |
|--|---------------|---------------------|---------------|--------------------|
| Device   | N500          | SN                  | SN 65 / SN 76 |                    |
| Testside / month   | Cologne, June | Limit value         | 505           | ppb                |
| Component  | NO            | allowed uncertainty | 15            | %                  |
| Results of the regression analysis   |               |                     |               |                    |
| Slope b  | <b>0.97</b>   |                     |               | <b>significant</b> |
| Uncertainty of b   | <b>0.00</b>   |                     |               |                    |
| Ordinate intercept a   | <b>-1.35</b>  |                     |               | <b>significant</b> |
| Uncertainty of a   | <b>0.06</b>   |                     |               |                    |
| Results of the equivalence test  |               |                     |               |                    |
| Deviation at limit value   | <b>-18.61</b> |                     |               | <b>ppb</b>         |
| Uncertainty $u_{c,s}$ at limit value   | <b>18.64</b>  |                     |               | <b>ppb</b>         |
| Combined measurement uncertainty $w_{CM}$  | <b>3.69</b>   |                     |               | <b>%</b>           |
| Expanded uncertainty $W_{CM}$  | <b>7.38</b>   |                     |               | <b>%</b>           |
| Status equivalence test  | <b>passed</b> |                     |               |                    |

Table 91: Comparison of the tested and reference instruments, December, NO<sub>2</sub>

| Comparison candidate with reference according to guidance "Demonstration of Equivalence Of Ambient Air Monitoring Methods" |                   |                          |               |     |
|--|-------------------|--------------------------|---------------|-----|
| Device   | N500              | SN                       | SN 65 / SN 76 |     |
| Testside / month   | Cologne, December | Limit value              | 104.6         | ppb |
| Component  | NO <sub>2</sub>   | allowed uncertainty      | 15            | %   |
| Results of the regression analysis   |                   |                          |               |     |
| Slope b  | <b>1.00</b>       | <b>significant</b>       |               |     |
| Uncertainty of b   | <b>0.00</b>       |                          |               |     |
| Ordinate intercept a   | <b>0.00</b>       | <b>nicht significant</b> |               |     |
| Uncertainty of a   | <b>0.01</b>       |                          |               |     |
| Results of the equivalence test  |                   |                          |               |     |
| Deviation at limit value   | <b>0.25</b>       | <b>ppb</b>               |               |     |
| Uncertainty $u_{c_s}$ at limit value   | <b>0.30</b>       | <b>ppb</b>               |               |     |
| Combined measurement uncertainty $w_{CM}$  | <b>0.29</b>       | <b>%</b>                 |               |     |
| Expanded uncertainty $W_{CM}$  | <b>0.57</b>       | <b>%</b>                 |               |     |
| Status equivalence test  | <b>passed</b>     |                          |               |     |

 Table 92: Comparison of the tested and reference instruments, January, NO<sub>2</sub>

| Comparison candidate with reference according to guidance "Demonstration of Equivalence Of Ambient Air Monitoring Methods" |                  |                     |               |     |
|--|------------------|---------------------|---------------|-----|
| Device   | N500             | SN                  | SN 65 / SN 76 |     |
| Testside / month   | Cologne, January | Limit value         | 104.6         | ppb |
| Component  | NO <sub>2</sub>  | allowed uncertainty | 15            | %   |
| Results of the regression analysis   |                  |                     |               |     |
| Slope b  | <b>1.00</b>      | <b>significant</b>  |               |     |
| Uncertainty of b   | <b>0.00</b>      |                     |               |     |
| Ordinate intercept a   | <b>-0.04</b>     | <b>significant</b>  |               |     |
| Uncertainty of a   | <b>0.01</b>      |                     |               |     |
| Results of the equivalence test  |                  |                     |               |     |
| Deviation at limit value   | <b>0.46</b>      | <b>ppb</b>          |               |     |
| Uncertainty $u_{c_s}$ at limit value   | <b>0.48</b>      | <b>ppb</b>          |               |     |
| Combined measurement uncertainty $w_{CM}$  | <b>0.46</b>      | <b>%</b>            |               |     |
| Expanded uncertainty $W_{CM}$  | <b>0.91</b>      | <b>%</b>            |               |     |
| Status equivalence test  | <b>passed</b>    |                     |               |     |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO<sub>2</sub> and NO<sub>x</sub>,  
Report No.: 936/21251100/A

Table 93: Comparison of the tested and reference instruments, May, NO<sub>2</sub>

| Comparison candidate with reference according to guidance "Demonstration of Equivalence Of Ambient Air Monitoring Methods" |                 |                          |               |     |
|--|-----------------|--------------------------|---------------|-----|
| Device   | N500            | SN                       | SN 65 / SN 76 |     |
| Testside / month   | Cologne, May    | Limit value              | 104.6         | ppb |
| Component  | NO <sub>2</sub> | allowed uncertainty      | 15            | %   |
| Results of the regression analysis   |                 |                          |               |     |
| Slope b  | <b>1.00</b>     | <b>nicht significant</b> |               |     |
| Uncertainty of b   | <b>0.00</b>     |                          |               |     |
| Ordinate intercept a   | <b>-1.63</b>    | <b>significant</b>       |               |     |
| Uncertainty of a   | <b>0.05</b>     |                          |               |     |
| Results of the equivalence test  |                 |                          |               |     |
| Deviation at limit value   | <b>-1.37</b>    | <b>ppb</b>               |               |     |
| Uncertainty $u_{c_s}$ at limit value   | <b>1.47</b>     | <b>ppb</b>               |               |     |
| Combined measurement uncertainty $w_{CM}$  | <b>1.41</b>     | <b>%</b>                 |               |     |
| Expanded uncertainty $W_{CM}$  | <b>2.81</b>     | <b>%</b>                 |               |     |
| Status equivalence test  | <b>passed</b>   |                          |               |     |

Table 94: Comparison of the tested and reference instruments, June, NO<sub>2</sub>

| Comparison candidate with reference according to guidance "Demonstration of Equivalence Of Ambient Air Monitoring Methods" |                 |                          |               |     |
|--|-----------------|--------------------------|---------------|-----|
| Device   | N500            | SN                       | SN 65 / SN 76 |     |
| Testside / month   | Cologne, June   | Limit value              | 104.6         | ppb |
| Component  | NO <sub>2</sub> | allowed uncertainty      | 15            | %   |
| Results of the regression analysis   |                 |                          |               |     |
| Slope b  | <b>1.00</b>     | <b>nicht significant</b> |               |     |
| Uncertainty of b   | <b>0.00</b>     |                          |               |     |
| Ordinate intercept a   | <b>-1.67</b>    | <b>significant</b>       |               |     |
| Uncertainty of a   | <b>0.06</b>     |                          |               |     |
| Results of the equivalence test  |                 |                          |               |     |
| Deviation at limit value   | <b>-1.70</b>    | <b>ppb</b>               |               |     |
| Uncertainty $u_{c_s}$ at limit value   | <b>1.82</b>     | <b>ppb</b>               |               |     |
| Combined measurement uncertainty $w_{CM}$  | <b>1.74</b>     | <b>%</b>                 |               |     |
| Expanded uncertainty $W_{CM}$  | <b>3.48</b>     | <b>%</b>                 |               |     |
| Status equivalence test  | <b>passed</b>   |                          |               |     |




Figure 26: Reference vs. tested instrument, December, NO




Figure 27: Reference vs. tested instrument, January, NO






Figure 28: Reference vs. tested instrument, May, NO




Figure 29: Reference vs. tested instrument, June, NO




Figure 30: Reference vs. tested instrument, December, NO<sub>2</sub>




Figure 31: Reference vs. tested instrument, January, NO<sub>2</sub>




Figure 32: Reference vs. tested instrument, May, NO<sub>2</sub>




Figure 33: Reference vs. tested instrument, June, NO<sub>2</sub>

Table 95: Meteorological data (daily averages), November 2020

| November 2020 | Date       | Avg. Ambient temperature [°C] | Rel. Humidity [%] | Air pressure [hPa] |
|---------------|------------|-------------------------------|-------------------|--------------------|
| 1             | 01.11.2020 | 15.5                          | 63.5              | 1004.9             |
| 2             | 02.11.2020 | 18.6                          | 67.7              | 1002.3             |
| 3             | 03.11.2020 | 11.7                          | 65.2              | 1013.9             |
| 4             | 04.11.2020 | 7.7                           | 58.5              | 1023.0             |
| 5             | 05.11.2020 | 5.0                           | 54.3              | 1030.3             |
| 6             | 06.11.2020 | 7.0                           | 52.6              | 1025.3             |
| 7             | 07.11.2020 | 10.3                          | 49.4              | 1016.8             |
| 8             | 08.11.2020 | 11.1                          | 51.7              | 1014.6             |
| 9             | 09.11.2020 | 13.2                          | 57.5              | 1015.1             |
| 10            | 10.11.2020 | 11.8                          | 58.4              | 1016.9             |
| 11            | 11.11.2020 | 10.4                          | 58.1              | 1017.0             |
| 12            | 12.11.2020 | 11.2                          | 58.0              | 1010.7             |
| 13            | 13.11.2020 | 10.0                          | 55.8              | 1010.1             |
| 14            | 14.11.2020 | 14.1                          | 58.8              | 1008.6             |
| 15            | 15.11.2020 | 14.5                          | 59.7              | 1000.6             |
| 16            | 16.11.2020 | 11.1                          | 59.1              | 1005.0             |
| 17            | 17.11.2020 | 11.8                          | 59.0              | 1016.4             |
| 18            | 18.11.2020 | 13.0                          | 58.4              | 1015.4             |
| 19            | 19.11.2020 | 10.6                          | 57.0              | 1013.3             |
| 20            | 20.11.2020 | 6.1                           | 53.6              | 1025.7             |
| 21            | 21.11.2020 | 5.6                           | 49.0              | 1023.9             |
| 22            | 22.11.2020 | 9.7                           | 51.5              | 1018.4             |
| 23            | 23.11.2020 | 8.7                           | 54.1              | 1019.9             |
| 24            | 24.11.2020 | 6.4                           | 51.2              | 1016.6             |
| 25            | 25.11.2020 | 7.0                           | 49.1              | 1010.4             |
| 26            | 26.11.2020 | 6.0                           | 48.3              | 1013.0             |
| 27            | 27.11.2020 | 3.9                           | 46.8              | 1012.1             |
| 28            | 28.11.2020 | 4.9                           | 46.0              | 1012.4             |
| 29            | 29.11.2020 | 1.9                           | 42.7              | 1017.6             |
| 30            | 30.11.2020 | -0.4                          | 39.4              | 1016.8             |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO2 and NOx,  
Report No.: 936/21251100/A

Table 96: Meteorological data (daily averages), December 2020

| December 2020 | Date       | Avg. Ambient temperature [°C] | Rel. Humidity [%] | Air pressure [hPa] |
|---------------|------------|-------------------------------|-------------------|--------------------|
| 1             | 01.12.2020 | 5.0                           | 46.0              | 1006.8             |
| 2             | 02.12.2020 | 4.6                           | 47.6              | 1009.3             |
| 3             | 03.12.2020 | 4.2                           | 46.4              | 994.2              |
| 4             | 04.12.2020 | 6.3                           | 45.4              | 978.7              |
| 5             | 05.12.2020 | 5.4                           | 46.5              | 988.4              |
| 6             | 06.12.2020 | 3.9                           | 45.4              | 992.0              |
| 7             | 07.12.2020 | 4.5                           | 47.7              | 990.8              |
| 8             | 08.12.2020 | 4.2                           | 46.4              | 995.2              |
| 9             | 09.12.2020 | 3.5                           | 45.5              | 1001.3             |
| 10            | 10.12.2020 | 0.8                           | 42.1              | 999.6              |
| 11            | 11.12.2020 | 3.1                           | 41.0              | 989.2              |
| 12            | 12.12.2020 | 6.2                           | 45.8              | 988.5              |
| 13            | 13.12.2020 | 7.8                           | 49.7              | 1001.1             |
| 14            | 14.12.2020 | 8.3                           | 47.6              | 1000.9             |
| 15            | 15.12.2020 | 8.4                           | 50.4              | 1002.2             |
| 16            | 16.12.2020 | 8.5                           | 53.5              | 1007.9             |
| 17            | 17.12.2020 | 6.9                           | 49.8              | 1002.3             |
| 18            | 18.12.2020 | 5.3                           | 52.7              | 1001.8             |
| 19            | 19.12.2020 | 6.4                           | 53.2              | 1005.8             |
| 20            | 20.12.2020 | 6.8                           | 50.8              | 999.7              |
| 21            | 21.12.2020 | 7.9                           | 54.2              | 998.6              |
| 22            | 22.12.2020 | 8.2                           | 55.8              | 1002.7             |
| 23            | 23.12.2020 | 7.7                           | 53.6              | 1006.9             |
| 24            | 24.12.2020 | 10.2                          | 54.2              | 1010.4             |
| 25            | 25.12.2020 | 7.4                           | 51.4              | 1013.9             |
| 26            | 26.12.2020 | 9.3                           | 51.4              | 1007.7             |
| 27            | 27.12.2020 | 10.2                          | 53.0              | 1009.8             |
| 28            | 28.12.2020 | 7.6                           | 51.6              | 1009.8             |
| 29            | 29.12.2020 | 13.3                          | 56.8              | 1004.5             |
| 30            | 30.12.2020 | 11.8                          | 60.0              | 1003.1             |
| 31            | 31.12.2020 | 7.4                           | 57.5              | 1001.7             |

Table 97: Meteorological data (daily averages), January 2021

| January 2021 | Date       | Avg. Ambient temperature [°C] | Rel. Humidity [%] | Air pressure [hPa] |
|--------------|------------|-------------------------------|-------------------|--------------------|
| 1            | 01.01.2021 | 3.8                           | 51.5              | 1015.7             |
| 2            | 02.01.2021 | 4.7                           | 48.0              | 1010.2             |
| 3            | 03.01.2021 | 5.3                           | 44.4              | 977.3              |
| 4            | 04.01.2021 | 3.7                           | 43.7              | 993.0              |
| 5            | 05.01.2021 | 2.9                           | 41.9              | 1007.3             |
| 6            | 06.01.2021 | 3.1                           | 43.5              | 1007.0             |
| 7            | 07.01.2021 | 2.3                           | 44.0              | 1004.8             |
| 8            | 08.01.2021 | 1.9                           | 45.9              | 1007.0             |
| 9            | 09.01.2021 | 1.9                           | 46.2              | 1015.1             |
| 10           | 10.01.2021 | 1.8                           | 46.0              | 1016.4             |
| 11           | 11.01.2021 | 1.7                           | 43.9              | 1011.9             |
| 12           | 12.01.2021 | 5.0                           | 48.2              | 1001.8             |
| 13           | 13.01.2021 | 3.5                           | 47.3              | 1007.9             |
| 14           | 14.01.2021 | 2.4                           | 43.5              | 1010.7             |
| 15           | 15.01.2021 | 1.6                           | 40.0              | 1018.4             |
| 16           | 16.01.2021 | 0.8                           | 38.8              | 1016.5             |
| 17           | 17.01.2021 | 1.1                           | 40.3              | 1011.0             |
| 18           | 18.01.2021 | 5.3                           | 44.3              | 1014.2             |
| 19           | 19.01.2021 | 6.4                           | 46.9              | 1002.9             |
| 20           | 20.01.2021 | 8.3                           | 43.8              | 993.0              |
| 21           | 21.01.2021 | 9.3                           | 45.3              | 984.5              |
| 22           | 22.01.2021 | 5.8                           | 41.1              | 985.1              |
| 23           | 23.01.2021 | 4.5                           | 37.0              | 987.1              |
| 24           | 24.01.2021 | 1.5                           | 36.1              | 989.6              |
| 25           | 25.01.2021 | 1.3                           | 34.4              | 996.5              |
| 26           | 26.01.2021 | 3.3                           | 35.2              | 1007.8             |
| 27           | 27.01.2021 | 3.2                           | 35.8              | 1005.0             |
| 28           | 28.01.2021 | 5.6                           | 43.0              | 995.3              |
| 29           | 29.01.2021 | 8.9                           | 48.2              | 988.5              |
| 30           | 30.01.2021 | 3.2                           | 38.8              | 992.8              |
| 31           | 31.01.2021 | 1.3                           | 31.3              | 992.8              |

Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO2 and NOx,  
Report No.: 936/21251100/A

Table 98: Meteorological data (daily averages), February 2021

| February 2021 | Date       | Avg. Ambient temperature [°C] | Rel. Humidity [%] | Air pressure [hPa] |
|---------------|------------|-------------------------------|-------------------|--------------------|
| 1             | 01.02.2021 | 1.2                           | 33.6              | 993.8              |
| 2             | 02.02.2021 | 4.1                           | 36.6              | 986.8              |
| 3             | 03.02.2021 | 7.4                           | 42.4              | 992.7              |
| 4             | 04.02.2021 | 10.3                          | 48.5              | 990.1              |
| 5             | 05.02.2021 | 8.6                           | 44.0              | 1004.4             |
| 6             | 06.02.2021 | 8.9                           | 43.3              | 1003.6             |
| 7             | 07.02.2021 | 5.8                           | 44.3              | 999.0              |
| 8             | 08.02.2021 | 0.9                           | 36.6              | 989.5              |
| 9             | 09.02.2021 | -3.3                          | 28.2              | 991.6              |
| 10            | 10.02.2021 | -5.9                          | 22.8              | 998.4              |
| 11            | 11.02.2021 | -5.3                          | 22.2              | 1005.3             |
| 12            | 12.02.2021 | -3.0                          | 21.6              | 1020.3             |
| 13            | 13.02.2021 | -4.1                          | 19.4              | 1024.9             |
| 14            | 14.02.2021 | -2.9                          | 18.5              | 1030.5             |
| 15            | 15.02.2021 | -0.1                          | 18.4              | 1028.5             |
| 16            | 16.02.2021 | 0.9                           | 22.9              | 1018.0             |
| 17            | 17.02.2021 | 9.8                           | 36.2              | 1008.2             |
| 18            | 18.02.2021 | 7.8                           | 41.3              | 1009.1             |
| 19            | 19.02.2021 | 6.0                           | 41.8              | 1004.1             |

Table 99: Meteorological data (daily averages), May 2021

| May 2021 | Date       | Avg. Ambient temperature [°C] | Rel. Humidity [%] | Air pressure [hPa] |
|----------|------------|-------------------------------|-------------------|--------------------|
| 1        | 01.05.2021 | 9.9                           | 62.2              | 1007.3             |
| 2        | 02.05.2021 | 8.2                           | 62.2              | 1011.4             |
| 3        | 03.05.2021 | 9.4                           | 53.9              | 1012.7             |
| 4        | 04.05.2021 | 11.8                          | 52.7              | 996.4              |
| 5        | 05.05.2021 | 7.9                           | 69.2              | 1000.0             |
| 6        | 06.05.2021 | 8.2                           | 69.7              | 1000.9             |
| 7        | 07.05.2021 | 7.8                           | 71.8              | 1007.6             |
| 8        | 08.05.2021 | 11.3                          | 60.6              | 1008.6             |
| 9        | 09.05.2021 | 20.2                          | 51.7              | 998.6              |
| 10       | 10.05.2021 | 19.7                          | 51.2              | 999.5              |
| 11       | 11.05.2021 | 14.6                          | 86.0              | 1002.5             |
| 12       | 12.05.2021 | 13.8                          | 76.9              | 1004.9             |
| 13       | 13.05.2021 | 12.8                          | 74.7              | 1001.5             |
| 14       | 14.05.2021 | 12.6                          | 75.5              | 1001.6             |
| 15       | 15.05.2021 | 11.5                          | 80.8              | 998.8              |
| 16       | 16.05.2021 | 11.2                          | 81.5              | 996.2              |
| 17       | 17.05.2021 | 12.0                          | 79.7              | 998.3              |
| 18       | 18.05.2021 | 11.6                          | 81.5              | 1006.5             |
| 19       | 19.05.2021 | 11.8                          | 76.1              | 1010.6             |
| 20       | 20.05.2021 | 14.4                          | 66.0              | 1012.7             |
| 21       | 21.05.2021 | 14.7                          | 55.9              | 1000.3             |
| 22       | 22.05.2021 | 11.5                          | 69.5              | 999.5              |
| 23       | 23.05.2021 | 12.7                          | 59.5              | 1007.3             |
| 24       | 24.05.2021 | 13.3                          | 61.5              | 1004.2             |
| 25       | 25.05.2021 | 9.8                           | 79.3              | 1007.3             |
| 26       | 26.05.2021 | 10.7                          | 85.3              | 1007.0             |
| 27       | 27.05.2021 | 11.4                          | 83.4              | 1011.5             |
| 28       | 28.05.2021 | 13.7                          | 69.2              | 1017.1             |
| 29       | 29.05.2021 | 14.3                          | 65.2              | 1018.7             |
| 30       | 30.05.2021 | 15.7                          | 61.6              | 1019.8             |
| 31       | 31.05.2021 | 18.0                          | 54.4              | 1015.4             |



Report on the performance test of the N500 ambient air quality measuring system manufactured by Teledyne API for the components NO, NO2 and NOx,  
Report No.: 936/21251100/A

Table 100: Meteorological data (daily averages), June 2021

| June 2021 | Date       | Avg. Ambient temperature [°C] | Rel. Humidity [%] | Air pressure [hPa] |
|-----------|------------|-------------------------------|-------------------|--------------------|
| 1         | 01.06.2021 | 16.7                          | 54.8              | 1012.3             |
| 2         | 02.06.2021 | 20.8                          | 49.1              | 1008.8             |
| 3         | 03.06.2021 | 19.9                          | 78.8              | 1013.0             |
| 4         | 04.06.2021 | 21.2                          | 80.5              | 1013.8             |
| 5         | 05.06.2021 | 17.7                          | 88.8              | 1014.4             |
| 6         | 06.06.2021 | 15.8                          | 87.1              | 1017.1             |
| 7         | 07.06.2021 | 18.0                          | 78.9              | 1016.7             |
| 8         | 08.06.2021 | 19.3                          | 70.6              | 1016.3             |
| 9         | 09.06.2021 | 20.6                          | 66.5              | 1015.2             |
| 10        | 10.06.2021 | 22.1                          | 55.8              | 1014.6             |
| 11        | 11.06.2021 | 22.5                          | 59.7              | 1013.1             |
| 12        | 12.06.2021 | 20.1                          | 67.2              | 1014.3             |
| 13        | 13.06.2021 | 17.6                          | 58.2              | 1020.4             |
| 14        | 14.06.2021 | 21.5                          | 53.4              | 1015.4             |
| 15        | 15.06.2021 | 22.1                          | 60.1              | 1012.8             |
| 16        | 16.06.2021 | 24.8                          | 54.7              | 1008.4             |
| 17        | 17.06.2021 | 28.0                          | 51.4              | 1004.9             |
| 18        | 18.06.2021 | 27.7                          | 54.9              | 1006.2             |
| 19        | 19.06.2021 | 26.3                          | 53.8              | 1006.9             |
| 20        | 20.06.2021 | 23.1                          | 69.0              | 1002.3             |
| 21        | 21.06.2021 | 19.6                          | 77.6              | 1001.0             |
| 22        | 22.06.2021 | 15.0                          | 84.5              | 1008.8             |
| 23        | 23.06.2021 | 16.4                          | 86.5              | 1013.2             |
| 24        | 24.06.2021 | 17.5                          | 81.0              | 1012.9             |
| 25        | 25.06.2021 | 19.1                          | 63.5              | 1012.4             |
| 26        | 26.06.2021 | 19.9                          | 64.1              | 1012.1             |
| 27        | 27.06.2021 | 21.0                          | 69.7              | 1009.5             |
| 28        | 28.06.2021 | 22.1                          | 79.8              | 1007.5             |
| 29        | 29.06.2021 | 21.0                          | 74.7              | 1005.8             |
| 30        | 30.06.2021 | 16.7                          | 87.1              | 1005.6             |

## Annex 2 Certificate of Accreditation to EN ISO/IEC 17025:2005



### Deutsche Akkreditierungsstelle GmbH

Beliehene gemäß § 8 Absatz 1 AkkStelleG i.V.m. § 1 Absatz 1 AkkStelleGBV  
Unterzeichnerin der Multilateralen Abkommen  
von EA, ILAC und IAF zur gegenseitigen Anerkennung

## Akkreditierung



Die Deutsche Akkreditierungsstelle GmbH bestätigt hiermit, dass das Prüflaboratorium

### TÜV Rheinland Energy GmbH

mit seinen in der Urkundenanlage aufgeführten Messstellen und Standorten


die Kompetenz nach DIN EN ISO/IEC 17025:2018 besitzt, Prüfungen in folgenden Bereichen durchzuführen:

Bestimmung (Probenahme und Analytik) von anorganischen und organischen gas- oder partikel-förmigen Luftinhaltsstoffen im Rahmen von Emissions- und Immissionsmessungen; Probenahme von luftgetragenen polyhalogenierten Dibenzo-p-Dioxinen und Dibenzofuranen bei Emissionen und Immissionen; Probenahme von faserförmigen Partikeln bei Emissionen und Immissionen; Ermittlung von gas- oder partikel-förmigen Luftinhaltsstoffen mit kontinuierlich arbeitenden Messgeräten; Bestimmung von Geruchsstoffen in Luft; Kalibrierungen und Funktionsprüfungen kontinuierlich arbeitender Messgeräte für Luftinhaltsstoffe einschließlich Systemen zur Datenauswertung und Emissionsfernüberwachung; Feuerraummessungen; Eignungsprüfungen von automatisch arbeitenden Emissions- und Immissionsmessenrichtungen einschließlich Systemen zur Datenauswertung und Emissionsfernüberwachung; Ermittlung der Emissionen und Immissionen von Geräuschen; Bestimmung von Geräuschen in der Nachbarschaft; Ermittlung der Emissionen und Immissionen von Geräuschen; Bestimmung von Geräuschen Messungen im Eisenbahnbereich; Bestimmung von Schalleistungspegeln von zur Verwendung im Freien vorgesehenen Geräten und Maschinen nach Richtlinie 2000/14/EG und Konformitätsbewertungsverfahren; Schornsteinhöhenberechnung und Immissionsprognose auf der Grundlage der Technischen Anleitung zur Reinhaltung der Luft und der Geruchsimmisions-Richtlinie und der VDI 3783 Blatt 13; Windenergieanlagen: Bestimmung von Windpotential, Energieerträgen, Standorterträgen und Standortgüte nach EEG, standortbezogenen Turbulenzcharakteristika und Extremwinde; Schallimmissionsprognosen, Schattenwurfimmissionsberechnung und Sichtbarkeitsbestimmung; Probenahme und mikrobiologische Untersuchungen von Nutzwasser gemäß § 3 Absatz 8 42. BImSchV; physikalische, physikalisch-chemische und mikrobiologische Untersuchungen von Wasser (Abwasser, Wasser aus Rückkühlwerken sowie raumlufttechnischen Anlagen); Probenahme von Abwasser; mikrobiologische und ausgewählte chemische Untersuchungen gemäß Trinkwasserverordnung; Probenahme von Roh- und Trinkwasser; ausgewählte mikrobiologische Untersuchungen von Bedarfsgegenständen und kosmetischen Mitteln; Probenahme anorganischer faserförmiger Partikel sowie von partikel- und gasförmigen luftverunreinigenden Stoffen in der Innenraumluft; ausgewählte mikrobiologische Untersuchungen in Innenräumen; Ermittlung von Aerosolen und Faserstäuben, anorganischen und organischen Gasen und Dämpfen sowie ausgewählten Parametern und/oder in ausgewählten Gebieten bei Arbeitsplatzmessungen gemäß Gefahrstoffverordnung § 7, Abs. 10; Modul Immissionsschutz

Die Akkreditierungsurkunde gilt nur in Verbindung mit dem Bescheid vom 17.06.2020 mit der Akkreditierungsnummer D-PL-11120-02. Sie besteht aus diesem Deckblatt, der Rückseite des Deckblatts und der folgenden Anlage mit insgesamt 48 Seiten.

Registrierungsnummer der Urkunde: D-PL-11120-02-00

Berlin, 17.06.2020

  
Im Auftrag Dipl.-Ing. Andrea Valbuena  
Abteilungsleiterin

Die Urkunde samt Urkundenanlage gibt den Stand zum Zeitpunkt des Ausstellungsdatums wieder. Der jeweils aktuelle Stand des Geltungsbereiches der Akkreditierung ist der Datenbank akkreditierter Stellen der Deutschen Akkreditierungsstelle GmbH (DAkkS) zu entnehmen. <https://www.dakks.de/content/datenbank-akkreditierter-stellen>

Siehe Hinweise auf der Rückseite

Figure 34: Certificate of accreditation according to EN ISO/IEC 17025:2005

## Deutsche Akkreditierungsstelle GmbH

Standort Berlin  
Spittelmarkt 10  
10117 Berlin

Standort Frankfurt am Main  
Europa-Allee 52  
60327 Frankfurt am Main

Standort Braunschweig  
Bundesallee 100  
38116 Braunschweig

Die auszugsweise Veröffentlichung der Akkreditierungsurkunde bedarf der vorherigen schriftlichen Zustimmung der Deutsche Akkreditierungsstelle GmbH (DAkkS). Ausgenommen davon ist die separate Weiterverbreitung des Deckblattes durch die umseitig genannte Konformitätsbewertungsstelle in unveränderter Form.

Es darf nicht der Anschein erweckt werden, dass sich die Akkreditierung auch auf Bereiche erstreckt, die über den durch die DAkkS bestätigten Akkreditierungsbereich hinausgehen.

Die Akkreditierung erfolgte gemäß des Gesetzes über die Akkreditierungsstelle (AkkStelleG) vom 31. Juli 2009 (BGBl. I S. 2625) sowie der Verordnung (EG) Nr. 765/2008 des Europäischen Parlaments und des Rates vom 9. Juli 2008 über die Vorschriften für die Akkreditierung und Marktüberwachung im Zusammenhang mit der Vermarktung von Produkten (Abl. L 218 vom 9. Juli 2008, S. 30). Die DAkkS ist Unterzeichnerin der Multilateralen Abkommen zur gegenseitigen Anerkennung der European co-operation for Accreditation (EA), des International Accreditation Forum (IAF) und der International Laboratory Accreditation Cooperation (ILAC). Die Unterzeichner dieser Abkommen erkennen ihre Akkreditierungen gegenseitig an.

Der aktuelle Stand der Mitgliedschaft kann folgenden Webseiten entnommen werden:

EA: [www.european-accreditation.org](http://www.european-accreditation.org)

ILAC: [www.ilac.org](http://www.ilac.org)

IAF: [www.iaf.nu](http://www.iaf.nu)

Figure 34: Certificate of accreditation according to EN ISO/IEC 17025:2005 - page 2

**Annex 3:**

# **Annex 3:**

## **Manual**



**TELEDYNE API**  
Everywhereyoulook™



**User Manual**  
***Model N500***  
***CAPS NO<sub>x</sub> Analyzer***

*with NumaView™ software*

© TELEDYNE API (TAPI)  
9970 CARROLL CANYON ROAD  
SAN DIEGO, CALIFORNIA 92131-1106  
USA

Toll-free Phone: +1 800-324-5190  
Phone: +1 858-657-9800  
Fax: +1 858-657-9816  
Email: [api-sales@teledyne.com](mailto:api-sales@teledyne.com)  
Website: <http://www.teledyne-api.com/>



## NOTICE OF COPYRIGHT

© 2020 Teledyne API (TAPI). All rights reserved.

## TRADEMARKS

All trademarks, registered trademarks, brand names or product names appearing in this document are the property of their respective owners and are used herein for identification purposes only.

## SAFETY MESSAGES

Important safety messages are provided throughout this manual for the purpose of avoiding personal injury or instrument damage. Please read these messages carefully. Each safety message is associated with a safety alert symbol and is placed throughout this manual; the safety symbols are also located inside the instrument. It is imperative that you pay close attention to these messages, the descriptions of which are as follows:



WARNING: Electrical Shock Hazard



HAZARD: Strong oxidizer



GENERAL WARNING/CAUTION: Read the accompanying message for specific information.



CAUTION: Hot Surface Warning



Do Not Touch: Touching some parts of the instrument without protection or proper tools could result in damage to the part(s) and/or the instrument.



Technician Symbol: All operations marked with this symbol are to be performed by qualified maintenance personnel only.



Electrical Ground: This symbol inside the instrument marks the central safety grounding point for the instrument.

### CAUTION



This instrument should only be used for the purpose and in the manner described in this manual. If you use this instrument in a manner other than that for which it was intended, unpredictable behavior could ensue with possible hazardous consequences.

NEVER use any combustible/explosive gas with this instrument!

For Technical Assistance regarding the use and maintenance of this instrument or any other Teledyne API product, contact Teledyne API's Technical Support Department:

Telephone: +1 800-324-5190  
Email: [api-techsupport@teledyne.com](mailto:api-techsupport@teledyne.com)

or access any of the service options on our website at <http://www.teledyne-api.com/>



## CONSIGNES DE SÉCURITÉ

Des consignes de sécurité importantes sont fournies tout au long du présent manuel dans le but d'éviter des blessures corporelles ou d'endommager les instruments. Veuillez lire attentivement ces consignes. Chaque consigne de sécurité est représentée par un pictogramme d'alerte de sécurité; ces pictogrammes se retrouvent dans ce manuel et à l'intérieur des instruments. Les symboles correspondent aux consignes suivantes :



AVERTISSEMENT : Risque de choc électrique



DANGER : Oxydant puissant



AVERTISSEMENT GÉNÉRAL / MISE EN GARDE : Lire la consigne complémentaire pour des renseignements spécifiques



MISE EN GARDE : Surface chaude



Ne pas toucher : Toucher à certaines parties de l'instrument sans protection ou sans les outils appropriés pourrait entraîner des dommages aux pièces ou à l'instrument.



Pictogramme « technicien » : Toutes les opérations portant ce symbole doivent être effectuées uniquement par du personnel de maintenance qualifié.



Mise à la terre : Ce symbole à l'intérieur de l'instrument détermine le point central de la mise à la terre sécuritaire de l'instrument.

### MISE EN GARDE



Cet instrument doit être utilisé aux fins décrites et de la manière décrite dans ce manuel. Si vous utilisez cet instrument d'une autre manière que celle pour laquelle il a été prévu, l'instrument pourrait se comporter de façon imprévisible et entraîner des conséquences dangereuses.

**NE JAMAIS** utiliser de gaz explosive ou combustible avec cet instrument!

## WARRANTY

### WARRANTY POLICY (02024J)

Teledyne API (TAPI), a business unit of Teledyne Instruments, Inc., provides that: Prior to shipment, TAPI equipment is thoroughly inspected and tested. Should equipment failure occur, TAPI assures its customers that prompt service and support will be available. (For the instrument-specific warranty period, please refer to the “Limited Warranty” section in the Terms and Conditions of Sale on our website at the following link: [http://www.teledyne-api.com/terms\\_and\\_conditions.asp](http://www.teledyne-api.com/terms_and_conditions.asp)).

### COVERAGE

After the warranty period and throughout the equipment lifetime, TAPI stands ready to provide on-site or in-plant service at reasonable rates similar to those of other manufacturers in the industry. All maintenance and the first level of field troubleshooting are to be performed by the customer.

### NON-TAPI MANUFACTURED EQUIPMENT

Equipment provided but not manufactured by TAPI is warranted and will be repaired to the extent and according to the current terms and conditions of the respective equipment manufacturer’s warranty.

### PRODUCT RETURN

All units or components returned to Teledyne API should be properly packed for handling and returned freight prepaid to the nearest designated Service Center. After the repair, the equipment will be returned, freight prepaid.

The complete Terms and Conditions of Sale can be reviewed at [http://www.teledyne-api.com/terms\\_and\\_conditions.asp](http://www.teledyne-api.com/terms_and_conditions.asp)

#### CAUTION – Avoid Warranty Invalidation



Failure to comply with proper anti-Electro-Static Discharge (ESD) handling and packing instructions and Return Merchandise Authorization (RMA) procedures when returning parts for repair or calibration may void your warranty. For anti-ESD handling and packing instructions please refer to the manual, Fundamentals of ESD, PN 04786, in its “Packing Components for Return to Teledyne API’s Customer Service” section. The manual can be downloaded from our website at <http://www.teledyne-api.com>. RMA procedures can also be found on our website.

## ABOUT THIS MANUAL

**Note**

We recommend that all users read this manual in its entirety before operating the instrument.

## CONVENTIONS USED

In addition to the safety symbols as presented in the *Safety Messages* page, this manual provides *special notices* related to the careful and effective use of the instrument and related, pertinent information.

**ATTENTION**

**COULD DAMAGE INSTRUMENT AND VOID WARRANTY**  
This special notice provides information to avoid damage to your instrument and possibly invalidate the warranty.

**Important**

**IMPACT ON READINGS OR DATA**  
Provides information about that which could either affect accuracy of instrument readings or cause loss of data.

**Note**

Provides information pertinent to the proper care, operation or maintenance of the instrument or its parts.

# TABLE OF CONTENTS

|  |           |
|--|-----------|
| Safety Messages .....  | ii        |
| Warranty .....   | iv        |
| Table of Contents .....  | vi        |
| List of Figures .....  | ix        |
| List of Tables .....   | x         |
| <b>1. INTRODUCTION, SPECIFICATIONS, APPROVALS, &amp; COMPLIANCE .....</b>  | <b>11</b> |
| 1.1. Specifications .....  | 11        |
| 1.2. EPA Designation .....   | 12        |
| 1.3. Safety .....  | 12        |
| 1.4. EMC .....   | 12        |
| <b>2. GETTING STARTED .....</b>  | <b>13</b> |
| 2.1. Unpacking .....   | 13        |
| 2.1.1. Ventilation Clearance .....   | 14        |
| 2.2. Instrument Layout .....   | 15        |
| 2.2.1. Front Panel .....   | 15        |
| 2.2.2. Rear Panel .....  | 16        |
| 2.2.3. Internal Chassis .....  | 17        |
| 2.3. Connections and Startup .....   | 18        |
| 2.3.1. Electrical Connections .....  | 18        |
| 2.3.1.1. Connecting Power .....  | 18        |
| 2.3.1.2. Connecting Analog Outputs Option .....                            | 19        |
| 2.3.1.3. Connecting the Digital I/O Expansion Board Option .....           | 20        |
| 2.3.1.4. Connecting Communications Interfaces .....                        | 21        |
| 2.3.2. Pneumatic Connections .....   | 22        |
| 2.3.2.1. Critical Tubing, Pressure, Venting and Exhaust Requirements ..... | 23        |
| 2.3.2.2. Basic Connections from Calibrator .....                           | 24        |
| 2.3.3. Pneumatic Flow Diagrams .....                                       | 25        |
| 2.3.4. Startup, Functional Checks and Calibration .....                    | 27        |
| 2.3.4.1. Startup .....   | 27        |
| 2.3.4.2. Alerts: Warnings and Other Messages .....                         | 28        |
| 2.3.4.3. Functional Checks .....   | 29        |
| 2.3.4.4. Calibration .....   | 29        |
| 2.4. Menu Overview .....   | 30        |
| 2.4.1. Home Page .....   | 31        |
| 2.4.2. Dashboard .....   | 33        |
| 2.4.3. Alerts .....  | 34        |
| 2.4.4. Calibration .....   | 35        |
| 2.4.5. Utilities .....   | 36        |
| 2.4.6. Setup .....   | 36        |
| 2.5. Setup Menu: Features/Functions Configuration .....                    | 36        |
| 2.5.1. Setup>Data Logging (Data Acquisition System, DAS) .....             | 37        |
| 2.5.1.1. Configuring Trigger Types: Periodic .....                         | 39        |
| 2.5.1.2. Configuring Trigger Types: Conditional .....                      | 40        |
| 2.5.1.3. Downloading DAS (Data Acquisition System) Data .....              | 40        |
| 2.5.2. Setup>Events .....  | 41        |
| 2.5.2.1. Editing or Deleting Events .....                                  | 43        |
| 2.5.2.2. Using Events as Triggers for Data Logging .....                   | 44        |
| 2.5.3. Setup>Dashboard .....   | 44        |
| 2.5.4. Setup>AutoCal (with Valve Option) .....                             | 44        |
| 2.5.5. Setup>Vars .....  | 45        |
| 2.5.6. Setup>Homescreen .....  | 46        |
| 2.5.7. Setup>Digital Outputs (Option) .....                                | 47        |

|   |    |
|---|----|
| 2.5.8. Setup>Analog Outputs (Option).....                                 | 48 |
| 2.5.8.1. Manual Calibration of Voltage Range Analog Outputs Option .....  | 50 |
| 2.5.8.2. Manual Adjustment of Current Range Analog Outputs Option .....   | 50 |
| 2.5.9. Setup>Instrument.....  | 52 |
| 2.5.10. Setup>Comm (Communications).....                                  | 52 |
| 2.5.10.1. COM1 .....  | 52 |
| 2.5.10.2. TCP Port1 .....   | 53 |
| 2.5.10.3. TCP Port2 .....   | 53 |
| 2.5.10.4. Network Settings .....  | 53 |
| 2.6. Transferring Configuration to Other Instruments .....                | 55 |
| 3. COMMUNICATIONS AND REMOTE OPERATION .....                              | 56 |
| 3.1. Serial Communication .....   | 56 |
| 3.2. Ethernet.....  | 56 |
| 3.3. NumaView™ Remote.....  | 56 |
| 4. CALIBRATION .....  | 57 |
| 4.1. Important Precalibration Information .....                           | 57 |
| 4.1.1. Calibration Requirements .....                                     | 57 |
| 4.1.2. Zero Air .....   | 58 |
| 4.1.3. Calibration (Span) Gas .....                                       | 58 |
| 4.1.4. Span Gas for Multipoint Calibration .....                          | 58 |
| 4.1.5. NO <sub>2</sub> Permeation Tubes.....                              | 59 |
| 4.1.6. Data Recording Devices .....                                       | 59 |
| 4.2. Calibration Procedures.....  | 60 |
| 4.2.1. Calibration and Check Procedures for Basic Configuration .....     | 60 |
| 4.2.1.1. Zero Calibration Check and Actual Calibration .....              | 61 |
| 4.2.1.2. Span Calibration Check and Actual Calibration .....              | 61 |
| 4.2.2. Calibration and Check Procedures with Valve Options Installed..... | 62 |
| 4.2.2.1. Use of Zero/Span Valve with Digital Expansion Board Option.....  | 62 |
| 4.3. Automatic Zero/Span Cal/Check (Auto Cal) .....                       | 63 |
| 4.4. Titration Efficiency (TE) Check/Calibration .....                    | 65 |
| 4.5. EPA Protocol Calibration.....  | 66 |
| 5. MAINTENANCE AND SERVICE .....  | 67 |
| 5.1. Maintenance Schedule.....  | 67 |
| 5.2. Predictive Diagnostics.....  | 68 |
| 5.3. Operational Health Checks .....                                      | 69 |
| 5.4. Software/Firmware Updates.....                                       | 70 |
| 5.4.1. Remote Updates .....   | 70 |
| 5.4.2. Manual Reload/Update Procedures.....                               | 70 |
| 5.5. Time Zone Changes.....   | 72 |
| 5.6. Hardware Maintenance Procedures .....                                | 73 |
| 5.6.1. Replacing the Sample Filter.....                                   | 73 |
| 5.6.2. Replacing the AREF Scrubber and Filter Assembly.....               | 73 |
| 5.6.3. Replacing the Internal Pump.....                                   | 74 |
| 5.6.4. Changing the IZS Permeation Tube Option.....                       | 75 |
| 5.6.5. Checking for Pneumatic Leaks .....                                 | 76 |
| 5.6.5.1. Detailed Pressure Leak Check.....                                | 76 |
| 5.6.5.2. Performing a Sample Flow Check .....                             | 77 |
| 5.7. Service and Troubleshooting .....                                    | 77 |
| 5.7.1. Fault Diagnosis with Alerts.....                                   | 78 |
| 5.7.2. Fault Diagnosis With Dashboard Functions.....                      | 79 |
| 5.7.3. Using the Diagnostic Signal I/O Functions .....                    | 79 |
| 5.7.4. Fault Diagnosis with LEDs .....                                    | 80 |
| 5.7.5. Calibration Problems.....  | 83 |
| 5.7.5.1. Negative Concentrations .....                                    | 83 |
| 5.7.5.2. Absence of Analyzer Response to Sample Gas .....                 | 83 |

|   |    |
|---|----|
| 5.7.5.3. Unstable Zero and Span.....                        | 83 |
| 5.7.5.4. Inability to Span - Deactivated SPAN Button .....  | 83 |
| 5.7.5.5. Inability to Zero - Deactivated ZERO Button .....  | 84 |
| 5.7.5.6. Non-Linear Response.....                           | 84 |
| 5.7.6. Other Performance Problems .....                     | 84 |
| 5.7.6.1. Excessive Noise .....                              | 85 |
| 5.7.6.2. Slow Response.....                                 | 85 |
| 5.7.6.3. AREF Warnings .....                                | 85 |
| 5.7.7. Subsystem Check for Troubleshooting .....            | 86 |
| 5.7.7.1. AC Main Power.....                                 | 86 |
| 5.7.7.2. LCD/Display Module.....                            | 87 |
| 5.7.7.3. O <sub>3</sub> Generator Module.....               | 87 |
| 5.7.7.4. O <sub>3</sub> Generator Override.....             | 87 |
| 5.7.7.5. Internal Span Gas Generator and Valve Options..... | 87 |
| 5.7.7.6. RS-232 Communications.....                         | 87 |
| 5.7.8. Service Procedures.....                              | 88 |
| 5.7.8.1. Fuse Replacement Procedure.....                    | 88 |
| 5.7.8.2. O <sub>3</sub> Generator Replacement .....         | 89 |
| 5.7.8.3. Sample Conditioner Assembly Replacement.....       | 90 |
| 5.8. Frequently Asked Questions .....                       | 91 |
| 5.9. Technical Assistance .....                             | 92 |
| 6. PRINCIPLES OF OPERATION .....                            | 93 |
| 6.1. Pneumatic Operation .....                              | 95 |
| 6.2. Electronic Operation.....                              | 96 |
| 6.2.1. Modules.....   | 96 |
| 6.2.2. Power Switches.....                                  | 96 |
| Glossary .....  | 97 |

## LIST OF FIGURES

|  |    |
|--|----|
| Figure 2-1. Front Panel Layout .....   | 15 |
| Figure 2-2. Rear Panel Layout.....   | 16 |
| Figure 2-3. Internal Chassis Layout.....   | 17 |
| Figure 2-4. Analog Outputs Connectors Panel Option .....   | 19 |
| Figure 2-5. Digital I/O Connector Panel Option .....   | 20 |
| Figure 2-6. Mainboard JP1 Location and Pin Arrangements.....                                       | 21 |
| Figure 2-7. Gas Line Connections from Calibrator – Basic Configuration.....                        | 24 |
| Figure 2-8. Pneumatic Flow Diagram including Zero/Span Valve Option .....                          | 25 |
| Figure 2-9. Pneumatic Flow Diagram including IZS Valve Option.....                                 | 26 |
| Figure 2-10. Status Screens at Startup.....  | 27 |
| Figure 2-11. Home Page Example.....  | 28 |
| Figure 2-12. Viewing Active Alerts Page.....   | 28 |
| Figure 2-13. Sample Dashboard Page .....   | 29 |
| Figure 2-14. User Interface Orientation .....  | 31 |
| Figure 2-15. Concentration and Stability Graph (top) and Meter Graph (bottom) .....                | 32 |
| Figure 2-16. Dashboard Page.....   | 33 |
| Figure 2-17. Active Alerts Page .....  | 34 |
| Figure 2-18. Active Alerts Cleared .....   | 35 |
| Figure 2-19. Utilities>Alerts Log of Active and Past Alerts and Events .....                       | 35 |
| Figure 2-20. Datalog Configuration, New Log Page .....   | 37 |
| Figure 2-21. Datalog Configuration, Existing Log .....   | 37 |
| Figure 2-22. Creating a New Data Log .....   | 38 |
| Figure 2-23. Datalog Periodic Trigger Configuration .....  | 39 |
| Figure 2-24. Datalog - Conditional Trigger Configuration .....                                     | 40 |
| Figure 2-25. DAS Download Page .....   | 40 |
| Figure 2-26. Events List .....   | 41 |
| Figure 2-27. Event Configuration .....   | 42 |
| Figure 2-28. Configured Event Sample.....  | 43 |
| Figure 2-29. Edit or Delete an Event.....  | 43 |
| Figure 2-30. Dashboard Display and Configuration.....  | 44 |
| Figure 2-31. Homescreen Configuration.....   | 46 |
| Figure 2-32. Digital Outputs Setup.....  | 47 |
| Figure 2-33. Analog Output Option Configuration Example .....                                      | 48 |
| Figure 2-34. Analog Outputs Group Calibration Screen .....   | 49 |
| Figure 2-35. Analog Outputs Option Manual Calibration Screen (AOUT2 Example).....                  | 49 |
| Figure 2-36. Setup for Checking / Calibrating DCV Analog Output Signal Levels .....                | 50 |
| Figure 2-37. Setup for Checking / Calibrating Current Output Signal Levels .....                   | 51 |
| Figure 2-38. Alternative Setup Using 250Ω Resistor for Checking Current Output Signal Levels ..... | 51 |
| Figure 2-39. Communications Configuration, Network Settings .....                                  | 54 |
| Figure 2-40. Configuration Transfer.....   | 55 |
| Figure 4-1. Multi-Point Calibration Page .....   | 61 |
| Figure 4-2. Zero and Span Calibration Screens .....  | 62 |
| Figure 4-3. Auto Cal Page.....   | 63 |
| Figure 5-1: Report Generation Page.....  | 69 |
| Figure 5-2. Remote Update Page .....   | 70 |
| Figure 5-3. Manual Update Page (and other utilities).....  | 70 |
| Figure 5-4. Time Zone Change Requirements .....  | 72 |
| Figure 5-5. Replacing the Sample Filter .....  | 73 |
| Figure 5-6. Pneumatic Layout of AREF Filter and Charcoal Filter.....                               | 74 |
| Figure 5-7. Internal Pump .....  | 75 |
| Figure 5-8. N500 Mainboard .....   | 80 |
| Figure 5-9. N500 CAPS DAQ Smart Board .....  | 81 |
| Figure 5-10. N500 Ozone Tower Smart Board .....  | 82 |
| Figure 5-11. Fuse Access .....   | 89 |

Figure 6-1. N500 Optical Absorption Cell ..... 94  
 Figure 6-2. Phase Shift Representation of Increased Concentration of NO<sub>2</sub> ..... 94

## LIST OF TABLES

Table 1-1. Specifications ..... 11  
 Table 2-1. Ventilation Clearance ..... 14  
 Table 2-2. Rear Panel Description ..... 16  
 Table 2-3. Analog Output Pin Assignments ..... 19  
 Table 2-4. Digital Input/Output Pin Assignments ..... 20  
 Table 2-5. JP1 Configurations for Serial Communication ..... 21  
 Table 2-6. Menu Overview ..... 30  
 Table 2-7. Utilities Submenu Descriptions ..... 36  
 Table 2-8. List of Variables with Descriptions ..... 45  
 Table 2-9. Analog Output Option Voltage/Current Range ..... 49  
 Table 2-10. Current Loop Output Check ..... 51  
 Table 2-11. Setup>Instrument Menu ..... 52  
 Table 2-12. COM1 Setup ..... 52  
 Table 2-13. LAN/Ethernet Configuration Properties ..... 54  
 Table 3-1. Ethernet Status Indicators ..... 56  
 Table 4-1. AUTO CAL States ..... 64  
 Table 4-2. Auto Cal Setup Combinations ..... 64  
 Table 4-3. Auto Cal Programming Sequence Execution ..... 65  
 Table 5-1. Maintenance Schedule ..... 68  
 Table 5-2. Predictive Uses for Dashboard Functions ..... 69  
 Table 5-3. Warning Alerts, Fault Conditions and Possible Causes ..... 78  
 Table 6-1. Sensor Signal Processing States ..... 95

### APPENDIX A – MODBUS REGISTERS

### APPENDIX B – WIRING DIAGRAM



# 1. INTRODUCTION, SPECIFICATIONS, APPROVALS, & COMPLIANCE

Teledyne API's Model N500 CAPS NO<sub>x</sub>-NO<sub>2</sub>-NO Analyzer uses Cavity-Attenuated Phase-Shift (CAPS) spectroscopy to render direct, true measurements of nitrogen dioxide (NO<sub>2</sub>). By combining direct NO<sub>2</sub> measurements with highly efficient gas phase titration (GPT), the N500 converts and measures the NO sample gas component as NO<sub>2</sub>, thus allowing for the measurement of total NO<sub>x</sub>. Furthering its accuracy and dependability is an auto-reference cycle to account and compensate for any potential baseline drift in the measured loss, that may be caused by varying environmental conditions.

CAPS technology is superior to the traditional chemiluminescence method in both speed and precision since the sampled NO<sub>2</sub> does not require catalytic conversion to calculate a difference measurement.

Economically, the Model N500 CAPS NO<sub>x</sub> instrument is less costly to operate and maintain than traditional analyzers in that it uses less power and fewer components.

The section on Principles of Operation provides more detail on the behavior and technique of the CAPS method for NO<sub>x</sub> measurement.

## 1.1. SPECIFICATIONS

**Table 1-1. Specifications**

| Parameter               | Description  |                           |
|-------------------------|--|---------------------------|
| Ranges                  | Min: 0-5 ppb Max: 0-1 ppm NO <sub>x</sub> (full scale user-selectable) |                           |
| Measurement Units       | ppb, ppm, µg/m <sup>3</sup> , mg/m <sup>3</sup> (user-selectable)      |                           |
| Zero Noise              | < 0.05 ppb (RMS)   |                           |
| Span Noise              | < 0.2% of reading + 50 ppt (RMS)                                       |                           |
| Zero Drift              | < 0.2 ppb / 24 hours   |                           |
| Span Drift              | < 0.5% of reading / 24 hours   |                           |
| Lower Detectable Limit  | < 0.1 ppb  |                           |
| Response Time           | < 60 Seconds to 95%  |                           |
| Linearity               | 1% Full Scale  |                           |
| Precision               | 0.5% of reading above 5 ppb  |                           |
| Sample Flow Rate        | 1000 cc/min ±10%   |                           |
| AC Power                | Rating   | Typical Power Consumption |
|                         | 100 - 240 V~ 47-63 Hz 3.0 A  | 110 W                     |
| Power Entry Module Fuse | 5.0 A, 250 V AC, 5 mm x 20 mm, SLO-BLO                                 |                           |

| Parameter   | Description  |
|---|--|
| Communications  |  |
| Standard I/O  | 1 Ethernet: TCP/IP<br>1 RS-232 or RS-485 (user-configurable) (300 – 115,200 baud)<br>2 front panel USB device ports  |
| Optional I/O  | Universal Analog Output Board (all user-definable):<br>4 x isolated voltage outputs (5 V, 10 V)<br>3 x individually isolated current outputs (4-20 mA)<br>Digital I/O Expansion Board includes:<br>3 x isolated digital input controls (fixed)<br>5 x isolated digital output controls (user-definable)<br>3 x form C relay alarm outputs (user-definable) |
| Operating Temperature                                     | 0-40 °C (with US EPA approval)   |
| Humidity Range  | 0-95% RH, Non-Condensing   |
| Dimensions HxWxD  | 7" x 17" x 23.5" (178 x 432 x 597 mm)  |
| Weight  | 33 lbs (15 kg)   |
| Environmental Conditions                                  | <ul style="list-style-type: none"> <li>• Installation Category (Over Voltage Category ) II Pollution Degree 2</li> <li>• Intended for Indoor Use Only at Altitudes ≤ 2000m</li> </ul>  |
| Note: All specifications are based on constant conditions |  |

## 1.2. EPA DESIGNATION

Teledyne API's Model N500 CAPS NO<sub>x</sub> analyzer is officially designated as US EPA Automated Equivalent Method, Designation Number EQNA- 0320-256.

The official List of Designated Reference and Equivalent Methods is published in the U.S. Federal Register, <http://www3.epa.gov/tn/amtic/criteria.html>, and specifies the instrument's settings and configurations.

## 1.3. SAFETY

IEC/EN 61010-1:2010 (3<sup>rd</sup> Edition), Safety requirements for electrical equipment for measurement, control, and laboratory use.

CE: 2014/95/EU, Low-Voltage Directive

## 1.4. EMC

IEC/EN 61326-1, Class A Emissions/Industrial Immunity

FCC 47 CFR Part 15B, Class A Emissions

CE: 2014/108/EU, Electromagnetic Compatibility Directive

## 2. GETTING STARTED

This section addresses unpacking, connecting, and initializing the instrument, getting an overview of the menu system, and setting up/configuring the functions.

### 2.1. UNPACKING



#### CAUTION - GENERAL SAFETY HAZARD

To avoid personal injury, always use two persons to lift and carry the analyzer.

#### ATTENTION

#### COULD DAMAGE INSTRUMENT AND VOID WARRANTY

Printed Circuit Assemblies (PCAs) are sensitive to electro-static discharges too small to be felt by the human nervous system. Failure to use Electro-Static Discharge (ESD) protection when working with electronic assemblies will void the instrument warranty. Refer to the manual, Fundamentals of ESD, PN 04786, which can be downloaded from our website at <http://www.teledyne-api.com>.

#### ATTENTION

#### COULD DAMAGE INSTRUMENT AND VOID WARRANTY

Do not operate this instrument without first removing dust plugs from SAMPLE and EXHAUST ports on the rear panel.

#### Note

Teledyne API recommends that you store shipping containers and materials for future use if/when the instrument should be returned to the factory for repair and/or calibration service. See Warranty statement in this manual and Return Merchandise Authorization (RMA) on our Website at <http://www.teledyne-api.com>.

Verify that there is no apparent external shipping damage. If damage has occurred, please advise the shipper first, then Teledyne API.

Included with your instrument is a printed record of the final performance characterization performed on your instrument at the factory. This record, titled Final Test and Validation Data Sheet, is an important quality assurance and calibration record and should be placed in the quality records file for this instrument.

With no power to the unit, carefully remove the top cover of the instrument and check for internal shipping damage by carrying out the following steps:

1. Carefully remove the top cover and check for internal shipping damage.
  - a. Remove the screws located on the instrument's sides.
  - b. Slide the cover backward until it clears the instrument's front bezel.
  - c. Lift the cover straight up.
2. Inspect the interior of the instrument to ensure all circuit boards and other components are intact and securely seated.
3. Check the connectors of the various internal wiring harnesses and pneumatic hoses to ensure they are firmly and securely seated.
4. Verify that all of the optional hardware ordered with the unit has been installed. These are listed on the paperwork accompanying the instrument.



**WARNING – ELECTRICAL SHOCK HAZARD**

**Never disconnect PCAs, wiring harnesses or electronic subassemblies while under power.**

### 2.1.1. VENTILATION CLEARANCE

Whether the instrument is set up on a bench or installed in a rack, be sure to leave sufficient ventilation clearance.

**Table 2-1. Ventilation Clearance**


| AREA                           | MINIMUM REQUIRED CLEARANCE |
|--------------------------------|----------------------------|
| Back of the instrument         | 10 cm / 4 in               |
| Sides of the instrument        | 2.5 cm / 1 in              |
| Above and below the instrument | 2.5 cm / 1 in              |

## 2.2. INSTRUMENT LAYOUT

Instrument layout includes front panel, rear panel connectors, and the internal chassis layout.

### 2.2.1. FRONT PANEL

The front panel (Figure 2-1) includes two USB ports for peripheral device connections, which can be used with mouse and keyboard as alternatives to the touchscreen interface, or with flash drive for uploads/downloads (devices not included).



**Figure 2-1. Front Panel Layout**

## 2.2.2. REAR PANEL

Figure 2-2 shows the layout of the rear panel.






Figure 2-2. Rear Panel Layout

Table 2-2. Rear Panel Description

| COMPONENT   | FUNCTION   |
|---|--|
| <b>Model/specs label</b>  | Identifies the analyzer model number and provides power specs.   |
| <b>SAMPLE</b>   | Connect a gas line from the source of sample gas here. Calibration gases can also enter here on units without zero/span/shutoff valve options installed.   |
| <b>EXHAUST</b>  | Connect an exhaust gas line of not more than 10 meters long here that leads outside the shelter or immediate area surrounding the instrument. The line must be ¼" tubing or greater.   |
| <b>SPAN</b>   | On units with zero/span valve option installed, connect a gas line to the source of calibrated span gas here.  |
| <b>ZERO AIR</b>   | On units with zero/span valve option installed, attach a gas line to the source of zero air here. If a permeation oven, also known as internal zero/span valve (IZS), option is installed attach the zero air scrubber here.     |
| <b>Cooling fan</b>  | Pulls ambient air into chassis through side vents and exhausts through rear.   |
|  | <b>Power cable connector</b><br>Connector for three-prong cord to apply AC power to the analyzer.<br><b>CAUTION! The cord's power specifications (specs) MUST comply with the power specs on the analyzer's rear panel label</b> |
|   | <b>Power On/Off Switch</b><br>Rocker switch to power unit on or off.<br><b>CAUTION! Prior to powering OFF, use front panel button for preliminary internal "soft" power-down to protect components from damage.</b>              |
| <b>Fuse drawer</b>  | For circuit protection.  |
| <b>ANALOG OUT Option</b>  | For voltage or current loop outputs to a strip chart recorder and/or a data logger.  |
| <b>DIGITAL I/O Option</b>   | For remotely activating the zero and span calibration modes.   |
| <b>ETHERNET</b>   | Connector for network or Internet remote communication, using Ethernet cable.  |
| <b>COM</b>  | Serial communications port for RS-232 or RS-485.   |

## 2.2.3. INTERNAL CHASSIS

Figure 2-3 shows the internal chassis layout.



**Figure 2-3. Internal Chassis Layout**

## 2.3. CONNECTIONS AND STARTUP

This section presents the electrical (Section 2.3.1) and pneumatic (Section 2.3.2) connections for setting up and preparing the instrument for operation (Section 2.3.3).

### 2.3.1. ELECTRICAL CONNECTIONS

**Note**

To maintain compliance with EMC standards, cable length must be no greater than 3 meters for all I/O connections.



**WARNING – Electrical Shock Hazard**

- High Voltages are present inside the instrument's case.
- Power connection must have functioning ground connection.
- Do not defeat the ground wire on power plug.
- Turn off instrument power before disconnecting or connecting electrical subassemblies.
- Do not operate with cover off.



**CAUTION – Avoid Damage to the Instrument**

Ensure that the AC power voltage matches the voltage indicated on the instrument's rear panel before plugging it into line power.

#### 2.3.1.1. CONNECTING POWER

**ATTENTION**

**COULD DAMAGE INSTRUMENT AND VOID WARRANTY**

Never power off the instrument from the rear panel Hard Power switch before first placing the internal computerized components into deep sleep mode through the front panel Soft Power switch. Press and momentarily hold the front panel Soft Power switch, which triggers the Supervisory chip to safely shut down the internal components. The LED state then changes from solid lit to blinking, at which time the rear panel Hard Power switch can be used to power off the instrument.

Attach the power cord between the instrument's AC power connector and a power outlet capable of carrying at least the rated current at your AC voltage range and ensure that it is equipped with a functioning earth ground. It is important to adhere to all safety and cautionary messages.



### 2.3.1.2. CONNECTING ANALOG OUTPUTS OPTION

The optional rear panel Analog Output board offers several channels that can be mapped to reflect various operating values in the analyzer, including concentration values, temperatures, pressures, etc. These mappings are not configured by default and must be set by the user.

The four **voltage** outputs (0-5 V or 0-10 V) are isolated from the instrument but share a common ground. The three **current** outputs are individually isolated from each other and from the instrument.

To access these signals, attach a strip chart recorder and/or data-logger to the appropriate analog output connections, and configure through the Setup>Analog Outputs menu.




Figure 2-4. Analog Outputs Connectors Panel Option

Table 2-3. Analog Output Pin Assignments

| PIN                      | OUTPUT  | DESCRIPTION   |
|--------------------------|---------|---|
| Isolated Voltage Outputs |         |   |
| V1                       | V +     | User definable through the Setup>Analog Outputs menu. |
| RTN                      | Ground  |   |
| V2                       | V +     |   |
| RTN                      | Ground  |   |
| V3                       | V +     |   |
| RTN                      | Ground  |   |
| V4                       | V +     |   |
| RTN                      | Ground  |   |
| Isolated Current Outputs |         |   |
| I-1                      | I Out + | User definable through the Setup>Analog Outputs menu. |
| RTN                      | I Out - |   |
| I-2                      | I Out + |   |
| RTN                      | I Out - |   |
| I-3                      | I Out + |   |
| RTN                      | I Out - |   |

### 2.3.1.3. CONNECTING THE DIGITAL I/O EXPANSION BOARD OPTION

The connections on this board include three relay alarms, five digital outputs, and three isolated digital input controls. The **Relays** can be mapped to reflect various internal instrument conditions and states. The **Outputs** are isolated from the instrument and consist of open collector transistors with a common ground; they can be mapped to reflect various internal instrument conditions and states; they can be used to interface with devices that accept logic-level digital inputs, such as Programmable Logic Controllers (PLCs). The **Inputs** are also isolated but share the same ground as the Outputs; they will work with relays, open collectors, or 3.3 V – 24 V logic. Pull low to activate. DI1 and DI2 are fixed (not mappable) for remote zero and span calibrations.




Figure 2-5. Digital I/O Connector Panel Option

Table 2-4. Digital Input/Output Pin Assignments

| PIN                        | DESCRIPTION  |   |
|----------------------------|--|---|
| Relays                     |  |   |
| NO                         | RLY 1  | Relay Alarms, user-configurable through the Setup>Digital Outputs menu. |
| COM                        |  |   |
| NC                         |  |   |
| NO                         | RLY 2  |   |
| COM                        |  |   |
| NC                         |  |   |
| NO                         | RLY 3  |   |
| COM                        |  |   |
| NC                         |  |   |
| Digital Outputs and Inputs |  |   |
| DO1                        | Digital Outputs mappable in the Setup>Digital Outputs menu, and viewable in the Utilities>Diagnostics>Digital Outputs menu |   |
| DO2                        |  |   |
| DO3                        |  |   |
| DO4                        |  |   |
| DO5                        |  |   |
| GND                        | Ground   |   |
| DI1                        | Digital Input1 = Remote Zero Cal   |   |
| DI2                        | Digital Input2 = Remote Span Cal   |   |
| DI3                        | (Digital Input3 not used)  |   |
| DI3                        | View status in Utilities>Diagnostics>Digital Inputs menu   |   |
| GND                        | Ground   |   |

## 2.3.1.4. CONNECTING COMMUNICATIONS INTERFACES

### ETHERNET CONNECTION

For network or Internet communication with the analyzer, connect an Ethernet cable from the analyzer's rear panel Ethernet interface connector to an Ethernet port. Although the analyzer is shipped with DHCP enabled by default, it should be manually configured with a static IP address.

### SERIAL CONNECTION

Received from the factory, the analyzer COM port is set up for RS-232 communications with data communication equipment (DCE). This port can be reconfigured for RS-232 communications with data terminal equipment (DTE) or with RS-485 by jumpering the pins on JP1 as indicated in Table 2-5 (view/edit software settings Table 2-12).



#### WARNING – ELECTRICAL SHOCK HAZARD

Disconnect power before performing any operation that requires entry into the interior of the analyzer.




Figure 2-6. Mainboard JP1 Location and Pin Arrangements

Table 2-5. JP1 Configurations for Serial Communication

| Function            | Jumpers          | DSub Pins |         |
|---------------------|------------------|-----------|---------|
|                     |                  | 2         | 3       |
| DCE RS232 (default) | 1-2, 4-5, 9-12   | 232Tx     | 232Rx   |
| DTE RS232           | 2-3, 5-6, 9-12   | 232Rx     | 232Tx   |
| RS485               | 7-8, 10-11, 9-12 | 485A(-)   | 485B(+) |

View/edit the Communications parameters in the Setup>Comm>COM1 menu.

**RS-232**

- **Baud rate:** 115200 bits per second (baud)
- **Data Bits:** 8 data bits with 1 stop bit
- **Parity:** None

**RS-485**

- **Baud rate:** 19200 bits per second (baud)
- **Data Bits:** 8 data bits with 1 stop bit
- **Parity:** None

## 2.3.2. PNEUMATIC CONNECTIONS

This section provides pneumatic connection and setup instructions for basic and valve option configurations. Pneumatic flow diagrams are shown in Section 2.3.3. Calibration instructions are provided in Section 4.

Before making the pneumatic connections, carefully note the following cautionary and special messages:



**CAUTION – General Safety Hazard**

**Do not vent calibration gas, exhaust gas or sample gas into enclosed areas.**



**CAUTION – General Safety Hazard**

**In units with a permeation tube option installed, vacuum pump must be connected and powered on to maintain constant gas flow through the analyzer at all times. Insufficient gas flow allows gas to build up to levels that will contaminate the instrument or present a safety hazard to personnel.**

**Remove the permeation tube when taking the analyzer out of operation and store in sealed container (use the original shipping packaging).**

**(See Section 5.6.4 for instructions on how to remove the permeation tube when the unit is not in operation).**

**ATTENTION**

**COULD DAMAGE INSTRUMENT AND VOID WARRANTY**

**VENT PRESSURIZED GAS:**

When any gas (span, zero air, sample) is received from a pressurized manifold, always provide a vent to equalize the pressure with the ambient atmosphere before it enters the instrument to ensure that the gases input do not exceed the instrument's maximum inlet pressure, as well as to prevent back diffusion and pressure effects

**REMOVE DUST PLUGS:**

Remove dust plugs from rear panel exhaust and supply line fittings before powering on the instrument.

Keep dust plugs for reuse in future storage or shipping to prevent debris from entering the pneumatics.

**Important**

**IMPACT ON READINGS OR DATA**

- Sample and calibration gases should only come into contact with PTFE tubing.
- Do NOT place any mufflers or filters downstream of the pump, i.e., external to the instrument.
- Run a leak check once the appropriate pneumatic connections have been made; check all pneumatic fittings for leaks per Section 5.4.12.1 (or Section 5.4.12.2 for detailed check if any leaking is suspected).

### 2.3.2.1. CRITICAL TUBING, PRESSURE, VENTING AND EXHAUST REQUIREMENTS

The requirements presented in this section apply to all pneumatic connection instructions.

**Tubing:**

- PTFE material
- Outer diameter (OD) minimum ¼"
- Min/max length 2 meters to 10 meters

**Pressure:**

- All Sample gas pressure must be at ambient atmospheric pressure, no greater than 1.0 psig.

**Venting** (to prevent back diffusion and pressure effects):

- Run tubing outside the enclosure or at least away from immediate area surrounding the instrument.


**Exhaust Outlet:**

- Run tubing outside the enclosure.

**Calibration Gas Sources:**

- The source of calibration gas or zero air is also attached to the **SAMPLE** inlet, but only when a calibration operation is actually being performed.

### 2.3.2.2. BASIC CONNECTIONS FROM CALIBRATOR



**Figure 2-7. Gas Line Connections from Calibrator – Basic Configuration**

For the analyzer’s basic configuration, in addition to tubing, pressure, venting, and exhaust requirements set out in Section 2.3.2.1, attach the following pneumatic lines:

#### **SAMPLE GAS SOURCE**

Connect a sample gas line to the SAMPLE inlet.

#### **CALIBRATION GAS SOURCES**

**NO<sub>2</sub> CAL GAS & ZERO AIR SOURCES:** The source of calibration gas is attached to the SAMPLE inlet, but only when a calibration operation is actually being performed.


#### **VENTING**

Vent the output of the calibrator if calibrator not already vented.


#### **EXHAUST OUTLET**

Attach an exhaust line to the EXHAUST outlet fitting and vent outside the shelter.

### 2.3.3. PNEUMATIC FLOW DIAGRAMS



**Figure 2-8. Pneumatic Flow Diagram including Zero/Span Valve Option**



**Figure 2-9. Pneumatic Flow Diagram including IZS Valve Option**



## 2.3.4. STARTUP, FUNCTIONAL CHECKS AND CALIBRATION

We recommend reading Section 6 to become familiar with the principles of operation. .

**Note** It is expected that all cautionary messages are being followed.

The front panel Soft Power switch has a status LED that indicates whether:

- instrument is powered down (LED off)
- instrument powered on and internal components in deep sleep mode (LED blinking, achieved by pressing and momentarily holding the soft power button)
- instrument powered on and internal components are operating (LED solid lit, achieved when first powered up; must place in deep sleep mode before power off)

When the instrument is first started (Section 2.3.4.1), check its functionality (Section 2.3.4.3) and run an initial calibration (Section 2.3.4.4). Section 2.4 introduces the menu system, and Section 2.5 provides setup/customization instructions.

### 2.3.4.1. STARTUP

Upon initial startup, a sequence of status screens (Figure 2-10) appear prior to the Home page (Figure 2-11).




Figure 2-10. Status Screens at Startup

Upon any startup, this instrument should warm up for approximately one hour before reliable measurements can be taken.




Figure 2-11. Home Page Example

### 2.3.4.2. ALERTS: WARNINGS AND OTHER MESSAGES

Because internal temperatures and other conditions may be outside the specified limits during the warm-up period, the software will suppress most Alerts for 45 minutes after power up. The Alerts page (Figure 2-12) shows the status of any active warning conditions or user-configured Events. (Section 2.4.3 provides more detailed information about Alerts, and Section 2.5.2 addresses Events).

Alerts can be viewed and cleared via either the Alerts menu or the Alerts shortcut (Caution symbol, bottom right corner of the screen). Although these alerts can be cleared from the Active Alerts page, a history of all alerts remains in the Utilities>Alerts Log page.

Navigate to the Active Alert page via the Alerts menu on Home screen.

(Also view a list of all active and past Alerts and Events via Utilities>Alerts Log).




Figure 2-12. Viewing Active Alerts Page

If Alerts about warning conditions persist after the warm-up period or after being cleared, investigate their cause using the troubleshooting guidelines in Section 5.7.1.

### 2.3.4.3. FUNCTIONAL CHECKS

After warm-up, verify that the software properly supports any hardware options that are installed (Setup>Instrument menu), and that the instrument is functioning within allowable operating parameters. View the Dashboard page to check that parameters show expected/reasonable values. (If any functional parameters are not displayed, configure the Dashboard through the Setup>Dashboard menu to add them; see Section 2.4.2).

These functions are also useful tools for diagnosing problems (information provided in Section 5.7.2).




Figure 2-13. Sample Dashboard Page

### 2.3.4.4. CALIBRATION

Before operation begins, the analyzer requires zero and span calibrations followed by a Titration Efficiency (TE) check. Also, any time an analyzer is moved, or its configuration changed, it must be calibrated. The method for performing a calibration differs slightly depending on whether or not any of the available internal valve options are installed. Follow the appropriate calibration instructions presented in Section 4.

## 2.4. MENU OVERVIEW


Table 2-6 describes the main menus and provides cross-references to the respective sections with configuration details.

**Table 2-6. Menu Overview**


| MENU               | DESCRIPTION   | LOCATION  |                |
|--------------------|---|---|----------------|
| <b>Home</b>        | View and plot concentration readings and other selectable parameter values (Figure 2-15).   | Section 2.4.1   |                |
| <b>Dashboard</b>   | View user-selected parameters and their values, some of which can be displayed in a live-plot graph (Figure 2-16).  | Section 2.4.2   |                |
| <b>Alerts</b>      | View and clear active Alerts that were triggered by factory-defined Events as well as user-defined Events. (Active and past Alerts are recorded in the Utilities>Alerts Log). | Section 2.4.3   |                |
| <b>Calibration</b> | Run calibrations on the NO <sub>2</sub> and NO <sub>x</sub> channels.   | Sections 2.4.4 and 4  |                |
| <b>Utilities</b>   | View logs, download data and firmware updates, copy configurations between instruments, and run diagnostics.  | Section 2.4.5   |                |
| <b>Setup</b>       | Configure a variety of features and functions through these submenus for customized operation.  | Section 2.5   |                |
|                    | Datalogging   | Track and record concentration and calibration data and selectable diagnostic parameters, the reports for which can be viewed in the Utilities>Datalog View menu (Section 2.4.5) and downloaded to a flash drive via the Utilities>USB Utilities menu (Section 2.4.5).<br>Also, select configured Events (Section 2.5.2) and create customized triggers for data logging functions. | Section 2.5.1  |
|                    | Events  | Select parameters and define the conditions by which they are to be flagged and recorded in the Alerts log (Section 2.4.3) when they are triggered. Once configured, Events can be used to trigger Datalogs. (Section 2.5.1). Note that some Events are predefined and are not editable.  | Section 2.5.2  |
|                    | Dashboard   | Monitor instrument functionality (Figure 2-13) via selectable parameters.   | Section 2.5.3  |
|                    | Auto Cal<br>(with valve options)  | When zero/span valve options installed, configure sequences for automatic calibration functions.  | Section 4.3    |
|                    | Vars  | Manually adjust several software variables that define specific operational parameters.   | Section 2.5.5  |
|                    | Homescreen  | Select up to three parameters to be displayed in the meters (Figure 2-14).  | Section 2.5.6  |
|                    | Digital I/O<br>(option)   | Map the rear-panel digital outputs to a variety of voltage or current signals present in the instrument to monitor the status of operating conditions or custom Events.   | Section 2.5.7  |
|                    | Analog<br>Outputs<br>(option)   | Send user-selected parameter readings in the form of user-defined voltage or current loop signals as outputs to a strip chart recorder and/or the data logger.  | Section 2.5.8  |
|                    | Instrument  | View product and system information, including list of options, if any; view network settings; view/adjust Date and Time settings*; and check for firmware updates when connected to a network that is connected to the Internet.<br>*Time Zone change requires special procedures (Section 5.5).   | Section 2.5.7  |
|                    | Comm  | View and configure network and serial communications.   | Section 2.5.10 |

## 2.4.1. HOME PAGE

Figure 2-14 presents an orientation to the main display screen; Figure 2-15 shows that pressing the gas name or its concentration value or a meter below displays a live plot of their respective readings. Section 2.5.6 provides configuration instructions.



**Figure 2-14. User Interface Orientation**



**Figure 2-15. Concentration and Stability Graph (top) and Meter Graph (bottom)**

## 2.4.2. DASHBOARD

The Dashboard displays an array of user-selectable parameters and their values (Section 2.5.3 provides configuration instructions). If there is a graphing icon in the upper right corner of a parameter, pressing that parameter displays a live plot of its readings as in Figure 2-16.

Select a graphable parameter in the Dashboard page to view a live plot.





Figure 2-16. Dashboard Page

## 2.4.3. ALERTS

Alerts are notifications triggered by specific criteria having been met by either factory-defined conditions (standard and not editable) or user-defined Events (Section 2.5.2). The Active Alerts page shows the status of any active warning conditions or Events that have been triggered.

When Alerts are triggered, a caution symbol appears in both the Alerts menu tab and in the bottom right corner of the software interface, which serves as a shortcut to the Alerts page from any other page. View a list of currently active Alerts by pressing either the Alerts menu on the Home screen or by pressing the Alerts shortcut (Figure 2-17).

While Alerts can be cleared from the Active Alerts page, they remain recorded in the Utilities>Alerts Log menu.




**Figure 2-17. Active Alerts Page**

Alerts can be configured as either latching (appears in Active Alerts screen when Event is triggered and must be cleared by the user) or non-latching (Active Alerts screen continuously updates based on the Event criteria, clearing on its own). See Section 2.5.2.

To clear Alerts from the Active Alerts page, either check individual boxes to choose specific Alerts, or check the Select All box to choose all Alerts, then press the Clear Selected button.




When all Alerts are cleared, the Alerts menu tab no longer shows the caution symbol, and a green LED replaces the caution symbol in the bottom right corner of the interface (Figure 2-18). However, Alerts can reappear if the conditions causing them are not resolved. For troubleshooting guidance, refer to Section 5.7.



**Figure 2-18. Active Alerts Cleared**

Alerts and Events remain recorded in the Utilities>Alerts Log (Figure 2-19).



**Figure 2-19. Utilities>Alerts Log of Active and Past Alerts and Events**

## 2.4.4. CALIBRATION

The Calibration menu is used for zero/span/multipoint calibrations and for external calibration with valve options installed. Calibration procedures are presented in Section 4.

## 2.4.5. UTILITIES

The Utilities menu has a variety of functions as described next in Table 2-7.

**Table 2-7. Utilities Submenu Descriptions**

| UTILITIES MENU          | DESCRIPTION   |
|-------------------------|---|
| <b>Datalog View</b>     | Displays the data logs that were configured via the Setup>Data Logging menu. From this list a log can be selected and filters applied to view the desired data. (For details on setting up and running the Data Logger, see Section 2.5.1).   |
| <b>Alerts Log</b>       | Displays a history of alerts that are triggered by factory-defined and user-defined Events, such as warnings and alarms (See Section 2.5.2 for Events configuration).   |
| <b>USB Utilities</b>    | Serves multiple purposes using a flash drive connected to the instrument's front panel USB port: <ul style="list-style-type: none"> <li>• download data from the instrument's Data Acquisition System (DAS), the Data Logger, to a flash drive (Section 2.5.1.3)</li> <li>• update firmware (Section 5.3)</li> <li>• transfer instrument configuration from/to other same-model instruments (Section 2.6)</li> <li>• download a basic operation functionality report (Section 5.3).</li> </ul>  |
| <b>Diagnostics</b>      | Provides access to various pages that facilitate troubleshooting.   |
| Analog Outputs (Option) | Show the voltage or current signals for the functions selected and configured in the Setup>Analog Outputs menu. (Section 2.3.1.2 presents the rear panel connections).  |
| Digital I/O (Option)    | Show whether specific available features are turned ON or OFF; input controls are fixed; output controls and relays are configurable in the Setup>Digital Outputs menu. (Section 2.3.1.3 presents the rear panel connections).  |
| Manual AREF             | At 20 minutes after power up and periodically thereafter when in Sample mode, the analyzer conducts a background measurement, known as an auto reference (AREF), whereby the sample is routed through an internal charcoal scrubber prior to the AREF valve. This measurement accounts for drift in the baseline loss. A manual AREF is recommended prior to initial calibration (Section 4.2) or after general maintenance.<br><br>Please note that AREF becomes disabled during calibration if the instrument stays in calibration mode (CALZ or CALS) for longer than it would normally take to run a calibration. To ensure that AREF is enabled, return to Sample mode after conducting a calibration. |
| O3 Gen Override         | Used to override the Ozone Generator state when needed, such as for service (Section 5.7.7.3).  |

## 2.4.6. SETUP

The Setup menu is used to configure the instrument's various features, functions, and data log. Section 2.5 provides details for the menus under Setup.

## 2.5. SETUP MENU: FEATURES/FUNCTIONS CONFIGURATION


Use the Setup menu to configure the instrument's software features, to gather information on the instrument's performance, and to configure and access data from the Datalogger, the instrument's internal data acquisition system (DAS). Once the setups are complete, the saved configurations can be downloaded to a USB drive through the Utilities>USB Utilities menu and uploaded to other instruments of the same model (Section 2.6).

## 2.5.1. SETUP>DATA LOGGING (DATA ACQUISITION SYSTEM, DAS)


The Datalogger can be configured to capture and store user-defined data, which then can be viewed in the Alerts page, if elected, as well as downloaded from the instrument to a USB flash drive or using NumaView™ Remote software for examination and analysis.

Figure 2-20 shows a new log; Figure 2-21 shows a sample existing log, which can be edited or deleted, and Figure 2-22 provides illustrated instructions for setting up a new log, with Sections 2.5.1.1 and 2.5.1.2 providing additional details.



To transfer captured instrument data to a flash drive, see Section 2.5.1.3.




**Figure 2-20. Datalog Configuration, New Log Page**




**Figure 2-21. Datalog Configuration, Existing Log**



When **Periodic** is selected as the Trigger Type, the field below it is populated with the Interval and Date/Time windows.



When **Conditional** is selected as the Trigger Type, the field below it is populated with the Trigger Tag and Condition definition windows.




(Please refer to the section on Configuring Trigger Types for details).

**Figure 2-22. Creating a New Data Log**

The parameters available in the list of Log Tags include the names of Events configured in the Events page (Section 2.5.2).

### 2.5.1.1. CONFIGURING TRIGGER TYPES: PERIODIC

The Periodic trigger is a timer-based trigger that is used to log data at a specific time interval. Periodic Trigger requires an interval that is set to number of minutes and a start time that is set to date and clock time.



Press the Interval field and use the keypad pop-up to edit the amount of time in seconds (default 60) between capturing data for the selected Log Tags.

Press the Start Time field and use the Time and Date pop-ups to select the time and the date to start capturing data for the parameters that were selected in the Log Tags field.

(toggle)


**Figure 2-23. Datalog Periodic Trigger Configuration**

### 2.5.1.2. CONFIGURING TRIGGER TYPES: CONDITIONAL

Conditional Trigger tracks/records data for user-selected parameters that meet specified conditions.

Press the Trigger Tag field and select a parameter to be tracked/logged. A default Condition associated with the selected Tag will populate the Condition field.

Press the Condition field to select a different choice from the condition list.



Either the Threshold field appears, or the Low and High fields appear if a condition requires either a threshold value or range values. Press a Value... field and use the keypad.

(Other condition lists include True/False and Enabled/Disabled)

Figure 2-24. Datalog - Conditional Trigger Configuration

### 2.5.1.3. DOWNLOADING DAS (DATA ACQUISITION SYSTEM) DATA

To download DAS data collected by the Datalogger from the instrument to a flash drive, navigate to the Utilities>USB Utilities>DAS Download menu.

1. Insert a flash drive into a front panel USB port and wait for the Status field to Indicate that the drive has been detected; available buttons will be enabled.




Figure 2-25. DAS Download Page

2. Select all or define a period from which to download the collected data.
3. Press the Download button, and when complete, as indicated in the Status field, press the Done button (changed from “Cancel”) and remove the flash drive.


## 2.5.2. SETUP>EVENTS

Events are occurrences that relate to any operating function and are used to define the conditions that can be set to trigger Alerts (Section 2.4.3). Events can provide diagnostic information about the instrument, typically referred to as “Warnings”, or they can provide other information on instrument functionality, such as concentration alarms. Some Events are standard and not editable while others are user-configurable, described here. Existing Events are listed in the Events page (Figure 2-26) under the Setup menu.



**Figure 2-26. Events List**

Access the Events Configuration page either from the Active Alerts page (Alerts Menu) by pressing the configuration button, or through the Home>Setup>Events menu (Figure 2-26). Press ADD to create a new Event (refer to Figure 2-27 for details), or select an existing Event to either Edit or Delete it (Figure 2-29).



**Figure 2-27. Event Configuration**

- **Enabled**  allows the choice of whether to track and record the Event (uncheck this box to “turn off” or deactivate the Event without deleting it). An Event must be enabled in order to use the Visible and the Latching options.
- **Visible**  allows the choice of whether or not to display the Event in the Alerts page when it is triggered (it will still be recorded and can be viewed in the Utilities>Alerts Log). To use this option, the Event must be enabled.
- **Latching**  allows the choice of whether or not to keep an Event visible even if the conditions that triggered it were to correct themselves. (Latching requires that the user interact with the Active Alerts screen to manually clear the Alert and internal Event state. Non-latching allows the entry in the Active Alerts screen and the internal Event state to continuously update based on the Event criteria, requiring no user interaction to clear the Alert or Event state).






Figure 2-28. Configured Event Sample

### 2.5.2.1. EDITING OR DELETING EVENTS

Select an Event from the list (Figure 2-26) and press the Edit button to view or edit the details (Figure 2-28), or press the Delete button to delete the Event.




Figure 2-29. Edit or Delete an Event

### 2.5.2.2. USING EVENTS AS TRIGGERS FOR DATA LOGGING


Events can also be used to create customized triggers for data logging functions. The name entered in the Name field of the Events Configuration page will appear in the list of Log Tags of the Datalog Configuration page. The Data Logger is presented in Section 2.5.1.

### 2.5.3. SETUP>DASHBOARD

To navigate to Dashboard Configuration, either press the configuration shortcut in the Dashboard page, or use the Setup>Dashboard menu.



Configuration Shortcut



Configuration editing:

To add a parameter to the Dashboard, select it in the “Available Tags” list and press the right-pointing arrows button. (Checkmarks in the “Available Tags” list indicate parameters that are currently in the Dashboard).

To remove a parameter from the Dashboard, select it from the “Dashboard” list and press the left-pointing arrows button.



Figure 2-30. Dashboard Display and Configuration

### 2.5.4. SETUP>AUTOCAL (WITH VALVE OPTION)

Auto Cal is available with installed valve options (see Section 4.3).

## 2.5.5. SETUP>VARS

Vars are user-adjustable software variables that define operational parameters automatically set by the instrument's firmware. Access the menu to see the list of variables; select a variable to view its description; use the Edit button to change its setting(s).

**Table 2-8. List of Variables with Descriptions**

| VARIABLE  | DESCRIPTION   |
|---|---|
| <p>This list includes several of the most common Vars; selecting any Var in the NumaView™ software interface will display its description in the information field to its right. Depending on configuration, some, all, or more of these variables appear in your instrument's Vars menu.</p> |   |
| Background Periodic Report Upload   | Allows/disallows uploading of basic functionality reports to a Web services "cloud" for TAPI Technical Support to view. (Frequency can be edited in Setup>Vars>Report Upload Interval).   |
| Conc Precision (or PRIGAS/SECGAS Precision)   | Sets the number of significant digits to the right of the decimal point display of concentration and stability values. ("PRIGAS" = primary gas with two or more other gases; "SECGAS" = secondary gas)  |
| Daylight Savings Enable   | Enable or disable Daylight Savings Time (also see Setup>Instrument>Date/Time Settings)  |
| Dilution Factor (option)  | <p>Sets the instrument to compensate for diluted sample gas, such as in continuous emission monitoring (CEM) where the quality of gas in a smokestack is being tested and the sampling method used to remove the gas from the stack dilutes the gas. Once the degree of dilution is known, this feature allows the user to add an appropriate scaling factor to the analyzer's NO, NO<sub>2</sub> and NO<sub>x</sub> concentration calculations so that the undiluted values for measurement range and concentration are shown on the instrument's front panel display and reported via the instrument's various outputs.</p> <p>Set the appropriate units of measure (Setup&gt;Vars&gt;User Units).<br/>           Select the reporting range mode (Setup&gt;Vars&gt;Range Mode) and set the reporting range upper limit (Setup&gt;Analog Output). Ensure that the upper span limit entered for the reporting range is the maximum expected concentration of the undiluted gas.<br/>           Set the dilution factor as a gain, e.g., a value of 20 means 20 parts diluent and 1 part sample gas (Setup&gt;Vars&gt;Dilution Factor).<br/>           Calibrate the analyzer; ensure that the calibration span gas is either supplied through the same dilution system as the sample gas or has an appropriately lower actual concentration.</p> |
| Dynamic Span Enable   | Dynamic span automatically adjusts the offsets and slopes of the NO <sub>2</sub> and NO <sub>x</sub> response when performing a span point calibration during an AutoCal (Section 4.3).   |
| Dynamic Zero Enable   | Dynamic zero automatically adjusts offset and slope of the NO <sub>2</sub> and NO <sub>x</sub> response when performing a zero point calibration during an AutoCal (Section 4.3).   |
| Instrument ID   | Set unique identifier number for the instrument when it is connected with other instruments on the same Ethernet LAN, or when applying MODBUS protocol.   |
| Measure Mode  | Set gas measure mode.   |
| Ozone Lamp Setpoint   | (Adjustment is dependent on low Titration Efficiency, Section 4.4).   |
| System Hours  | Total system runtime hours  |
| Titration Efficiency  | Used for checking or calibrating efficiency of ozone titrating NO (Section 4.4).  |

## 2.5.6. SETUP>HOMESCREEN

To select a parameter (“tag”) for display in each of the three meters at the bottom of the Home page, navigate to the Homescreen configuration page through either the Setup>Homescreen menu or from Home page using the configuration icon (Figure 2-31).



**Figure 2-31. Homescreen Configuration**

An orientation to the Homescreen was presented in Section 2.4.1, including Figure 2-14 and Figure 2-15.

## 2.5.7. SETUP>DIGITAL OUTPUTS (OPTION)

Specify the function of each digital output (connected through the rear panel Digital I/O connector, Figure 2-5) by mapping the outputs to a selection of “Signals” present in the instrument. Create custom “Signals” in the Setup>Events menu (Section 2.5.2). The three Relays can also be connected, and functions assigned.

To map Digital Outputs to Signals, select a pin in the Outputs list (DO1 thru DO5), then make a selection from the Signals list and press the Map button; if/as needed, change the polarity by pressing the Polarity button. Save any changes by pressing the Apply button, or discard the changes by pressing the Home or the back button (a pop-up provides a warning that the changes will be lost, and will prompt for confirmation to apply changes or not). Map the Digital Relays in the same manner.

Navigate to the Utilities>Diagnostics>Digital Outputs menu to change the state (ON/OFF) of individual digital outputs.





Figure 2-32. Digital Outputs Setup

## 2.5.8. SETUP>ANALOG OUTPUTS (OPTION)

Map the four user-configurable Analog Outputs to any of a wide variety of “Signals” present in the instrument and customize their respective configurations.



**Figure 2-33. Analog Output Option Configuration Example**

Refer to Figure 2-33 for the following settings in each of the fields:

- Signal Out: select a Signal for the output.
- Min/Max: edit minimum and maximum values associated with the selected Signal.
- Calibration Type:
  - AUTO for group calibration (Figure 2-34) of the analog outputs (cannot be selected when Current is selected for the Range)
  - MANUAL for individual calibration (Figure 2-35) of analog outputs where manual adjustments can be made (the only calibration type allowed when Current is selected for the Range). See Sections 2.5.8.1 and 2.5.8.2 .
- Range: assign a voltage or select Current (refer to Table 2-9).
- Recorder Offset: add a zero offset for recording slightly negative readings from noise around the zero point.
- Allow Overrange: check to allow a  $\pm 5\%$  over-range; uncheck to disable over-range if the recording device is sensitive to excess voltage or current.



Figure 2-34. Analog Outputs Group Calibration Screen




Figure 2-35. Analog Outputs Option Manual Calibration Screen (AOUT2 Example)

Table 2-9. Analog Output Option Voltage/Current Range

| RANGE <sup>1</sup> | RANGE SPAN | MINIMUM OUTPUT | MAXIMUM OUTPUT |
|--------------------|------------|----------------|----------------|
| 5V                 | 0-5 VDC    | - 1 VDC        | 6 VDC          |
| 10V                | 0-10 VDC   | - 2 VDC        | 12 VDC         |
| Current            | 4-20 mA    | 3 mA           | 21 mA          |

<sup>1</sup> Each range is usable from -5% to +5% of the rated span.

For manual calibration adjustments, see Section 2.5.8.1 for voltage and Section 2.5.8.2 for current.

### 2.5.8.1. MANUAL CALIBRATION OF VOLTAGE RANGE ANALOG OUTPUTS OPTION

It is possible to manually calibrate the voltages by using a voltmeter connected across the output terminals (Figure 2-36) and changing the output signal level when Manual is selected in the Calibration Type field of the Analog Output screen (Figure 2-35).




Figure 2-36. Setup for Checking / Calibrating DCV Analog Output Signal Levels

### 2.5.8.2. MANUAL ADJUSTMENT OF CURRENT RANGE ANALOG OUTPUTS OPTION

This option changes the normal DC voltage output to a 4-20 milliamp signal. Adjusting the signal zero and span levels of the current loop output is done by raising or lowering the voltage output of the D-to-A converter circuitry. This raises or lowers the signal level produced by the current loop option circuitry.


The software allows this adjustment to be made in 100, 10 or 1 count increments. Since the exact amount by which the current signal is changed per D-to-A count varies from output-to-output and instrument-to-instrument, you will need to measure the change in the signal levels with a separate, current meter placed in series with the output circuit. See Figure 2-4 for pin assignments and diagram of the analog output connector.



#### CAUTION!


Do not exceed 60 V peak voltage between current loop outputs and instrument ground.





**Figure 2-37. Setup for Checking / Calibrating Current Output Signal Levels**

An alternate method for measuring the output of the Current Loop converter is to connect a 250 ohm  $\pm 1\%$  resistor across the current loop output in lieu of the current meter (see Figure 2-4 for pin assignments and diagram of the analog output connector). This allows the use of a voltmeter connected across the resistor to measure converter output as VDC or mVDC.



**Figure 2-38. Alternative Setup Using 250Ω Resistor for Checking Current Output Signal Levels**

In this case, follow the procedure above but adjust the output for the following values:

**Table 2-10. Current Loop Output Check**

| % FS | Voltage across Resistor for 4-20 mA |
|------|-------------------------------------|
| 0    | 1000 mVDC                           |
| 100  | 5000 mVDC                           |

## 2.5.9. SETUP>INSTRUMENT

As presented in Table 2-11, view product and system information and network settings, edit network settings, and perform certain maintenance tasks.

**Table 2-11. Setup>Instrument Menu**

| MENU               | DESCRIPTION   |
|--------------------|---|
| Product Info       | View Model, Part, and Serial Numbers and Package and Driver Versions, and options information.  |
| System Info        | View Windows and RAM information.   |
| Network Settings   | View the network settings (configurable through the Setup>Comm>Network Settings menu).  |
| Date/Time Settings | Adjust date, hour, and minutes, select a time zone*, and set the system clock to automatically adjust for Daylight Savings Time or not. (Also see Setup>Vars>Daylight Savings Enable). *Time Zone change requires a special procedure; see Maintenance Section 5.5. |
| NTP Time Settings  | Configure Network Time Protocol settings for clock synchronization.   |
| Language           | Select an available language.   |
| Remote Update      | When an instrument is connected to a network that is connected to the Internet, follow the instructions on this Remote Update page to check for and activate software/firmware updates. (Also refer to Section 5.3).  |

## 2.5.10. SETUP>COMM (COMMUNICATIONS)

This menu is for specifying the various communications configurations.

### 2.5.10.1. COM1

Configure the instrument's COM port to operate in modes listed in Table 2-12.

**Table 2-12. COM1 Setup**

| MODE                    | DESCRIPTION  |
|-------------------------|--|
| Baud Rate               | Set the baud rate.   |
| Command Prompt Display  | Enable/disable a command prompt to be displayed when in terminal mode.   |
| Data Bits               | Set the data bits to 7 or 8 (typically set in conjunction with Parity and Stop bits).  |
| Echo and Line Editing   | Enable/disable character echoing and line editing.   |
| Handshaking Mode        | Choose SOFTWARE handshaking for data flow control (do NOT use SOFTWARE handshaking mode when using MODBUS RTU for Protocol mode; select only HARDWARE or OFF for MODBUS RTU), or HARDWARE for CTS/RTS style hardwired transmission handshaking. (This style of data transmission handshaking is commonly used with modems or terminal emulation protocols). Or choose to turn OFF handshaking. |
| Hardware Error Checking | Enable/disable hardware error checking.  |
| Hardware FIFO           | Enable/disable the hardware First In – First Out (FIFO) for improving data transfer rate for that COM port.  |
| Modem Connection        | Select either a modem connection or a direct cable connection.   |

| MODE              | DESCRIPTION   |
|-------------------|---|
| Modem Init String | Input an initialization string to enable the modem to communicate.  |
| Parity            | Select odd, or even, or no parity (typically set in conjunction with Data Bits and Stop Bits).  |
| Protocol          | If selecting a MODBUS protocol, see Handshaking Mode notes, this table; MODBUS Registers are presented in Appendix A, this manual. Also see <a href="http://www.modbus.org">www.modbus.org</a> .  |
| Quiet Mode        | Enable/disable Quiet mode, which suppresses any feedback from the analyzer (such as warning messages) to the remote device and is typically used when the port is communicating with a computer program where such intermittent messages might cause communication problems.<br>Such feedback is still available, but a command must be issued to receive them. |
| Security          | Enable/disable the requirement for a password for this serial port to respond. The only command that is active is the request-for-help command, ? <b>CR</b> .   |
| Stop bits         | Select either 0 or 1 stop bit (typically set in conjunction with Parity and Data bits).   |

### 2.5.10.2. TCP PORT1

TCP Port1 allows choosing whether or not to display the command prompt, editing the Port 1 number for defining the terminal control port by which terminal emulation software addresses the instrument, such as Internet or NumaView™ Remote software, and enabling or disabling security on this port.

### 2.5.10.3. TCP PORT2

TCP Port2 is configured with the port number for MODBUS (Registers provided in Appendix A).

### 2.5.10.4. NETWORK SETTINGS

The Setup>Comm>Network Settings menu is for Ethernet configuration. The address settings default to automatic configuration by Dynamic Host Configuration Protocol (DHCP). Most users will want to configure the instrument with a static IP address: click the Static radio button to manually assign a static IP address (consult your network administrator and see Table 2-13 for information).





Figure 2-39. Communications Configuration, Network Settings

Table 2-13. LAN/Ethernet Configuration Properties

| PROPERTY               | DESCRIPTION   |
|------------------------|---|
| <b>IP address</b>      | A string of four packets of 1 to 3 numbers each (e.g. 192.168.76.55.) is the internet protocol address of the instrument itself.  |
| <b>Subnet Mask</b>     | A string of four packets of 1 to 3 numbers each (e.g. 255.255.252.0) number that masks an IP address, and divides the IP address into network address and host address and identifies the LAN to which the device is connected. All addressable devices and computers on a LAN must have the same subnet mask. Any transmissions sent to devices with different subnets are assumed to be outside of the LAN and are routed through the gateway computer onto the Internet. |
| <b>Default Gateway</b> | A string of numbers very similar to the Instrument IP address (e.g. 192.168.76.1.) that is the address of the computer used by your LAN and serves as a router to access the Internet or another network.   |

## 2.6. TRANSFERRING CONFIGURATION TO OTHER INSTRUMENTS

Once an instrument is configured, the same configuration can be copied to other instruments of the same Model. This encompasses essentially anything the user can configure and does not apply to instrument-specific settings such as those that are configured at the factory for calibration.



**Figure 2-40. Configuration Transfer**

1. In the source instrument, go to the Home>Utilities>USB Utilities>General page.
2. Insert a flash drive into either of the two front panel USB ports.
3. When the Status field Indicates that the USB drive has been detected, press the “Download Configuration from Instrument” Start button.
4. When the Status field Indicates that the download is complete, remove the flash drive.
5. In the target instrument, go to the Home>Utilities>USB Utilities>General page.
6. Insert a flash drive into either of the two front panel USB ports.
7. When the Status field Indicates that the USB drive has been detected, press the “Upload Configuration to Instrument” Start button.
8. When the Status field Indicates that the upload is complete, remove the flash drive.

### 3. COMMUNICATIONS AND REMOTE OPERATION

This instrument's rear panel connections include an Ethernet port and a serial communications port. Connection instructions were provided in Section 2.3.1.4. Configuration information was provided in Section 2.5.10.

Data acquisition is set up through the Datalogger (Section 2.5.1).

#### 3.1. SERIAL COMMUNICATION

The rear panel COM port operates on the RS-232 protocol (default configuration is DCE RS-232), or it can be configured for DTE RS-232 or for half-duplex RS-485 operation (Section 2.3.1.4).

Referring to Table 2-12, use the SETUP>COMM menu to view/edit the communications settings for the COM port.

#### 3.2. ETHERNET

When using the Ethernet interface, the analyzer can be connected to any Ethernet network via low-cost network hubs, switches or routers. The interface operates as a standard TCP/IP device on port 3000. This allows a remote computer to connect through the network to the analyzer using NumaView™ Remote, terminal emulators or other programs.

The Ethernet connector has two LEDs that are on the connector itself, indicating its current operating status.

**Table 3-1. Ethernet Status Indicators**

| LED              | FUNCTION                                 |
|------------------|--|
| green (link)     | On when connection to the LAN is valid.  |
| amber (activity) | Flickers during any activity on the LAN. |

The analyzer is shipped with DHCP enabled by default. This allows the instrument to be connected to a network or router with a DHCP server; however, it should be configured with a Static IP address as soon as practical. See Section 2.5.10.4 for configuration details.

#### 3.3. NUMAVIEW™ REMOTE

For remote operation and data capture through an Ethernet connection, please refer to the NumaView™ Remote Software User Guide, PN 08492, available on our website.

## 4. CALIBRATION

This section is organized into subsections as follows:

**SECTION 4.1 – Important Precalibration Information:** contains important information you should know before calibrating the instrument.

**SECTION 4.2 – Calibration Procedures:** describes the procedure for manually checking calibration and performing actual calibration of the instrument.

**SECTION 4.3 – Automatic Zero/Span Cal/Check (Auto Cal):** describes the procedure for using the AutoCal feature to check calibration or to calibrate the instrument. (The AutoCal feature requires that either the zero/span valve option or the internal span gas generator option be installed and operating).

**SECTION 4.4 – Titration Efficiency (TE) Check/Calibration:** describes corrections made for titration effects.

**SECTION 4.5 – EPA Protocol Calibration:** provides links to the US EPA website for references regarding calibration with EPA protocols.

### 4.1. IMPORTANT PRECALIBRATION INFORMATION

#### Note

A start-up period of 4-5 hours is recommended prior to calibrating the analyzer.

#### 4.1.1. CALIBRATION REQUIREMENTS

The following equipment, supplies, and expendables are required for calibration:

- Zero Air Generator (e.g., T701)
- Span gas source
- Gas lines - all gas line materials should be stainless steel or Teflon-type (PTFE or FEP).
- High-concentration NO<sub>2</sub> gas transported over long distances may require stainless steel lines to avoid oxidation of NO<sub>2</sub> due to the possibility of O<sub>2</sub> diffusing into the tubing.

Optional equipment:

Calibrator with photometer option (e.g., T700U) for use with NO and GPT to span the analyzer.

A recording device such as a strip-chart recorder and/or data logger.

For electronic documentation, the internal data acquisition system (DAS) can be used by configuring the Datalogger through the Setup>Data Logging menu; Section 2.5.1).

The method for performing an initial calibration for the analyzer differs between the standard instrument and those with options.

- See Section 4.2.1 for instructions for initial calibration of the analyzer in its base configuration.
- See Section 4.2.2 for information regarding setup and calibration of the analyzer with Z/S Valve options.

**Note**

**Zero air and span gases must be supplied at twice the instrument's specified gas flow rate.**

### 4.1.2. ZERO AIR

Zero air or zero calibration gas is similar in chemical composition to the measured medium but without the gas to be measured by the analyzer. A zero generator, such as the Teledyne API Model T701, can be used.

### 4.1.3. CALIBRATION (SPAN) GAS

NO<sub>2</sub> cylinder Calibration gas is diluted, using a T700 calibrator, to match the chemical composition of the type of gas being measured at near full scale of the desired reporting range. Thus, it is recommended that the span gas be of a concentration equal to 80% of the measurement range for your application.

Alternatively, if a calibrator is available that contains a trusted source of stable ozone, e.g., Teledyne API Model T700U with certified photometer, it is possible to use that O<sub>3</sub> output directly to obtain the NO<sub>2</sub> concentration using GPT.



**CAUTION!**

**If the presence of ozone is detected at any time, power down the instrument and contact Teledyne API Technical Support as soon as possible:**

**+1 800-324-5190 or email: [api-techsupport@teledyne.com](mailto:api-techsupport@teledyne.com)**

### 4.1.4. SPAN GAS FOR MULTIPOINT CALIBRATION

Some applications, such as EPA monitoring, require a multipoint calibration where span gases of different concentrations are needed. We recommend using an NO<sub>2</sub> gas cylinder of higher concentration combined with a gas dilution calibrator such as the Teledyne API Model T700. Calibrators mix high concentration span gas with zero air to accurately produce span gas of the desired concentration. Linearity profiles can be automated with these models and run unattended overnight.

If a dynamic dilution system is used to dilute high concentration gas standards to low, ambient concentrations, ensure that the NO<sub>2</sub> concentration of the reference gas matches the dilution range of the calibrator. (Section 2.5.5 contains information about the dilution option).



Choose the NO<sub>2</sub> gas concentration so that the dynamic dilution system operates in its mid-range and not at the extremes of its dilution capabilities.

EXAMPLE:

- A dilution calibrator with 10-10000 dilution ratio will not be able to accurately dilute a 5000 ppm NO<sub>2</sub> gas to a final concentration of 500 ppb, as this would operate at the very extreme dilution setting.
- A 100 ppm NO<sub>2</sub> gas in nitrogen is much more suitable to calibrate the analyzer (dilution ratio of 200, in the mid-range of the system's capabilities).

#### 4.1.5. NO<sub>2</sub> PERMEATION TUBES

Teledyne API offers an optional internal span gas module that utilizes an NO<sub>2</sub> permeation tube as a span gas source. The accuracy of these devices is only about  $\pm 5\%$ . Whereas this may be sufficient for quick, daily calibration checks, we recommend using certified NO<sub>2</sub> gases for accurate calibration.

##### CAUTION!



**Insufficient gas flow allows gas to build up to levels that will contaminate the instrument or present a safety hazard to personnel.**

**In units with a permeation tube installed, either the tube must be removed and stored in a sealed container (use original container that tube was shipped in) during periods of non-operation, or vacuum pump must be connected and powered on to maintain constant gas flow through the analyzer at all times.**

(See Section 5.6.4 for removal instructions).

#### 4.1.6. DATA RECORDING DEVICES

A strip chart recorder, data acquisition system or digital data acquisition system should be used to record data from either the Ethernet, serial or analog outputs.

- If analog readings are used, the response of the recording system should be checked against a NIST traceable voltage source or meter.
- Data recording devices should be capable of bi-polar operation so that negative readings can be recorded.

For electronic data recording, the analyzers provide an internal data logger, which is configured through the Setup>Data Logger menu (Section 2.5.1).

NumaView™ Remote is a remote control program, which is also available as a convenient and powerful tool for data viewing and handling, download, storage, quick check and plotting.

## 4.2. CALIBRATION PROCEDURES

Check that the pneumatic connections for the specific instrument configuration are as instructed in Section 2.3.2. Calibration procedures include setting the expected span gas concentration (see Note below).

Verify User Units setting and Titration Efficiency (TE) value:

- User Units (unit of Measure): PPB
- TE is 96% or greater (check Titrator Eff in Dashboard; if less than .96, adjust in Setup>Vars – see Section 4.4. No adjustment needed if  $\geq .96$ ).

Then perform the calibration:

- Perform a Zero calibration using zero air on both NO<sub>2</sub> & NO<sub>x</sub> channels.
- Perform a Span calibration on both the NO<sub>2</sub> & NO<sub>x</sub> channels using a known concentration of NO<sub>2</sub> span gas.
- Perform the Titration Efficiency (TE) check using a known concentration of NO span gas. No adjustment of TE value is required if the efficiency is good.

### Note

**The span gas concentration should be 80% of range of concentration values likely to be encountered in your application.**

To calibrate or to perform a calibration check for basic configuration instruments, see Section 4.2.1.


To calibrate or to perform a calibration check for instruments with valve options, see Section 4.2.2.

To perform a calibration check for instruments with the internal span gas generator, see Section 4.3.

### 4.2.1. CALIBRATION AND CHECK PROCEDURES FOR BASIC CONFIGURATION

Although this section uses the Calibration menu for both calibration check and actual calibration, a check does not require the Calibration menu. Instead, while in Home page, simply flow the zero air or the NO<sub>2</sub> span gas through the Sample port, and check the reading after the Stability falls below 1.0 PPB (either in the gas graph or in the Dashboard).

Otherwise, follow the steps presented in Sections 4.2.1.1 and 4.2.1.2.



**Figure 4-1. Multi-Point Calibration Page**

#### 4.2.1.1. ZERO CALIBRATION CHECK AND ACTUAL CALIBRATION

1. Go to the Calibration>M-P menu.
2. Input Zero air through the Sample port and press the Start button.
3. Either check or calibrate as follows:

**CHECK ONLY:**

- a. Wait for reading to stabilize.
- b. Press Stop and check the reading.

**ACTUAL CALIBRATION:**

- a. Press the Zero button.
- b. Press Stop and check the reading.

#### 4.2.1.2. SPAN CALIBRATION CHECK AND ACTUAL CALIBRATION

1. While still in the Calibration>M-P menu, input NO<sub>2</sub> Span gas through the Sample port and press the Start button.
2. Either check or calibrate as follows:

**CHECK ONLY:**


- a. Wait to reach stability, then press Stop.
- b. Record the reading(s).

**ACTUAL CALIBRATION:**

- a. Press the Set Span Target button and enter the NO<sub>2</sub> and NO<sub>x</sub> concentrations.
- b. Verify the concentration reading is the same as the NO<sub>2</sub> concentration being supplied.
- c. If correct, wait to reach stability, then press the Span button.
- d. In the Cal Result window, press OK.

3. Press the Stop button and return to Home screen.
4. In the Dashboard, check and record the Slope(s) and the Offset(s).

## 4.2.2. CALIBRATION AND CHECK PROCEDURES WITH VALVE OPTIONS INSTALLED



**Figure 4-2. Zero and Span Calibration Screens**

Follow the instructions in Section 4.2.1, except instead of the M-P menu, go to the Calibration>Zero menu for Zero cal and to the Calibration>Span menu for NO<sub>2</sub> Span cal.

### 4.2.2.1. USE OF ZERO/SPAN VALVE WITH DIGITAL EXPANSION BOARD OPTION

Digital inputs are available for controlling calibration and calibration checks, when the Digital I/O Board option is installed. Instructions for setup and use of this option are outlined in Section 2.3.1.3.

When the Digital Inputs are activated for at least 5 seconds, the instrument switches into zero, low span or high span mode and the internal zero/span valves will be automatically switched to the appropriate configuration.

- The remote calibration Digital Inputs may be activated in any order.
- It is recommended that the Digital Inputs remain closed for at least 10 minutes to establish a reliable reading.
- The instrument will stay in the selected mode for as long as the Input remains closed.

If Digital Inputs are being used in conjunction with the analyzer’s AutoCal (see Section 4.3) feature and the AutoCal attribute “Calibrate” is enabled (selection box is checked), the analyzer will not recalibrate the analyzer UNTIL the Input is opened. At this point, the new calibration values will be recorded before the instrument returns to SAMPLE mode.

If the AutoCal attribute “Calibrate” is disabled (selection box is unchecked), the instrument will return to SAMPLE mode, leaving the instrument’s internal calibration variables unchanged.

### 4.3. AUTOMATIC ZERO/SPAN CAL/CHECK (AUTO CAL)

The Auto Cal feature allows unattended periodic operation of the ZERO/SPAN valve options by using the instrument’s internal time of day clock. Auto Cal operates by executing preprogrammed calibrations or calibration checks set up by the user to initiate the various calibration states of the analyzer and to open and close valves appropriately. It is possible to set up and run up to three separate preprogrammed calibrations or calibration checks (labeled # 1, 2 and 3). Each calibration or check can operate in one of three modes (Zero, Low or High), or be disabled.

Table 4-1 and Table 4-2 show how to set up the operating states of each calibration or check, and Table 4-3 shows how to program the execution of each.




Figure 4-3. Auto Cal Page

**Table 4-1. AUTO CAL States**

| MODE NAME        | ACTION   |
|------------------|--|
| <b>Enabled</b>   | <input checked="" type="checkbox"/> enables the sequence.<br><input type="checkbox"/> disables the sequence.   |
| <b>Calibrate</b> | <input checked="" type="checkbox"/> enables an actual calibration when the Enabled box is also <input checked="" type="checkbox"/><br><input type="checkbox"/> allows a calibration check when the Enabled box is also <input checked="" type="checkbox"/> .   |
| <b>Zero</b>      | <input checked="" type="checkbox"/> causes the sequence to perform a Zero calibration when both the Calibrate and Enabled boxes are also <input checked="" type="checkbox"/><br><input checked="" type="checkbox"/> causes a Zero check when the Enabled box is also <input checked="" type="checkbox"/> and the Calibrate box is unchecked ( <input type="checkbox"/> )<br><input type="checkbox"/> disables Zero calibration and check                                   |
| <b>Low</b>       | <input checked="" type="checkbox"/> causes the sequence to perform a Low Span calibration when both the Calibrate and Enabled boxes are also <input checked="" type="checkbox"/><br><input checked="" type="checkbox"/> causes a Low Span check when the Enabled box is also <input checked="" type="checkbox"/> and the Calibrate box is unchecked ( <input type="checkbox"/> )<br><input type="checkbox"/> disables Low Span calibration and check                       |
| <b>High</b>      | <input checked="" type="checkbox"/> causes the sequence to perform a High Span concentration calibration when both the Calibrate and Enabled boxes are also <input checked="" type="checkbox"/><br><input checked="" type="checkbox"/> causes a High Span check when the Enabled box is also <input checked="" type="checkbox"/> and the Calibrate box is unchecked ( <input type="checkbox"/> )<br><input type="checkbox"/> disables the High Span calibration and check. |

Table 4-2 shows how the selection boxes would be enabled/disabled for calibration checks and calibrations.

**Table 4-2. Auto Cal Setup Combinations**

| MODE          | ACTION    | STATE                               |                                     |                                     |                                     |                                     |
|---------------|-----------|-------------------------------------|-------------------------------------|-------------------------------------|-------------------------------------|-------------------------------------|
|               |           | Enabled                             | Calibrate                           | Zero                                | Low                                 | High                                |
| Zero          | Check     | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
|               | Calibrate | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| Low           | Check     | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            | <input checked="" type="checkbox"/> | <input type="checkbox"/>            |
|               | Calibrate | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input checked="" type="checkbox"/> | <input type="checkbox"/>            |
| High          | Check     | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            | <input type="checkbox"/>            | <input checked="" type="checkbox"/> |
|               | Calibrate | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            | <input checked="" type="checkbox"/> |
| Zero Low High | Check     | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> |
|               | Calibrate | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> |

For each sequence, there are four parameters that control operational details: Date, Time (both in the Start field), Interval, and Duration, as presented in Table 4-3.

**Table 4-3. Auto Cal Programming Sequence Execution**

| ATTRIBUTE | ACTION   |
|-----------|--|
| Start     | When the Enabled box is “on” <input checked="" type="checkbox"/> , the Sequence (identified by its number) begins on the date and time shown in the configurable Start field. (Click the field for the pop-up window and toggle between the Time (Hour/Minutes) and the Date (Year/Month/Day) attributes to edit as needed). |
| Interval  | Number of minutes to skip between each Sequence execution. (Click the field to input the number of minutes in the pop-up window).  |
| Duration  | Number of minutes that each Sequence execution is to run. (Click the field to input the number of minutes in the pop-up window).   |

**Important**

**IMPACT ON READINGS OR DATA**

- The programmed **STARTING\_TIME** must be a minimum of 5 minutes later than the real time clock for setting real time clock (Setup>Instrument, Section 2.5.7).
- Avoid setting two or more sequences at the same time of the day.
- Any new sequence that is initiated whether from a timer, the COM ports or the contact closure inputs will override any sequence that is in progress.
- It is recommended that calibrations be performed using external sources of Zero Air and Span Gas whose accuracy is traceable to EPA standards.

## 4.4. TITRATION EFFICIENCY (TE) CHECK/CALIBRATION

The efficiency at which the O<sub>3</sub> titrates NO into NO<sub>2</sub> in the sample gas can change over time, thereby affecting the accuracy of the instrument’s NO<sub>x</sub> measurement. To compensate for this and maintain accuracy, the firmware includes a Titration Efficiency (TE) gain factor that is used to adjust the calculated NO gas concentration. This gain factor is stored in the analyzer’s memory.

The method is to deliver a known concentration of NO span gas, having first calibrated the instrument with NO<sub>2</sub>, typically 80%-90% of the desired range, and after the instrument stabilizes, perform a TE adjustment in the Vars menu, if required (Setup>Vars>Titrate Eff). This adjusts the calculated NO concentration, which is then added to the measured NO<sub>2</sub> concentration to calculate the corrected NO<sub>x</sub> value.

## 4.5. EPA PROTOCOL CALIBRATION

When running this instrument for U.S. EPA compliance, always calibrate prior to use, adhering to the EPA designation requirements for this instrument. (The official List of Designated Reference and Equivalent Methods is published in the U.S. Federal Register: <http://www3.epa.gov/ttn/amtic/criteria.html>; this List specifies the settings and configurations for EPA calibration protocol). Pay strict attention to the built-in warning features, periodic inspection, regular zero/span checks, regular test parameter evaluation for predictive diagnostics and data analysis, and routine maintenance. Any instrument(s) supplying the zero air and span calibration gasses used must themselves be calibrated, and that calibration must be traceable to an EPA/NIST primary standard.

Comply with Code of Federal Regulations, Title 40 (downloadable from the U.S. Government Publishing Office at <http://www.gpo.gov/fdsys/>) and with Quality Assurance Guidance documents (available on the EPA website: <http://www3.epa.gov/ttn/amtic/qalist.html>). Give special attention to specific regulations regarding the use and operation of ambient NO<sub>x</sub> analyzers using Cavity Attenuated Phase Shift Spectroscopy method.



## 5. MAINTENANCE AND SERVICE

Although the Model N500 analyzer requires little service, a few simple procedures should be performed regularly to ensure that it continues to operate accurately and reliably over its lifetime. In general, the exterior can be wiped down with a lightly damp cloth. Service and troubleshooting are covered in Section 5.7.

### ATTENTION

**COULD DAMAGE INSTRUMENT AND VOID WARRANTY**  
Avoid spraying anything directly onto any part of the analyzer.

### 5.1. MAINTENANCE SCHEDULE

Table 5-1 shows a typical maintenance schedule. The actual frequency of performing these procedures can vary depending on the operating environment. Additionally, in some cases, there are local regulations or standards that also need to be considered.

In certain environments (e.g., dusty, very high ambient pollutant levels) some maintenance procedures may need to be performed more often than shown.



#### WARNING – ELECTRICAL SHOCK HAZARD

Disconnect power before performing any of the following operations that require entry into the interior of the analyzer.



#### CAUTION – QUALIFIED PERSONNEL

These maintenance procedures must be performed by qualified technicians only.

### Important

#### IMPACT ON READINGS OR DATA

A span and zero calibration check (see CAL CHECK REQ'D Column of Table 5-1) must be performed following some of the maintenance procedures listed herein. To perform a CHECK of the instrument's Zero or Span Calibration, refer to Sections 4.2.1.1 and 4.2.1.2, respectively.

**DO NOT** press the Zero or Span buttons at the end of each operation (actual calibration), as this will reset the stored values for OFFSET and SLOPE and alter the instrument's calibration.

Alternatively, use the Auto Cal feature described in Section 4.3 with the CALIBRATE attribute set to OFF (not enabled).

**Table 5-1. Maintenance Schedule**

| ITEM                               | ACTION                            | FREQ  | CAL CHECK REQ'D            | DATE PERFORMED |  |  |  |  |  |  |  |  |  |  |  |  |  |
|------------------------------------|-----------------------------------|---|----------------------------|----------------|--|--|--|--|--|--|--|--|--|--|--|--|--|
|                                    |                                   |   |                            |                |  |  |  |  |  |  |  |  |  |  |  |  |  |
| Dashboard functions                | Review and evaluate               | Weekly  | No                         |                |  |  |  |  |  |  |  |  |  |  |  |  |  |
| Zero/span check                    | Evaluate offset and slope         | Weekly  | No                         |                |  |  |  |  |  |  |  |  |  |  |  |  |  |
| Zero/span calibration              | Zero and span calibration         | Every 3 months  | Yes                        |                |  |  |  |  |  |  |  |  |  |  |  |  |  |
| Sample filter                      | Change sample filter              | Annually (may need more frequently in a high dust load environment) | No                         |                |  |  |  |  |  |  |  |  |  |  |  |  |  |
| AREF filter assembly               | Change                            | Annually  | Yes                        |                |  |  |  |  |  |  |  |  |  |  |  |  |  |
| Spectrometer mirrors               | Contact Technical Support         | As necessary due to excessive Measured Loss                         | Yes                        |                |  |  |  |  |  |  |  |  |  |  |  |  |  |
| Sample conditioner filter assembly | Change                            | Only if necessary (contact Technical Support)                       | Yes                        |                |  |  |  |  |  |  |  |  |  |  |  |  |  |
| Pneumatic sub-system               | Check for leaks in gas flow paths | Annually or after repairs involving pneumatics                      | Yes, if a leak is repaired |                |  |  |  |  |  |  |  |  |  |  |  |  |  |
| Internal Pump                      | Replace                           | Measured Flow less than 800 cm <sup>3</sup> /min                    | Yes                        |                |  |  |  |  |  |  |  |  |  |  |  |  |  |

## 5.2. PREDICTIVE DIAGNOSTICS

Predictive diagnostic functions, including failure warnings and alarms built into the analyzer’s firmware, aid in determining whether and when repairs are necessary.

Dashboard Functions can also be used to predict failures by looking at how their values change over time, compared to the values recorded on the printed record of the *Final Test and Validation Data Sheet*. The internal data logger is a convenient way to record and track these changes (set up through the Data Logger, Section 2.5.1). Use NumaView™ Remote to download and review this data from a remote location.


The following table, checked weekly, can be used as a basis for taking action as these values change with time.

**Table 5-2. Predictive Uses for Dashboard Functions**

| FUNCTION                           | EXPECTED  | ACTUAL                   | INTERPRETATION & ACTION   |
|------------------------------------|---|--------------------------|---|
| <b>Cell Press</b><br>(pressure)    | Constant within atmospheric changes               | Fluctuating              | Developing leak in pneumatic system. Check for leaks.             |
|                                    |   | Slowly increasing        | Developing leak in pneumatic system. Check for leaks. Pump aging. |
|                                    |   | Slowly decreasing        | Sample filter getting clogged with dust. Replace sample filter.   |
| <b>AREF</b>                        | Constant within $\pm 100$ Mm-1 of check-out value | Significantly increasing | Developing AREF valve failure. Replace valve.                     |
|                                    |   |                          | Developing leak in pneumatic system. Check for leaks.             |
|                                    |   |                          | Debris on mirrors. Contact Technical Support.                     |
| <b>Gas Conc</b><br>(concentration) | Constant for known gas concentration              | Decreasing over time     | Developing leak in pneumatic system. Check for leaks.             |

### 5.3. OPERATIONAL HEALTH CHECKS

Navigate to the Utilities>USB Utilities>Report menu (Figure 5-1) to download a report on the basic operations of the instrument. To download the report for your own viewing on a computer or to send to others, insert a flash drive into a front panel USB port and press the Download button, which is enabled when the instrument detects the flash drive.



**Figure 5-1: Report Generation Page**

The report can also be set to generate periodically and sent to a Web services “cloud” where it is available for viewing by Teledyne API technical support personnel. Set this function with two Vars:

Setup>Vars>Upload Report to Cloud: set to True.

Setup>Vars>Report Upload>Interval: edit the number of hours between report uploads.

## 5.4. SOFTWARE/FIRMWARE UPDATES

There are two ways to check for and acquire updates: either remotely or manually.

### 5.4.1. REMOTE UPDATES

The instrument must be connected to a network that is connected to the Internet. In the Setup>Instrument menu, select the Remote Update menu and press the Check for Updates button. If an update is available, it can be downloaded through this page.




Figure 5-2. Remote Update Page

### 5.4.2. MANUAL RELOAD/UPDATE PROCEDURES

To reload or update firmware, first contact Technical Support to obtain the applicable file(s): [api-techsupport@teledyne.com](mailto:api-techsupport@teledyne.com) /+1 800-324-5190.

1. Follow Technical Support's instructions for copying the firmware files to a flash drive.
2. Go to the Utilities>USB Utilities>General menu.




Figure 5-3. Manual Update Page (and other utilities)

3. Insert a flash drive into a front panel USB port and wait for the Status field to indicate that the drive has been detected.


4. In the Update Firmware field, press the Check button for the instrument to determine whether the firmware on the flash drive is more recent than what is currently installed. Once it's been determined that the firmware is new, the Install button will be enabled; if the firmware version on the flash drive is the same as or older than the current firmware of the instrument, the Install button will not be enabled.
5. Press the Install button and note the messages in the Status field at the bottom of the page. Use the Cancel button if necessary.
6. When complete, as indicated in the Status field, press the Done button, which replaces the Cancel button, and remove the flash drive.
7. Power off and restart the instrument to complete the new firmware installation.

## 5.5. TIME ZONE CHANGES

There is an option to change between 12-hour and 24-hour format in the Setup>Vars menu (System Time Format). Effectively changing the Time Zone requires a specific procedure as follows:

1. In Setup>Instrument>Date/Time Settings select the applicable Time Zone.
2. Allow adequate time for the selected Time Zone to be properly accepted.
3. Verify: return to Home page then back to the Date/Time Settings page and check that the selected Time Zone is now highlighted.
4. Without making any other changes, power OFF the instrument and power ON again.
5. Once restarted, return to the Date/Time Settings page where the newly selected Time Zone should be highlighted. (If not, it means that not enough time had passed for the instrument to accept the change before the power was cycled OFF).
6. After the Time Zone is implemented first (Steps 1 through 5), then other changes to the date and/or time can be made, and recycling the power is not necessary.

- ① Time zone change must be set **first**.
- ② **Wait**. Allow sufficient time to accept new Time Zone.
- ③ **Verify**. Return to Home page, then return to Date/Time Settings page.
- ④ After correct Time Zone is displayed, **power recycle** the instrument.
- ⑤ Only after Time Zone is selected and instrument rebooted, can other changes to date and/or time be made effectively.



Changes to date and/or time do **not** require a reboot.

**Figure 5-4. Time Zone Change Requirements**

## 5.6. HARDWARE MAINTENANCE PROCEDURES

Perform the following procedures as standard maintenance per Table 5-1.

### ATTENTION

#### COULD DAMAGE INSTRUMENT AND VOID WARRANTY

Ensure there is no power to the unit so that there is no vacuum of any kind when changing anything in the pneumatic flow path. Dust entering the Optical Cell erroneously will contaminate the mirrors.

### 5.6.1. REPLACING THE SAMPLE FILTER

Inspect the particulate filter often for signs of plugging or contamination.

To change the filter:

1. Turn OFF the analyzer to prevent drawing debris into the instrument.
2. Open the hinged rear panel and disconnect the pneumatic fittings, using the appropriate wrenches, and remove the disposable sample filter.
3. Insert new filter and reconnect pneumatic fittings.
4. Close the front panel




Figure 5-5. Replacing the Sample Filter

### 5.6.2. REPLACING THE AREF SCRUBBER AND FILTER ASSEMBLY

The AREF filter assembly consists of the AREF DFU filter, which is a backup particulate filter, and the charcoal filter, which is an NO<sub>x</sub> gas scrubber.

1. Turn OFF the analyzer (this is important to avoid potential for contamination of the measurement cell), unplug the power cord and remove the cover.
2. Using snips, remove the zip tie holding the pair of fittings to the mounting bracket.

3. Using a wrench, disconnect the fitting that is connected to the AREF filter. Then disconnect the fitting that is connected to the charcoal filter. This will free up the AREF filter assembly.
4. Remove the AREF filter assembly from the mounting brackets.
5. Remove the fittings that connect the two filters to one another.
6. Install replacement filters and reconnect AREF assembly and fittings in reverse order.




Figure 5-6. Pneumatic Layout of AREF Filter and Charcoal Filter

### 5.6.3. REPLACING THE INTERNAL PUMP

To replace the internal pump:

1. Turn OFF the analyzer (this is important to avoid potential for contamination of the measurement cell).
2. Remove the top cover of the analyzer.
3. Locate the vacuum pump and using snips, remove the zip ties that secure the four hoses that connect to the pump (two on each side)

(It may be a good idea to mark the tubes with color codes to be sure they are reconnected properly).

4. Remove the four tubes from the hose barb connections on the pump.
5. Remove the four screws that hold the pump/bracket assembly to the bottom of the chassis.
6. Disconnect the power connector labeled “pump” located about 4 inches down the black/red cable coming from the top of the pump.
7. Pull pump assembly out and set aside.
8. Install replacement pump/bracket assembly.
9. Be sure to reconnect tubes in the proper orientation and zip tie to secure.



10. Replace screws for the pump mounting bracket into the bottom of the chassis, using caution against pinching wires or pneumatic tubes under bracket.
11. Connect the power connector.
12. Once complete, replace the instrument cover and perform a system leak check procedure.




Figure 5-7. Internal Pump

#### 5.6.4. CHANGING THE IZS PERMEATION TUBE OPTION

1. Turn off the analyzer, unplug the power cord and remove the cover.
2. Locate the oven with the internal zero span (IZS) gas permeation tube.
3. Remove the top layer of insulation if necessary.
4. Unscrew the black aluminum cover of the oven (3 screws) using a medium Phillips-head screwdriver.
  - Leave the fittings and tubing connected to the cover.
  - **Never apply power to the analyzer when the internal pneumatics are disconnected.**
5. Remove the old permeation tube and replace it with the new tube (or store the permeation tube in its original container if the instrument will not be operated for several or more hours).
  - Ensure that the tube is placed into the larger of two holes and that the open permeation end of the tube (plastic) is facing up.
6. Re-attach the cover with three screws and return the insulation if removed.
  - Ensure that the three screws are tightened evenly.
7. Replace the analyzer cover, plug the power cord back in and turn on the analyzer.
8. Carry out a span check to see if the new permeation tube works properly (see Section 4).
9. The permeation rate may need several days to stabilize.

#### ATTENTION

#### COULD DAMAGE INSTRUMENT AND VOID WARRANTY

**Do not leave instrument turned off for more than 8 hours without removing the permeation tube. Do not ship the instrument without removing the permeation tube. The tube continues to emit NO<sub>2</sub>, even at room temperature and will contaminate the entire instrument.**

## 5.6.5. CHECKING FOR PNEUMATIC LEAKS

This section covers a simple leak check and a detailed leak check.



### CAUTION - TECHNICAL INFORMATION

**Do not exceed 15 psi when pressurizing the system during either Simple or Detailed checks.**

### 5.6.5.1. DETAILED PRESSURE LEAK CHECK

Obtain a leak checker that contains a small pump, shut-off valve, and pressure gauge. Alternatively, a tank of pressurized gas, with the two-stage regulator adjusted to  $\leq 10$  psi, a shutoff valve and a pressure gauge may be used.

#### ATTENTION

#### COULD DAMAGE INSTRUMENT AND VOID WARRANTY

**Once tube fittings have been wetted with soap solution under a pressurized system, do not apply vacuum as this will cause soap solution to be sucked into the instrument, contaminating inside surfaces.**

1. Turn OFF power to the instrument and remove the instrument cover.
2. Install a leak checker on the **sample inlet** at the rear panel.
3. Cap rear panel ports and cap the pump port.
  - If zero/span valves are installed, disconnect the tubing from the zero and span gas ports and cap the ports (Figure 2-2).
4. Pressurize the instrument with the leak checker, allowing enough time to fully pressurize the instrument.
  - When performing a pressure leak test, always pressurize through the **Sample port (not Exhaust)**.
  - Do not exceed 10 psi pressure.
  - Relieve pressure by removing cap fitting at **Exhaust port**.
5. Once the leak has been located and repaired, the leak-down rate of the indicated pressure should be less than 1 psi in 5 minutes after the pressure is turned off. Replace the instrument cover and restart the analyzer.
6. If the leak still cannot be found, check each tube connection (fittings, hose clamps) with soap bubble solution, looking for fine bubbles.
  - Pressurize the instrument with the leak checker first, allowing enough time to fully pressurize the instrument.
  - Do not exceed 10 psi pressure.
  - Wet the bench with soap solution last.
7. Clean surfaces from soap solution, reconnect the sample and pump lines and replace the instrument cover.
8. Restart the analyzer.

### 5.6.5.2. PERFORMING A SAMPLE FLOW CHECK

#### Important

#### IMPACT ON READINGS OR DATA

Use an external calibrated flow meter capable of measuring flows between 0 and 1000 cm<sup>3</sup>/min to measure the gas flow rate through the analyzer.

Sample flow checks are useful for monitoring the actual flow of the instrument. A decreasing sample flow may point to slowly clogging pneumatic paths, most likely the sample flow restrictor or the sample filter. To perform a sample flow check:

1. Disconnect the sample inlet tubing from the rear panel SAMPLE port.
2. Attach the outlet port of a flow meter to the sample inlet port on the rear panel.
  - Ensure that the inlet to the flow meter is at atmospheric pressure.
3. Check that the sample flow measured with the external flow meter is within specification.
4. If sample flow is out of specification, first check for kinks in the tubing, then check for leaks.
5. Once kinks and leaks are ruled out, first replace the sample flow restrictor, and if the sample flow is still out of specification, replace the sample filter.
6. If flow is still out of spec, call Technical Support.

## 5.7. SERVICE AND TROUBLESHOOTING

This section contains methods to identify the source of performance problems with the analyzer and procedures to service the instrument.



#### CAUTION

The operations outlined in this section must be performed by qualified maintenance personnel only.



#### WARNING – RISK OF ELECTRICAL SHOCK

Some operations need to be carried out with the analyzer open and running.

Exercise caution to avoid electrical shocks and electrostatic or mechanical damage to the analyzer.

Do not drop tools into the analyzer or leave them after your procedures.

Do not short or touch electric connections with metallic tools while operating inside the analyzer.

Use common sense when operating inside a running analyzer.

The analyzer has been designed so that problems can be rapidly detected, evaluated and repaired. During operation, it continuously performs diagnostic tests and provides the ability to evaluate its key operating parameters without disturbing monitoring operations.

A systematic approach to troubleshooting will generally consist of the following five steps:

1. Note any Alerts and take corrective action as necessary (see Table 5-3).
2. Examine the values of all basic functions in the Dashboard and compare them to factory values. Note any major deviations from the factory values and take corrective action.
3. Use the internal electronic status LEDs to determine whether the electronic communication channels are operating properly. Refer to Figure 5-8, Figure 5-9, and Figure 5-10
4. Suspect a leak first!
  - Customer service data indicates that the majority of all problems are eventually traced to leaks in the internal pneumatics of the analyzer or the diluent gas and source gases delivery systems.
  - Check for gas flow problems such as clogged or blocked internal/external gas lines, damaged seals, punctured gas lines, damaged / malfunctioning pumps, etc.
5. Follow the procedures defined in Section 2.3.4.3 to confirm that the analyzer's vital functions are working.

### 5.7.1. FAULT DIAGNOSIS WITH ALERTS

Table 5-3 lists brief descriptions of warning Alerts that may occur during start up and describes their possible causes for diagnosis and troubleshooting.

It should be noted that if more than two or three warning Alerts occur at the same time, it is often an indication that some fundamental sub-system (power supply, smart module, mainboard) has failed rather than an indication of the specific failures referenced by the warnings.

**Table 5-3. Warning Alerts, Fault Conditions and Possible Causes**

| WARNING                             | FAULT CONDITION   | POSSIBLE CAUSES  |
|-------------------------------------|---|--|
| <b>AUTO REF WARNING</b>             | AREF value outside allowable limit.   | Drift in baseline loss due to large leak.<br>Sample filter bypassed. Mirrors may be dirty – contact Technical Support to confirm |
| <b>CANNOT DYN SPAN <sup>1</sup></b> | Dynamic Span operation failed. (Contact closure span calibration failed while <i>DYN_SPAN</i> was set to <i>ON</i> ). | Measured concentration value is too high or low<br>Concentration Slope value to high or too low                                  |
| <b>CANNOT DYN ZERO <sup>2</sup></b> | Dynamic Zero operation failed. (Contact closure zero calibration failed while <i>DYN_ZERO</i> was set to <i>ON</i> ). | Measured concentration value is too high<br>Concentration Offset value to high   |

| WARNING                   | FAULT CONDITION  | POSSIBLE CAUSES   |
|---------------------------|--|---|
| <b>CELL PRESS WARN</b>    | Cell pressure is too high or too low for accurate NO, NO <sub>2</sub> , NO <sub>x</sub> readings.<br>(<5 in-Hg or > 35 in-Hg).<br>Normally 29.92 in-Hg at sea level decreasing at 1 in-Hg per 1000 ft of altitude (with no flow – pump disconnected).    | If Cell Pressure is < 5 in-Hg:<br>•Failed Pressure Sensor/circuitry<br><br>If Cell Pressure is > 35 in-Hg:<br>•Bad Pressure Sensor/circuitry<br>•Pressure too high at Sample Inlet.   |
| <b>CONFIG INITIALIZED</b> | Configuration and Calibration data reset to original Factory state or erased.  | Failed Disk on Module<br>User erased data   |
| <b>DATA INITIALIZED</b>   | Data Storage in DAS was erased before the last power up occurred.  | Failed Disk-on-Module<br>User cleared data.   |
| <b>SAMPLE PRESS WARN</b>  | Sample pressure is too high or too low for accurate NO, NO <sub>2</sub> , NO <sub>x</sub> readings.<br>(<15 in-Hg or > 35 in-Hg).<br>Normally 29.92 in-Hg at sea level decreasing at 1 in-Hg per 1000 ft of altitude (with no flow – pump disconnected). | If Sample Pressure is < 15 in-Hg:<br>•Blocked Particulate Filter<br>•Blocked Sample Inlet/Gas Line<br>•Failed Pressure Sensor/circuitry<br><br>If Sample Pressure is > 35 in-Hg:<br>•Bad Pressure Sensor/circuitry<br>•Pressure too high at Sample Inlet. |
| <b>SYSTEM RESET</b>       | The computer has rebooted.   | This message occurs at power on.<br>If it is confirmed that power has not been interrupted:<br>Failed +5 VDC power<br>Fatal Error caused software to restart<br>Loose connector/wiring  |

<sup>1</sup> Clears the next time successful zero calibration is performed.

<sup>2</sup> Clears the next time successful span calibration is performed.

## 5.7.2. FAULT DIAGNOSIS WITH DASHBOARD FUNCTIONS

In addition to being useful as predictive diagnostic tools, the functions viewable in the Dashboard can be used to isolate and identify many operational problems.

The acceptable ranges for these functions are listed in the “Nominal Range” column of the analyzer’s *Final Test and Validation Data Sheet* shipped with the instrument. Values outside these acceptable ranges Indicates a failure of one or more of the analyzer’s subsystems. Functions whose values are still within acceptable ranges but have significantly changed from the measurement recorded on the factory data sheet may also Indicates a failure.

Make note of these values for reference in troubleshooting.

## 5.7.3. USING THE DIAGNOSTIC SIGNAL I/O FUNCTIONS

The signal I/O functions in the Utilities>Diagnostics menu allows access to the digital and analog functions in the analyzer. Some of the digital signals can be controlled through the Setup menu. These signals are useful for troubleshooting in three ways:

- The technician can view the raw, unprocessed signal level of the analyzer’s critical inputs and outputs.

- Many of the components and functions that are normally under algorithmic control of the CPU can be manually exercised.
- The technician can directly control the signal level Analog and Digital Output signals.

This allows the technician to observe systematically the effect of directly controlling these signals on the operation of the analyzer. Use the Utilities>Diagnostics menu to view the raw voltage of an input signal or the Setup menu to control the state of an output voltage or control signal.

### 5.7.4. FAULT DIAGNOSIS WITH LEDS

The following illustrations show connectors and LEDs that can indicate where issues may lie. Figure 5-8 shows the layout for the mainboard; Figure 5-9 shows the layout for the CAPS DAQ smart board, and Figure 5-10 shows the layout for the ozone tower smart module.






Figure 5-8. N500 Mainboard



**Figure 5-9. N500 CAPS DAQ Smart Board**



\*DS4 stays off until the end of a 20-minute warm-up period after initial power-on, unless it is forced on through the O3 Gen Override menu.

\*\*DS2 and DS4: an audible clicking sound must also be present to indicate that NO<sub>x</sub> mode is functioning.

**Figure 5-10. N500 Ozone Tower Smart Board**



## 5.7.5. CALIBRATION PROBLEMS

This section describes possible causes of calibration problems.

### 5.7.5.1. NEGATIVE CONCENTRATIONS

Negative concentration values can be caused by calibration error: If the zero air were contaminated, and the analyzer was calibrated at “zero”, the analyzer may report a negative value when measuring air that contains little or no NO<sub>x</sub>.

### 5.7.5.2. ABSENCE OF ANALYZER RESPONSE TO SAMPLE GAS

If the instrument shows no response (display value is near zero), even though sample gas is supplied properly, and the instrument seems to perform correctly:

1. Confirm the lack of response by supplying NO<sub>2</sub> span gas of about 80% of the range value to the analyzer.
2. Check the sample pressure and cell pressure for proper value.
3. Check for disconnected cables to the sensor module.
4. If delivering NO gas, check that the Lamp Current is 7-10 mA and the NO<sub>x</sub> valve is Pulsing in NO<sub>x</sub> mode. There is an audible clicking sound when the valve activates.

### 5.7.5.3. UNSTABLE ZERO AND SPAN

Leaks in the instrument or in the external gas supply and vacuum systems are the most common source of unstable and non-repeatable concentration readings.

1. Check for leaks in the pneumatic systems as described in Section 5.6.5.
2. Consider pneumatic components in the gas delivery system outside the analyzer such as a change in zero air source (ambient air leaking into zero air line) or a change in the span gas concentration due to zero air or ambient air leaking into the span gas line.
3. Once the instrument passes a leak check, do a flow check (this chapter) to ensure that the instrument is supplied with adequate sample and ozone air.
4. Confirm the sample pressure, cell pressure, and sample flow are correct and steady.
5. Verify that the sample filter element is clean and does not need to be replaced.

### 5.7.5.4. INABILITY TO SPAN - DEACTIVATED SPAN BUTTON

In general, the analyzer will deactivate certain buttons whenever the actual value of a parameter is outside of the expected range for that parameter. If the Span is grayed out, the actual concentration must be outside of the range of the expected span gas concentration, which can have several causes.

1. Verify that the expected concentration is set properly to the actual span gas concentration in the CONC sub-menu.
2. Confirm that the NO<sub>2</sub> span gas source is accurate.
  - This can be done by comparing the source with another calibrated analyzer.

3. Check for leaks in the pneumatic systems as described in Section 5.6.5.
  - Leaks can dilute the span gas and, hence, the concentration that the analyzer measures may fall short of the expected concentration defined in the CONC sub-menu.

#### 5.7.5.5. INABILITY TO ZERO - DEACTIVATED ZERO BUTTON

In general, the analyzer will deactivate certain buttons whenever the actual value of a parameter is outside of the expected range for that parameter. If the Zero button is grayed out, the actual gas concentration must be significantly different from the actual zero point (as per last calibration), which may be for any of several reasons.

1. Confirm that there is a good source of zero air.
2. Check to ensure that there is no ambient air leaking into zero air line.
3. Check for leaks in the pneumatic systems as described in Section 5.6.5.

#### 5.7.5.6. NON-LINEAR RESPONSE

The analyzer was factory calibrated to a high level of NO<sub>2</sub> and should be linear to within 1% of full scale. Common causes for non-linearity are:

- Leaks in the pneumatic system:
  - Leaks can add ambient air, zero air or span gas to the current sample gas stream, which may be changing in concentrations as the linearity test is performed.
  - Check for leaks as described in Section 5.6.5.
- The calibration device is in error:
  - Check flow rates and concentrations, particularly when using low concentrations.
  - If a mass flow calibrator is used and the flow is less than 10% of the full scale flow on either flow controller, you may need to purchase lower concentration standards.
- The standard gases may be mislabeled as to type or concentration.
  - Labeled concentrations may be outside the certified tolerance.
- The sample delivery system may be contaminated.
  - Check for dirt in the sample lines or reaction cell.
- Dilution air contains sample or span gas.
- Incoming concentrations may not be linear.
  - Check input bottles.
- Span gas overflow is not properly vented and creates a back-pressure on the sample inlet port. Also, if the span gas is not vented at all and does not supply enough sample gas, the analyzer may be evacuating the sample line.
  - Ensure to create and properly vent excess span gas.

### 5.7.6. OTHER PERFORMANCE PROBLEMS

Dynamic problems (i.e., problems that only manifest themselves when the analyzer is monitoring sample gas) can be the most difficult and time consuming to isolate and resolve. The following section provides an itemized list of the most common dynamic problems with recommended troubleshooting checks and corrective actions.

### 5.7.6.1. EXCESSIVE NOISE

Excessive noise levels under normal operation usually indicates leaks in the sample supply or the analyzer itself.

- Ensure that the sample or span gas supply is leak-free, and carry out a detailed leak check as described earlier in this chapter.

Other sources of measurement noise may be related to cabling issues or large ambient Temperature changes.

- Gain access to the instrument, when powered down, and reset the cable connectors.
- Do not locate the analyzer in the vicinity of the air conditioning ducting.

### 5.7.6.2. SLOW RESPONSE

If the analyzer starts responding too slow to any changes in sample, zero or span gas, check for the following:

- Dirty or plugged sample filter or sample lines.
- Sample inlet line is too long.
- Dirty or plugged flow restrictor. Check flows, pressures and, if necessary, change restrictor.
- Wrong materials in contact with sample - use glass, stainless steel or Teflon materials only.
- Insufficient time for purging lines upstream of the analyzer. Wait for stability is reached.
- Insufficient time for calibration gas source to stabilize. Wait until stability is reached.

### 5.7.6.3. AREF WARNINGS

Auto Reference (AREF) warnings occur if the signal measured during an AREF cycle is higher than 1100 Mm<sup>-1</sup>.

#### Note

**The AREF warning displays the value of the Auto Reference reading when the warning occurs.**

**(Also note that there will not be an AREF warning if the AREF feature was disabled due to prolonged time in CAL mode.**

**Ensure the instrument is returned to SAMPLE mode as soon as a calibration has been completed).**

- If this value is higher than 1100 Mm<sup>-1</sup>, check that the Auto Reference valve is operating properly.
  - To do so, use the Utilities>Diagnostics>Digital Outputs menu to toggle the valve on and off.
  - Listen to hear whether the valve is switching, and see if its D7 LED (associated with J8 connector) on the CAPS DAQ board lights accordingly
  - Check the power supply to the valve (24 V to the valve should turn on and off when measured with a voltmeter).

**Note**

**It takes only a small leak across the ports of the valve to show excessive Auto Zero values when supplying high concentrations of span gas.**

If the Auto Reference valve is working properly, then the problem could be due to dirty mirrors. Please contact Technical Support (Section 5.9) to confirm this and to rule out other possibilities.

## 5.7.7. SUBSYSTEM CHECK FOR TROUBLESHOOTING

The preceding sections of this manual discussed a variety of methods for identifying possible sources of failures or performance problems within the analyzer. This section describes how to determine if a certain component or subsystem is actually the cause of the problem being investigated.

### 5.7.7.1. AC MAIN POWER



**WARNING – ELECTRICAL SHOCK HAZARD**

**Should the AC power circuit breaker trip, investigate and correct the condition causing this situation before turning the analyzer back on.**

The instrument's electronic systems will operate with any of the specified power within the 100 VAC to 240 VAC, at 47 Hz to 63 Hz. Using the properly rated power cord, the instrument will power on when the rear panel Hard Power switch is placed in the ON position. (If the power source is disrupted, the instrument will turn on once the power is restored. If the instrument doesn't start, check the following possible causes and possible solutions:

- Check the power cord for damage, such as whether it's cut or burned.
- Check that the power cord is adequately rated for the instrument's specified power rating.
- Check that the power source is of the proper voltage for the instrument's specified power rating.
- If there are no findings in the preceding steps, then note whether the instrument had been opened for maintenance; if so, place the rear panel Hard Power switch in the OFF position, and disconnect the power cord; then reopen the instrument and check that no wiring had been dislodged, and no tools were left inside.
- If no other reason can be found for the instrument not powering on, then check the fuse with an ohmmeter to determine its viability: carefully follow the instructions in Section 5.7.8.1 to remove the fuse for testing.
  - If the fuse is blown, replace it with a fuse of the correct specifications as instructed in Section 5.7.8.1.
- If the fuse is not blown, or if the replacement fuse blows, then call Technical Support (Section 5.9).

### 5.7.7.2. LCD/DISPLAY MODULE

Assuming that there are no wiring problems and that the DC power supply is operating properly, the display screen should light and show the splash screen and other indications of its state as the CPU goes through its initialization process.

### 5.7.7.3. O<sub>3</sub> GENERATOR MODULE

The ozone generator can fail in three ways, electronically (printed circuit board) and functionally - UV Lamp issues or NO<sub>x</sub> valve issues. Assuming that air is supplied properly to the generator, the generator should automatically turn on 30 minutes after the instrument is powered up. However, if the generator appears to be working properly but the sensitivity or calibration of the instrument is reduced, suspect a leak in the ozone generator pneumatics or a NO<sub>x</sub> valve that isn't cycling On/Off in NO<sub>x</sub> mode.

A leak in the dryer or between the dryer and the generator can reduce sensitivity and cause performance drift. Carry out a leak check (Section 5.6.5).

### 5.7.7.4. O<sub>3</sub> GENERATOR OVERRIDE

This feature in the Utilities>Diagnostics menu is used to manually turn the ozone generator off and on. This should be done before disconnecting the generator, to prevent ozone from leaking out, and after a system restart if the user does not want to wait the 30-minute warm-up time.

### 5.7.7.5. INTERNAL SPAN GAS GENERATOR AND VALVE OPTIONS

The zero/span valves and internal span gas generator options need to be enabled in the software (contact the factory on how to do this).

The semi-permeable PTFE membrane of the permeation tube is severely affected by humidity. Variations in humidity between day and night are usually enough to yield highly variable output results. If the instrument is installed in an air-conditioned shelter, the air is usually dry enough to produce good results. If the instrument is installed in an environment with variable or high humidity, variations in the permeation tube output will be significant. In this case, a dryer for the supply air is recommended (dewpoint should be -20° C or less).

The permeation tube of the internal span gas generator option is heated with a proportional heater circuit, and the temperature is maintained at 50°C ±1°C. Check the IZS Temp in the Dashboard or the IZS Temp Raw signal in the Utilities>Diagnostics>Analog Outputs menu (configurable in the Setup>Analog Outputs menu). At 50° C, the temperature signal from the IZS thermistor should be around 2500 mV.

### 5.7.7.6. RS-232 COMMUNICATIONS

Teledyne API's analyzers use the RS-232 communications protocol to allow the instrument to be connected to a variety of computer-based equipment. Problems with RS-232 connections usually center around such things as incorrect connector configuration, incorrect software settings, improper/incomplete seating of the internal cable. Do not do anything inside the instrument without first contacting Technical Support (Section 5.9). For additional information, see Section 2.3.1.4 under "Serial Connection."

## 5.7.8. SERVICE PROCEDURES

This section contains some procedures that may need to be performed when a major component of the analyzer requires repair or replacement.

### Note

Regular maintenance procedures are discussed in Section 5.5 and are not listed here). Also, there may be more detailed service notes for some of the below procedures. Contact Teledyne API's Technical Support Department.



### WARNING – ELECTRICAL SHOCK HAZARD

Unless the procedure being performed requires the instrument to be operating, turn it off and disconnect power before opening the analyzer and removing, adjusting or repairing any of its components or subsystems.



### CAUTION – QUALIFIED TECHNICIAN

The operations outlined in this chapter are to be performed by qualified maintenance personnel only.

### 5.7.8.1. FUSE REPLACEMENT PROCEDURE

### ATTENTION

#### COULD DAMAGE INSTRUMENT AND VOID WARRANTY

Fuses do not typically fail without definite cause. Do not attempt to replace until after all measures to detect the cause of a power failure, per Section 5.7.7.1, have been carried out, including Soft Power switch LED not lit (neither solid nor blinking), but Hard Power switch is in ON position and instrument's power cord properly connected at both ends. If an ohmmeter shows that the fuse is good, or if a new fuse blows, call Technical Support (Section 5.9).



### WARNING – ELECTRICAL SHOCK HAZARD

Never pull out fuse drawer without ensuring that the Hard Power switch is in OFF position and power cord disconnected, to ensure there is no power to the instrument before checking/changing fuse.




Figure 5-11. Fuse Access

### 5.7.8.2. O<sub>3</sub> GENERATOR REPLACEMENT

The ozone tower has a printed circuit board (see Figure 2-3). The board has a blue LED that, when lit, indicates ozone is being generated. To replace the ozone generator:

1. Turn off the analyzer power; remove the power cord and the analyzer cover.
2. Disconnect the tubing from the ozone generator. **Never apply power to the analyzer when the internal pneumatics are disconnected.**
3. Unplug the CANBUS cable that is connected to the PCA.
4. Unscrew the two mounting screws that attach the ozone generator to the chassis and take out the entire assembly.
5. With a complete replacement generator with circuit board and mounting bracket attached, simply reverse the above steps to replace the current generator.

**Note**

**Ensure to carry out a leak check (Section 5.6.5) and a recalibration after the analyzer has warmed up for about 60 minutes.**

### 5.7.8.3. SAMPLE CONDITIONER ASSEMBLY REPLACEMENT

The sample conditioner assembly includes a permeation dryer, a Charcoal scrubber (AREF), and a DFU filter.

Change the dryer component as follows:

1. Turn off power to the analyzer and pump and remove the power cord and the analyzer cover.
2. Disconnect tubing to and from assembly. ***Never apply power to the analyzer when the internal pneumatics are disconnected.***
3. Noting the orientation, unscrew the mounting screws and remove the assembly.
4. Install the new assembly in the same orientation as noted in the preceding Step, and secure in place with the mounting screws.
5. Reattach tubing to and from assembly.
6. Carry out a detailed leak check (see Section 5.6.5.1).
7. Close the analyzer and reconnect the power cord.
8. Power up pump and analyzer and re-calibrate the instrument after it stabilizes.



## 5.8. FREQUENTLY ASKED QUESTIONS

The following list was compiled from the Teledyne API's Technical Support Department's ten most commonly asked questions relating to the analyzer.

| QUESTION   | ANSWER  |
|--|---|
| Why does the Enter button sometimes disappear in some of the menus?                              | Sometimes the Enter button will disappear if you input a setting that is invalid or out of the allowable range for that parameter, such as trying to set the 24-hour clock to 25:00:00 or a range outside its allowable limits. Once you adjust the setting to an allowable value, the Enter button will re-appear.   |
| Why is the ZERO or SPAN button deactivated during calibration?                                   | This happens when the measured gas concentration differs significantly from the span or zero gas concentration value entered by the user. This prevents accidental recalibration of the analyzer to an out-of-range response curve.<br>EXAMPLE: The span set point is 400 ppb but gas concentration being measured is only 50 ppb.  |
| How do I enter or change the value of my Span Gas?   | See Section 4.2.1.2.  |
| Can I automate the calibration of my analyzer?   | Any analyzer with zero/span valve or IZS option can be automatically calibrated using the instrument's AutoCal feature (Section 4.3).   |
| Can I use the IZS option to calibrate the analyzer?  | Yes. However, the accuracy of the IZS option's permeation tube is only $\pm 5\%$ . To achieve highest accuracy, it is recommended to use cylinders of calibrated span gases in combination with a zero air source.  |
| How do I measure the sample flow?  | For accurate measurement, attach a calibrated volumetric flow meter to the sample inlet port, and get a reading while the instrument is operating. The sample flow should be as specified in Table 1-1. (To calibrate, use the Utilities>Diagnostics menu; refer to Section 5.6.5.2).   |
| How often do I need to change the particulate filter?  | Refer to the Maintenance Schedule in Table 5-1. Keep in mind that highly polluted sample air may require more frequent changes.   |
| How long does the sample pump last?  | The sample pump should last one to two years and the pump head should be replaced when necessary. (Refer to Maintenance Schedule Table 5-1).  |
| Why does my RS-232 serial connection not work?   | There are several possible reasons: <ul style="list-style-type: none"> <li>• The wrong cable: please use the provided or a generic "straight-through" cable (do not use a "null-modem" type cable) and ensure the pin jumpers are correct (Section 2.3.1.4 under Serial Connection).</li> <li>• The baud rate of the analyzer's COM port does not match that of the serial port of your computer/data logger (Table 2-12).</li> </ul> |
| How do I set up and use the contact closures (Control Inputs) on the rear panel of the analyzer? | See Section 2.3.1.3.  |

## 5.9. TECHNICAL ASSISTANCE

For spare parts, or if this manual and its troubleshooting & service section do not solve your problems, technical assistance may be obtained from:

**Teledyne API Technical Support**  
**9970 Carroll Canyon Road**  
**San Diego, California 92131-1106 USA**

**Toll-free Phone:** +1 800-324-5190

**Phone:** +1 858-657-9800

**Fax:** +1 858-657-9816

**Email:** [api-techsupport@teledyne.com](mailto:api-techsupport@teledyne.com)

**Website:** <http://www.teledyne-api.com/>

## 6. PRINCIPLES OF OPERATION

The Cavity Attenuated Phase Shift (CAPS) NO<sub>x</sub> monitor operates as an optical absorption spectrometer that yields both reliable and accurate measurements down to sub ppb concentrations, with lower noise levels than chemiluminescence-based monitors. The CAPS method uses light from a blue Ultraviolet (UV) light emitting diode (LED) centered at 405 nm, a measurement cell with high reflectivity mirrors located at either end to provide an extensive optical path length, and a vacuum photodiode detector. These components are assembled into the optical cell which resides in a temperature-controlled oven. The oven raises the ambient temperature of the sample gas to 45 degrees Celsius. This mitigates the formation of moisture on the surfaces of the mirrors while also minimizing changes in the absorption coefficient due to temperature fluctuations.

Optical absorption is well-defined and is described by Beer's Law, where the Absorbance (lost light) is directly proportional to both the path-length and concentration of the absorbing gas.

$$A = \epsilon l c$$

*(A = Absorbance,  $\epsilon$  = Molar absorptivity,  $l$  = Mean path Length,  $c$  = concentration)*

The CAPS method employed in the N500 is unique in that it applies this fundamental optical absorption law in the frequency domain, rather than using relative changes in light intensity as the primary signal. Ultraviolet light (UV) from the modulating high intensity LED enters a near confocal optical cell (Figure 6-1) through the rear of mirror A. The intensity of the light, as observed by the detector, which is also modulating at a slightly different frequency, located behind Mirror B, builds exponentially in the cell while the LED is ON. The opposite is true when the LED is OFF. Because both mirrors are highly reflective at 405 nm, a prominent absorption band for NO<sub>2</sub>, the light takes a considerable amount of time to plateau in the absence of the absorbing gas. However, when NO<sub>2</sub> is present, the mean path length traveled by the light is significantly reduced. This has two effects on the observed intensity as measured by the detector:


- The light plateau intensity level is lower, more importantly for the N500
- The light intensity plateaus sooner.

Thus, an observed phase shift from the modulating LED is detected (Figure 6-2). The phase shift is largest when measuring zero air, decreases when NO<sub>2</sub> is present, and is proportionate to the concentration of the NO<sub>2</sub>.


Both the LED and the Detector are modulated ON and OFF such that the observed signal has a much lower frequency, equal to the difference between the modulated frequencies and is referred to as a beat frequency. The system hardware and software take advantage of this, as it makes it easier to post process the signal using a micro controller. The technique is known as heterodyning.

The instrument translates the phase shift from the presence of absorbing gas into a concentration measurement. Typical absorption techniques of other analyzers take a reference and measure value of the light intensity "level" in order to derive concentration and compensate for source drift. Using the CAPS technique, the amount of phase shift remains constant for a given concentration, even if the LED drifts over time. This measurement approach offers many advantages over traditional (or "Chemi") analyzers,

such as faster response (single gas stream), lower noise at span, and more importantly, greater specificity.



**Figure 6-1. N500 Optical Absorption Cell**



**Figure 6-2. Phase Shift Representation of Increased Concentration of NO<sub>2</sub>**

(Black = LED State, Blue = Light build up in the absence of NO<sub>2</sub>, Green = phase shifted/attenuated light).

## 6.1. PNEUMATIC OPERATION

An internal vacuum pump, which is located pneumatically downstream from all the instrument's other components, pulls the sample gas through the instrument's pneumatic pathway. The flow rate is controlled by a flow restrictor placed upstream of the sensor.

After being pulled through a series of filtering and conditioning components, the sample gas is routed through the sensor for a true NO<sub>2</sub> reading; the sample gas is alternately titrated with higher levels of ozone where the sensor reads a second, higher NO<sub>2</sub> concentration (NO<sub>x</sub> Mode), after which the two readings are used in calculations to provide a NO<sub>x</sub> value and an NO concentration. The analyzer alternates between NO<sub>2</sub> mode for 20 seconds and then NO<sub>x</sub> mode for 20 seconds.

The alternating valve toggling for ozone titration is very rapid and follows a pattern of signal processing states that the Dashboard displays in the "SigProc State" parameter as described in Table 6-1. To ensure accuracy, the analyzer periodically goes into AREF mode to run a background measurement while in Sample mode and adjusts for any baseline drift that may occur.

**Table 6-1. Sensor Signal Processing States**

| STATE      | DESCRIPTION  |
|------------|--|
| IDLE       | Transitioning between measurement modes (very brief state, typically not visible)                                    |
| NO2_WAIT   | Flushing the cell of the prior NO <sub>x</sub> sample before taking the NO <sub>2</sub> measurement.                 |
| NO2_SAMPLE | Reading NO <sub>2</sub> concentration.   |
| NOx_WAIT   | Filling the cell with titrated NO <sub>x</sub> before taking the NO <sub>x</sub> measurement.                        |
| NOx_SAMPLE | Reading NO <sub>2</sub> concentration in titrated sample (which is representative of NO <sub>x</sub> concentration). |
| AUTOREF    | In Auto Reference mode. Sample pulled through Charcoal scrubber.   |

Board LEDs shown in Figure 5-8, Figure 5-9, and Figure 5-10 also indicate what state the analyzer is in.

Briefly summarized, the Model N500 analyzer first measures NO<sub>2</sub> directly by way of optical absorption to yield a true reading of NO<sub>2</sub> concentration. Next, it uses a precisely-timed pulsation to mix the sample with an excessive concentration of O<sub>3</sub> to create NO<sub>x</sub>. The software subtracts the NO<sub>2</sub> concentration from the resulting NO<sub>x</sub> concentration to calculate an NO reading. The NO concentration is then corrected for titration efficiency (TE) which is then added to the measured NO<sub>2</sub> value to produce a corrected NO<sub>x</sub> measurement.

Pressure transducers monitor the sample pressure and the cell pressure to verify pressure levels, which are used for a variety of important calculations and diagnostics.

## 6.2. ELECTRONIC OPERATION

The electronic platform is based on a Controller Area Network (CAN) bus modular system. CAN is the central networking system that enables communication among all the parts and facilitates centralized diagnoses of errors, as well as configuration of all the parts. CAN bus technology allows for a uniform cable architecture with interchangeable 6-pin connectors configured for power (5 V and 24 V) and communications (CAN high and CAN low serial lines).

The Mainboard is the main hub, which not only contains the Central Processing Unit (CPU) that communicates with other modules, but also directs power and communication distribution. The Mainboard includes an altitude sensor, a temperature sensor, and the Supervisory Chip.

The Supervisory Chip monitors power and the sensors, and when the front panel Soft Power switch is pressed (see Power Switches, Section 6.2.2), the Supervisory Chip directs the soft power down of the internal components, to safely shut down processes and close connections to prevent damage.

### 6.2.1. MODULES

Each module consists of its own board controlled by a microprocessor that receives messages from and sends information to the Mainboard on the CAN network. Depending on the signal line, CAN Low or CAN High, the modules can determine whether a message is intended for them and what the priority is, and then act on the applicable messages. These are called “Smart Modules,” which conduct local operations, such as activating the zero/span valves, toggling the NO<sub>x</sub>/NO<sub>2</sub> valve, or controlling manifold temperature. There is also the Sensor Module; it is comprised of the Optical Bench Unit which has the gas sensor inside the main oven and the CAPS data acquisition (DAQ) board with logic device, microcontroller and LED driver mounted on the outside. The Sensor Module calculates gas concentrations and may command the Smart Modules.

### 6.2.2. POWER SWITCHES

The front panel Soft Power switch is used to protect the internal components from damage. When the instrument is initially powered on, the Supervisory Chip spins up the internal computer components and places them in operational mode (indicated by LED’s solid-lit state). However, before powering off the instrument, pressing and momentarily holding the solid-lit Soft Power switch tells the Supervisory Chip to put the internal computer components through a soft-shutdown process and into deep sleep mode (indicated by LED’s blinking state) in order to protect them from damage when fully turning off power.

The rear panel Hard Power switch is used to turn on or off the instrument; however, before turning off the instrument, the Soft Power switch must be used first as described above. If there is an unexpected loss of source power while the instrument is running, it will power up in the ON state when source power is restored.

## GLOSSARY

| TERM   | DESCRIPTION/DEFINITION   |
|--|--|
| ASSY   | <i>Assembly</i>  |
| CAS  | <i>Code-Activated Switch</i>   |
| CD   | <i>Corona Discharge</i> , a frequently luminous discharge, at the surface of a conductor or between two conductors of the same transmission line, accompanied by ionization of the surrounding atmosphere and often by a power loss          |
| CE   | <i>Converter Efficiency</i> , the percentage of the total amount that is actually converted (e.g., light energy into electricity; NO <sub>2</sub> into NO, etc.)   |
| CEM  | <i>Continuous Emission Monitoring</i>  |
| Chemical elements that may be included in this document: |  |
| CO <sub>2</sub>  | carbon dioxide   |
| C <sub>3</sub> H <sub>8</sub>                            | propane  |
| CH <sub>4</sub>  | methane  |
| H <sub>2</sub> O   | water vapor  |
| HC   | general abbreviation for hydrocarbon   |
| HNO <sub>3</sub>   | nitric acid  |
| H <sub>2</sub> S   | hydrogen sulfide   |
| NO   | nitric oxide   |
| NO <sub>2</sub>  | nitrogen dioxide   |
| NO <sub>x</sub>  | nitrogen oxides, here defined as the sum of NO and NO <sub>2</sub>   |
| NO <sub>y</sub>  | nitrogen oxides, often called odd nitrogen: the sum of NO <sub>x</sub> plus other compounds such as HNO <sub>3</sub> (definitions vary widely and may include nitrate (NO <sub>3</sub> ), PAN, N <sub>2</sub> O and other compounds as well) |
| NH <sub>3</sub>  | ammonia  |
| O <sub>2</sub>   | molecular oxygen   |
| O <sub>3</sub>   | ozone  |
| SO <sub>2</sub>  | sulfur dioxide   |
| cm <sup>3</sup>  | metric abbreviation for <i>cubic centimeter</i> (replaces the obsolete abbreviation "cc")  |
| CPU  | <i>Central Processing Unit</i>   |
| DAC  | <i>Digital-to-Analog Converter</i>   |
| DAS  | <i>Data Acquisition System</i>   |
| DCE  | <i>Data Communication Equipment</i>  |
| DFU  | <i>Disposable Filter Unit</i>  |
| DHCP   | <i>Dynamic Host Configuration Protocol</i> . A protocol used by LAN or Internet servers to automatically set up the interface protocols between themselves and any other addressable device connected to the network                         |
| DOM  | <i>Disk On Module</i> , a 44-pin IDE flash drive with up to 128MB storage capacity for instrument's firmware, configuration settings and data  |
| DOS  | <i>Disk Operating System</i>   |
| DRAM   | <i>Dynamic Random Access Memory</i>  |
| DR-DOS   | <i>Digital Research DOS</i>  |
| DTE  | <i>Data Terminal Equipment</i>   |
| EEPROM   | <i>Electrically Erasable Programmable Read-Only Memory</i> also referred to as a FLASH chip or drive   |
| ESD  | <i>Electro-Static Discharge</i>  |

| TERM        | DESCRIPTION/DEFINITION  |
|-------------|---|
| ETEST       | <i>Electrical Test</i>  |
| Ethernet    | a standardized (IEEE 802.3) computer networking technology for local area networks (LANs), facilitating communication and sharing resources   |
| FEP         | <i>Fluorinated Ethylene Propylene</i> polymer, one of the polymers that Du Pont markets as <i>Teflon</i> <sup>®</sup>   |
| Flash       | non-volatile, solid-state memory  |
| FPI         | <i>Fabry-Perot Interface</i> : a special light filter typically made of a transparent plate with two reflecting surfaces or two parallel, highly reflective mirrors   |
| GFC         | <i>Gas Filter Correlation</i>   |
| IC          | <i>Integrated Circuit</i> , a modern, semi-conductor circuit that can contain many basic components such as resistors, transistors, capacitors, etc., in a miniaturized package used in electronic assemblies   |
| IP          | <i>Internet Protocol</i>  |
| IZS         | <i>Internal Zero Span</i>   |
| LAN         | <i>Local Area Network</i>   |
| LCD         | <i>Liquid Crystal Display</i>   |
| LED         | <i>Light Emitting Diode</i>   |
| LPM         | <i>Liters Per Minute</i>  |
| MFC         | <i>Mass Flow Controller</i>   |
| M/R         | <i>Measure/Reference</i>  |
| NDIR        | <i>Non-Dispersive Infrared</i>  |
| MOLAR MASS  | <p>the mass, expressed in grams, of 1 mole of a specific substance. Conversely, one mole is the amount of the substance needed for the molar mass to be the same number in grams as the atomic mass of that substance.</p> <p>EXAMPLE: The atomic weight of Carbon is 12 therefore the molar mass of Carbon is 12 grams. Conversely, one mole of carbon equals the amount of carbon atoms that weighs 12 grams.</p> <p>Atomic weights can be found on any Periodic Table of Elements.</p> |
| NDIR        | <i>Non-Dispersive Infrared</i>  |
| NIST-SRM    | <i>National Institute of Standards and Technology - Standard Reference Material</i>   |
| PC          | <i>Personal Computer</i>  |
| PCA         | <i>Printed Circuit Assembly</i> , the PCB with electronic components, ready to use  |
| PC/AT       | <i>Personal Computer / Advanced Technology</i>  |
| PCB         | <i>Printed Circuit Board</i> , the bare board without electronic component  |
| PFA         | <i>Per-Fluoro-Alkoxy</i> , an inert polymer; one of the polymers that <i>Du Pont</i> markets as <i>Teflon</i> <sup>®</sup>  |
| PLC         | <i>Programmable Logic Controller</i> , a device that is used to control instruments based on a logic level signal coming from the analyzer  |
| PLD         | <i>Programmable Logic Device</i>  |
| PLL         | <i>Phase Lock Loop</i>  |
| P/N (or PN) | <i>Part Number</i>  |
| PSD         | <i>Prevention of Significant Deterioration</i>  |
| PTFE        | <i>Poly-Tetra-Fluoro-Ethylene</i> , a very inert polymer material used to handle gases that may react on other surfaces; one of the polymers that <i>Du Pont</i> markets as <i>Teflon</i> <sup>®</sup>  |
| PVC         | <i>Poly Vinyl Chloride</i> , a polymer used for downstream tubing   |
| Rdg         | Reading   |



| TERM   | DESCRIPTION/DEFINITION  |
|--------|---|
| RS-232 | specification and standard describing a serial communication method between DTE (Data Terminal Equipment) and DCE (Data Circuit-terminating Equipment) devices, using a maximum cable-length of 50 feet   |
| RS-485 | specification and standard describing a binary serial communication method among multiple devices at a data rate faster than RS-232 with a much longer distance between the host and the furthest device  |
| SAROAD | <i>Storage and Retrieval of Aerometric Data</i>   |
| SLAMS  | <i>State and Local Air Monitoring Network Plan</i>  |
| SLPM   | <i>Standard Liters Per Minute</i> of a gas at standard temperature and pressure   |
| STP    | <i>Standard Temperature and Pressure</i>  |
| TCP/IP | <i>Transfer Control Protocol / Internet Protocol</i> , the standard communications protocol for Ethernet devices  |
| TEC    | <i>Thermal Electric Cooler</i>  |
| TPC    | <i>Temperature/Pressure Compensation</i>  |
| USB    | <i>Universal Serial Bus</i> : a standard connection method to establish communication between peripheral devices and a host controller, such as a mouse and/or keyboard and a personal computer or laptop |
| VARs   | <i>Variables</i> , the variable settings of the instrument  |
| V-F    | <i>Voltage-to-Frequency</i>   |
| Z/S    | <i>Zero / Span</i>  |

## APPENDIX A – MODBUS REGISTERS

| ADDR                   | DESCRIPTION                      | UNITS  |
|------------------------|----------------------------------|--|
| <b>Discrete Inputs</b> |                                  |  |
| 1                      | SYS_WARN_CAPS_COM_WARNING        | Indicates status of CAPS board communication warning       |
| 2                      | SYS_WARN_CELL_PRESSURE           | Indicates status of cell pressure warning                  |
| 3                      | SYS_WARN_MANUAL_AREF             | Indicates status of manual auto reference warning          |
| 4                      | SYS_WARN_OZONE_PRESSURE          | Indicates status of Ozone pressure warning                 |
| 5                      | SYS_WARN_OZONE_TOWER_COM_WARNING | Indicates status of Ozone tower communication warning      |
| 6                      | SYS_WARN_RESET                   | Indicates status of system reset warning                   |
| 7                      | SYS_WARN_SAMPLE_FLOW             | Indicates status of sample flow warning                    |
| 8                      | SYS_WARN_SAMPLE_PRESSURE         | Indicates status of sample pressure warning                |
| 9                      | SYS_WARN_SAMPLE_TEMP_WARN        | Indicates status of sample temperature warning             |
| 10                     | SYS_WARN_TIME_NOT_SYNCED         | Indicates status of time-since warning                     |
| 11                     | SYS_WARN_ACAL1_CAL_FAIL          | Indicates Auto-Cal Seq 1 failed                            |
| 12                     | SYS_WARN_ACAL2_CAL_FAIL          | Indicates Auto-Cal Seq 2 failed                            |
| 13                     | SYS_WARN_ACAL3_CAL_FAIL          | Indicates Auto-Cal Seq 3 failed                            |
| <b>Coils</b>           |                                  |  |
| 20                     | MB_ZERO_CAL_RANGE1               | Trigger zero calibration of NO <sub>2</sub> range 1        |
| 21                     | MB_SPAN_CAL_RANGE1               | Trigger span calibration of NO <sub>2</sub> range 1        |
| <b>Input Registers</b> |                                  |  |
| 0                      | NO2_SLOPE1                       | Slope for range 1  |
| 4                      | NO2_OFFSET1                      | Offset for range 1 in PPB                                  |
| 8                      | NO2_PRE_CAL_CONC_1               | Concentration just prior to last zero/span cal of range #1 |
| 12                     | NO2_CONC                         | NO <sub>2</sub> concentration for range 1                  |
| 16                     | NO2_STABILITY                    | NO <sub>2</sub> stability                                  |
| 18                     | AI_REALTIME_LOSS                 | Realtime loss value in Mm-1                                |
| 20                     | AI_OVEN_TEMP                     | Oven temperature in degree C                               |
| 22                     | AI_OVEN_DUTY_CYCLE               | Oven temperature duty cycle                                |
| 24                     | AI_MANIFOLD_TEMP                 | Manifold temperature in degree C                           |
| 26                     | AI_MANIFOLD_CYCLE                | Manifold temperature duty cycle                            |
| 28                     | AI_IZS_TEMP                      | IZS lamp temperature in degree C                           |
| 30                     | AI_IZS_CYCLE                     | IZS temperature duty cycle                                 |
| 32                     | AI_SAMPLE_TEMP                   | Sample temperature in degree C                             |
| 34                     | AI_SAMPLE_PRESSURE               | Sample pressure in inches Hg-A                             |
| 36                     | AI_BOX_TEMP                      | Box temperature in degree C                                |
| 42                     | AI_PHASE_ANGLE                   | Phase angle value  |
| 44                     | NO_SLOPE1                        | NO slope for range 1                                       |
| 48                     | NO_OFFSET1                       | NO Offset for range 1                                      |
| 52                     | NO_PRE_CALC_CONC_1               | NO concentration for range 1 during zero/span calibration  |
| 56                     | NO_CONC                          | NO concentration for range 1                               |
| 60                     | NO_STABILITY                     | NO stability   |
| 62                     | NOX_SLOPE1                       | NOX slope for range 1                                      |

| ADDR                     | DESCRIPTION            | UNITS  |
|--------------------------|------------------------|--|
| 68                       | NOX_OFFSET1            | NOX Offset for range 1                                     |
| 72                       | NOX_PRE_CALC_CONC_1    | NOX concentration for range 1 during zero/span calibration |
| 76                       | NOX_CONC               | NOX concentration for range 1                              |
| 80                       | NOX_STABILITY          | Concentration stability                                    |
| <b>Holding Registers</b> |                        |  |
| 0                        | NO2_TARGET_SPAN_CONC_1 | NO2 Target span concentration for range #1                 |
| 4                        | NOX_TARGET_SPAN_CONC_1 | NOX Target span concentration for range #1                 |

